

HiMedia Laboratories Private Limited

23, Vadhani Industrial Estate, L.B.S. Marg,
Mumbai - 400086 , Website : www.himedialabs.com,
Email : info@himedialabs.com

Certificate of Analysis, Quality and Conformity

Material Code : M880	Material Name : Rappaport Vassiliadis Medium	Lot No : 0000409725
Report No.: 040000968248	Date of Release & Report : 2019-11-11	Expiry Date : 2024-10

Appearance

Light yellow to light blue homogeneous free flowing powder . Observed : Light blue

Colour and Clarity of prepared medium

Bluish green coloured, clear to slightly opalescent solution with slight precipitate

Reaction

Reaction of 4.92% w/v aqueous solution at 25°C.

pH

pH Range :5.00-5.40 Observed : 5.26

Cultural Response

Cultural characteristics observed after an incubation at 42-43°C for 18- 24 hours. After incubation, subculture on selective agar media like MacConkey Agar (M081) or XLD Agar (M031) and incubate at 35-37°C for 18-24 hours.

Organism	Inoculum (CFU)	Growth at 42±1°C	Recovery	Colour of Colony on M081
Cultural Response				
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	50-100	none-poor	<=10%	pink-red
<i>Salmonella enterica</i> serovar <i>Enteritidis</i> ATCC 13076 (WDCM 00030)	50-100	good-luxuriant	>=50%	colourless
<i>Salmonella enterica</i> serovar <i>Typhi</i> ATCC 6539	50-100	good-luxuriant	>=50%	colourless
<i>Salmonella enterica</i> serovar <i>Typhimurium</i> ATCC 14028 (WDCM 00031)	50-100	good-luxuriant	>=50%	colourless

- . ATCC is a registered trade mark of the American Type Culture Collection
- . NCTC and National Collection of Type Culture are registered trade mark of the Health Protection Agency

Control Media :

- . For Bacteria : Soyabean Casein Digest Agar / Columbia Blood Agar base enriched with 5% v/v Sheep/Horse blood.
- . For Yeast & Mold : Sabouraud Dextrose Agar.

- . All ISO 11133 : 2014/Amd.1:2018(E) control strains are included in the Quality parameter
- . HiMedia Laboratories Pvt Ltd is Certified for ISO 9001:2015, ISO 13485:2012 , WHO GMP

HiMedia Laboratories Private Limited

23, Vadhani Industrial Estate, L.B.S. Marg,
Mumbai - 400086 , Website : www.himedialabs.com,
Email : info@himedialabs.com

Certificate of Analysis, Quality and Conformity

Material Code : M880	Material Name : Rappaport Vassiliadis Medium	Lot No : 0000409725
Report No.: 040000968248	Date of Release & Report : 2019-11-11	Expiry Date : 2024-10

STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets. The results reported were obtained at the time of release.

This document has been produced electronically and is valid


Sheetal Shewale

**Microbiologist/Sr.Executive
Microbiologist**


Ujwala M. Kokate

Asst./Dy/QC Manager


Dr. Santosh Kaul

Dy/QA Manager

2019-11-11