

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Urinalysis Reagent Strips

PURCHASE NO.: 37740068B

BATCH NO.: URS0080061

EXPIRATION DATE: 2022-09-04

QUANTITY: 60,000pcs

ANALYSIS		SPECIFICATIONS (Color card)	TEST RESULT
Glucose	Neg. urine	Negative	Complies
	250mg/dL	>100mg/dL <500mg/dL	Complies
Ketone	Neg. urine	Negative	Complies
	Yihua Urine controls	>40mg/dL <160mg/dL	Complies
S.G	1.000	≥1.000 <1.005	Complies
	1.025	>1.020 <1.030	Complies
Blood	Neg. urine	Negative	Complies
	Yihua Urine controls	≥++ ≤+++	Complies
pH	5.0	≥5.0 <6.0	Complies
	9.0	>8.0 ≤9.0	Complies
Protein	Neg. urine	Negative	Complies
	Yihua Urine controls	>+, ≤+++	Complies
Nitrite	Neg. urine	Negative	Complies
	Yihua Urine controls	+	Complies
Leukocyte	Neg. urine	Negative	Complies
	Yihua Urine controls	>+ <+++	Complies

Quality Department: Hedy Zhang

Date: Sep. 23, 2020

