

Manufacturer's Authorization

Date: March 11, 2019

TO: WHEREAS

We Biobase Biodustry (Shandong) Co., Ltd, who are official manufacturers of *Biological safety cabinet , Laminar flow cabinet, Fume hood, Incubator, Drying Oven, Refrigerator&freezer, Autoclave, Steam Generator, Pressure Sterilizer, Centrifuge-complete range etc.* having factories at *Mingshui Economic Developing Zone, Zhangqiu, Jinan, Shandong, China*, do hereby authorize SRL ECOCHIMIE to submit a bid the purpose of which is to provide the following Goods, manufactured by us *Biological safety cabinet , Laminar flow cabinet, Fume hood, Incubator, Drying Oven, Refrigerator&freezer, Autoclave, Steam Generator, Pressure Sterilizer, Centrifuge-complete range etc.* and to subsequently negotiate and sign the Contract.

We hereby extend our full guarantee and warranty in accordance with the Conditions of Contract, with respect to the Goods offered by the above firm.

Ecochimie SRL is authorized to register all equipment at the Agency of Drugs and Medical Devices

Signed:

Ruth yang



Name: *Ruthyang*

Title: Sales Manager

Duly authorized to sign this Authorization on behalf of: SRL ECOCHIMIE

Dated on March 11, 2019



This is to certify that the Quality Management System of

Biobase Biodustry (Shandong) Co., Ltd.

Unified Social Credit Code : 9137018179889855X7

Operation Address : Biobase Biodustry Park, Crossing of East Jingshi Road and Mingbu Road, Mingshui Economic Development Zone, Zhangqiu, Jinan City, Shandong Province, China(Production); No.51, South Gongye Road, Jinan City, Shandong Province, China(Sale)

Registered Address : Biobase Biodustry Park, Crossing of East Jingshi Road and Mingbu Road, Mingshui Economic Development Zone, Zhangqiu, Jinan City, Shandong Province, China

applicable to

R & D, production and sales of in vitro diagnostic kit(see attachment for details), biosafety cabinet, medical aerosol absorber, pressure steam sterilizer, blood bank refrigerator, microbial incubator and automated ELISA processor(within the scope of license qualification)

has been assessed and registered by NQA against the provisions of

ISO 13485: 2016

This registration is subject to the company maintaining a quality management system, to the above standard, which will be monitored by NQA.

Certified Clients shall accept regular surveillance assessments, the validity of certificates shall be maintained for the positive result of audit.

The information of this certificate can be checked on CNCA's website (www.cnca.gov.cn)
SNQA's website : www.snqa.com.cn

M. G. Smith

Managing Director



Certificate Number **45014**

Date: 09 July 2018
Valid Until: 09 July 2021
EAC Code: 19



XSB-Series Laboratory Biological Microscope



	Specification	Model
		XSB-301A
Viewing Head		
	Sliding Trinocular Head inclined at 45°	√
Nosepiece	Ball Beating Quadruple Nosepiece	√
Eyepiece	WF10X18mm	√
Objective		
	Achromatic 4X,10X,40X(S)100X(S,Oil)	√
Stage	Double Layer Mechanical Stage 135*140mm	√
Condenser	ABBE NA1.25 condenser with Iris Diaphragm &filter,rack&pinion adjustable	√
Illumination	Build-in Illumination,Halogen lamp 6V/20W,	√
Power Supply	AC 110/220V±10%,50/60Hz	√
Accessories	Photo attachment, Videoattachment	√
	Dark fieldcondenser,Polarization set ,Phase contrast kit	√
Package	Each set in styrofoam,to cartoon for safety	√
Package	Cartoon with foam	√
Package size	240*340*400mm	√
Gross weight	7kg	√

Note:110V/220V for LED lamp ,220V only for Halogen lamp

