

## Sırt Tahtası (Plastik) (ES-510)

### Spine Board (Plastic)



- Vakumlama sistemi ile sert plastikten imal edilmiştir.
- 160 Kg. taşıma kapasitesine sahiptir.
- Çocuk ve yetişkinler için 2 farklı tiptedir.
- Yetişkin tipinde 3 adet, çocuk tipinde 2 adet emniyet kemeri vardır.

#### Text

- *Made of hard plastic with vacuum system.*
- *160 Kg. maximum carrying capacity.*
- *Two different types for children and adults.*
- *3 seat belts for adult type, 2 for children type.*
- *X ray&EMR compatible*

#### Yetişkin için / For adults

Ağırlık Weight	Uzunluk Lenght	Genişlik Width
5.1 Kg.	1840 mm.	425 mm.

#### Çocuk için / For child

Ağırlık Weight	Uzunluk Lenght	Genişlik Width
3.5 Kg.	1250 mm.	425 mm.

## Kurtarma Yeleği (ET-370)

### Body Splint

- Omurga, baş, boyun kırık zedelenmelerinde, hastanın güvenli ve hareket ettirilmeden kurtarma operasyonları ve taşınmasında kullanılır.
- Renk kodlu, kolay açılıp kilitlenebilir kayışları sayesinde karanlıkta ve acil müdahale gerektiren anlarda bile hatasız olarak kullanılabilir.
- Taşıma çantası bulunmaktadır.



- *Usage for rescue operations and carrying of patient safely in cases of spine, head, neck fractures and injuries.*
- *Can be used faultlessly by means of easy opening and lockable belts with color codes even in the darkness and in the events requiring emergency intervention.*
- *Accompanying carrying case.*

# DECLARATION OF CONFORMITY



EMS MOBİL SİSTEMLER A.Ş..

1. Organize Sanayi Bölgesi Kırmızı Hanlığı Caddesi No: 9

Sincan-Ankara /TURKEY

Tel : +90 312 267 18 67 - Faks : +90 312 267 18 66

**Product** : Spine Board (Plastic Adult)

**Product Code** : ES 510

**Classification** : Class I- Rule I

We herewith declare with this document that compliance evaluation was applied upon the above mentioned products as per the Council Directive 93/42/EEC Medical Devices Regulation Annex 7 for Medical Device Directive


## General Applicable Directives

Medical Device Directive council directive 93/42/EEC of June 1993 concerning medical devices. (MDD 93/42/EEC) Compliance Declaration evaluation was applied as per Annex 7.

## Standarts

- |                      |   |
|----------------------|---|
| TS EN 1865-1+A1:2015 | Patient handling equipment used in road ambulances - Part 1: General stretcher systems and patient handling equipment |
| TS EN 980:2011       | Symbols for use in the labeling of medical devices  |
| TS EN 1041+A1:2015   | Information supplied by the manufacturer of medical devices   |
| TS EN ISO 14971:2013 | Medical devices - Application of risk management to medical devices   |

## Signature

  
EMS MOBİL SİSTEMLER A.Ş.  
1. Organize Sanayi Bölgesi  
Kırmızıhanlığı Caddesi No: 9 Sincan/ANKARA  
Tel: 0312 267 18 67 - Fax: 0312 267 18 66  
Merkezi No: 0475005500500012  
Sincan Vefi Dairesi: 475 005 5005

**Name** : Mevlüt Şahin

**Position** : General Manager

**Date** : 15.05.2016