




neodisher endo[®] SEPT GA



High-level disinfectant for the automated reprocessing of flexible endoscopes

Liquid concentrate

Field of application:

- High-level disinfection of flexible endoscopes in washer disinfectors for flexible endoscopes (WD-E) of all leading manufacturers

Activity spectrum:

- Active against bacteria (incl. MRSA, tuberculosis bacteria and *Helicobacter pylori*), yeasts and viruses (incl. hepatitis A, B and C, HIV, rotaviruses, noroviruses)
- The disinfecting activity has been tested and confirmed according to DIN EN 14885. neodisher endo SEPT GA conforms to the requirements on disinfectants for the disinfection of flexible endoscopes according to DIN EN ISO 15883-4
- The automated reprocessing process with neodisher endo CLEAN and neodisher endo SEPT GA fulfils the requirements of DIN EN ISO 15883-4 with regard to a germ reduction of $> 9 \log 10$ in the total process. Additionally, it is active against the spores of *Clostridium difficile*
- On the list of the IHO¹ for disinfectants with virucidal activity

Special Properties:

- Particularly user-friendly due to its low-odour formulation
- Excellent material compatibility, suitable for the endoscopes of all leading manufacturers
- Free of formaldehyde and quaternary ammonium compounds (QAV)

Application and dosage:

- neodisher endo SEPT GA is used in washer disinfectors for flexible endoscopes (WD-E). neodisher endo SEPT GA is dosed via the machine's integrated dosing device at the beginning of the disinfection step. The following parameters must be observed to achieve the activity spectrum mentioned below:

bactericidal, fungicidal, mycobactericidal and virucidal activity	10 ml/l (1.0 %), 55 °C, 5 min.
---	-----------------------------------

- For the cleaning step we recommend the alkaline-enzymatic detergent neodisher endo CLEAN.
- The products neodisher endo CLEAN and neodisher endo SEPT GA are perfectly matched. A possible carry-over of neodisher endo CLEAN into the disinfection step has no impact on the disinfection performance of neodisher endo SEPT GA.

General instructions on use:

- For professional use only.
- The neodisher endo SEPT GA working solution has to be rinsed off completely with water (preferably deionised water).
- Rinse out dosing system including suction hose with water before changing product.



neodisher endo[®] SEPT GA

- Reprocessing should comply with all ordinances pursuant to the Medical Device Directive and should be performed with appropriate validated processes.
- Please observe the instrument manufacturer's recommendations for reprocessing according to the requirements of DIN EN ISO 17664.
- The instructions given by the manufacturer of the washer disinfector are to be observed.
- Do not mix with other products.

Expert reports:

The disinfecting activity has been confirmed by certification. Expert reports are available on request.

Technical data:

pH-value	approx . 4.3 (1.0 %, determined in deionised water, 20 °C)
viscosity	< 10 mPa s (concentrate, 20 °C)
density	1.0 g/cm ³ (20 °C)

Ingredients:


Disinfecting agents in 100 g: 10.5 g glutaral.

CE-mark: 0297

neodisher endo SEPT GA conforms to the European directive 93/42/EEC, Annex I, concerning medical devices.

Storage information:

Always store at a temperature between 0 and 30 °C.

Usable for 3 years when stored as recommended.
For expiry date refer to the stamp mark on the label behind the hourglass symbol .

Hazard and precautionary statements:

For safety information see EC Safety Data Sheets. These are available at www.drweigert.com under the category "Service".

Dispose only when container is empty and closed. For disposal of product residues, refer to Safety Data Sheet.

MB 4071/3-2
05/2015

¹ Industrieverband Hygiene und Oberflächenschutz (Industry Association Surface Protection and Hygiene, Germany)

With the above information, to our current knowledge we describe our product regarding safety necessities, but we do not involve any quality description or promise certain properties.