

Quality Control Report


SEM ELISA kit

(Cat No: KA00307H)

Lot No.: H260113C10

Date of Production: 2026-01-14

Expired: 2027-01-13



Standards						
Ser. No.	Concentration ppt	Absorbance (Mean)	(CV)	B/B0 (%)	calculated ppt	Deviation (%)
1	0.0	2.168	0.8	100.0		
2	25.0	1.659	2.3	76.5	25.0	0.0
3	75.0	1.130	2.0	52.1	75.1	0.1
4	225.0	0.646	1.8	29.8	224.8	0.1
5	675.0	0.265	5.3	12.2	676.5	0.2
6	2025.0	0.090	0.8	4.2	2019.8	0.3

Please notice: This QC report is only for reference, for the absorbance of the standard solutions may fluctuate in a certain range.

