

BOEN HEALTHCARE CO., LTD


CERTIFICATE OF ANALYSIS

Report No.: GM03-BN23

Product Name	Cover Glass	Specification	24x24mm
Quantity	1200bxs	Test Quantity	25bxs
Lot. No.	20230401	Mfg. Date	20230401
Exp. Date	20260331	Report Date	2023.05.15
Inspection Standard	JB/T 8230.4-1997		
Technical Requirement			Conclusion
The size of cover glass is 24x24mm.			Meet the requirement
The cover glass is colorless and transparent.			Meet the requirement
The surface should be smooth and thickness should be uniformity.			Meet the requirement
The cover glass should be clean, no mist, fingerprint or water stains.			Meet the requirement
The edges should be smooth and without damage.			Meet the requirement
After boiled test, the microscope slide should be no blur, bubble or crack.			Meet the requirement
Conclusion: According to the requirements of JB/T 8230.4-1997, concluded that this batch pass.			

Checker: Leo Sun Date: 2023.05.15

Approved: Irene Qiao Date: 2023.05.15



BOEN HEALTHCARE CO., LTD


CERTIFICATE OF ANALYSIS

Report No.: GM01-BN24

Product Name: Eppendorf Pipette Tips		Lot. No.: 20240301					
MFG. Date: 2024-03		EXP. Date: 2029-02					
Raw Material: PP		Specification: 1000μL					
Standard: Enterprise Standard							
Package: 500pcs/bag, 15000pcs/ctn							
Quantity: 500000pcs			Report Date: 2024.03.26				
Test Items	Technical Requirement	Inspection Level	Sample	AQL		Defect Qty	Result
Dimension	L: 69±1mm	5pcs/lot	5pcs	0		0	Pass
				Ac	Re		
				0	1		
Appearance	No burr, black dot, dust, scratch, no oil spot, breakage, shrink color correct, blocking is not permitted	GB/T 2828.1-2003 Normal test level II	800pcs	1.5		0	Pass
				Ac	Re		
				21	22		
Function	Can be assembled with pipettor well	10pcs/lot	10pcs	0		0	Pass
				Ac	Re		
				0	1		
Package	Label and marking should be clear, correct and the label should be on correct location.	1 outer box/lot	1 outer box	0		0	Pass
				Ac	Re		
				0	1		
Conclusion		The product meets all the requirements by our test.					

Inspected by: Leo Sun

Checked by: Irene Qiao



BOEN HEALTHCARE CO., LTD


CERTIFICATE OF ANALYSIS

Report No.: GM03-BN23

Product Name: Gilson Pipette Tips		Lot. No.: 20230401					
MFG. Date: 2023-04		EXP. Date: 2028-03					
Raw Material: PP		Specification: 200μL					
Standard: Enterprise Standard							
Package: 1000pcs/bag, 50000pcs/ctn							
Quantity: 1500000pcs			Report Date: 2023.05.15				
Test Items	Technical Requirement	Inspection Level	Sample	AQL		Defect Qty	Result
Dimension	L: 50±1mm	5pcs/lot	5pcs	0		0	Pass
				Ac	Re		
				0	1		
Appearance	No burr, black dot, dust, scratch, no oil spot, breakage, shrink color correct, blocking is not permitted	GB/T 2828.1-2003 Normal test level II	800pcs	1.5		0	Pass
				Ac	Re		
				21	22		
Function	Can be assembled with pipettor well	10pcs/lot	10pcs	0		0	Pass
				Ac	Re		
				0	1		
Package	Label and marking should be clear, correct and the label should be on correct location.	1 outer box/lot	1 outer box	0		0	Pass
				Ac	Re		
				0	1		
Conclusion	The product meets all the requirements by our test.						

Inspected by: Leo Sun

Checked by: Irene Qiao



BOEN HEALTHCARE CO., LTD.
CERTIFICATE OF ANALYSIS

Report No.: GM01-BN24

Product Name	Microscope Slide	Specification	7105
Order Quantity	4000BXS	Test Quantity	13bxs
Lot. No.	20240301	Mfg. Date	20240301
Exp. Date	20270228	Report Date	2024.03.26
Inspection Standard	JB/T 8230.3-1997		
No.	Technical Requirement	Conclusion	
1	The size of microscope slide is 25.4 x 76.2mm (1"x 3").	Meet the requirement	
2	The Thickness of microscope slide is 1.1mm±0.05mm.	Meet the requirement	
3	The microscope slide should be colorless and transparent.	Meet the requirement	
4	The surface of microscope slide should be smooth and thickness should be uniformity.	Meet the requirement	
5	The microscope slide should be clean, no mist, fingerprint or water stains.	Meet the requirement	
6	The edges of microscope slide should be smooth and should be polished complete.	Meet the requirement	
7	After boiled test, the microscope slide should be no blur, bubble or crack.	Meet the requirement	
Conclusion	According to the requirements of JB/T 8230.3-1997, concluded that this batch pass.		

Inspector: Leo Sun
Date: 2024.03.26



Checker: Irene Qiao
Date: 2024.03.26

CERTIFICATE OF ANALYSIS

Report No.: GM03-BN23

Product Name	Stool Container	Specification	30mL, with label
Lot. No.	20230401	Quantity	100000pcs
Mfg. Date	2023-04	Exp. Date	2028-03
Test Standard	Factory Standard	Report Date	2023.05.15
Item	Technical Requirement	Result	Conclusion
Appearance	The surface of stool container should be smooth, no burr, no flying edges, no crack, no oil stains or other injection defects.	Comply	Qualified
	The color of the container is transparent and the cap is red, there is no different color difference.	Comply	Qualified
Performance	The cap and the container can be screwed well.	Comply	Qualified
	When the cap and the container is screwed, there is no leakage.	Comply	Qualified
Spoon	The spoon should be smooth, no burr, no flying edges, no crack and other injection defects.	Comply	Qualified
Conclusion	Meet the requirement		

Inspector: Leo Sun
Date: 2023.05.15Checker: Irene Qiao
Date: 2023.05.15

BOEN HEALTHCARE CO., LTD.

Certificate of Analysis

Product Name:	Visible Blood Collection Needle	Report Date:	2023.05.15
Specification:	21G, 22G	Test Qty:	200pcs
Produce Date:	20230330	Batch Qty:	20000pcs
Lot No.:	20230330	Expiry Date:	20260329
Inspection Item	Technique request	Result	
Contamination	Observe with normal vision or rectified vision, the surface is cleaning, no contaminate.	pass	
Color	The nominal outside diameter color should be same with the color of neilsbed and/or Protection Covers or needle handle.	pass	
Connection Fastness	No abruption between needle and neilsbed, after 20N Axial static tension for 10s on the connection	pass	
Needle Lumen Smoothness	Lumen should be smooth, under the pressure of 20Kpa, water flow ≥ 21 ml	28.9ml	
Neilsbed	no apparent burr, bubble or any other defect	pass	
Needlepoint	Good puncture force, after enlarge 2.5 times, eyeballing, the needlepoint should be sharp, no burr, no flat top, no hook or any other defect	pass	
Rubber Sheath	Good sealing performance, continuously puncture its top for 10 times, there should not be water flows down. No adhesion and break.	pass	
Lubricant	There should be no visible accumulation of the lubricant on the needle appearance	pass	
Handle of Needle	The handle of Needle should be complete with clear mark. And it should be on the same angle with needlepoint, the gradient shouldn't larger than 30° .	pass	
Plastic Protection Covers	straight, no visible bend; cover should be well fitted with neilsbed or the other protection cover, it shouldn't be easy to fall off; Under axial drawing without impact, the separation power ≤ 15 N	pass	
Puncture Force	≤ 0.85 N	0.52N	
Size	The length of connected needle should be no less than 12mm; The length of puncture end should be accord with the technique requirement.	pass	
Sterilization	The products should be sterilized by Ethylene oxide.	pass	
Ethylene oxide leftover	the leftover per pc ≤ 0.1 mg	0.014mg	
Conclusion	Pass		

Approver: Zhang Li

Inspector: Li Shanshan

