

# Certificate of Analysis

## Cathode Buffer Container 3500 Series

Product No. 4408256  
Lot No. 2501061  
Date of Manufacture 05MAR2025  
Expiration Date 29APR2026

TEST	SPECIFICATION	RESULT
<b>VISUAL INSPECTION</b>		
Clear, particulate free, viscous solution		Pass
<b>pH TEST</b> Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	7.911
<b>CONDUCTIVITY</b> Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	1772 µS/cm

***For Research Use Only.  
Not for use in diagnostic procedures.***

**Manufactured in compliance with our ISO 13485 certified quality management system.  
Site: Bedford, MA, USA**

Corporate Legal Entity  
Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

*Elsa Qirollari*  
**Quality Assurance  
Issued 06MAR2025**