

SOFIA™

Catheter Family

EMEA

Trusted Trackability.



SOFIA™ Catheter

Category	Product Name	Catalog Number	Usable Length (cm)	Distal/Proximal ID (in) (mm)		Distal OD (F) (in) (mm)			Proximal OD (F) (in) (mm)			Distal Soft Length (cm)	Tip Shape
5F	SOFIA Distal Access Catheter	DA5115ST	115	0.055	1.40	5.1	0.067	1.7	5.2	0.068	1.7	17	Straight
	SOFIA Aspiration Catheter	DA5125ST	125										
6F	SOFIA Distal Access Catheter	DA6115ST	115	0.070	1.78	6.2	0.082	2.1	6.3	0.083	2.1	19	Straight
	SOFIA Plus Aspiration Catheter	DA6125ST	125										
		DA6131ST	131										

SOFIA™ EX Catheter

Category	Product Name	Catalog Number	Usable Length (cm)	Distal/Proximal ID (in) (mm)		Distal OD (F) (in) (mm)			Proximal OD (F) (in) (mm)			Distal Soft Length (cm)	Tip Shape
5F	SOFIA EX Intracranial Support Catheter	ISC5105ST	105	0.058	1.47	5.2	0.068	1.73	5.4	0.071	1.80	9	Straight
		ISC5115ST	115										

INDICATIONS FOR USE

The SOFIA Plus and SOFIA Distal Access Catheter is indicated for general intravascular use, including the neuro and peripheral vasculature. The SOFIA Plus and SOFIA Distal Access Catheter can be used to facilitate introduction of diagnostic or therapeutic agents. The SOFIA Plus and SOFIA Distal Access Catheter is not intended for use in coronary arteries. Moreover, the SOFIA Plus and SOFIA Distal Access Catheter is intended for use in the removal/aspiration of emboli and thrombi from selected blood vessels in the arterial system, including the peripheral and neuro vasculatures.

The SOFIA EX Catheter is indicated for general intravascular use, including the neuro and peripheral vasculature. The SOFIA EX Catheter can be used to facilitate introduction of diagnostic agents or therapeutic devices. The SOFIA EX Catheter is not intended for use in coronary arteries. This device should be used only by physicians trained in percutaneous, intravascular techniques and procedures at medical facilities with the appropriate fluoroscopy equipment.

RX Only: Federal (USA) law restricts this device to sale by or on the order of a physician. This device should be used only by physicians trained in percutaneous, intravascular techniques and procedures at medical facilities with the appropriate fluoroscopy equipment.

Class III-CE0297

For healthcare professionals intended use.

MICROVENTION and SOFIA are registered trademarks of MicroVention, Inc. in the United States and other jurisdictions.
© 2022 MicroVention, Inc. MM1193(i) EMEA 02/22

MicroVention Worldwide
Innovation Center (Manufacturer) +1.714.247.8000

35 Enterprise
Aliso Viejo, CA 92656 USA

MicroVention UK Limited +44 (0) 191 258 6777

MicroVention Europe, S.A.R.L. (Legal Manufacturer) +33 (1) 39 21 77 46

MicroVention Deutschland GmbH +49 211 210 798-0

Website microvention.com

