Page: 1



CERTIFICATE OF ANALYSIS

Honeywell Specialty Chemicals Seelze GmbH Wunstorferstrasse 40 Seelze, 30926 Telefon: +49 5137 999-0 <u>lab.honeywell.com</u>

Brand: Product: Lot No: Production date: Expiry Date: Honeywell Fluka 35256 N3490 15.Dec.2023 04.Dec.2025

Sodium hydroxide solution 1 mol/l, volumetric solution, Reag. Ph. Eur.

Parameter	Specification Units	Result Units
factor	0.999-1.001	1.001
BAM/EMPA refmaterial	effektiv	60357
BAM/EMPA Certificate no.	effektiv	U2-0823
Traceable to NIST SRM	effektiv	84L
Produced using quelity grade "purice, p.e." tested row metericle		

Produced using quality grade "puriss. p.a." tested raw materials.

Test method: Alkalimetric titration related to BAM/EMPA reference material Potassium hydrogen phthalate

The factor (titer) is measured at 20°C.

BAM: Federal Institute for Materials Research and Testing Germany EMPA: Swiss Federal Laboratories for Materials Testing and Research NIST: National Institute of Standards and Technology USA

QC Release Date: 04.Jan.2024

The minimum shelf life is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.

We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.

Particular properties of the products or the suitability for a particular area of application are not assured.

We guarantee a proper quality within our General Conditions of Sales.

Page: 2



Quality Management

This document was produced electronically and is valid without a signature