



Oxidase Test Disc

Rapid test for detection of cytochrome oxidase enzymatic activity.

INTENDED PURPOSE

Rapid test for detection of cytochrome oxidase enzyme on bacterial isolates. This test is intended as an aid in the diagnosis, requiring further tests to complete the diagnostic results.

DESCRIPTION

Oxidase Test Disc are filter paper discs impregnated with a reagent for differentiation and microbial identification, particularly of Gram-negative bacteria, based on the presence of enzyme cytochrome oxidase. This enzyme is characteristically produced by a number of microorganisms, including *Neisseria* and *Pseudomonas* species while *Enterobacteriaceae* are oxidase negative (with the exception of *Plesiomonas*, which is the only oxidase-positive genus in this family).

KIT CONTENT

- 1 cartridge of 30 discs.

METHOD PRINCIPLE

Oxidase-positive bacteria produces the enzyme cytochrome oxidase (indophenol oxidase) that catalyzes the transport of electrons from donor compounds (NADH) to electron acceptors (usually oxygen). Tetramethyl-p-phenylenediamine dihydrochloride contained in Oxidase Test Disc acts as an artificial electron donor and is oxidized by oxidase-positive bacteria forming the coloured compound indophenol blue.

MATERIALS REQUIRED BUT NOT PROVIDED

Standard microbiological supplies and equipment such as: inoculating loop, pipettes, culture media, quality control organisms.

REAGENTS

- Each disc is impregnated with a solution of N,N,N',N'-tetramethyl-p-phenylenediamine dihydrochloride.

SPECIMEN

Collect specimens in sterile containers or with sterile swabs and transport to the laboratory. Process each specimen using procedures appropriate for that sample. This product is recommended for use only with pure cultures.

Refer to specific guidelines for more detailed information.

TEST PROCEDURE

1. Allow product to reach room temperature before use, for minimizing condensation on the disc.
2. Pick up one or more than one well isolated colony from a fresh culture plate (18-24 hours old) and smear on the disc. Alternatively, make a heavy suspension of the test organism in saline and place an oxidase disc in the tube containing the microbial suspension.
3. Observe for color development within 60 seconds.

Note: Positive and negative controls should be run simultaneously with the organism to be tested (see QUALITY CONTROL).

INTERPRETING RESULTS

A positive oxidase reaction is indicated by development of a blue-purple color.

No color change corresponds to a negative test, i.e. the organism under investigation does not produce the enzyme cytochrome oxidase.

STORAGE

Store at 2-8°C away from light. Do not use the product beyond its expiry date on the label or if product shows any evidence of contamination or any sign of deterioration.

SHELF LIFE

1 year.

QUALITY CONTROL

Control strain		Characteristic reactions
<i>Escherichia coli</i>	WDCM 00013 ATCC® 25922 NCTC 12241	Negative reaction: No color change
<i>Cronobacter sakazakii</i>	WDCM 00214 ATCC® 29544 NCTC 11467	
<i>Cronobacter muytjensii</i>	WDCM 00213 ATCC® 51329	
<i>Pseudomonas aeruginosa</i>	WDCM 00025 ATCC® 27853 NCTC 12903	Positive reaction: Colour change to blue-purple
<i>Pseudomonas fluorescens</i>	WDCM 00115 ATCC® 13525 NCTC 10038	

PERFORMANCE CHARACTERISTICS

Performance testing of Oxidase Test Disc was carried out using the QC strains listed above. The results obtained met the established criteria.

LIMITATIONS

The most suitable cultures for the oxidase test are those from culture media without dyes, indicators or inhibitors.

Bacterial colonies taken from media with pH values below 5.5 (e.g. after the metabolism of carbohydrates with subsequent acidification of the culture medium) can give a false negative oxidase reaction.

Colonies taken from media containing nitrate may give unreliable results.

Use of nickel-chrome loops may produce false positive results from oxidation because of the iron contained in the loops. A platinum or plastic loop, or wooden applicator stick is recommended.

The colonies taken from media containing nitrates may give unreliable results.

Oxidase-cytochrome production might be inhibited by the generation of acids and falsely negative reactions can be obtained by bacteria cultured on media containing fermentable carbohydrates such as MacConkey, TCBS Agar etc.

WARNING AND PRECAUTIONS

- 1) **For *in vitro* diagnostic use (IVD).**
- 2) **For laboratory professional use only.**
- 3) Operators must be trained and have certain experience. Please read the instructions carefully before using the product. Reliability of assay results cannot be guaranteed if there are any deviations from the instructions in this document.
- 4) Consult the Safety Data Sheet (SDS) for information regarding hazards and safe handling practices.
- 5) Do not use if the product or packaging appears to be damaged.
- 6) Follow standard precautions. All patient specimens should be considered potentially infectious and handled accordingly.
- 7) Handle all specimens as if infectious using safe laboratory procedures. Dispose of hazardous or biologically contaminated materials according to the practices of your institution.
- 8) Avoid cross-contamination of samples by using disposable tips and changing them after each sample.
- 9) Do not mix reagents of different batches. Please use the product within the validity period.
- 10) Do not eat, drink, smoke, apply cosmetics or handle contact lenses in areas where reagents and human specimens are handled.
- 11) Results should be interpreted by a trained professional in conjunction with the patient's history and clinical signs and symptoms, and epidemiological risk factors.
- 12) Ensure laboratory equipment is calibrated and maintained in accordance with the laboratory's procedure.
- 13) When test results are transmitted from the laboratory to an informatics centre, attention has to be done to avoid erroneous data transfer.

DISPOSAL OF WASTE

Disposal of waste must be carried out according to national and local regulations in force.

BIBLIOGRAPHY

See the references at the end of this document.

TABLE OF SYMBOLS

See the table of symbols at the end of this document.

ORDER INFORMATION

Product	Packaging	Ref.
Oxidase Test Disc	30 tests	88004

Revision History

Revision	Release Date	Change Summary
4	2024-09-16	Added notice to report any malfunction, defect or incident.
3	2023-05-15	Updated layout and content in compliance with IVDR 2017/746.

In case of malfunctions or defects, contact immediately Liofilchem (*) or the local representative.

In case of incident associated with the device, notify immediately Liofilchem (*) or its local representative and the National Competent Authority.

*Please login to <https://www.liofilchemstore.it/login.php> (user ID and password required) and click on Complaint.

This IFU document and the SDS are available from the online Support Center:

[liofilchem.com/ifu-sds](https://www.liofilchem.com/ifu-sds)



Oxidase Test Disc

Test rapido per la rilevazione dell'attività enzimatica della citocromo ossidasi.

DESTINAZIONE D'USO

Test rapido per la ricerca dell'enzima citocromo ossidasi da isolati batterici. Il test è inteso come ausilio alla diagnosi, e sono necessari ulteriori test per completare i risultati diagnostici.

DESCRIZIONE

Oxidase Test Disc sono dischi di carta da filtro impregnati di un reagente per la differenziazione e l'identificazione microbica, in particolare di batteri Gram-negativi, basata sulla presenza dell'enzima citocromo ossidasi. Questo enzima è tipicamente prodotto da un certo numero di microrganismi, tra cui le specie *Neisseria* e *Pseudomonas*, mentre le *Enterobacteriaceae* sono ossidasi negative (con l'eccezione di *Plesiomonas*, che è l'unico genere ossidasi positivo in questa famiglia).

CONTENUTO DEL KIT

- 1 cartuccia da 30 dischi

PRINCIPIO DEL METODO

I batteri ossidasi positivi producono l'enzima citocromo ossidasi (indofenolo ossidasi) che catalizza il trasporto degli elettroni da un composto donatore (NADH) ad uno accettore (di solito l'ossigeno).

Il tetrametil-p-fenilenediammina dicloridrato contenuto in Oxidase Test Disc agisce come un donatore artificiale di elettroni e viene ossidato dai batteri ossidasi positivi formando il composto colorato indofenolo blu.

MATERIALI RICHIESTI MA NON FORNITI

Forniture e apparecchiature microbiologiche standard come: ansa da inoculo, pipette, provette, microrganismi per il controllo qualità.

REAGENTI

- Ciascun disco è impregnato con una soluzione di N,N,N',N'-tetrametil-p-fenilendiammina dicloridrato.

CAMPIONI

Raccogliere i campioni prelevati in contenitori sterili o con tamponi sterili e trasportarli in laboratorio. Trattare i campioni secondo la procedura appropriata per ognuno di loro. Questo prodotto è consigliato per l'uso esclusivamente con colture pure.

Fare riferimento alle linee guida specifiche per informazioni più dettagliate.

PROCEDURA DEL TEST

1. Lasciare che il prodotto raggiunga la temperatura ambiente prima dell'uso, per ridurre al minimo la formazione di condensa sul disco.
2. Prelevare una o più colonie ben isolate da una piastra di coltura fresca (di 18-24 ore) e spalmare sul disco. In alternativa, preparare una sospensione pesante del microrganismo in esame in soluzione fisiologica e inserire un disco di ossidasi nella provetta contenente la sospensione microbica.
3. Osservare lo sviluppo del colore entro 60 secondi.

Nota: i controlli positivi e negativi devono essere eseguiti simultaneamente con l'organismo da testare (vedere CONTROLLO DI QUALITÀ).

INTERPRETAZIONE DEI RISULTATI

Un risultato positivo per il Test Oxidase è indicato dallo sviluppo di un colore blu-viola. Nessun sviluppo di colore corrisponde ad un test negativo, ciò significa che il microrganismo esaminato non produce l'enzima citocromo ossidasi.

CONSERVAZIONE

Conservare a 2-8°C al riparo dalla luce. Non usare il prodotto dopo la sua data di scadenza indicata sull'etichetta o se il prodotto mostra segni di contaminazione o deterioramento.

VALIDITÀ

1 anno.

CONTROLLO QUALITÀ

Control strain		Characteristic reactions
<i>Escherichia coli</i>	WDCM 00013 ATCC® 25922 NCTC 12241	Reazione negativa: Nessun cambio di colore
<i>Cronobacter sakazakii</i>	WDCM 00214 ATCC® 29544 NCTC 11467	
<i>Cronobacter muytjensii</i>	WDCM 00213 ATCC® 51329	
<i>Pseudomonas aeruginosa</i>	WDCM 00025 ATCC® 27853 NCTC 12903	Reazione positiva: Colore cambia al blu viola
<i>Pseudomonas fluorescens</i>	WDCM 00115 ATCC® 13525 NCTC 10038	

CARATTERISTICHE PRESTAZIONALI

Il controllo delle prestazioni dell'Oxidase Test Disc è stato eseguito utilizzando i ceppi sopra elencati. I risultati ottenuti hanno soddisfatto i criteri stabiliti.

LIMITAZIONI

Le colture più idonee per il test dell'ossidasi sono quelle provenienti da terreni di coltura senza coloranti, indicatori o inibitori.

Le colonie batteriche prelevate da terreni con valori di pH inferiori a 5,5 (ad es. dopo il metabolismo dei carboidrati con successiva acidificazione del terreno di coltura) possono dare una reazione di ossidasi falsamente negativa.

Le colonie prelevate da terreni contenenti nitrati possono dare risultati inaffidabili.

L'uso di anse in nichel-cromo può produrre, a causa dell'ossidazione del ferro contenuto nelle anse, risultati falsi positivi. Si consiglia un anello di platino o di plastica o un bastoncino applicatore in legno.

Le colonie prelevate da terreni contenenti nitrati possono dare risultati inaffidabili.

La produzione della citocromo-ossidasi potrebbe essere inibita dalla generazione di acidi; reazioni falsamente negative possono essere ottenute da batteri coltivati su terreni contenenti carboidrati fermentabili come MacConkey, TCBS Agar ecc.

AVVERTENZE E PRECAUZIONI

- 1) **Per uso diagnostico in vitro (IVD).**
- 2) **Solo per uso professionale di laboratorio.**
- 3) Gli operatori devono essere formati e avere una certa esperienza. Si prega di leggere attentamente le istruzioni prima di utilizzare il prodotto. L'affidabilità dei risultati del test non può essere garantita in caso di deviazioni dalle istruzioni contenute in questo documento.
- 4) Consultare la scheda di sicurezza (SDS) per informazioni sui pericoli e sulle pratiche di manipolazione sicure.
- 5) Non utilizzare se il prodotto o la confezione sembrano danneggiati.
- 6) Seguire le precauzioni standard. Tutti i campioni dei pazienti devono essere considerati potenzialmente infetti e maneggiati di conseguenza.
- 7) Maneggiare tutti i campioni come infetti utilizzando procedure di laboratorio sicure. Smaltire materiali pericolosi o biologicamente contaminati secondo le pratiche del proprio istituto.
- 8) Evitare la contaminazione incrociata dei campioni utilizzando puntali monouso e sostituendole dopo ogni campione.
- 9) Non mescolare reagenti di lotti diversi. Si prega di utilizzare il prodotto entro il periodo di validità.
- 10) Non mangiare, bere, fumare, applicare cosmetici o maneggiare lenti a contatto nelle aree in cui vengono manipolati reagenti e campioni umani.

- 11) I risultati devono essere interpretati da un professionista qualificato insieme alla storia del paziente, ai segni e sintomi clinici e ai fattori di rischio epidemiologici.
- 12) Assicurarsi che le apparecchiature di laboratorio siano calibrate e mantenute in conformità con la procedura del laboratorio.
- 13) Quando i risultati dei test vengono trasmessi dal laboratorio a un centro informatico, è necessario prestare attenzione per evitare trasferimenti di dati errati.

SMALTIMENTO DEI RIFIUTI

Lo smaltimento dei rifiuti deve essere effettuato in conformità alle normative nazionali e locali in vigore.

INFORMAZIONI PER L'ORDINE

Prodotto	Confezionamento	Rif.
Oxidase Test Disc	30 test	88004

In caso di malfunzionamenti o difetti, contattare immediatamente Liofilchem (*) o il rappresentante locale.

In caso di incidente associato al dispositivo, avvisare immediatamente Liofilchem (*) o il suo rappresentante locale e l'Autorità Nazionale Competente.

*Si prega di effettuare il login su <https://www.liofilchemstore.it/login.php> (user ID e password richiesti) e cliccare su "Complaint".

Questo documento IFU e la SDS sono disponibili dal Support Center online:

[liofilchem.com/ifu-sds](https://www.liofilchem.com/ifu-sds)

References / Riferimenti

1. EN ISO 11133:2014+Amd1:2018+Amd2:2020. Microbiology of food, animal feed and water – Preparation, production, storage and performance testing of culture media.
2. Public Health England. (2019). Oxidase test. UK Standards for Microbiology Investigations. TP 26 Issue 4. <https://www.gov.uk/guidance/uk-standards-for-microbiology-investigations-smi-quality-and-consistency-in-clinical-laboratories>
3. Public Health England. (2015). Identification of Enterobacteriaceae. UK Standards for Microbiology Investigations. ID 16 Issue 4.
4. ISO 9308-1:2014. Water quality – Enumeration of *Escherichia coli* and coliform bacteria – Part 1: Membrane filtration method for waters with low bacterial background flora.
5. ISO 16266:2006. Water quality – Detection and Enumeration of *Pseudomonas aeruginosa* – Method by membrane filtration.
6. Steel KJ (1962) J. Appl. Bact. 25:445-447.

Table of Symbols / Tabella dei Simboli

	Batch code / Codice del lotto
	Catalogue number / Numero di catalogo
	<i>In Vitro</i> Diagnostic Medical Device / Dispositivo Medico Diagnostico <i>in vitro</i>
	Manufacturer / Fabbrikante
	Use by / Utilizzare entro
	Fragile, handle with care / Fragile, maneggiare con cura
	Temperature limitation / Limiti di temperatura
	Contains sufficient for <n> tests / Contenuto sufficiente per <n> saggi
	Consult instructions for use / Consultare le istruzioni per l'uso
	Do not reuse / Non riutilizzare
	Keep away from sunlight / Tenere al riparo dalla luce solare



Liofilchem® s.r.l.

Via Scozia, 64026 Roseto degli Abruzzi (TE) Italy

Tel. +39 0858930745 Fax +39 0858930330 www.liofilchem.com liofilchem@liofilchem.com

