



Add: No.9 Bofu Road, Luhe District, Nanjing, 211505, China
Tel: 86-25-68568508 Email: overseas@geteincom.cn Web: www.bio-GP.com.cn

Document No.: GP-GMSQ-2024121101

Letter of Authorization

To whom it may concern,

We, Getein Biotech, Inc. (No.9 BoFu Road, Luhe District, Nanjing, 211505, China), hereby authorize Sanmedico SRL (Address: Republic of Moldova, Chisinau, MD-2059, Petricani street, 88/1, office 10) as our official and non-exclusive distributor for registration, promoting, selling, distributing and providing after-sale services of under-mentioned product in the territory of Moldova only:

FIA8000 Quantitative Immunoassay Analyzer and Reagents

Getein1100 Immunofluorescence Quantitative Analyzer and Reagents

Getein1160 Immunofluorescence Quantitative Analyzer and Reagents

Getein 1600 Immunofluorescence Quantitative Analyzer and Reagents

Sanmedico SRL will comply with the laws and regulations of the countries and regions where they are located in and where they are selling mentioned product.

Sanmedico SRL will carry out marketing efforts to fulfill service and maintenance for above mentioned products and will provide with users benefits of having a local stock of above mentioned products and on time delivery with every order, supported by a local service in local language.

This authorization starts from Jan 1, 2025 and will be valid to December 31 2025.

Getein Biotech, Inc. has the right to terminate the authorization before validity and will inform **Sanmedico SRL** with 10 days in advance.

Getein Biotech, Inc.
基蛋生物科技股份有限公司
Name: Steven Zhou
Position: Overseas Sales Director
GETEIN BIOTECH, INC.

A handwritten signature in black ink, appearing to read 'Steven Zhou'.

Certificate of Registration

QUALITY MANAGEMENT SYSTEM - ISO 13485:2016

This is to certify that: Getein Biotech, Inc.
No.9 Bofu Road
Luhe District
Nanjing
Jiangsu
211505
China

基蛋生物科技股份有限公司
中国
江苏省
南京市
六合区
沿江工业开发区
博富路9号
邮编: 211505

Holds Certificate No: **MD 728432**

and operates a Quality Management System which complies with the requirements of ISO 13485:2016 for the following scope:

Please see scope page.

For and on behalf of BSI:



Graeme Tunbridge, Senior Vice President Medical Devices

Original Registration Date: 2020-05-29

Effective Date: 2023-07-26

Latest Revision Date: 2023-04-26

Expiry Date: 2026-07-25



Page: 1 of 3

...making excellence a habit.™

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract.
An electronic certificate can be authenticated [online](#).
Printed copies can be validated at www.bsi-global.com/ClientDirectory or telephone +86 10 8507 3000.

Information and Contact: BSI, John M. Keynesplein 9, 1066 EP Amsterdam The Netherlands. Tel: +31 (0) 20 3460 780
BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A Member of the BSI Group of Companies.

Certificate No: **MD 728432**

Registered Scope:

Design & Development, Manufacture and Distribution of Chemiluminescence Immunoassay, Biochemistry Assay, Point of Care Assay (including Colloidal Gold Assay, Immunofluorescence Assay, Dry Chemistry Assay), PCR Assay and Colloidal Gold self-testing Assay to detect infectious disease. Design & Development, Manufacture and Distribution of Analyzers in use of Chemiluminescence Immunoassay, Biochemistry Assay, Point of Care Assay (including Colloidal Gold Assay, Immunofluorescence Assay, Dry Chemistry Assay), PCR Assay to detect infectious disease, Immunofluorescence self-testing Assay to detect dyslipidemia disease, Blood Coagulation Assay to detect thrombotic disease.

研发,生产和销售化学发光法试剂,生化试剂,即时诊断(包括胶体金法,免疫荧光法,干式化学法)试剂,传染病相关PCR分子诊断试剂和胶体金自测试剂。研发,生产和销售用于化学发光法试剂,生化试剂,即时诊断(包括胶体金法,免疫荧光法,干式化学法)试剂,传染病相关PCR分子诊断试剂,血脂异常疾病相关免疫荧光自测试剂,血栓疾病相关血凝试剂配套使用的分析仪。



Original Registration Date: 2020-05-29

Effective Date: 2023-07-26

Latest Revision Date: 2023-04-26

Expiry Date: 2026-07-25

Page: 2 of 3

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract.

An electronic certificate can be authenticated [online](#).

Printed copies can be validated at www.bsi-global.com/ClientDirectory or telephone +86 10 8507 3000.

Information and Contact: BSI, John M. Keynesplein 9, 1066 EP Amsterdam The Netherlands. Tel: +31 (0) 20 3460 780
BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A Member of the BSI Group of Companies.

Certificate No: **MD 728432**

Location

Registered Activities

Getein Biotech, Inc. No.9 Bofu Road Luhe District Nanjing Jiangsu 211505 China 基蛋生物科技股份有限公司 中国 江苏省 南京市 六合区 沿江工业开发区 博富路9号 邮编: 211505	Design & Development, Manufacture and Distribution of Chemiluminescence Immunoassay, Biochemistry Assay, Point of Care Assay (including Colloidal Gold Assay, Immunofluorescence Assay, Dry Chemistry Assay), PCR Assay and Colloidal Gold self-testing Assay to detect infectious disease. Design & Development, Manufacture and Distribution of Analyzers in use of Chemiluminescence Immunoassay, Biochemistry Assay, Point of Care Assay (including Colloidal Gold Assay, Immunofluorescence Assay, Dry Chemistry Assay), PCR Assay to detect infectious disease, Immunofluorescence self-testing Assay to detect dyslipidemia disease, Blood Coagulation Assay to detect thrombotic disease. 研发, 生产和销售化学发光法试剂, 生化试剂, 即时诊断(包括胶体金法, 免疫荧光法, 干式化学法)试剂, 传染病相关PCR分子诊断试剂和胶体金自测试剂。研发, 生产和销售用于化学发光法试剂, 生化试剂, 即时诊断(包括胶体金法, 免疫荧光法, 干式化学法)试剂, 传染病相关PCR分子诊断试剂, 血脂异常疾病相关免疫荧光自测试剂, 血栓疾病相关血凝试剂配套使用的分析仪。
Getein Biotech, Inc. No. 6 KeFeng Road Jiangbei New District Nanjing Jiangsu 211505 China 基蛋生物科技股份有限公司 中国 江苏省 南京 江北新区 科丰路6号 邮编: 211505	Manufacture of Chemiluminescence Immunoassay, Biochemistry Assay, Point of Care Assay (including Colloidal Gold Assay, Immunofluorescence Assay, Dry Chemistry Assay), Colloidal Gold self-testing Assay to detect infectious disease. Manufacture of Analyzers in use of Chemiluminescence Immunoassay, Biochemistry Assay, Point of Care Assay (including Colloidal Gold Assay, Immunofluorescence Assay, Dry Chemistry Assay), PCR Assay to detect infectious disease, Immunofluorescence self-testing Assay to detect dyslipidemia disease, Blood Coagulation Assay to detect thrombotic disease. 生产化学发光法试剂, 生化试剂, 即时诊断(包括胶体金法, 免疫荧光法, 干式化学法)试剂和传染病相关胶体金自测试剂。生产用于化学发光法试剂, 生化试剂, 即时诊断(包括胶体金法, 免疫荧光法, 干式化学法)试剂, 传染病相关PCR分子诊断试剂, 血脂异常疾病相关免疫荧光自测试剂, 血栓疾病相关血凝试剂配套使用的分析仪。

Original Registration Date: 2020-05-29

Effective Date: 2023-07-26

Latest Revision Date: 2023-04-26

Expiry Date: 2026-07-25

Page: 3 of 3

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract.

An electronic certificate can be authenticated [online](#).

Printed copies can be validated at www.bsi-global.com/ClientDirectory or telephone +86 10 8507 3000.

Information and Contact: BSI, John M. Keynesplein 9, 1066 EP Amsterdam The Netherlands. Tel: +31 (0) 20 3460 780

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands

A Member of the BSI Group of Companies.

EC Declaration of Conformity

according to Directive 98/79/EC, on in vitro diagnostic medical devices

Ref. No.:20220513-A05

Manufacturer
(Name, Address)

Getein Biotech, Inc.
No. 9 Bofu Road, Luhe District, Nanjing, 211505, China

**Authorized
Representative**
(Name, Address)

CMC Medical Devices & Drugs S.L.
Add: C/ Horacio Lengo Nº 18, CP 29006, Málaga, Spain

Medical device

No.	Product Name
1	Getein 1100 Immunofluorescence Quantitative Analyzer
2	Cardiac Troponin I Fast Test Kit (Immunofluorescence Assay)
3	NT-proBNP Fast Test Kit (Immunofluorescence Assay)
4	hs-CRP+CRP Fast Test Kit (Immunofluorescence Assay)
5	NT-proBNP/cTnI Fast Test Kit (Immunofluorescence Assay)
6	CK-MB/cTnI/Myo Fast Test Kit (Immunofluorescence Assay)
7	D-Dimer Fast Test Kit (Immunofluorescence Assay)
8	PCT Fast Test Kit (Immunofluorescence Assay)
9	CysC Fast Test Kit (Immunofluorescence Assay)
10	mAlb Fast Test Kit (Immunofluorescence Assay)
11	NGAL Fast Test Kit (Immunofluorescence Assay)
12	β 2-MG Fast Test Kit (Immunofluorescence Assay)
13	CK-MB/cTnI Fast Test Kit (Immunofluorescence Assay)
14	HCG+ β Fast Test Kit (Immunofluorescence Assay)
15	H-FABP Fast Test Kit (Immunofluorescence Assay)
16	PCT/CRP Fast Test Kit (Immunofluorescence Assay)
17	CK-MB/cTnI/H-FABP Fast Test Kit (Immunofluorescence Assay)
18	HbA1c Fast Test Kit (Immunofluorescence Assay)
19	NT-proBNP/NGAL Fast Test Kit (Immunofluorescence Assay)
20	CK-MB Fast Test Kit (Immunofluorescence Assay)
21	hs-cTnI Fast Test Kit (Immunofluorescence Assay)
22	T3 Fast Test Kit (Immunofluorescence Assay)
23	T4 Fast Test Kit (Immunofluorescence Assay)
24	TSH Fast Test Kit (Immunofluorescence Assay)
25	Scr Fast Test Kit (Immunofluorescence Assay)
26	PLGF Fast Test Kit (Immunofluorescence Assay)

27	HCY Fast Test Kit (Immunofluorescence Assay)
28	Anti-CCP Fast Test Kit (Immunofluorescence Assay)
29	25-OH-VD Fast Test Kit (Immunofluorescence Assay)
30	Lp-PLA2 Fast Test Kit (Immunofluorescence Assay)
31	FOB Fast Test Kit (Immunofluorescence Assay)
32	SAA Fast Test Kit (Immunofluorescence Assay)
33	H. pylori Fast Test Kit (Immunofluorescence Assay)
34	PRL Fast Test Kit (Immunofluorescence Assay)
35	Transferrin Fast Test Kit (Immunofluorescence Assay)
36	Insulin Fast Test Kit (Immunofluorescence Assay)
37	PG I /PG II Fast Test Kit (Immunofluorescence Assay)
38	LH Fast Test Kit (Immunofluorescence Assay)
39	FSH Fast Test Kit (Immunofluorescence Assay)
40	Anti-TP Fast Test Kit (Immunofluorescence Assay)
41	AFP/CEA Fast Test Kit (Immunofluorescence Assay)
42	AMH Fast Test Kit (Immunofluorescence Assay)
43	fT3 Fast Test Kit (Immunofluorescence Assay)
44	fT4 Fast Test Kit (Immunofluorescence Assay)
45	Total IgE Fast Test Kit (Immunofluorescence Assay)
46	Vit-B12 Fast Test Kit (Immunofluorescence Assay)
47	Prog Fast Test Kit (Immunofluorescence Assay)
48	Testosterone Fast Test Kit (Immunofluorescence Assay)
49	E2 Fast Test Kit (Immunofluorescence Assay)
50	RF Fast Test Kit (Immunofluorescence Assay)
51	ASO Fast Test Kit (Immunofluorescence Assay)
52	Ferritin Fast Test Kit (Immunofluorescence Assay)
53	ST2 Fast Test Kit (Immunofluorescence Assay)
54	CA125 Fast Test Kit (Immunofluorescence Assay)
55	CA19-9 Fast Test Kit (Immunofluorescence Assay)
56	CA15-3 Fast Test Kit (Immunofluorescence Assay)
57	RSV/Influenza A/B Fast Test Kit (Immunofluorescence Assay)
58	Influenza A/B Fast Test Kit (Immunofluorescence Assay)
59	RSV Fast Test Kit (Immunofluorescence Assay)
60	IL-6 Fast Test Kit (Immunofluorescence Assay)
61	BNP Fast Test Kit (Immunofluorescence Assay)
62	SAA/CRP Fast Test Kit (Immunofluorescence Assay)
63	Folate acid Fast Test Kit (Immunofluorescence Assay)
64	hs-CRP Fast Test Kit (Immunofluorescence Assay)
65	TnT Fast Test Kit (Immunofluorescence Assay)
66	PCT/IL-6 Fast Test Kit (Immunofluorescence Assay)

67	HBP Fast Test Kit (Immunofluorescence Assay)
68	S100- β Fast Test Kit (Immunofluorescence Assay)
69	CK-MB/hs-cTnI/Myo Fast Test Kit (Immunofluorescence Assay)
70	Cortisol Fast Test Kit (Immunofluorescence Assay)
71	CEA Fast Test Kit (Immunofluorescence Assay)
72	AFP/CEA Fast Test Kit (Immunofluorescence Assay)

Classification Other device (according to Annex II of the directive 98/79/EC)

Conformity assessment route Annex III of the 98/79/EC

Applicable coordination standards	EN 13612:2002	EN ISO 14971:2019	EN ISO15223-1:2016
	EN ISO 18113-1:2011	EN ISO 18113-2:2011	EN ISO 18113-3:2011
	EN ISO 23640:2015	EN ISO 13485:2016	ISO 780:2015
	EN 61326-2-6:2006	IEC 61326-1:2013	
	EN 61010-2-101:2002	IEC 61010-1:2010	

Signatory representative declares herein the above-mentioned device meets the basic requirements of the European Parliament and the Council's in vitro diagnostic medical devices directive: 98/79/EC Annex I.

This declaration of conformity is based on European Parliament and the Council's 98/79/EC directive Annex III. The compiled technical file and quality system document according to 98/79/EC directive Annex III are testified and the quality system certificate has issued by BSI Group The Netherlands B. V. The manufacturer is exclusively responsible for the declaration of conformity.

General Manager Enben Su

Nanjing
13th, May, 2022

(place and date of issue)

(name and signature or equivalent
marking of authorized person)





Stock Code: 603387

OPTIMIZED POINT-OF-CARE SOLUTION

MAKING TEST EASY

Getein 1100

Immunofluorescence Quantitative Analyzer



Getein 1100

Immunofluorescence Quantitative Analyzer



HIGHLY EFFICIENT & ACCURATE

Advanced fluorescence immunoassay

Multiple quality control



REAL-TIME AND RAPID TEST


One-step test

3-15 min/test


5 sec/test for multiple tests

OPERATION MODES

Inside Mode (single sample rapid test mode)



Quick mode (mass samples rapid test mode)





CONVENIENT OPERATION

RFID card calibration

Keyboard and mouse connectivity through USB port

Handwriting input available

Continuous test for 3 hours with optional lithium battery



USER-FRIENDLY INTERFACE

Android system

7-inch touch screen



① 7-inch Touch Screen

② SD Card Recognition Zone

③ Test Card Slot

④ SD Card Slot

⑤ USB Slot

⑥ Built-in Thermal Printer



PORTABLE DESIGN

Small in size: 261 × 241 × 115 mm

Light in weight: 2.0 kg



LARGE MEMORY

Up to 10,000 results storage capacity

TECHNICAL PARAMETERS

Methodology	Immunofluorescence	Screen	7-inch touch screen
Result	Quantitative	Power Supply	100-240 V~50 Hz/60 Hz, 60 VA
Sample Type	WB, Plasma, Serum, Urine, Stool, Nasal swab, Saliva, Capillary blood	Working Environment	Temperature: 10-35°C Relative humidity ≤70% Air pressure 70.0~106.0 kpa
Storage Capacity	10000 data	Dimensions	261 mm×241 mm×115 mm (D×W×H)
Language	English/Chinese/Spanish/Portuguese	Weight	2.0 kg

TEST ITEMS

Cat.#	TEST ITEMS	DISEASES	CUT-OFF VALUE	SAMPLE TYPES	MEASURING RANGE	SAMPLE VOLUME	REACTION TIME	QUALIFICATION
Cardiac Markers								
IF1001	cTnI	Myocardial infarction	0.10 ng/mL	S/P/WB	0.10-50.00 ng/mL	100 µL	10 min	NMPA CE
NEW IF1098	TnT	Myocardial infarction	14.0 pg/mL	S/P/WB	10.0-10000.0 pg/mL	100 µL	15 min	NMPA CE
IF1089	BNP	Heart failure	100.0 pg/mL	P/WB	5.0-5000.0 pg/mL	100 µL	10 min	NMPA CE
IF1002	NT-proBNP	Heart failure	300 pg/mL	S/P/WB	100-35000 pg/mL	100 µL	10 min	NMPA CE
IF1005	CK-MB/cTnI/Myo	Myocardial damage /infarction	CK-MB: 5.00 ng/mL cTnI: 0.10 ng/mL Myo: 70.0 ng/mL	S/P/WB	2.50-80.00 ng/mL 0.10-50.00 ng/mL 30.0-600.0 ng/mL	100 µL	10 min	NMPA CE
IF1012	CK-MB/cTnI	Myocardial damage /infarction	CK-MB: 5.00 ng/mL cTnI: 0.10 ng/mL	S/P/WB	2.50-80.00 ng/mL 0.10-50.00 ng/mL	100 µL	10 min	CE
IF1014	H-FABP	Myocardial damage	6.36 ng/mL	S/P/WB	1.00-120.00 ng/mL	100 µL	3 min	NMPA CE
IF1016	CK-MB/cTnI/H-FABP	Myocardial damage /infarction	CK-MB: 5.00 ng/mL cTnI: 0.10 ng/mL H-FABP: 6.36 ng/mL	S/P/WB	2.50-80.00 ng/mL 0.10-50.00 ng/mL 2.00-100.00 ng/mL	100 µL	10 min	NMPA CE
IF1018	CK-MB	Myocardial injury	5.00 ng/mL	S/P/WB	2.50-80.00 ng/mL	100 µL	10 min	CE
Coagulation Marker								
IF1006	D-Dimer	Venous thromboembolism	0.50 mg/L	P/WB	0.10-10.00 mg/L	100 µL	10 min	NMPA CE
Inflammation								
IF1003	hs-CRP+CRP	Cardiovascular inflammation /normal inflammation	3.0 mg/L 10.0 mg/L	S/P/WB/ Fingertip blood	0.5-200.0 mg/L	10 µL	3 min	NMPA CE
IF1007	PCT	Sepsis, bacterial infection	0.10 ng/mL	S/P/WB	0.05-50.00 ng/mL	100 µL	15 min	NMPA CE
IF1015	PCT/CRP	Sepsis, bacterial infection	PCT: 0.10 ng/mL CRP: 3.0 mg/L	S/P/WB	0.10-50.00 ng/mL 0.5-200.0 mg/L	100 µL	15 min	NMPA CE
IF1044	SAA	Bacterial/Virus infection	10.0 mg/L	S/P/WB/ Fingertip blood	5.0-200.0 mg/L	10 µL	5 min	NMPA CE
IF1090	SAA/CRP	Neonatal sepsis, Bacterial/virus infection	SAA: 10.0 mg/L CRP: 10.0 mg/L	S/P/WB/ Peripheral blood	5.0-200.0 mg/L 0.5-200.0 mg/L	10 µL	5 min	NMPA CE
IF1088	IL-6	Acute inflammation	7.0 pg/mL	S/P/WB/ Peripheral blood	1.5-4000.0 pg/mL	100 µL	15 min	NMPA CE
Renal Function								
IF1008	CysC	Acute and chronic renal diseases	0.51-1.09 mg/L	S/P/WB	0.50-10.00 mg/L	10 µL	3 min	NMPA CE
IF1009	mAlb	Diabetic nephropathy, hypertensive nephropathy	20.0 mg/L	Urine	10.0-200.0 mg/L	100 µL	3 min	NMPA CE
IF1010	NGAL	Acute kidney injury	Serum: 200.0 ng/mL Urine: 100.0 ng/mL	S/Urine	50.0-5000.0 ng/mL	10 µL	10 min	NMPA CE
IF1011	β ₂ -MG	Acute and chronic kidney diseases/tumours	0.80-3.00 mg/L	S/P/WB	0.50-20.00 mg/L	10 µL	3 min	NMPA CE
Diabetes Mellitus								
IF1017	HbA1c	Diabetes mellitus	3.80%-5.80%	WB	2.00%-14.00%	10 µL	5 min	NGSP IFCC NMPA CE
Metabolic Marker								
IF1031	25-OH-VD	Osteomalacia, osteoporosis	30.00-50.00 ng/mL	S/P	8.00-70.00 ng/mL	40 µL	15 min	NMPA CE
Thyroid Function								
IF1024	TSH	Thyroid malfunction	0.27-4.20 µIU/mL	S/P	0.10-50.00 µIU/mL	100 µL	15 min	NMPA CE
IF1022	T3	Hyperthyroidism, hypothyroidism	1.30-3.10 nmol/L	S/P	0.30-10.00 nmol/L	40 µL	15 min	NMPA CE
IF1023	T4	Hyperthyroidism, hypothyroidism	59.00-154.00 nmol/L	S/P	5.40-320.00 nmol/L	100 µL	15 min	NMPA CE
IF1067	fT3	Hyperthyroidism, hypothyroidism	3.10-6.80 pmol/L	S/P/WB	0.60-50.00 pmol/L	100 µL	15 min	CE
IF1068	fT4	Hyperthyroidism, hypothyroidism	12.00-22.00 pmol/L	S/P/WB	0.30-100.00 pmol/L	100 µL	15 min	CE

Cat. #	TEST ITEMS	DISEASES	CUT-OFF VALUE	SAMPLE TYPES	MEASURING RANGE	SAMPLE VOLUME	REACTION TIME	QUALIFICATION	
Reproduction/Fertility									
IF1013	HCG+β	Fertility	5.1 mIU/mL	S/P	5.0-100000.0 mIU/mL	100 µL	10 min	NMPA CE	
IF1055	LH	PCOS, infertility evaluation	Refer to User Manual	S/P	0.20-150.00 mIU/mL	100 µL	15 min	NMPA CE	
IF1056	FSH	PCOS, infertility evaluation and pituitary disorders	Refer to User Manual	S/P	0.20-150.00 mIU/mL	100 µL	15 min	NMPA CE	
IF1066	AMH	Fertility, PCOS, gonadal function, precocious/late puberty	Refer to User Manual	S/P	0.10-20.00 ng/mL	200 µL	15 min	CE	
IF1048	PRL	Infertility, gonadal disorders	Refer to User Manual	S/P	0.50-200.00 ng/mL	100 µL	15 min	NMPA CE	
IF1071	Prog	Infertility, evaluation of ovulation	Refer to User Manual	S/P	0.10-40.00 ng/mL	100 µL	15 min	CE	
NEW IF1073	Testosterone	Female polycystic ovary syndrome, Male: 1.75-7.81 ng/mL male testosterone insufficiency Female: 0.10-0.75 ng/mL	Refer to User Manual	S/P	0.10-16.00 ng/mL	100 µL	15 min	CE	
NEW IF1074	E2	Ovarian function	Refer to User Manual	S/P	40.0-4800.0 pg/mL	100 µL	15 min	CE	
Tumor Markers									
IF1053	tPSA	Prostate cancer	4.00 ng/mL	S/P	0.50-100.00 ng/mL	100 µL	15 min	NMPA	
IF1072	fPSA	Prostate cancer	1.00 ng/mL	S/P	0.05-30.00 ng/mL	100 µL	10 min	NMPA	
IF1050	AFP	Liver cancer, cancer of ovaries or testicles, etc.	7.0 ng/mL	S/P	2.0-500.0 ng/mL	100 µL	15 min	CE	
IF1051	CEA	Cancer marker: colon cancer etc.	4.7 ng/mL	S/P	2.0-500.0 ng/mL	100 µL	15 min	CE	
Infectious Disease									
IF1057	Anti-HCV	Hepatitis C	1.00 S/CO	S/P	1.00-20.00 S/CO	100 µL	15 min		
IF1058	Anti-TP	Syphilis	1.00 S/CO	S/P	1.00-50.00 S/CO	100 µL	15 min	CE	
IF1059	Anti-HIV	AIDS	1.00 S/CO	S/P	1.00-1000.00 S/CO	100 µL	15 min		
IF1064	HBsAg	Hepatitis B	1.00 IU/mL	S/P	1.00-100.00 IU/mL	100 µL	15 min		
NEW IF1063	Anti-HBs	Hepatitis B	10.00 mIU/mL	S/P/WB	10.00-1000.00 mIU/mL	100 µL	15 min	CE	
IF1084	2019-nCoV IgM/IgG	COVID-19	1.00 COI	S/P/WB		100 µL	10 min	CE	
NEW IF1091	SARS-CoV-2 Antigen	COVID-19	1.00 COI	Nasal swab/Saliva			100 µL	15 min	CE
NEW IF1095	SARS-CoV-2 Neutralizing Antibody	COVID-19	Refer to User Manual	S/P/WB/ Fingertip blood			40 µL	15 min	CE
IF1047	<i>H. pylori</i>	<i>H. pylori</i> infection	5.0 ng/mL	Stool	1.0-200.0 ng/mL (about 100 µL)	3 drops	10 min	CE	
NEW IF1086	Influenza A/B	Respiratory viral infection	1.00 S/CO	Nasal swab			100 µL	15 min	CE
NEW IF1136	Dengue NS1 Ag	Dengue virus infection	1.00 S/CO	S/P/WB	0.50-50.00 S/CO	100 µL	15 min	CE	
Specific Protein and Rheumatism									
NEW IF1075	RF	Rheumatoid arthritis	15.9 IU/mL	S/P/WB	10.0-640.0 IU/mL	10 µL	10 min	CE	
NEW IF1076	ASO	Rheumatic fever, acute glomerulonephritis, group A streptococcal infection	408.0 IU/mL	S/P/WB	60.0-1370.0 IU/mL	10 µL	10 min	CE	
NEW IF1029	Anti-CCP	Rheumatoid arthritis	25.0 U/mL	S/P/WB	10.0-400.0 U/mL	10 µL	15 min	CE	
Others									
NEW IF1077	Ferritin	Anemia/tumors	Male: 30.00-400.00 ng/mL Female: 13.00-150.00 ng/mL	S/P	0.50-1000.00 ng/mL	10 µL	15 min	CE	
NEW IF1069	Total IgE	Allergic disorders	Refer to User Manual	S/P/WB	1.00-2000.00 IU/mL	100 µL	15 min	CE	
NEW IF1052	PG I/PG II	Atrophic gastritis, stomach cancer	PG I<70.0 ng/mL PG I/PG II<3.0 ng/mL	S/P	PG I: 1.0-200.0 ng/mL PG II: 1.0-100.0 ng/mL	100 µL	15 min		

Coming Soon: FOB, Folate...

 Getein Biotech, Inc.

Add.: No.9 Bofu Road, Luhe District, Nanjing, 211505, China
Tel.: +86-25-68568508/68568594
Fax: +86-25-68568500
E-mail: sales@getein.com.cn; overseas@getein.com.cn
Web.: www.getein.com

 FSC CE NMPA NGSP IFCC IVD





CK-MB/cTnI/Myo Fast Test Kit (Immunofluorescence Assay)

User Manual

CE IVD

IF1005 for Getein1100
IF3005 for Getein1180
IF4005 for Getein1200
IF2005 for Getein1600
IF5005 for Getein1660
IF6005 for Getein208

REF

chest pain, reaching peak concentrations in approximately 8 to 28 hours, and remain elevated for 3 to 10 days following AMI. Due to the high myocardial specificity and the long duration of elevation, cTnI has become an important marker in the diagnosis and evaluation of patients suspected of having an AMI.

Myoglobin is a small monomeric protein which serves as an intracellular oxygen storage site. It is found in abundance in the muscle and can get through into the blood circulation directly when myocardial cell is damaged mildly, and can be elevated 1~2 hours after myocardial injury. Therefore, myoglobin has been advocated as a sensitive marker for early acute myocardial injury by American College of Cardiology Committee.

PRINCIPLE

Mixed monoclonal antibodies against human CK-MB, cTnI and Myo are conjugated with fluorescence latex and another set of anti-human CK-MB/cTnI/Myo monoclonal antibodies were coated on different test lines respectively. After the sample has been applied to the test strip, the fluorescence latex-labelled anti-human CK-MB, cTnI and Myo monoclonal antibodies will bind with the CK-MB, cTnI and Myo in sample respectively and form marked antigen-antibody complexes. These complexes move to the test card detection zone by capillary action. Then marked antigen-antibody complexes will be captured on different test lines by another set of monoclonal antibodies against human CK-MB, cTnI or Myo respectively resulting in the accumulation of fluorescence particles on the test lines. The fluorescence intensity of each test line increases in proportion to the amount of CK-MB, cTnI or Myo in sample.

Then insert test card into Getein1100/Getein1160/Getein1180 Immunoassay Quantitative Analyzer/Getein208 Hand-held Integrated System (automatically inserted by Getein1200/Getein1600 Immunoassay Quantitative Analyzer (hereafter referred to as Getein1100, Getein1160, Getein1180, Getein208, Getein1200 and Getein1600), the concentrations of CK-MB, cTnI and Myo in sample will be measured and displayed on the screen. The value will be stored in Getein1100/Getein1160/Getein1180/Getein208/Getein1200/Getein1600 and available for downloading. The result can be easily transmitted to LIS and HIS.

CONTENTS

1. A kit for Getein1100/Getein1160/Getein1180 contains:

Package specifications: 25 tests/kit, 10 tests/kit

1) CK-MB/cTnI/Myo test card in a sealed pouch with desiccant

2) Disposable pipet

3) User manual: 1 piece/kit

4) SD card: 1 piece/kit

5) Whole blood buffer: 1 bottle/kit

2. A kit for Getein208 contains:

Package specifications: 25 tests/kit, 10 tests/kit

1) CK-MB/cTnI/Myo test card in a sealed pouch with desiccant

2) Disposable pipet

3) Sample diluent

4) User manual: 1 piece/kit

5) SD card: 1 piece/kit

3. A kit for Getein1200/Getein1600 contains:

Package specifications: 2x24 tests/kit, 2x48 tests/kit

1) Sealed cartridge with 24/48 Getein CK-MB/cTnI/Myo test cards

2) User manual: 1 piece/kit

Materials required for Getein1200/Getein1600:

1) Sample diluent: 1 bottle/kit

2) Box with pipette tips: 96 tips/kit

3) Mixing plate: 1 piece/kit

4. Sample diluent/Whole blood buffer composition:

Phosphate buffered saline, proteins, detergent, preservative, stabilizer.

5. A test card consists of:

A plastic shell and a reagent strip which is composed of a sample pad, nitrocellulose membrane (one end of the membrane is coated with fluorescence latex-labelled anti-human CK-MB, cTnI and Myo monoclonal antibodies, these three lines are coated with another anti-human CK-MB, another anti-human cTnI and another anti-human Myo monoclonal antibody, respectively, and the control line C is coated with rabbit anti-mouse IgG antibody), absorbent paper and liner.

Note: Do not mix or interchange different batches of kits.

APPLICABLE DEVICE

Getein1100 Immunoassay Quantitative Analyzer

Getein1180 Immunoassay Quantitative Analyzer

Getein1600 Immunoassay Quantitative Analyzer/Getein208 Hand-held Integrated System (automatically inserted by Getein1200/Getein1600 Immunoassay Quantitative Analyzer (hereafter referred to as Getein1100, Getein1160, Getein1180, Getein208, Getein1200 and Getein1600), the concentrations of CK-MB, cTnI and Myo in sample will be measured and displayed on the screen. The value will be stored in Getein1100/Getein1160/Getein1180/Getein208/Getein1200/Getein1600 and available for downloading. The result can be easily transmitted to LIS and HIS.

STORAGE AND STABILITY

Store the test kit at 4~30°C with a valid period of 24 months. Use the test card for Getein1100/Getein1160/Getein1180/Getein-208 within 1 hour once the foil pouch is opened. For test card of Getein1200/Getein1600: if the cartridge is opened, it could be stable within 24 hours once exposed to air. If the test cards can't be used up at a time, please put the cartridge back to the foil pouch and reseal along the entire edge of zip-seal. The remaining test cards should be used up within 7 days.

PRECAUTIONS

1. *In vitro* diagnostic use only.

2. Do not use the kit beyond the expiration date.

3. Do not use the test card if the foil pouch or the cartridge is damaged.

4. Do not open pouches or the cartridge until ready to perform the test.

5. Do not reuse the test card.

6. Do not reuse the pipet.

7. Handle all specimens as potentially infectious. Proper handling and disposal methods should be followed in accordance with local regulations.

8. Carefully read and follow user manual to ensure proper test performance.

SPECIMEN COLLECTION AND PREPARATION

1. This test can be used for *serum, plasma and whole blood samples*. **Heparin** and **EDTA** should be used as the anticoagulant for plasma and whole blood. Samples should be free of hemolysis.

2. Suggest using serum or plasma for better results.

3. Serum or plasma can be used directly. For whole blood sample, one drop of whole blood buffer must be added before testing.

4. If testing is delayed, serum and plasma samples may be stored up to 7 days at 2~8°C or stored at -20°C for 6 months before testing (whole blood sample may be stored up to 3 days at 2~8°C).

5. Refrigerated or frozen sample should reach room temperature and be homogeneous before testing. Avoid multiple freeze-thaw cycles.

6. Do not use heat-inactivated samples.

7. **SAMPLE VOLUME (for Getein1100/Getein1160/Getein1180): 100 µL.**
(for Getein208): 70 µL.

TEST PROCEDURE

1. Collect specimens according to user manual.

2. Test card, sample and reagent should be brought to room temperature before testing.

For Getein1100:

1. Confirm SD card lot No. in accordance with test kit lot No.. Perform SD card calibration when necessary.

2. Remove the test card from the sealed pouch immediately before use. Label the test card with patient or control identification.

3. Put the test card on a clean table, horizontally placed.

4. Using sample transfer pipette, deliver **100 µL** of sample into the sample well on the test card (for whole blood sample, one drop of whole blood buffer must be added after loading **100 µL** sample on the test card).

5. **Reaction time: 10 minutes.** Insert the test card into Getein1100 and press "ENT" button or click on "Start" icon (for Android Getein 1100) after reaction time is elapsed. The result will be shown on the screen and printed automatically. For Getein1160/Getein1180:

1. Confirm SD card lot No. in accordance with test kit lot No.. Perform "SD card" calibration when necessary.

2. Enter testing interface of Getein1160/Getein1180.

3. Remove the test card from the sealed pouch immediately before use. Label the test card with patient or control identifi-

Clinical studies have demonstrated the release of cTnI into the blood stream within hours following acute myocardial infarctions (AMI) or ischemic damage. Elevated levels of cTnI are detectable in blood within 4 to 6 hours after the onset of

cation.

4. Put the test card on a clean table, horizontally placed.
5. Using sample transfer pipette, deliver **100 µL** of sample into the sample well on the test card (for whole blood sample, one drop of whole blood buffer must be added after loading **100 µL** sample on the test card).

6. Insert the test card into Getein1160/Getein1180 immediately after sample loading. The analyzer will count down the reaction time (10 minutes) and automatically test the card after reaction time is elapsed. The result will be shown on the screen and printed automatically.

For Getein208:

1. Long press the Power Button to start the analyzer.
2. The system will enter (Test) menu.
3. Confirm SD card lot No. in accordance with test kit lot No.. Read the relevant information in the SD card for calibration.
4. Insert test card according to the analyzer prompts.

Note: Do not move the test card after it is inserted.

5. Add sample according to the analyzer prompts. Then draw **70 µL** of sample and drop it into sample diluent. Then drop **70 µL** of sample mixture into the sample port on the test card.

6. After sample adding, the system starts react-time countdown automatically.

7. After the countdown is over, the result will be shown on the screen.

For Getein1200/Getein1600:

1. Each cartridge for Getein1200/Getein1600 contains a specific RFID card which can calibrate automatically.

2. Place the sample diluent at the correct position in Getein1200/Getein1600.

3. Place samples in the designed area of the sample holder, insert the holder and select the right test item, Getein1200/Getein1600 will do the testing and print the result automatically.

Notes:

1. It is required to perform SD card calibration when using a new batch of kits for Getein1100/Getein1160/Getein1180/Getein208.

2. It is suggested to calibrate once for one batch of kits for Getein1100/Getein1160/Getein1180/Getein208.

3. Make sure the test card and the sample insertion is correct and complete.

TEST RESULTS

Getein1100/Getein1160/Getein1180/Getein208/Getein1200/Getein1600 can scan the test card automatically and display the result on the screen. For additional information, please refer to the user manual of Getein1100/Getein1160/Getein1180/Getein208/Getein1200/Getein1600.

EXPECTED VALUE

The expected normal value for CK-MB was determined by testing samples from 500 apparently healthy individuals. The 99th percentile of the concentration for CK-MB is 5.00 ng/ml.

(The probability that value of a normal person below 5.00 ng/ml is 99%).

The expected normal value for cTnI was determined by testing samples from 500 apparently healthy individuals. The 99th percentile of the concentration for cTnI is 0.10 ng/ml. (The probability that value of a normal person below 0.10 ng/ml is 99%).

The expected normal value for Myo was determined by testing samples from 500 apparently healthy individuals. The 95th percentile of the concentration for Myo is 50.0 ng/ml. The 97.5th percentile of the concentration for Myo is 70.0 ng/ml. (According to different Statistics method, the probability that value of a normal person below 50.0 ng/ml is 95% or below 70.0 ng/ml is 97.5%).

It is recommended that each laboratory establish its own expected values for the population it serves.

PERFORMANCE CHARACTERISTICS

	CK-MB	cTnI	Myo
Measuring Range	2.50~80.00 ng/ml	0.10~50.00 ng/ml	30.0~600.0 ng/ml
Lower Detection Limit	≤ 2.50 ng/ml	≤ 0.10 ng/ml	≤ 30.0 ng/ml
Within-Run Precision		≤10%	
Between-Run Precision		≤15%	

LIMITATIONS

1. As with all diagnostic tests, a definitive clinical diagnosis should not be made based on the result of a single test. The test results should be interpreted considering all other test results and clinical information such as clinical signs and symptoms.
2. Interferents in samples may influence the results. The table below listed the maximum allowance of these potential interferents.

Interferent	Hemoglobin	Triglyceride	Bilirubin
Concentration (Max)	5 g/L	10 g/L	0.2 g/L

REFERENCES

1. Mauro Pantaghi; Undefined International Federation of Clinical Chemistry and Laboratory Medicine (IFCC). Scientific Division Committee on Standardization of Markers of Cardiac Damage. Clin Chem Lab Med,1998, 36:887~893.
2. Antman EM, Anbe DT, Armstrong PW, et al. ACC/AHA guidelines for the management of patients with ST-elevation myocardial infarction: a report of the American College of Cardiology/American Heart Association Task Force on Practice Guidelines (Committee to Revise the 1999 Guidelines for the Management of the Patient with ST-Elevation Myocardial Infarction). Circulation, 2004, 109:1031~1045.
3. EN ISO 18113-1:2011 In vitro diagnostic medical devices -

Information supplied by the manufacturer (labelling) - Part 1: Terms, definitions and general requirements.

4. EN ISO 18113-2:2011 In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 2: In vitro diagnostic reagents for professional use.

DESCRIPTION OF SYMBOLS USED

Key to symbols used		
	Manufacturer	
	Do not re-use	
	Consult instructions for use or consult electronic instructions for use	
	Temperature limit	
	Contains sufficient for <n> tests	
	CE mark	
	Catalogue number	

Thank you for purchasing CK-MB/cTnI/Myo Fast Test Kit (Immunofluorescence Assay). Please read this user manual carefully before operating to ensure proper use.

Version: WIF09-S-13

Getein Biotech, Inc.
Add.: No.9 Bofu Road, Luhe District, Nanjing, 211505, China.
Tel: +86-25-68568508
Fax: +86-25-68568500
E-mail: tech@getein.com.cn, overseas@getein.com.cn
Website: www.getein.com

CMC Medical Devices & Drugs S.L.
Add.: C/ Horacio Lengo N° 18, CP 29006, Málaga, Spain
Tel: +34951214054



Cardiac Troponin I Fast Test Kit (Immunofluorescence Assay)

Instructions for Use

REF

IF1001 for Getein1100
IF2001 for Getein1600
IF5001 for Getein1160
IF3001 for Getein1180
IF4001 for Getein1200
IF6001 for Getein208

INTENDED USE

Cardiac Troponin I Fast Test Kit (Immunofluorescence Assay) is intended for *in vitro* quantitative determination of Cardiac Troponin I (cTnI) in human serum, plasma or whole blood samples. This test is used as an aid in the diagnosis of myocardial injury such as Acute Myocardial Infarction (AMI), Unstable Angina, Acute Myocarditis and Acute Coronary Syndrome (ACS).

SUMMARY

Troponin, a molecular complex that is bound to the thin filament (actin) of striated muscle fibers, acts with intracellular calcium to control the interaction of the thin filament with the thick filament (myosin), thus regulating muscle contraction. Troponin consists of three regulatory proteins: T, which connects the troponin complex and tropomyosin (another cardiac muscle regulatory protein); I, which prevents muscle contraction in the absence of calcium; C, which binds calcium. Cardiac Troponin I (MW 22.5 kDa) and the two skeletal muscle isoforms of Troponin I have considerable amino acid sequence homology, but cTnI contains an additional N-terminal sequence and is highly specific for myocardium.

Clinical studies have demonstrated the release of cTnI into the blood stream within hours following acute myocardial infarction (AMI) or ischemic damage. Elevated levels of cTnI are detectable in blood within 4 to 6 hours after the onset of chest pain, reaching peak concentrations in approximately 8 to 28 hours, and remain elevated for 3 to 10 days following AMI. Due to the high myocardial specificity and the long duration of elevation, cTnI has become an important marker in the diagnosis and evaluation of patients suspected of having an AMI.

The current guideline of The Joint European Society of Cardiology/American College of Cardiology Committee support the use of cTnI as a preferred marker of myocardial injury. Several major studies have shown that cTnI is also a predictor of cardiac risk in patients with unstable angina. The American College of Cardiology and the American Heart Association's current guidelines recommend using troponin results when making treatment decisions regarding unstable angina and

non-ST segment elevation MI (NSTEMI).

PRINCIPLE

The test uses an anti-human cTnI monoclonal antibody conjugated with fluorescence latex and another anti-human cTnI monoclonal antibody coated on the test line. After the sample has been applied to the test strip, the fluorescence latex-labelled anti-human cTnI monoclonal antibody binds with the cTnI in sample and forms a marked antigen-antibody complex. This complex moves to the test card detection zone by capillary action. Then marked antigen-antibody complex is captured on the test line by the anti-human cTnI monoclonal antibody. The fluorescence intensity of the test line increases in proportion to the amount of cTnI in sample.

Then insert test card into Getein1100/Getein1160/Getein1180 Immunofluorescence Quantitative Analyzer/Getein208 Hand-held Integrated System/automatically inserted by Getein1200/Getein1600 Immunofluorescence Quantitative Analyzer (hereinafter referred to as Getein1100, Getein1160, Getein1180, Getein208, Getein1200 and Getein1600), the concentration of cTnI in sample will be measured and displayed on the screen. The value will be stored in Getein1100/Getein1160/Getein1180/Getein208/Getein1200/Getein1600 and available for download. The result can be easily transmitted to the laboratory or hospital information system.

CONTENTS

1. A kit for Getein1100/Getein1160/Getein1180 contains:

- 1) Package specifications: 25 tests/kit, 10 tests/kit
- 1) cTnI test card in a sealed pouch with desiccant
- 2) Disposable pipet
- 3) User manual: 1 piece/kit
- 4) SD card: 1 piece/kit
- 5) Whole blood buffer: 1 bottle/kit

2. A kit for Getein208 contains:

- 1) Package specifications: 25 tests/kit, 10 tests/kit
- 1) cTnI test card in a sealed pouch with desiccant
- 2) Disposable pipet
- 3) Sample diluent
- 4) User manual: 1 piece/kit
- 5) SD card: 1 piece/kit

3. A kit for Getein1200/Getein1600 contains:

- 1) Package specifications: 2x24 tests/kit, 2x48 tests/kit
- 1) Sealed cartridge with 24/48 Getein cTnI test cards
- 2) User manual: 1 piece/kit

Materials required for Getein1200/Getein1600:

- 1) Sample diluent: 1 bottle/kit
- 2) Box with pipette tips: 96 tips/kit
- 3) Mixing plate: 1 piece/kit

4. Sample diluent/Whole blood buffer composition:

Phosphate buffered saline, proteins, detergent, preservative, stabilizer.

5. A test card consists of:

A plastic shell and a reagent strip which is composed of a sample pad, nitrocellulose membrane (one end of the membrane is coated with a fluorescence latex-labelled anti-human cTnI monoclonal antibody, the test line is coated with another anti-human cTnI monoclonal antibody and the control line is coated with rabbit anti-mouse IgG antibody), absorbent paper and liner.

Note: Do not mix or interchange different batches of kits.

APPLICABLE DEVICE

Getein1100 Immunofluorescence Quantitative Analyzer
Getein1160 Immunofluorescence Quantitative Analyzer
Getein208 Hand-held Integrated System
Getein1200 Immunofluorescence Quantitative Analyzer
Getein1600 Immunofluorescence Quantitative Analyzer

STORAGE AND STABILITY

Store the test kit at 4~30°C with a valid period of 24 months. Use the test card for Getein1100/Getein1160/Getein1180/Getein-208 within 1 hour once the foil pouch is opened.

For test card of Getein1200/Getein1600: if the cartridge is opened, it could be stable within 24 hours once exposed to air. If the test cards can't be used up at a time, please put the cartridge back to the foil pouch and reseal along the entire edge of zip-seal. The remaining test cards should be used up within 7 days.

PRECAUTIONS

- 1. For *in vitro* diagnostic use only.
- 2. Do not use the kit beyond the expiration date.
- 3. Do not use the test card if the foil pouch or the cartridge is damaged.
- 4. Do not open pouches or the cartridge until ready to perform the test.
- 5. Do not reuse the test card.
- 6. Do not reuse the pipet.
- 7. Handle all specimens as potentially infectious. Proper handling and disposal methods should be followed in accordance with local regulations.
- 8. Carefully read and follow user manual to ensure proper test performance.

SPECIMEN COLLECTION AND PREPARATION

1. This test can be used for **serum, plasma and whole blood samples**. **Heparin and EDTA** should be used as the anticoagulant for plasma and whole blood. Samples should be free of hemolysis.
2. Suggest using serum or plasma for better results.
3. Serum or plasma can be used directly. For whole blood sample, one drop of whole blood buffer must be added before

testing.

4. If testing is delayed, serum and plasma samples may be stored up to 7 days at 2~8°C or stored at -20°C for 6 months before testing (whole blood sample may be stored up to 3 days at 2~8°C).
5. Refrigerated or frozen sample should reach room temperature and be homogeneous before testing. Avoid multiple freeze-thaw cycles.
6. Do not use heat-inactivated samples.
7. **SAMPLE VOLUME**(for Getein1100/Getein1600/Getein1180): **100 µL**.
(for Getein208): **70 µL**.

TEST PROCEDURE

1. Collect specimens according to user manual.
2. Test card, sample and reagent should be brought to room temperature before testing.
For Getein1100:
 1. Confirm SD card lot No. in accordance with test kit lot No.. Perform "SD card" calibration when necessary.
 2. Remove the test card from the sealed pouch immediately before use. Label the test card with patient or control identification.
 3. Put the test card on a clean table, horizontally placed.
 4. Using sample transfer pipette, deliver **100 µL** of sample into the sample well on the test card (for whole blood sample, one drop of whole blood buffer must be added after loading **100 µL** sample on the test card).
5. **Reaction time: 10 minutes.** Insert the test card into Getein1100 and press "ENT" button or click on "Start" icon (for Android Getein 1100) after reaction time is elapsed. The result will be shown on the screen and printed automatically.
For Getein1160/Getein180:
 1. Confirm SD card lot No. in accordance with test kit lot No.. Perform "SD card" calibration when necessary.
 2. Enter testing interface of Getein1160/Getein180.
 3. Remove the test card from the sealed pouch immediately before use. Label the test card with patient or control identification.
 4. Put the test card on a clean table, horizontally placed.
 5. Using sample transfer pipette, deliver **100 µL** of sample into the sample well on the test card (for whole blood sample, one drop of whole blood buffer must be added after loading **100 µL** sample on the test card).
 6. Insert the test card into Getein1160/Getein1180 immediately after sample loading. The analyzer will count down the reaction time (10 minutes) and automatically test the card after reaction time is elapsed. The result will be shown on the screen and printed automatically.
6. **For Getein208:**
 1. Long press the Power Button to start the analyzer.
 2. The system will enter (Test) menu.
 3. Confirm SD card lot No. in accordance with test kit lot No..

Read the relevant information in the SD card for calibration.

- Insert test card according to the analyzer prompts.
- Note:** Do not move the test card after it is inserted.
- Add sample according to the analyzer prompts. Then draw **70 µL** of sample and drop it into sample diluent. Then drop **70 µL** of sample mixture into the sample port on the test card.
- After sample adding, the system starts react-time countdown automatically.
- After the countdown is over, the result will be shown on the screen.

For Getein1200/Getein1600:

- Each cartridge for Getein1200/Getein1600 contains a specific RFID card which can calibrate automatically.
- Place the sample diluent at the correct position in Getein1200/Getein1600.
- Place samples in the designed area of the sample holder, insert the holder and select the right test item, Getein1200/Getein1600 will do the testing and print the result automatically.

Notes:

- It is required to perform "SD card" calibration when using a new batch of kits for Getein1100/Getein1160/Getein1180/Getein208.
- It is suggested to calibrate once for one batch of kits for Getein1100/Getein1160/Getein1180/Getein208.
- Make sure the test card and the sample insertion is correct and complete.

TEST RESULTS

Getein1100/Getein1160/Getein1180/Getein208/Getein1200/Getein1600 can scan the test card automatically and display the result on the screen. For additional information, please refer to the user manual of Getein1100/Getein1160/Getein1180/Getein208/Getein1200/Getein1600.

EXPECTED VALUE

The expected normal value for Troponin I was determined by testing samples from 500 apparently healthy individuals. The 99th percentile of the concentration for cTnI is 0.10 ng/ml. (The probability that value of a normal person below 0.10 ng/ml is 99%.)

It is recommended that each laboratory establish its own expected values for the population it serves.

PERFORMANCE CHARACTERISTICS

Measuring Range	0.10~50.00 ng/ml
Lower Detection Limit	≤ 0.10 ng/ml
Within-Run Precision	≤10%
Between-Run Precision	≤15%

LIMITATIONS

- As with all diagnostic tests, a definitive clinical diagnosis should not be made based on the result of a single test. The test results should be interpreted considering all other test results and clinical information such as clinical signs and symptoms.
- Interferers in samples may influence the results. The table below listed the maximum allowance of these potential interferers.

Interferent	Hemoglobin	Triglyceride	Bilirubin
Concentration (Max)	5 g/L	10 g/L	0.2 g/L

REFERENCES

- Mauro Pantaghini. Undefined International Federation of Clinical Chemistry and Laboratory Medicine (IFCC). Scientific Division Committee on Standardization of Markers of Cardiac Damage. Clin Chem Lab Med, 1998, 36:887-893.
- Antman EM, Anbe DT, Armstrong PW, et al. ACC/AHA guidelines for the management of patients with ST-elevation myocardial infarction: a report of the American College of Cardiology/American Heart Association Task Force on Practice Guidelines (Committee to Revise the 1999 Guidelines for the Manage 2004).
- EN ISO 18113-1:2011 In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 1: Terms, definitions and general requirements.
- EN ISO 18113-2:2011 In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 2: In vitro diagnostic reagents for professional use.

DESCRIPTION OF SYMBOLS USED

The following graphical symbols used in or found on Cardiac Troponin I Fast Test Kit (Immuno fluorescence Assay) are the most common ones appearing on medical devices and their packaging. They are explained in more details in the European Standard EN ISO 15223-1:2021.

Key to symbols used		
	Manufacturer	
	Do not re-use	
	Consult instructions for use or consult electronic instructions for use	
	Temperature limit	
	Contains sufficient for <n> tests	
	CE mark	
	Catalogue number	

Thank you for purchasing Cardiac Troponin I Fast Test Kit (Immuno fluorescence Assay). Please read this user manual carefully before operating to ensure proper use.

Version: WIF02-S1-05



Getein Biotech, Inc.
Add.: No.9 Bofu Road, Luhe District, Nanjing, 211505, China
Tel: +86-25-68568508
Fax: +86-25-68568500
E-mail: tech@getein.com.cn
overseas@getein.com.cn
Website: www.getein.com



CMC Medical Devices & Drugs S.L.
Add.: C/ Horacio Lengo N° 18, CP 29006, Málaga, Spain
Tel: +34951214054

D-Dimer

Fast Test Kit

(Immunofluorescence Assay)

User Manual

INTENDED USE

D-Dimer Fast Test Kit (Immunofluorescence Assay) is intended for *in vitro* quantitative determination of D-Dimer in human plasma or whole blood samples. The test is used as an aid in the exclusion of deep vein thrombosis (DVT) and pulmonary embolism (PE) disease in patient suspected of DVT or PE, and an aid in the diagnosis of disseminated intravascular coagulation (DIC).

For professional and laboratory use only.

SUMMARY

Thrombin converts fibrinogen to soluble fibrin by cleaving the fibrinopeptides A and B. The fibrin monomers polymerize spontaneously. Active factor XIII links two D domains and generates a solid fibrin clot. A new plasmin resistant antigenic determinant (D-dimer) is produced. Fragments containing D-dimer are accordingly formed during the degradation of a fibrin clot by plasmin. D-dimer antigens are specific markers of fibrin clot formation and fibrinolysis and may be clinically useful markers for excluding venous thromboembolism.

The primary diagnostic application of D-dimer testing is to rule out thromboembolic events, such as deep vein thrombosis or pulmonary embolism. If the D-dimer test result is below the decision threshold, a thromboembolic event can be ruled out by the negative predictive value (NPV) of the test. The D-dimer test, in combination with a well-validated preclinical trial probability score, is an effective and safe screening tool to rule out thromboembolic events. However, the presence of symptoms over a certain period of time, such as more than one week, may yield normal D-dimer values. Fibrin degradation products are a sensitive marker in disseminated intravascular coagulation (DIC).

In addition to DVT, PE and DIC, D-dimers may reflect other causes associated with fibrin formation, such as age, pregnancy complications, malignant disease or vascular

abnormalities. Therefore, elevated D-dimer levels must be interpreted in the context of possible underlying disease and clinical symptoms.

PRINCIPLE

D-Dimer Fast Test Kit (Immunofluorescence Assay) is a lateral flow immunoassay in a sandwich design. After the sample has been applied to the test strip, the fluorescence labelled D-Dimer monoclonal antibody binds with the D-Dimer in sample and forms a marked antigen-antibody complex. This complex moves to the test card detection zone by capillary action. Then marked antigen-antibody complex is captured on the test line by another D-Dimer monoclonal antibody. The fluorescence intensity of the test line increases in proportion to the amount of D-Dimer in sample. Fluorescent signals intensity can be analyzed by applicable device thus the D-Dimer in sample be detected quantitatively.

CONTENTS

Materials provided	Getein 1100/ Getein 1160/ Getein 1180/ Getein 208		Getein 1150		Getein 1200/Getein 1600		
	10 T/kit	25 T/kit	10 T/kit	25 T/kit	2×12 T/kit	2×24 T/kit	2×48 T/kit
D-Dimer test card*	10 pcs	25 pcs	10 pcs	25 pcs	2 cartridges, 12 pcs in each	2 cartridges, 24 pcs in each	2 cartridges, 48 pcs in each
Disposable pipet	10 pcs	25 pcs	10 pcs	25 pcs	/	/	/
Sample diluent**	10 tubes	25 tubes	10 tubes	25 tubes	1 box	1 box	1 box
Instructions for use	1 pc	1 pc	1 pc	1 pc	1 pc	1 pc	1 pc
SD card	1 pc	1 pc	/	/	1 pc in each cartridge	1 pc in each cartridge	1 pc in each cartridge

* D-Dimer test card

A test card mainly consists of: Fluorescence labelled D-Dimer monoclonal antibody and D-Dimer monoclonal antibody.

** Sample diluent

(1) Sample diluent for Getein 1100/Getein 1150/Getein 1160/Getein 1180/Getein 208 in each tube mainly consists of: phosphate buffer, Na_3N ($< 0.1\%$).

(2) Sample diluent for Getein 1200/Getein 1600 is an independent packing box mainly consists of:

- phosphate buffer, Na_3N ($< 0.1\%$) (25 mL/bottle for Getein 1200, 40 mL/bottle for Getein 1600),
- Box with pipette tips (96 tips/box),
- Mixing plate (1 piece/box).

Note:

1. The SD card, also known as the standard curve data card, stores standard curve data for the specific test items and uses RFID technology to transfer the data to analyzers via touch.
2. The standard curve data for Getein 1150 is written to the QR code on the outer packaging box.
3. Do not mix or interchange different batches of kits.

APPLICABLE DEVICE

Getein 1100 Immunofluorescence Quantitative Analyzer
Getein 1150 Immunofluorescence Quantitative Analyzer
Getein 1160 Immunofluorescence Quantitative Analyzer
Getein 1180 Immunofluorescence Quantitative Analyzer
Getein 1200 Immunofluorescence Quantitative Analyzer
Getein 1600 Immunofluorescence Quantitative Analyzer
Getein 208 Hand-held Integrated System

STORAGE AND STABILITY

Realtime stability:

Store the kit at 4–30°C with a valid period of 24 months. The test kits are stable until the expiry date printed on the labels.

In-use stability:

For the test card of Getein 1100/Getein 1150/Getein 1160/Getein 1180/Getein 208: Use the test card within 1 hour once the foil pouch is opened.

For test card of Getein 1200/Getein 1600: if the cartridge is opened, it could be stable within 24 hours once exposed to air. If the test cards can't be used up at a time, please put the cartridge back to the foil pouch and reseal along the entire edge of zip-seal. The remaining test cards should be used up within 7 days.

PRECAUTIONS

1. For *in vitro* diagnostic use only.
2. For professional and laboratory use only, not for near-patient test and self-testing.
3. Do not use the test card if the foil pouch or the cartridge is damaged.
4. Do not open pouches until performing the test.
5. Do not reuse the test card and disposable pipet.
6. Handle all specimens as potentially infectious. The foil bag is non-degradable. Proper handling and disposal methods should be followed in accordance with local regulations.
7. It is recommended that operators take necessary self-protection measures (work clothes and disposable gloves, etc) when touching kits or samples.

SPECIMEN COLLECTION AND PREPARATION

1. This test can be used for **plasma and whole blood samples**. **Sodium citrate** can be used as the anticoagulant for plasma and whole blood. Samples should be free of hemolysis.
2. Suggest using plasma for better results.
3. Plasma are stable for 4 hours at room temperature (15–30°C), 3 days at 2–8°C, and 1 month at -20°C.
4. Whole blood and fingertip blood are stable for 4 hours at room temperature (15–30°C), 3 days at 2–8°C and avoid cryopreservation.
5. Refrigerated or frozen sample should reach room temperature and be homogeneous before testing. Avoid multiple freeze-thaw cycles.

TEST PROCEDURE

1. User must carefully read and operate in strict accordance with the instructions for use before testing, otherwise reliable results cannot be guaranteed.
2. Test kit and sample should be brought to room temperature before testing.

For Getein 1100:

- 1) Confirm SD card lot No. in accordance with test kit lot No. It is required to perform "SD card" calibration when using a new batch of kits.
- 2) Select the corresponding "Sample" on the analyzer according to the sample type (see the user manual of analyzer for details).
- 3) Remove the test card from the sealed pouch before use. Horizontally place the test card.
- 4) Deliver 100 μL of sample into one tube of sample diluent using disposable pipet or pipette, mix gently and thoroughly. Then drop 100 μL of sample mixture into the sample well on the test card.

- 5) **Reaction time: 10 minutes.** After reaction time is elapsed, insert the test card into Getein 1100 and press "ENT" button (click on "Start" icon for Android Getein 1100). The result will be shown on the screen and printed automatically.

For Getein 1160/Getein 1180:

- 1) Confirm SD card lot No. in accordance with test kit lot No. It is required to perform "SD card" calibration when using a new batch of kits.
- 2) Select the corresponding "Sample" on the analyzer according to the sample type (see the user manual of analyzer for details).
- 3) Remove the test card from the sealed pouch before use. Horizontally place the test card.

4) Deliver 100 μ L of sample into one tube of sample diluent using disposable pipet or pipette, mix gently and thoroughly. Then drop 100 μ L of sample mixture into the sample well on the test card.

5) Insert the test card into Getein 1160/Getein 1180 **immediately** after sample loading. The analyzer will count down the reaction time (10 minutes) and automatically test the card after reaction time is elapsed. The result will be shown on the screen and printed automatically.

For Getein 1150:

- 1) Turn on the instrument and enter the sample test interface. Insert the test card and scan the QR code (**On the outer packaging box**) to complete calibration as prompted by the instrument.
- 2) Select the corresponding "Sample" mode on the analyzer (refer to the analyzer user manual for details).
- 3) Deliver 100 μ L of sample into one tube of sample diluent using disposable pipet or pipette, mix gently and thoroughly. Then drop 100 μ L of sample mixture into the sample well on the test card.
- 4) Press the "Start" button immediately after sample loading. The analyzer will initiate a 10-minute reaction countdown, and the test results will be automatically displayed on the screen upon completion.

For Getein 208:

- 1) Enter testing interface of Getein 208. Confirm SD card lot No. in accordance with test kit lot No. Read the relevant information in the SD card for calibration.
- 2) Select the corresponding "Sample" mode on the analyzer (refer to the analyzer user manual for details). Insert test card according to the analyzer prompts.
- 3) Deliver 60 μ L of sample into one tube of sample diluent using disposable pipet or pipette, mix gently and thoroughly. Then drop 60 μ L of sample mixture into the sample well on the test card according to the analyzer prompts.
- 4) After sample adding, the analyzer will initiate a 10-minute reaction countdown, and the test results will be automatically displayed on the screen upon completion.

For Getein 1200/Getein 1600:

- 1) Each cartridge for Getein 1200/Getein 1600 contains a specific RFID card (SD card) which can calibrate automatically.
- 2) Place the sample diluent at the correct position in Getein 1200/Getein 1600.
- 3) Place samples in the designed area of the sample holder, insert the holder, set parameters (more operational details refer to the user manual of analyzer) and run the instrument.

Getein 1200/Getein 1600 will do the testing and print the result automatically.

LIMITATIONS

1. As with all diagnostic tests, a definitive clinical diagnosis should not be made based on the result of a single test. The test results should be interpreted considering all other test results and clinical information such as clinical signs and symptoms.
2. Interferents in samples may influence the results. The table below listed the maximum allowance of these potential interferents.

Interferent	Hemoglobin	Triglyceride	Bilirubin
Concentration (Max)	5 g/L	25 g/L	0.1 g/L

3. Patient samples may contain heterophilic antibodies (e.g. human anti-mouse antibodies (HAMA) and rheumatoid factors) that could react in immunoassays to give a falsely elevated or depressed result. This assay has been designed to minimize interference from heterophilic antibodies. Nevertheless, complete elimination of this interference from all patient specimens cannot be guaranteed.

EXPECTED VALUE

The expected value for D-Dimer was determined by testing samples from 500 apparently healthy individuals. The 95th percentile of the concentration for D-Dimer is 0.50 mg/L.

Each laboratory should verify the transferability of the expected values to its own population, and if necessary, determine its own expected values according to good laboratory practice.

PERFORMANCE CHARACTERISTICS

Measuring Range	0.10–10.00 mg/L
Limit of Detection	≤ 0.10 mg/L
Within-Lot Precision	$\leq 10\%$
Between-Lot Precision	$\leq 15\%$

REFERENCES

1. MedlinePlus [Internet]. Bethesda (MD): National Library of Medicine (US); [updated Jun 24; cited 2020 Jul 1]. Available from: <https://medlineplus.gov/>.
2. Sarig G, Kilil-Drori AJ, Chap-Marshak D, Brenner B, Drugan A. Activation of coagulation in amniotic fluid during normal human pregnancy. *Thromb Res*. 2011 Apr 18.
3. Roldán V, Marín F, Muiña B, Torregrosa JM, Hernández-Romero D, Valdés M, Vicente V, Lip GY. Plasma

von Willebrand Factor Levels Are an Independent Risk Factor for Adverse Events Including Mortality and Major Bleeding in Anticoagulated Atrial Fibrillation Patients. *J Am Coll Cardiol*. 2011 Apr 11.

4. Sakamoto K, Yamamoto Y, Okamatsu H, Okabe M. D-dimer is helpful for differentiating acute aortic dissection and acute pulmonary embolism from acute myocardial infarction. *Hellenic J Cardiol*. 2011 Mar-Apr; 52(2):123-127.

5. EN ISO 18113-1:2011 In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 1: Terms, definitions and general requirements.

6. EN ISO 18113-2:2011 In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 2: In vitro diagnostic reagents for professional use.



Getein Biotech, Inc.

Add.: No.9 Bofu Road, Luhe District, Nanjing, 211505, China

Tel: +86-25-68568508

Fax: +86-25-68568500

E-mail: tech@getein.com.cn
overseas@getein.com.cn

Website: www.getein.com



CMC Medical Devices & Drugs S.L.

Add.: C/ Horacio Lengo Nº 18, CP 29006, Málaga, Spain

Tel: +34951214054

Catalogue number	Applicable analyzer	Package specification
IF1006-10T	Getein 1100	10 T/kit
IF1006	Getein 1100	25 T/kit
IF8006-10T	Getein 1150	10 T/kit
IF8006	Getein 1150	25 T/kit
IF5006-10T	Getein 1160	10 T/kit
IF5006	Getein 1160	25 T/kit
IF3006-10T	Getein 1180	10 T/kit
IF3006	Getein 1180	25 T/kit
IF6006-10T	Getein 208	10 T/kit
IF6006	Getein 208	25 T/kit
IF4006-24T	Getein 1200	2×12 T/kit
IF4006	Getein 1200	2×24 T/kit
IF4006-96T	Getein 1200	2×48 T/kit
IF2006-24T	Getein 1600	2×12 T/kit
IF2006	Getein 1600	2×24 T/kit
IF2006-96T	Getein 1600	2×48 T/kit

Thank you for using D-Dimer Fast Test Kit (Immunofluorescence Assay). Please read the instructions for use carefully before operating to ensure proper use.