


# TS EN 868-8 STANDARD ANNEX-F/G TESTS VALIDATION REPORT

**Protocol Number** : V270521-01  
**Report Number** : AYA-VR080621-01  
**Date** : 08.06.2021  
**Revision Number** : 02  
**Revision Date** : 23.01.2024

**AYA Validasyon Ambalaj Sterilizasyon A.Ş.**  
Saray Mah. 106 Cad. No:14 Kahramankazan-Ankara • (+90 312) 800 01 05-06  
www.ayalab.com.tr • info@ayalab.com.tr

**This Report cannot be copied in whole or in part, or reproduced as controlled copies without the written consent of AYA.**

**Unauthorized Report is invalid.**



# TS EN 868-8 STANDARD ANNEX-F/G TESTS VALIDATION REPORT

Document No : 85AYA.PR.07-F03  
First Release Date: 10.07.2020  
Rev. No : 00  
Rev. Date :-

Report No: AYA-VR080621-01 Rev.02

08 June 2021

## 1. CONTENTS

1. CONTENTS .....	3
2. TABLE INDEX.....	3
3. REPORT APPROVAL PAGE .....	4
4. SCOPE.....	5
4.1. Products Covered by Validation.....	5
4.2. Steam Sterilization Unit and Biological Indicator Information .....	5
5. VALIDATION STUDY AND ITS RESULTS .....	6
5.1. Annex-F Determination of Sterilization Performance .....	6
5.2. Annex-G Dryness Determination Test.....	9
6. RESULT .....	10
7. REFERENCES .....	10
8. DEVIATIONS.....	10
9. REVISION HISTORY.....	10
10. ANNEXES .....	11

## 2. TABLE INDEX

Table 4.1 Product Groups

Table 4.2 Steam Sterilization Unit Information

Table 4.3 Biological Indicator Information

Table 5.1 Biological Indicator Sterility Test Results - Biobarrier Full Size Sterilization Container

Table 5.2 Biological Indicator Sterility Test Results - Full Size Container Bottom Non-Perforated

Table 5.3 Datalogger Temperature-Pressure Distribution Results - Biobarrier Full Size Sterilization Container

Table 5.4 Datalogger Temperature-Pressure Distribution Results - Full Size Container Bottom Non-Perforated

Table 5.5 Dryness Determination Test Results - Biobarrier Full Size Sterilization Container

Table 5.6 Dryness Determination Test Results - Full Size Container Bottom Non-Perforated



# TS EN 868-8 STANDARD ANNEX-F/G TESTS VALIDATION REPORT

Document No : 85AYA.PR.07-F03

First Release Date: 10.07.2020

Rev. No : 00

Rev. Date : -

Report No: AYA-VR080621-01 Rev.02

08 June 2021

### 3. REPORT APPROVAL PAGE

With this Report, it has been observed that the Performance Adequacy of CUSTOMER products in accordance with the TS EN 868-8 Standard Annex-F / Annex-G tests were evaluated, revised and fulfilled the requirements of the relevant standard.

AYA			
Approved By	Title	Signature	Date
MENE YAZ	Sterilization & Validation Manager		23.01.2024
Aysel Hürrem Yılmaz	Quality System Manager		23.01.2024
MENE YAZ	Management		23.01.2024

CUSTOMER			
Approved By	Title	Signature	Date
Ahmet AYDEMİR	General Manager		23.01.2024
Ufuk GÜLER	Deputy General Manager		23.01.2024
Melike Burcu Aydın	Quality Management Representative		23.01.2024



# TS EN 868-8 STANDARD ANNEX-F/G TESTS VALIDATION REPORT

Document No : 85AYA.PR.07-F03  
First Release Date: 10.07.2020  
Rev. No : 00  
Rev. Date : -

Report No: AYA-VR080621-01 Rev.02

08 June 2021

## 4. SCOPE

This Validation Report has been prepared on the basis of the Validation Protocol numbered V270521-01 published on 27.05.2021 and the TS EN 868-8 Standard Annex-F / Annex-G tests.

### 4.1. Products Covered by Validation

The CUSTOMER has requested the validation studies to be carried out on the product given in Table 4.1. (Validation Report ANNEX-1).

Table 4.1 Product Groups

Product Name	Lot No
Biobarrier Full Size Sterilization Container <i>(Sample Size: 580x280x260 mm)</i>	2102492
<i>Full Size Container Bottom Non-Perforated</i> <i>(Sample Size: 580x280x260 mm)</i>	2102498

### 4.2. Steam Sterilization Unit and Biological Indicator Information

In the TS EN 868-8 Standard Annex-F / Annex-G tests, the Performance Validation Studies have been carried out before and the AYA Steam Sterilization Unit documented in the Report No. AYA2020-302-001 was used.

Table 4.2 Steam Sterilization Unit Information

Brand	NÜVE
Model	OT 300
Serial Number	03.0034
Volume	300 L

Table 4.3 Biological Indicator Information

Test Microorganism	<i>Geobacillus stearothermophilus</i>
Biological Indicator Type	SCBI
Lot Number	DDN112022
D Value for Steam (134 °C)	1,5 min.
Bacteria Population	1.0 x 10 <sup>6</sup>

Biological Indicator Information and Test Reports are given in the Validation Report ANNEX-2.



# TS EN 868-8 STANDARD ANNEX-F/G TESTS VALIDATION REPORT

Document No : 85AYA.PR.07-F03  
First Release Date: 10.07.2020  
Rev. No : 00  
Rev. Date :-

Report No: AYA-VR080621-01 Rev.02

08 June 2021

## 5. VALIDATION STUDY AND ITS RESULTS

### 5.1. Annex-F Determination of Sterilization Performance

Validation Cycle	Number of Studies	Tests to Be Performed	Acceptance Criteria
Annex-F Determination of Sterilization Performance	3	Biological Indicator Sterility Test	No growth in biological indicators
		Datalogger Temperature-Pressure Distribution Test	Temperature difference between datalogger should not exceed 2 °C.

**Table 5.1 Biological Indicator Sterility Test Results - Biobarrier Full Size Sterilization Container**

Product Name	Biobarrier Full Size Sterilization Container			
Biological Indicator No	1st cycle	2nd cycle	3rd cycle	Acceptance Criteria
BI 1 (in container)	No Growth	No Growth	No Growth	<b>No Growth</b>
BI 2 (container top)	No Growth	No Growth	No Growth	
BI 3 (in autoclave)	No Growth	No Growth	No Growth	
Negative Control	No Growth	No Growth	No Growth	
Positive Control	Growth	Growth	Growth	<b>Growth Observed</b>

**Table 5.2 Biological Indicator Sterility Test Results - Full Size Container Bottom Non-Perforated**

Product Name	Full Size Container Bottom Non-Perforated			
Biological Indicator No	1st cycle	2nd cycle	3rd cycle	Acceptance Criteria
BI 1 (in container)	No Growth	No Growth	No Growth	<b>No Growth</b>
BI 2 (container top)	No Growth	No Growth	No Growth	
BI 3 (in autoclave)	No Growth	No Growth	No Growth	
Negative Control	No Growth	No Growth	No Growth	
Positive Control	Growth	Growth	Growth	<b>Growth Observed</b>

Biological Indicator Test Results Validation Report can be seen in ANNEX-3.1.



# TS EN 868-8 STANDARD ANNEX-F/G TESTS VALIDATION REPORT

Document No : 85AYA.PR.07-F03  
First Release Date: 10.07.2020  
Rev. No : 00  
Rev. Date : -

Report No: AYA-VR080621-01 Rev.02

08 June 2021

Table 5.3 Datalogger Temperature-Pressure Distribution Results - Biobarrier Full Size Sterilization Container

Product Name	Biobarrier Full Size Sterilization Container		
Cycle No	1st cycle	2nd cycle	3rd cycle
Autoclave Working Temperature and Time	134°C / 5 min.	134°C / 5 min.	134°C / 5 min.
Test Date	04.06.2021	04.06.2021	04.06.2021
Start Time	14:10:00	15:05:00	15:44:00
Exposure Start Time	14:26:30	15:17:00	15:57:30
Exposure Finish Time	14:31:30	15:22:00	16:02:30
Finish time	14:54:30	14:54:30	16:27:00
Exposure Time	00:05:00	00:05:00	00:05:00
<u>Vacuum Drying Pressure (Bar)</u>	-0.693	-0.695	-0.692
<u>Vacuum Drying Temperature (°C)</u>	95.4	94.5	95.5
<u>Vacuum Drying Time</u>	15 min.	15 min.	15 min.
Max. Temperature (After Stabilization)	136,48	136,69	136,37
Min. Temperature (After Stabilization)	134,56	134,83	135,18
Deviation Between Datalogger After Stabilization Period	1,92	1,86	1,19
Compliance with Acceptance Criteria	Suitable <input checked="" type="checkbox"/> Not Suitable <input type="checkbox"/>	Suitable <input checked="" type="checkbox"/> Not Suitable <input type="checkbox"/>	Suitable <input checked="" type="checkbox"/> Not Suitable <input type="checkbox"/>



# TS EN 868-8 STANDARD ANNEX-F/G TESTS VALIDATION REPORT

Document No : 85AYA.PR.07-F03  
First Release Date: 10.07.2020  
Rev. No : 00  
Rev. Date :-

Report No: AYA-VR080621-01 Rev.02

08 June 2021

Table 5.4 Datalogger Temperature-Pressure Distribution Results - Full Size Container Bottom Non-Perforated

Product Name	<u>Full Size Container Bottom Non-Perforated</u>		
Cycle No	1st cycle	2nd cycle	3rd cycle
Autoclave Working Temperature and Time	121°C / 15 min.	121°C / 15 min.	121°C / 15 min.
Test Date	26.05.2021	26.05.2021	26.05.2021
Start Time	16:00:00	17:25:00	18:55:00
Exposure Start Time	16:10:30	17:38:30	19:07:00
Exposure Finish Time	16:25:30	17:53:30	19:22:00
Finish time	16:50:00	18:15:00	19:45:00
Exposure Time	00:15:00	00:15:00	00:15:00
<u>Vacuum Drying Pressure (Bar)</u>	-0.698	-0.700	-0.701
<u>Vacuum Drying Temperature (°C)</u>	89.5	88.7	89.4
<u>Vacuum Drying Time</u>	15 min.	15 min.	15 min.
Max. Temperature (After Stabilization)	123,63	123,85	123,56
Min. Temperature (After Stabilization)	121,7	121,9	122,35
Deviation Between Datalogger After Stabilization Period	1,93	1,95	1,21
Compliance with Acceptance Criteria	Suitable <input checked="" type="checkbox"/> Not Suitable <input type="checkbox"/>	Suitable <input checked="" type="checkbox"/> Not Suitable <input type="checkbox"/>	Suitable <input checked="" type="checkbox"/> Not Suitable <input type="checkbox"/>

Datalogger Temperature-Pressure Distribution Results Validation Report can be seen in ANNEX-3.2.

## ANNEX-3: Annex-F Determination of Sterilization Performance

- 3.1. Biological Indicator Test Results
- 3.2. Datalogger Temperature-Pressure Distribution Results
- 3.3. Biological Indicator and Datalogger Layout
- 3.4. Device Outputs



# TS EN 868-8 STANDARD ANNEX-F/G TESTS VALIDATION REPORT

Document No : 85AYA.PR.07-F03  
First Release Date: 10.07.2020  
Rev. No : 00  
Rev. Date : -

Report No: AYA-VR080621-01 Rev.02

08 June 2021

## 5.2. Annex-G Dryness Determination Test

Validation Cycle	Number of Studies	Tests to Be Performed	Acceptance Criteria
Annex-G Dryness Determination Test	3	$\frac{m_3 - m_2}{m_2 - m_1} \times 100 \%$ <p><math>m_1</math>, is the mass of the Empty Container, in grams <math>m_2</math>, is the mass of the loaded Container before the cycle, in grams <math>m_3</math>, is the mass of the loaded Container at completion of the cycle, in grams.</p>	The calculated value should not exceed 1%.

**Table 5.5 Dryness Determination Test Results - Biobarrier Full Size Sterilization Container**

Product Name		Biobarrier Full Size Sterilization Container			
Sterilization No	$m_1$	$m_2$	$m_3$	Calculation	Result
1.Sterilization	4.881,17 g	5.181,09 g	5.178,2 g	$\frac{m_3 - m_2}{m_2 - m_1} \times 100 \%$	%0,9636
2.Sterilization	4.878,91 g	5.180,86 g	5.178,15 g	$\frac{m_3 - m_2}{m_2 - m_1} \times 100 \%$	%0,8975
3.Sterilization	4.877,58 g	5.179,87 g	5.177,56 g	$\frac{m_3 - m_2}{m_2 - m_1} \times 100 \%$	%0,7642

**Table 5.6 Dryness Determination Test Results - Full Size Container Bottom Non-Perforated**

Product Name		Full Size Sterilization Container			
Sterilization No	$m_1$	$m_2$	$m_3$	Calculation	Result
1.Sterilization	4.514,39 g	4.815,14 g	4.812,29 g	$\frac{m_3 - m_2}{m_2 - m_1} \times 100 \%$	%0,9476
2.Sterilization	4.513,04 g	4.815,17 g	4.812,55 g	$\frac{m_3 - m_2}{m_2 - m_1} \times 100 \%$	%0,8672
3.Sterilization	4.512,87 g	4.814,14 g	4.811,7 g	$\frac{m_3 - m_2}{m_2 - m_1} \times 100 \%$	%0,8099



# TS EN 868-8 STANDARD ANNEX-F/G TESTS VALIDATION REPORT

Document No : 85AYA.PR.07-F03  
First Release Date: 10.07.2020  
Rev. No : 00  
Rev. Date : -

Report No: AYA-VR080621-01 Rev.02

08 June 2021

## 6. RESULT

When the Validation Study Results were evaluated, it was observed that TS EN 868-8 Standard Annex-F / Annex-G Tests were met with all parameters.

## 7. REFERENCES

TS EN 868-8	:	Packaging for terminally sterilized medical devices - Part 8: Re-usable sterilization containers for steam sterilizers conforming to EN 285 - Requirements and test methods
TS EN 285	:	Sterilization - Steam Sterilizers - Large sterilizers
TS EN ISO 13485	:	Medical Devices-Quality Management Systems-Conditions for Legislative Purposes
TS EN ISO 11138-1	:	Sterilization of health care products-Biological Indicators-Part 1: General requirements
TS EN ISO 11138-3	:	Sterilization of Health Care Products - Biological Indicators - Part 3: Biological Indicators for Moist Heat Sterilization Processes
AYA.TL.3.43	:	BI Sterility Control Instruction

## 8. DEVIATIONS

Has there been a deviation?

YES

NO

Deviation No	Deviation Remedied?	Explanation

## 9. REVISION HISTORY

Revision No	Page Information	Revision Explained	Revision Date
01	5,7,8,11	1.) Added sample size, 2.) Vacuum drying pressure,temperature and time added. 3.) Device outputs added to attachments	31.08.2023
02	3,5,6,8,9	With the Customer Declaration, the product name Full Size Sterilisation Container 580x280x260 was revised as Full Size Container Bottom Non-Perforated 580x280x260.	23.01.2024

**AYA Validasyon Ambalaj Sterilizasyon A.Ş.**

Saray Mah. 106 Cad. No:14 Kahramankazan-Ankara • (+90 312) 800 01 05-06 • www.ayalab.com.tr • info@ayalab.com.tr

10/11



# TS EN 868-8 STANDARD ANNEX-F/G TESTS VALIDATION REPORT

Document No : 85AYA.PR.07-F03  
First Release Date: 10.07.2020  
Rev. No : 00  
Rev. Date :-

Report No: AYA-VR080621-01 Rev.02

08 June 2021

## 10. ANNEXES

**ANNEX-1 VALIDATION REQUEST FORM**

**ANNEX-2 BIOLOGICAL INDICATOR INFORMATION AND TEST REPORTS**

**ANNEX-3 ANNEX-F DETERMINATION OF STERILIZATION PERFORMANCE**

ANNEX-3.1 Biological Indicator Test Results

ANNEX-3.2 Datalogger Temperature-Pressure Distribution Results

ANNEX-3.3 Biological Indicator and Datalogger Layout

ANNEX-3.4 Device Outputs