

Detalii specifice și standarde tehnice

LOT: „Hepafiltre pentru hotel de siguranță biologică, Biobase”:

№	Name of product	Technical specifications	Filter dimensions, mm	Offered specifications, Model; Manufacturer
1	Filter for BSC class II A2, Biobase BSC-3FA2	<ul style="list-style-type: none"> BSC serial #EU3FAIA2010002 Biobase; model BSC-3FA2; man. year 2020 	Supply filter 470*925*69,910*480*69 Exhaust Filter 420*800*69, 600*430*69	HEPA FILTER (Supply filter 470*925*69mm,Exhaust Filter 420*800*69mm) ULPA FILTER (Supply filter 910*480*69mm,Exhaust Filter 600*430*69mm) Biobase Biodusty(Shandong), Co., Ltd.
Installation and Certification for all BSC where filters have been replaced			Installation and Certification for all BSC where filters have been replaced	
	Mandatory requirements (this certificates/letters must be included in the offer)	<ul style="list-style-type: none"> Manufacturer’s ISO 9001 for manufacturing and/or ISO 13485 certificate; Declaration of conformity (CE); European standards EN12469 or United States NSF/ANSI Standard 4-2008; All presented certificates should be valid; Original manufacturer packing for all items; Warranty: the remaining shelf life should be not less than 80% from the total one. 	<ul style="list-style-type: none"> Manufacturer’s ISO 9001 for manufacturing and ISO 13485 certificate; Attestation of conformity (CE); European standards EN12469 – 2000, EN 61010-1:2010/A1:2019 , NSF/ANSI Standard 4-2008 All presented certificates valid; Original manufacturer packing for all items; Warranty: the remaining shelf life should be not less than 80% from the total one. 	

- **Toate bunurile vor fi testate de beneficiar. Acceptarea bunurilor și semnarea actelor de recepție va avea loc după testarea pozitivă efectuată de beneficiar.**

Data completării 13.10.2023

Cu stimă,

Semnat: _____

Nume: Ermicev Alexandr

În calitate de: Director

Ofertantul: „LABROMED LABORATOR” SRL

Adresa: MD 2038, Chișinău, str. Trandafirilor, 15