



**Lorne Laboratories Ltd.**

Unit 1 Cutbush Park Industrial Estate  
Danehill, Lower Earley,  
Berkshire RG6 4UT, UK

**Document Number: F031**

**Revision Level: 01**

**Date: 31 AUG 2022**

**Title: CERTIFICATE OF ANALYSIS**

**Page Number: 1 of 1**

## Certificate of Analysis

(See LOR006 for Methodology)



<b>PRODUCT</b>	RPR Carbon Kit
<b>LOT NUMBER</b>	LO16287
<b>DATE OF EXPIRY</b>	2024-12
<b>STORAGE CONDITIONS</b>	Refrigerated at 2 – 8°C. Protect from light. Do not freeze.
<b>SHIPPING</b>	This product has data supporting stability tolerance during fluctuations in ambient shipping temperature

### QUALITY CONTROL:

REAGENT	SPECIFICATIONS			RESULT
	Appearance	Colour	Functionality	
RPR Carbon Kit	Liquid solution. Clear and transparent	Grey	Tested against kit (+) and (-) controls	<b>PASS</b>
Positive Control Negative Control	Liquid solution	Clear and transparent		<b>PASS</b>

Components from human origin have been tested and found negative for the presence of antibody to HIV as well as HBsAg and HCV. Handle with caution as potentially infectious  
This product was tested by methods described in the manufacturers package insert  
This product is intended for In Vitro Diagnostic use only

**We certify that this product has been released as meeting our acceptance criteria**

<b>QUALITY CONTROL STATEMENT</b>	This product is in compliance with <b>MEDICAL DEVICES REGULATIONS 2002</b> for Annex 3 and Self-Certification. This product was manufactured, packaged and tested in accordance with <b>LORNE QUALITY SYSTEMS</b> and meets all product specifications
----------------------------------	--

<b>Title</b>	<b>Quality Assurance Associate</b>
<b>Signature</b>	
<b>Date:</b>	<b>05 December 2022</b>