

Manufacturer Declaration Letter

Date: 06-May-2026

To Whom It May Concern:

We hereby clarify some specifications of electric hospital bed model YA-D5-11 as follow.

- Back section to be equipped with a frame for an X-ray cassette/detector $\geq 35 \times 43$ cm
- Maximum permitted load: 250 kg
- Frame for label on the headboard
- Height adjustable range: min. 45–80 cm (± 2 cm)
- Wheels: diameter ≥ 15 cm
- Orthopedic mattress, washable: yes (Model MK-M01); minimum thickness 20 cm
- Trendelenburg $15^\circ \pm 1$
- Anti-trendelenburg $15^\circ \pm 1$

Signed & Stamp:

Name: Eric CC

Title: Sales Manager

Dated: 06-MAR-2026

