

АНАЛИТИЧЕСКИЙ ПАСПОРТ

Набор реактивов для предстерилизационного контроля

Азопирам Ст

-- 0 5 2 4 ОКП 32.50.50.000

МАЙ 2024

Дата изготовления

Изготовитель НИИ ИТМ, г. Санкт-Петербург

№Показатель	Требования по ТУ	Результаты анализа
1. Внешний вид		
1.1. Реактив А амидофенил	Порошок белого цвета	00ТВЕТСТВ11
1.2. Реактив СА Солннокислый ангидрид	Порошок белого (от серого до светло-зеленого) цвета	00ТВЕТСТВ11
1.3. Гидроксид натрия	Порошок белого цвета	00ТВЕТСТВ11
2. Технические характеристики		
2.1. Чувствительность		
Азопирам Ст	1:100000	00ТВЕТСТВ11
Положительная реакция при разведении крови не менее		00ТВЕТСТВ11

Заключение: **требования ТУ**

00ТВЕТСТВ11

Начальник ОТМ  Гавриков К.Е.



СЕРТИФИКАТ КАЧЕСТВА


Настоящим удостоверяется, что товар, идентифицированный как суше компоненты реактива для определения скрытой крови при предстерилизационном контроле Азопирам Ст код:382200000 ТНВЭД

Страна происхождения: Россия
Изготовитель: НИИ ИТМ МАЙ 2024

Дата изготовления: МАЙ 2026
Годен до: МАЙ 2026
№ партии: -- 0 5 2 4

Соответствует требованиям:

- 1 ГОСТ 5822-78
- 2. СанПин № 8. 01. 013. 03
- 3. Методическим указаниям МЗ СССР 28-6/13 от 26,05,1988г.
- 4. ТУ 9398-004-90814321-2012

Подпись ответственного лица 



Instrument Training

Vital Scientific BV hereby declares that the participant has attended a four days seminar for service engineers and the participant is now a certified engineer for the declared instruments.

Participant: Mr. A. Legun

Company: Global Biomarketing Group-Moldova SRL
Moldova

Instrument: Vitalab: XL Series
E Series
Junior Series
Dry ISE
Micro Series
ProXS


Date of training: April 20th – April 23rd, 2010

System Support Manager:



Jan Oostendorp

System Support Engineer:



Frank v.d. Kerput

MEDICA

Medica Corporation
5 Oak Park Drive
Bedford, Massachusetts 01730
Tel 781 275 4892
Fax 781 275 2731
www.medicacorp.com

Declaration of Conformity

Product Name:

EasyLyte and accessories per attachment


EasyElectrolytes and accessories per attachment

Model/Type:

EasyLyte Na/K, Na/K/Cl, Na/K/Li, Na/K/Cl/Li,
Na/K/Ca/pH, Na/K/Cl/Ca/Li

EasyElectrolytes Na/K/Cl, Na/K/Li

Manufacturer

 Medica Corporation
5 Oak Park Drive, Bedford, Massachusetts, 01730, USA

Representative

 Emergo Europe, Prinsessegracht 20,
2514 AP The Hague, The Netherlands
Tel: +31 70 345 8570
Fax: +31 70 346 7299

Means of Conformity

Medica Corporation declares that the products listed are covered by Annex III of Directive 98/79/EC. These products are self-certified since they are for professional use only and are not listed on Annex II, List A or Annex II, List B of Directive 98/79/EC. In addition, they are in conformity with the Annex I, "Essential Requirements" and provisions of council Directive 98/79/EC for In Vitro Diagnostic Medical Devices, Directive 2011/65/EU Restriction of Hazardous Substance in Electrical and Electronic Equipment, and the corresponding national laws of the Member States.

Place and Date: Bedford, Massachusetts, USA, September 27, 2018

Signature:



Name: Photios Makris, Ph.D.
Title: VP, Regulatory Affairs

EasyLyte Accessories

Catalog No.	Accessory	EDMA Code
2004	EasyLyte Na/K Analyzer	21 07 11 02
2014	EasyLyte Plus Na/K/Cl Analyzer	21 07 11 02
2015	EasyLyte Lithium Na/K/Li Analyzer	21 07 11 02
2016	EasyLyte Calcium Na/K/Ca/pH Analyzer	21 07 11 02
2021	EasyLyte Na/K/Cl/Li Analyzer	21 07 11 02
2030	EasyLyte EXPAND Analyzer, Na/K/Cl/Ca-Li	21 07 11 02
2070	EasyLyte EasySampler	21 07 11 02
2101	EasyLyte K+ Electrode	11 04 01 06
2102	EasyLyte Na+ Electrode	11 04 01 07
2113	EasyLyte Cl- Electrode	11 04 01 03
2106	EasyLyte Li+ Electrode	11 04 01 04
2150	EasyLyte Ca++ Electrode	11 04 01 02
2151	EasyLyte pH Electrode	11 70 31 02
2152	EasyLyte Disposable Reference Electrode	11 04 04 01
2103	EasyLyte Reference Electrode	11 04 04 01
2258	EasyLyte Membrane Assembly	21 07 11 02
2120	EasyLyte Na/K 800 ml Solutions Pack	11 04 04 02
2121	EasyLyte Na/K/Cl 800mL Solutions Pack	11 04 04 02
2122	EasyLyte Na/K/Li 800mL Solutions Pack	11 04 04 02
2123	EasyLyte Na/K/Ca/pH 800mL Solutions Pack	11 04 04 02
2028	EasyLyte Na/K/Cl/Li 400mL Solution Pack	11 04 04 02
2109	EasyLyte Na/K 400mL Solutions Pack	11 04 04 02
2112	EasyLyte Na/K/Cl 400mL Solutions Pack	11 04 04 02
2115	EasyLyte Na/K/Li 400mL Solutions Pack	11 04 04 02
2114	EasyLyte Na/K/Ca/pH 400mL Solutions Pack	11 04 04 02
2026	EasyLyte Na/K/Cl/Li 800mL Solution Pack	11 04 04 02
2124	EasyLyte Na/K/Cl/Ca-Li 800ml Solutions Pack	11 04 04 02
2814	EasyQC Bi-Level Quality Control Kit	11 50 02 04
2815	EasyQC Tri-Level Quality Control Kit	11 50 02 04
2843	EasyLyte Quality Control Sample Cups (60)	21 07 11 02
2118	Daily Cleaning Solution Kit	11 01 01 27
2598	EasyLyte Daily Cleaner Cup	21 07 11 02
2108	EasyLyte Solutions Valve	21 07 11 02
2107	EasyLyte Sample Probe	21 07 11 02
2257	EasyLyte Sample Detector	21 07 11 02

EasyLyte Accessories, continued

Catalog No.	Accessory	EDMA Code
2104	EasyLyte Tubing Kit	21 07 11 02
2100	EasyLyte Calcium Tubing Kit	21 07 11 02
2492	EasyLyte Internal Filling Solution (125mL)	11 04 04 90
2309	EasyLyte Wash Solution (50mL)	11 04 04 90
2111	EasyLyte Urine Diluent (500mL)	11 04 04 90
2577	EasyLyte Standard Solution, Urine (50mL)	11 04 04 90
2323	EasyLyte Probe Wipers (6)	21 07 11 02
2541	EasyLyte Printer Paper (3 rolls)	21 07 11 02
2595	EasyLyte EasySampler Sample Cups, 500uL (500)	21 07 11 02
2596	EasyLyte Sample Cups 2.0mL (500)	21 07 11 02
10745	Anti-Evaporation Caps (500)	21 07 11 02
2293	EasyLyte Capillary Tubes	21 07 11 02
2590	EasyLyte Capillary Adaptor Kit	21 07 11 02
2292	EasyLyte Capillary Adaptor Cleaning Kit	21 07 11 02
2578	EasyLyte Red Dye Test Solution (50mL)	11 30 01 11
2572	EasyLyte Troubleshooting Kit	21 07 11 02
2571	EasyLyte Troubleshooting Kit (Na/K/Ca/pH and Na/K/Cl/Li)	21 07 11 02
2105	EasyLyte Quarterly Operating Kit	21 07 11 02
2095	EasyLyte Maintenance Kit	21 07 11 02
2076	EasyLyte Sample Tray	21 07 11 02
2074	EasyLyte Sample Cup Retainer Ring	21 07 11 02
7118	Daily Rinse/Cleaning Solution Kit	11 01 01 27
2544	EasyLyte C Series Printer Paper (5 rolls)	21 07 11 02
2934	EasyLyte Barcode Reader Kit	21 07 11 02

EasyElectrolytes Accessories

Catalog No.	Accessory	EDMA Code
4002	EasyElectrolyte Na/K/Cl Analyzer	21 07 11 02
4003	EasyElectrolyte Na/K/Li Analyzer	21 07 11 02
4102	Reagent Module, Na/K/Cl	11 04 04 02
4103	Reagent Module, Na/K/Li	11 04 04 02
7205	EasyElectrolyte/EasyStat Na+ Electrode	11 04 01 07
7206	EasyElectrolyte/EasyStat K+ Electrode	11 04 01 06
4203	EasyElectrolyte Cl- Electrode	11 04 01 03
4204	EasyElectrolyte Li+ Electrode	11 04 01 04
6204	EasyElectrolyte/EasyStat/EasyBloodGas Reference Electrode	11 04 04 01
4207	EasyElectrolyte Spacer Electrode	11 04 01 90
4301	EasyElectrolyte Troubleshooting Kit	21 07 11 02
2118	Daily Cleaning Solution Kit	11 01 01 27
4402	EasyStat/EasyBloodGas/EasyElectrolyte Red Test Dye Solution	11 30 01 11
4403	EasyElectrolyte Urine Diluent	11 04 04 90
2814	Bi-Level Quality Control Kit	11 50 02 04
2815	Tri-Level Quality Control Kit	11 50 02 04
4405	EasyElectrolyte Na/K/Cl Demonstration Kit	21 07 11 02
4406	EasyElectrolyte Na/K/Li Demonstration Kit	21 07 11 02
4404	EasyElectrolyte Capillary Tube Kit	21 07 11 02
4306	EasyElectrolyte Sampler	21 07 11 02
6504	EasyBloodGas/EasyElectrolyte Pump Tube	21 07 11 02
6505	EasyStat/EasyBloodGas/EasyElectrolyte Printer Paper	21 07 11 02
4506	EasyElectrolyte Sensor Module	21 07 11 02
4507	EasyElectrolyte Valve Module	21 07 11 02
4508	EasyStat/EasyBloodGas/EasyElectrolyte Compression Plate	21 07 11 02
7302	Probe Wipers	21 07 11 02
4522	EasyElectrolyte Daily Cleaner Sample Cups	21 07 11 02
4539	EasyElectrolyte Sensor Module, Li+	21 07 11 02
6537	EasyElectrolyte/EasyStat/EasyBloodGas Serial Cable, 9-pin	21 07 11 02
6520	EasyElectrolyte/EasyStat/EasyBloodGas Barcode Reader Kit	21 07 11 02

CERTIFICATE OF REGISTRATION

This is to certify that the management system of:

Medica Corporation

(FIN F002402)

Main Site: 5 Oak Park Drive, Bedford, Massachusetts, 01730, United States

Additional Site: 3 Oak Park Drive, Bedford, Massachusetts, 01730, United States

has been registered by Intertek, an MDSAP recognized auditing organization, as conforming to the requirements of:

ISO 13485:2016

Brazil: Federal Law n. 6360/76; RDC ANVISA n. 16/2013; RDC ANVISA n. 23/2012; RDC ANVISA n. 67/2009; RDC ANVISA n. 56/2001

Canada: Medical Devices Regulations – Part 1- SOR 98/282

United States: 21 CFR 820, 21 CFR 803, 21 CFR 806, 21 CFR 807 (Subparts A to D)

Japan: MHLW Ministerial Ordinance 169, Article 4 to Article 68; PMD Act

The management system is applicable to:

Design, Development, Manufacture, Service, Installation and Distribution of in-vitro diagnostic medical devices, in-vitro diagnostic test kits, in-vitro diagnostic reagents, in-vitro diagnostic analyzers/software used in diagnosis and management of cancer, immune status, disease status, autoimmune status, cardiac markers, protein metabolism, endocrine disorders, blood analytes, urinalysis, blood gases.

Certificate Number:

0089217-01

Initial Certification Date:

2019-04-19

Date of Certification Decision:

2022-03-24

Certification Effective Date:

2022-04-18

Certification Expiry Date:

2025-04-18



intertek

Calin Moldovean

President, Business Assurance

Intertek Testing Services NA, Inc.
900 Chelmsford Street
Lowell, MA, USA 01851



EasyBloodGas™ analyzer
EasyLyte® analyzer

EasyElectrolytes® analyzer
EasyStat® analyzer

Training Certificate

This is to certify that

Mr. Sergiu Sorocovici

Of GBG-MLD S.R.L.

has completed training for the operation and service of the

EasyBloodGas™ analyzer, EasyElectrolytes® analyzer, EasyLyte® analyzer and EasyStat® analyzer

04/22/2016
DATE



Medica Corporation

David Hagopian
Director of Technical Support

CERTIFICATE

ELECTROMAGNETIC COMPATIBILITY

Applicant : **Vital Scientific B.V.**
Contact person : **Mrs. C. v.d. Broek**
Address : **Van Rensselaerweg 4**
Postal code, Place : **6956 AV Spankeren/Dieren**
Country : **The Netherlands**

Manufacturer : **Vital Scientific B.V.**
Address : **Van Rensselaerweg 4**
Postal code, Place : **6956 AV Spankeren/Dieren**
Country : **The Netherlands**

Electrical apparatus : **Clinical Analyser**
Trademark : **Elitech Clinical Systems**
Type designations : **Flexor EL200, Selectra ProM**

Environment : **Laboratory**

EN 61326-1:2006 : Equipment for measurement, control and laboratory use
EN 61326-2-6:2006 : Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 2-6: particular requirements – In vitro diagnostic (IVD) medical equipment, from which:

EN 55011:2007 : Emission - Class A
+A2:2007

EN 61000-3-2:2006 : Limit for harmonic currents emissions
EN 61000-3-3:1995 : Limitation of voltage fluctuations and flicker
+A1:2001+A2:2005

EN 61000-4-2:1995 : Electrostatic discharge (ESD) immunity
A1:1998+A2:2001

EN 61000-4-3:2006 : Radiated Electro-Magnetic field immunity
+A1:2008

EN 61000-4-4:2004 : Electrical fast transient (EFT) immunity
EN 61000-4-5:2006 : Surge transient immunity
EN 61000-4-6:2007 : Conducted Radio-Frequency disturbances immunity
EN 61000-4-8:1993 : Power frequency magnetic field immunity
+A1:2001

EN 61000-4-11:2004 : Voltage dips and interruptions immunity

The undersigned declares that the described product meets the requirements of the mentioned standards, based on a non-recurrent examination. The test results lay down in our test reports with reference 2129388.0501-QUA/EMC and 2136226.0501-QUA/EMC.

KEMA Quality B.V.
(Notified Body EMC)
Arnhem, September 20, 2010

A.T. van der Meijden
Certification Manager EMC

Certificate nr. **2136226.0551-QUA/EMC**

Integral publication of this certificate and adjoining reports is allowed.

KEMA Quality B.V. Utrechtseweg 310, 6812 AR Arnhem P.O. Box 5185, 6802 ED Arnhem The Netherlands
T +31 26 3 56 20 00 F +31 26 3 52 58 00 www.kemaquality.com Registered Arnhem 09085396