

**ANALYSENZERTIFIKAT**  
**CERTIFICATE OF ANALYSIS**

**SERVALYT® PRECOTES® pH 4-6**  
**300 µm, 125x125 mm**

**Kat.Nr./Cat.No. : 42875**  
**Lot/Contr.No. : P230403**

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Beschreibung</b> description	Gebrauchsfertige Polyacrylamidgele mit Servalyt, die an eine Polyesterfolie gebunden sind. Gelformat: 125x125 mm, 300 µm Ready-to-use polyacrylamide gels with servalyte bound to polyester sheet. Gel format: 125x125 mm, 300 µm		
<b>pH-Servalyt</b>	2% in H <sub>2</sub> O, RT	5.20 ± 0.4	5.1
<b>Leitfähigkeit ( mS/cm )</b> conductivity	2% in H <sub>2</sub> O, RT	0.150 - 0.350	0.220
<b>Proteintrennung</b> protein separation	IEF, 300 µm-Gel, 5 °C, 5000 Vh	entspricht Standard conform reference	entspricht conform
<b>Haftung des Gels</b> adhesion of the gel	IEF, 300 µm-Gel, 5 °C, 5000 Vh	Haftung ja adhesion yes	ja yes
<b>Mindesthaltbarkeit</b> minimum shelf life			09.11.2026
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 13.11.2023**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.