Hanna Instruments SRL Romania, Salaj, Nusfalau, No. 1, Hanna Str. Tel.: +40/260/607700; Fax: +40/260/607739



Document No: RN-0147

## **CE DECLARATION OF CONFORMITY**

Hanna Instruments SRL - Nusfalau Science Park Romania, Salaj, Nusfalau, No. 1, Hanna Str. Tel.: +40/260/607700; Fax: +40/260/607739

This certifies that the following designated products

## HI 9829, HI98290

are in conformity with the EMC directive **2004/108/EC** and its standards, and Low Voltage Directive **2006/95/EC** and its standards:

EN 61326-1: 2013 Electrical Equipment for Measurement, Control and Laboratory use

- EMC requirements Part 1: General requirements and its standards:

EN 61000-4-2: 2009, EMC - Electrostatic Discharge Immunity Test

EN 61000-4-3: 2006, EMC - Radiated RF, Electromagnetic Field Immunity Test

EN 61000-4-4: 2012, EMC - Electrical Fast Transient/Burst Immunity Test

Class A

4 kV/8 kV contact/air 3 V/m (80 MHz to 1 GHz) 1 kV (5/50 ns, 5 kHz)

**EN 61010-1: 2010 (IEC 61010-1: 2010)** Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use Part 1: General Requirements.

Date of issue:

February, 03, 2014

Name and Function:

Marius Talpos elect. Eng.