

## **UKCA DECLARATION OF CONFORMITY**



We, Motic China Group Co., Ltd.

Declare that the product,

Microscope Model(s):

## Panthera C2 Series

We take full responsibility for the product's compliance on:

## **Electrical Equipment (Safety) Regulations 2016**

Standards to which conformity is declared:

BS EN 61010-1: 2010 Safety requirements for electrical equipment for measurement, control and laboratory use.

Part 1: General requirements

## **Electromagnetic Compatibility Regulations 2016**

Standards to which conformity is declared:

BS EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General

equirements

EN 61326-2-2:2013 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-2: Particular

requirements - Test configurations, operational conditions and performance criteria for portable test, measuring

and monitoring equipment used in low-voltage distribution Systems.

EN 61000-3-2:2018 Electromagnetic compatibility (EMC)-Part 3-2: Limits for harmonic current emissions (equipment input current

≤16 A per phase

EN 61000-3-3:2013 Electromagnetic compatibility (EMC)-Part 3: Limits-Section 3: Limitation of voltage changes, voltage fluctuations

and flicker in public low-voltage supply systems, for equipment with rated current≤16 A per phase and not

subject to conditional connection

Manufacturer: Motic China Group Co., Ltd.

Address: Motic building Torch Hi-Tech Industrial Development Zone, Xiamen, Fujian, 361006 China. Tel: (592) 5627866 Fax: (592) 5627855 Email: info@motic.com

Signature:

Name: Zhihe Zhang (Quality Manager)

Date: 13 September 2023