

Reinforced Gown Specifications

OPS ADVANCED	REINFORCED SURGICAL GOWN
Reference(s)	30.002.05
Description	<p>Reinforced Surgical Gown – XX Large</p> <p>1. Made of non-woven material - SMS;2. Repellent, SMS material thickness minimum 35 g/m²;3. Reinforced at the hands, the reinforcement material - polypropylene film and polyethylene, with a minimum thickness of 40 g/m²; reinforcement dimensions at sleeves with a minimum length of 40cm (±3cm);4. Reinforced in front the reinforcement material - polypropylene film and polyethylene, with a minimum thickness of 33g/m²; reinforcement dimensions in front from shoulders to thighs with a minimum length of 90cm (±3cm);5. Laser stitching to ensure impermeability to microbes and liquids, needle stitching, waterproofing in the area will be excluded reinforcement 100%;6. Sleeve length not less than 60cm (±3cm), length robe not less than 155cm (±3cm), from shoulders to thighs, the width of the robe in the axillary area minimum 75cm (±3cm);7. Individual packaging with 2 absorbent cellulose wipes;8. Packaging-sterilization with EO.</p>
Main Materials	- SMS (43 GR , %100 PP)
Packaging	Sterilization Pack 25x35
Manufacturer	<p>ZENMED TIBBİ CİHAZLAR VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD. ŞTİ.</p> <p>ISO 13485:2016</p> <p>Latex Free</p> <p>CE 1984</p> <p>STERILE EO</p>

ZENMED TIBBİ CİHAZLAR
VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD.ŞTİ.
Mahmutbey Mah. 2567. Sk. No:1 İş Yarı No:11,21
Bağcılar / İSTANBUL / TÜRKİYE
Tel: 0212 445 45 25 TİC. Sic. No: 794704
Güneşli Y.D.: 897 067 1942

Manufacturing Company:

ZENMED TIBBİ CİHAZLAR VE SAĞLIK ÜRÜNLERİ SAN.

Mahmutbey Mahallesi 2567. Sk. No: 1/11 Bağcılar-İstanbul

+90 212 445 45 25

+90 212 445 45 49

www.zenmed.com.tr




In compliance with Medical Devices Directive 93/42/EEC
HP: High performance according to EN 13795 standard
MDD class: I sterile



Standard Gown Specifications

OPS ADVANCED	STANDARD SURGICAL GOWN
Reference(s)	30.001.03
Description	Standart Surgical Gown –Large – 1 Unit 1.size: L2. non-woven material3 layers SMS resistance to penetration of liquids within the limit of 50-80 cm H2O according to the EN 20811:1981 test (test method of the European certification body)3. with elastic, knitted cuffs4. sterile5. individually packed6. neck adjustment7. Overlapping closure at the back with ties8. Length not less than 132 cm9. The robe corresponds to the EN 13795 standard, the stitching – with laser.
Main Materials	%100 PP SMS / HYDROPHOBIC
Packaging	Sterilization Pack 25x35
Manufacturer	ZENMED TIBBİ CİHAZLAR VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD. ŞTİ. ISO 13485:2016
Regulation	In compliance with Medical Devices Directive 93/42/EEC HP: High performance according to EN 13795 standard MDD class: I sterile

ZENMED TIBBİ CİHAZLAR
VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD.ŞTİ.
Mahmutbey Mah. 2567.Sk./No:1 İş Yeri No:11,21
Bağcılar / İSTANBUL / TÜRKİYE
Tel: 0212 445 45 25 Tic.Sic.No: 794704
Güneşli M.D.: 997 067 1942

Latex Free CE 1984   

Manufacturing Company:

ZENMED TIBBİ CİHAZLAR VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD. ŞTİ

Mahmutbey Mahallesi 2567. Sk. No: 1/11 Bağcılar-İstanbul/TURKEY

+90 212 445 45 25




+90 212 445 45 40

www.zenmed.com.tr



OPS ADVANCED	STANDARD SURGICAL GOWN
Reference(s)	30.001.02
Description	Impervious Gown -1 Unit 1.size: M2. Non-woven material3 layers SMS resistance to penetration of liquids within the limit of 50-80 cm H2O according to the EN 20811:1981 test (test method of the European certification body)3. with elastic, knitted cuffs4. sterile5. individually packed6. neck adjustment7. Overlapping closure at the back with ties8. Length not less than 125 cm9. The robe corresponds to the EN 13795 standard, the stitching – with laser.
Main Materials	-30 gsm Mavi Filik NW+ 2 gsm Tutkal + 24 gsm Mavi NBRT Film -30 gsm Blue Hydrophilic NW + 2 gsm Hotmelt + 24 gsm Blue NBRT Film
Packaging	Sterilization Pack 25x35
Manufacturer	ZENMED TIBBİ CİHAZLAR VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD. ŞTİ. ISO 13485:2016
Regulation	In compliance with Medical Devices Directive 93/42/EEC HP: High performance according to EN 13795 standard MDD class: I sterile

**ZENMED TIBBİ CİHAZLAR
VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD.ŞTİ.**
Mahmutbey Mah. 2567. Sk. No:1 İş Yeri No:11,21
Bağcılar / İSTANBUL / TÜRKİYE
Tel: 0212 445 45 25 Tic.Sic.No: 794704
Güneşli M.D.: 997 067 1942

Latex Free CE 1984   

Manufacturing Company:

ZENMED TIBBİ CİHAZLAR VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD. ŞTİ.

Mahmutbey Mahallesi 2567. Sk. No: 1/11 Bağcılar-İstanbul/TURKEY

+90 212 445 45 25

+90 212 445 45 40




www.zenmed.com.tr



Standard Gown Specifications

OPS ADVANCED	STANDARD SURGICAL GOWN
Reference(s)	30.001.01
Description	Surgical Surgical Gown –Small– 1 Unit 1.size: S2. Non-woven material 3 layers SMS resistance to penetration of liquids within the limit of 50-80 cm H2O according to the test EN 20811:1981 (test method of the certification bodyEuropean) 3. with elastic, knitted cuffs 4. sterile 5. individually packed 6. neck adjustment 7. Overlapping closure at the back with ties 8. Length not less than 117 cm 9. The robe corresponds to the EN 13795 standard, the stitching – with laser.
Main Materials	%100 PP SMS / HYDROPHOBIC
Packaging	Sterilization Pack 25x35
Manufacturer	ZENMED TIBBİ CİHAZLAR VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD. ŞTİ. ISO 13485:2016
Regulation	In compliance with Medical Devices Directive 93/42/EEC HP: High performance according to EN 13795 standard MDD class: I sterile

**ZENMED TIBBİ CİHAZLAR
VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD.ŞTİ.**
Mahmutbey Mah. 2567. Sk. No:1 İş Yeri No:11,21
Bağcılar / İSTANBUL / TÜRKİYE
Tel: 0212 445 45 25 Tic.Sic.No: 794704
Güneşli YD: 997 067 1942

Latex Free CE 1984   

Manufacturing Company:

ZENMED TIBBİ CİHAZLAR VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD. ŞTİ

Mahmutbey Mahallesi 2567. Sk. No: 1/11 Bağcılar-İstanbul/TURKEY

+90 212 445 45 25

+90 212 445 45 49




www.zenmed.com.tr



Standard Gown Specifications

OPS ADVANCED	STANDARD SURGICAL GOWN
Reference(s)	30.001.04
Description	Standart Surgical Gown –Extra Large– 1 Unit 1.size: XL2. Non-woven material 3 layers SMS resistance to penetration of liquids within the limit of 50-80 cm H2O according to the test EN 20811:1981 (test method of the certification body European)3. with elastic, knitted cuffs4. sterile5. individually packed6. neck adjustment7. Overlapping closure at the back with ties8. Length not less than 140 cm9. The robe must comply with the EN 13795 standard or the equivalent, the stitching - with laser.
Main Materials	%100 PP SMS / HYDROPHOBIC
Packaging	Sterilization Pack 25x35
Manufacturer	ZENMED TIBBİ CİHAZLAR VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD. ŞTİ. ISO 13485:2016
Regulation	In compliance with Medical Devices Directive 93/42/EEC HP: High performance according to EN 13795 standard MDD class: I sterile

**ZENMED TIBBİ CİHAZLAR
VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD. ŞTİ.**
Mahmutbey Mah. 2567. Sk. No: 1 İş Yeri No: 11,21
Bağcılar / İSTANBUL / TÜRKİYE
Tel: 0212 445 45 25 T.C.Sic.No: 794704
Güneşli YD.: 997 067 1942

Latex Free   

Manufacturing Company:

ZENMED TIBBİ CİHAZLAR VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD. ŞTİ

Mahmutbey Mahallesi 2567. Sk. No: 1/11 Bağcılar-İstanbul/TURKEY

+90 212 445 45 25

+90 212 445 45 40




www.zenmed.com.tr



Standard Gown Specifications

OPS ADVANCED	STANDARD SURGICAL GOWN
Reference(s)	30.001.05
Description	Surgical Surgical Gown –XX Large– 1 Unit 1.size: XXL2. Non-woven material 3 layers SMS resistance to penetration of liquids within 50-80 cm H2O according to the test EN 20811:1981 (test method of the European certification body) 3. with elastic, knitted cuffs4. sterile5. individually packed6. neck adjustment7. Overlapping closure at the back with ties8. Length not less than 150 cm9. The robe must comply with the EN 13795 standard or the equivalent, the stitching – with laser.
Main Materials	%100 PP SMS / HYDROPHOBIC
Packaging	Sterilization Pack 25x35
Manufacturer	ZENMED TIBBİ CİHAZLAR VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD. ŞTİ. ISO 13485:2016
Regulation	In compliance with Medical Devices Directive 93/42/EEC HP: High performance according to EN 13795 standard MDD class: I sterile

**ZENMED TIBBİ CİHAZLAR
VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD.ŞTİ.**
Mahmutbey Mah. 2567.Sk.No:1 İş Yeri No:11,21
Bağcılar / İSTANBUL / TÜRKİYE
Tel: 0212 445 45 25 E.Sic.No: 794704
Güneşli Y.D.: 997 067 1942

Latex Free   

Manufacturing Company:

ZENMED TIBBİ CİHAZLAR VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD. ŞTİ

Mahmutbey Mahallesi 2567. Sk. No: 1/11 Bağcılar-İstanbul/TURKEY

+90 212 445 45 25

+90 212 445 45 49

www.zenmed.com.tr

