

Specificații tehnice

[Acest tabel va fi completat de către ofertant în coloanele 2, 3, 4, 6, 7, iar de către autoritatea contractantă – în coloanele 1, 5,]

Numărul procedurii de achiziție ocds-b3wdp1-MD-1741858933887 din 13.03.2025
Obiectul achiziției: <u>Piese de schimb pentru Dispozitivele Medicale, pentru anul 2025</u>

Denumirea bunurilor	Denumirea modelului bunurilor	Țara de origine	Producătorul	Specificarea tehnică deplină solicitată de către autoritatea contractantă	Specificarea tehnică deplină propusă de către ofertant	Standarde de referință
1	2	3	4	5	6	7
Lotul nr. 23 - Piese de schimb pentru monitor de pacient Comen C50						
Piese de schimb pentru monitor de pacient Comen C50	Piese de schimb pentru monitor de pacient Comen C50	China	SHENZHEN COMEN MEDICAL INSTRUMENTS CO., LTD.	pompa de aer pompa de aer, mini pump CJP37 C12B1, DC:12V ON0815CC, WRF-370CA-12V15, compatibila cu monitor de pacient Comen C50	Pompa de aer, mini pump CJP37 C12B1, DC:12V ON0815CC, WRF-370CA-12V15, compatibila cu monitor de pacient Comen C50.	
Lotul nr. 56 - Piese de schimb pentru Generator de Oxigen IM-GO 400						
Piese de schimb pentru Generator de Oxigen IM-GO 400	Piese de schimb pentru Generator de Oxigen IM-GO 400	Turcia	IDEAL MAKINA END. UEUN. SAN. VETIC A.S.	Kit de mentenanță *Filtru de aer INVERSYS 55+(12067) Buc 1 *Filtru de ulei INVERSYS55+ (922) Buc 1 *Filtru panoului de admisie INVERSYS55+ (848) Buc 1 *Ulei pentru compresor / SmartOil INVERSYS55+(R46) (5 Litri canistra) Litri 40 *Separator INVERSYS55+(965) Buc 1 *ORING (4012) Buc 1 *IDEL 32075M FILTER ELEMENT FOR (AF0476XM FILTER) Buc 1 *IDEAL 32075S FILTER ELEMENT FOR(AF0476XS FILTER) Buc 1 *IDEAL 32075XS FILTER ELEMENT FOR (AF0306A2 ADSORBER FILTER) Buc 1 *IDEAL 32075XS FILTER ELEMENT FOR (AF0306A2 ADSORBER FILTER) Buc 1 *WALKER E30412SR FILTER ELEMENT (FOR A3051MS FILTER) Buc 1 *IDEAL 7050P FILTER ELEMENT(FOR AF0076P FILTER) Buc 1	Kit de mentenanță *Filtru de aer INVERSYS 55+(12067) Buc 1 *Filtru de ulei INVERSYS55+ (922) Buc 1 *Filtru panoului de admisie INVERSYS55+ (848) Buc 1 *Ulei pentru compresor / SmartOil INVERSYS55+(R46) (5Litri canistra) Litri 40 *Separator INVERSYS55+(965) Buc 1 *ORING (4012) Buc 1 *IDEL 32075M FILTER ELEMENT FOR (AF0476XM FILTER) Buc 1 *IDEAL 32075S FILTER ELEMENT FOR(AF0476XS FILTER) Buc 1 *IDEAL 32075XS FILTER ELEMENT FOR (AF0306A2 ADSORBER FILTER) Buc 1 *IDEAL 32075XS FILTER ELEMENT FOR (AF0306A2 ADSORBER FILTER) Buc 1 *WALKER E30412SR FILTER ELEMENT (FOR A3051MS FILTER) Buc 1	

				*AAF 0476 HOUSING+ 32075 FILTER ELEMENT Buc 1 *18558 MIN.PRES. VALVE KIT MPV. (TO9000)Buc 1 *12426 INTAKE VALVE KIT - 1 buc. (TO9000) Buc 1 *12464 THER. VALVE KIT YLF - 8 Buc 1 *12430 INTAKE VALVE S. VALVE - 1 buc. (TO9000) Buc 1	*IDEAL 7050P FILTER ELEMENT (FOR AF0076P FILTER) Buc 1 *AAF 0476 HOUSING + 32075 FILTER ELEMENT Buc 1 *18558 MIN.PRES. VALVE KIT MPV. (TO9000) Buc 1 *12426 INTAKE VALVE KIT - 1 Buc. (TO9000) *12464 THER. VALVE KIT YLF - 8 Buc 1 *12430 INTAKE VALVE S. VALVE - 1 Buc. (TO9000)	
TOTAL						

Semnat: _____

Numele, Prenumele: Dimitrovici Olga
 În calitate de: Administrator
 Ofertantul: "Romedcom" SRL
 Adresa: Mun.Chişinău, str.Sciusev 16 (of)6