

STORZ

KARL STORZ—ENDOSKOPE

en Instructions for use
ENDOFLATOR 40



09-2022

Copyright ©

All product illustrations, product descriptions, and texts are the intellectual property of
KARL STORZ SE & Co. KG.

Their use and reproduction by third parties require the express approval of KARL STORZ SE & Co. KG.

All rights reserved.

Table of contents

1 General information	6
1.1 Read the instructions for use	6
1.2 Read the instructions for use of compatible products	6
1.3 Scope	6
1.4 General signs and symbols	6
1.5 Description of warning messages	7
2 Normal use	8
2.1 Intended purpose	8
2.2 Indications	8
2.3 Contraindications	8
2.4 Application risks	8
2.5 Target user populations	8
2.6 Patient population	8
2.7 Intended conditions of use	9
3 Safety and warning	10
3.1 Serious incidents	10
3.2 Correct handling and product testing	10
3.3 Product not clean	10
3.4 Combination with other components	10
3.5 Damage due to ingress of liquid in electrical components	11
3.6 Dangers from electrical current	11
3.7 Risk of explosion and fire	11
3.8 Electromagnetic interference	12
3.9 Observing ambient conditions	12
3.10 Failure of products	12
3.11 Functionality of the touch screen	12
3.12 Safety features	12
3.12.1 Self-test	12
3.12.2 Monitoring during operation	13
3.12.3 Passive safety measures	13
4 Product description	14
4.1 Description of operation	14
4.2 Product overview	14
4.3 Possible combinations	16
4.4 Technical data	17
4.5 Symbols employed	17
4.5.1 Symbols on the packaging	17
4.5.2 Symbols on the product	19
4.5.3 Symbols on the type plate	19
4.5.4 Symbols on the user interface	20
4.6 Ambient conditions	22
5 Preparation	23
5.1 Unpacking the product	23
5.2 Setting up the product	23
5.3 Installing the bottle holder (option)	23

5.4	Connecting the product	24
5.5	Connecting a gas bottle	25
5.6	Putting the product into operation	27
5.7	Selecting the mode	29
5.8	Connecting the insufflation filter and insufflation tube	31
5.9	Checking the functioning and sealing before use	32
6	Application	34
6.1	Working area	34
6.2	Setting operating parameters	34
6.3	Performing insufflation	35
6.4	Ending the insufflation	36
6.5	Settings	36
6.5.1	Setting the language	37
6.5.2	Adjusting the volume	37
6.5.3	Selecting the source of gas supply	38
6.5.4	Activating the safety valve	39
6.5.5	Device information	39
6.5.6	System log	39
6.5.7	Administration	40
6.5.8	Auditor	43
6.5.9	Service	45
6.6	Alarm and information signals	45
6.6.1	Alarm specification	46
6.6.2	Information signals	49
6.6.3	Checking the excess pressure alarm	50
6.6.4	Checking the gas empty alarm	51
6.6.5	Checking information signals	51
7	Maintenance, servicing, repairs, and disposal	52
7.1	Maintaining the product	52
7.1.1	Maintenance	52
7.2	Changing a fuse	52
7.3	Safety inspection in accordance with IEC 62353	54
7.3.1	Visual inspection	54
7.3.2	Electric measurements	54
7.3.3	Functional test	54
7.4	Repairing the product	55
7.5	Disposing of the product	55
8	Accessories and spare parts	57
8.1	Accessories	57
8.2	Spare parts	57
9	Electromagnetic compatibility	58
9.1	General notes on the operating environment	58
9.2	Accessories and cables	58
9.3	Table 1 – Compliance level for immunity tests	59
9.4	Table 2 – Test levels for near fields from HF wireless communications equipment	60
9.5	Table 3 – Test levels for radiated and conducted immunity tests	61
9.6	Table 4 – Emission class and group	62

9.7 Table 5 – Recommended separation distances between portable and mobile HF communications devices and the product	62
10 Errors and messages	64
10.1 Troubleshooting	64
11 Overview of mitigating warnings	65
12 Subsidiaries	66

1 General information

1.1 Read the instructions for use

If the instructions for use are not followed, patients, users, and third parties may be injured or the product may be damaged.

- ▶ Read the instructions for use carefully and follow all the safety notes and warnings.
- ▶ Read the reprocessing instructions carefully and follow all the safety notes and warnings. The reprocessing instructions can be downloaded from www.karlstorz.com/ifu by entering the item number.
- ▶ Keep the instructions for use and reprocessing instructions in a safe place.

1.2 Read the instructions for use of compatible products

If the instructions for use of compatible products are not followed, patients, users, and third parties may be injured or the product may be damaged.

- ▶ Read the instructions for use of the compatible products carefully and follow all the safety notes and warnings.
- ▶ Read the reprocessing instructions of the compatible products carefully and follow all the safety notes and warnings.

1.3 Scope

This instruction manual is valid for:

Product name	Item number
ENDOFLATOR 40	UI400
Insufflation Tube	UI004

The products listed here may not yet be available in all countries due to differences in approval requirements.

1.4 General signs and symbols

The signs and symbols used in this document have the following meaning:

Practical tip

-  This sign refers to useful and important information.

Actions to be performed

Action to be carried out by several steps:

- ✓ Prerequisite that must be met before carrying out an action.
- 1. Step 1
 - ⇒ Interim result of an action
- 2. Step 2
 - ⇒ Result of a completed action

Actions in safety notes or in the case of a single step:

- ▶ Step 1

Lists

1. Numbered list
- Unnumbered list, 1st level
- Unnumbered list, 2nd level

1.5 Description of warning messages

To prevent any injury to persons or damage to property, the warnings and safety notes in the instructions for use must be observed. The warnings use the following levels of danger:

▲ WARNING**WARNING**

Designates a possible imminent risk. If this is not avoided, it could lead to death or serious injuries.

▲ CAUTION**CAUTION**

Designates a possible imminent risk. If this is not avoided, it could lead to minor injuries.

NOTICE**NOTICE**

Designates a possibly harmful situation. If this is not avoided, the products could be damaged.

2 Normal use

2.1 Intended purpose

ENDOFLATOR 40

Insufflators are intended to deliver CO₂ for insufflation (creating and maintaining a cavity) or replacement of ambient air in laparoscopy, thoracoscopy, transanal endoscopy, and endoscopic vessel harvesting. Insufflators are non-invasive and meant for short-term use.

Insufflation tube

Tubing sets for insufflation are intended for transfer of CO₂ from the insufflator to the patient. Tubing sets for insufflation are non-invasive and meant for short-term use.

2.2 Indications

The insufflator and its accessories are suitable for endoscopic examinations where CO₂ is required for insufflation.

This includes:

- Laparoscopy
- Thoracoscopy
- Transanal surgery
- Endoscopic Vessel Harvesting

2.3 Contraindications

The insufflator and its accessories must not be used for interventions in direct contact with the CNS (central nervous system) and central circulatory system. Furthermore, there are no contraindications for the use of the insufflator and its accessories directly associated with the product.

2.4 Application risks

The following risks have been identified in clinical literature in connection with insufflation:

- Mild hypothermia
- CO₂ embolism
- Emphysema requiring treatment
- Systemic acidosis
- Pneumothorax
- Hypotension
- Other pulmonary complications
- Other cardiovascular complications
- Postoperative ileus

2.5 Target user populations

The medical device may only be used by doctors and medical assistants with a relevant specialist qualification.

2.6 Patient population

There are no restrictions in terms of patient groups for this product.

2.7 Intended conditions of use

The product may only be used in hospitals and doctors' offices in suitable ambient conditions.

Condition	Operation
Frequency of use	One or more times a day
Length of use	Several minutes to several hours a day
Place of installation	Positioning on a level, vibration-free surface above patient height
Mobility	Can be moved if placed on a cart.
Combination	Can be used on the patient at the same time as other devices.
Control	Can be controlled via the KARL STORZ SCB.
Site	Placed outside of the sterile area.

3 Safety and warning

⚠ WARNING

Danger due to non-observance of warnings and safety notes

This chapter contains warnings and safety notes structured according to hazards and risks.

- ▶ 1. Carefully read and observe all warnings and safety notes.
- ▶ 2. Follow the instructions.

3.1 Serious incidents

A 'serious incident' includes incidents which, directly or indirectly, had, could have had or could have any of the following consequences:

- Death of a patient, user, or another person
- Temporary or permanent serious deterioration in the medical condition of a patient, user, or another person
- A serious threat to public health
- ▶ The manufacturer and appropriate authority must be notified of all serious incidents.

3.2 Correct handling and product testing

If the product is not handled correctly, patients, users, and third parties may be injured.

- ▶ Only persons with the necessary medical qualification and who are acquainted with the application of the product may work with it.
- ▶ Check that the product is suitable for the procedure prior to use.
- ▶ Check the product for the following properties, for example, before and after every use:
 - Functionality
 - Damage
 - Changes to the surface
 - In the case of several components: completeness and correct assembly
- ▶ Do not continue to use damaged products.
- ▶ Dispose of the product properly.

3.3 Product not clean

The product is not clean when delivered. The use of products that have not been cleaned poses a risk of infection for patients, users, and third parties.

- ▶ Reprocess the product in line with the reprocessing instructions before initial use and every subsequent use.

3.4 Combination with other components

The use of unauthorized devices and components may result in injuries.

- ▶ Ensure that any additional devices connected to electrical medical devices comply with the relevant IEC or ISO standards.
- ▶ Ensure that all configurations comply with the requirements for medical electrical systems.
- ▶ Only combine the product with devices and components that the manufacturer has approved for combined use, see chapter *Possible combinations*.

- ▶ Only use devices and components that have standardized interfaces and do not breach the normal use of the product.
- ▶ When using RF devices and accessories, it is important to verify normal operation in the configuration in which they will be used.
- ▶ Do not make any modifications to the product.
- ▶ Comply with national and local regulations.

3.5 Damage due to ingress of liquid in electrical components

In the case of electrical products, individual components or the product itself may be live. Liquid ingress into an electrical product may result in a short circuit or an unintentional transfer of current. The product is damaged as a result and patients, users and third parties may be injured.

- ▶ Do not store liquids near the product or on the product.
- ▶ If liquid has entered the product, pull out the plug and allow the product to dry completely.

3.6 Dangers from electrical current

An improper power supply may cause an electric shock and injure patients, users, or third parties.

- ▶ Ensure that all electrical installations of the operation room in which the product is connected and used conform with the applicable IEC standards.
- ▶ Use either the power cord supplied by KARL STORZ or a power cord which has the same properties and which bears a national mark of conformity.
- ▶ The product may only be operated with the line voltage stated on the rating plate.
- ▶ Position the product appropriately so that the power cord can be unplugged at any time. The product is only voltage-free when the mains plug has been disconnected.
- ▶ The product's ground line should be installed by a qualified electrician.
- ▶ Ensure potential equalization according to the applicable national rules and regulations.
- ▶ To ensure reliable protective earth grounding, connect the product to a properly installed socket that is approved for use in the operation room.
- ▶ Connect the product to a power supply with protective conductor.

In the case of electrical products, individual components or the product itself may be live. Live parts can cause electric shocks in the event of contact and injure patients, users, and third parties.

- ▶ Do not open the product.
- ▶ Have servicing carried out by KARL STORZ or a company authorized by KARL STORZ.
- ▶ Do not touch the output jacks of the product and the patient at the same time during use
- ▶ Always pull out the mains plug before carrying out any cleaning and maintenance work.


If several products supplied with energy are used simultaneously, the patient leakage currents accumulate. These leakage currents can exceed the limit values and injure patients.

- ▶ Only use products of the same type, for example, endotherapy device and application part of type CF.

3.7 Risk of explosion and fire

The product can generate sparks, which cause combustible or flammable gases and liquids to ignite or explode. This may cause injuries to patients, users, and third parties.

- ▶ When using explosive narcotic gases: Operate the product outside of the hazard zone.



- ▶ Do not use the product in the presence of flammable anesthetics.
- ▶ The product must not be operated in oxygenated environments.
- ▶ Only connect or disconnect the power plug to or from the power supply outside explosive atmospheres.

3.8 Electromagnetic interference

Medical electrical devices are subject to special precautions regarding electromagnetic compatibility. If other devices (e.g., MRT, CT, diathermy, electrocautery, or RFID) emit electromagnetic radiation, the product's functionality may be impaired. High-frequency (HF) communication equipment can affect electrical medical devices and impair their performance.

- ▶ During installation and operation of the product, please take note of the information on electromagnetic compatibility, see chapter *Electromagnetic compatibility*.

3.9 Observing ambient conditions

If the device is stored, transported, operated or reprocessed under unsuitable conditions, patients, users or third parties may be injured and the device can be damaged.

- ▶ Observe the ambient conditions listed in the instructions for use and reprocessing.

3.10 Failure of products

The product may fail during use.

- ▶ Have a replacement product ready for each application or plan for an alternative surgical technique.

3.11 Functionality of the touch screen

If the functionality of the touch screen is limited, the product cannot be used correctly. Patients, users, and third parties may be injured.

- ▶ Do not use the product if the touch screen is defective.
- ▶ Do not tap the touch screen in several places at the same time.

3.12 Safety features

3.12.1 Self-test

During power-up, the device performs a self-test when it is supplied with gas. If the self-test is negative, a corresponding error message is displayed and the device cannot be used properly.

The following tests are performed:

- Tightness test of the system (valves)

- Test of the pressure measurement
- Test of the flow measurement

3.12.2 Monitoring during operation

The following parameters are monitored during operation:

- Excess pressure

As soon as the intracorporeal pressure exceeds the specified setpoint value for more than 3 seconds during insufflation, an acoustic signal sounds and the gas supply is stopped. The excess pressure is decreased after 4 seconds when the safety valve is active.

- Operating mode-specific pressure
- Remaining gas supply

If there is no CO₂ gas when the device is switched on, a warning signal sounds. A visual and acoustic alarm is output when the gas bottle is empty and the insufflation stops.


- Gas preheater

The internal gas preheating is monitored by a second circuit, independently of the control. If a fault arises, the heating is switched off.

3.12.3 Passive safety measures

The product is equipped with the following safety measures:

- The pressure and flow recording are redundant.
- Both pressure control stages and the flow recording are equipped with safety pressure release valves.
- The function of the mechanical high pressure controller is monitored.

 If the monitoring system detects a fault that prevents the safe functioning of the product, the product is put into "safe mode". The product can only be reset from this mode when the power supply is switched off.

4 Product description

4.1 Description of operation

The ENDOFLATOR 40 is an insufflation device for universal application in laparoscopic and thoracoscopic examinations and operations as well as transanal endoscopy and endoscopic vessel harvesting. The pressure and flow measurements and the pressure and flow controls can be used in different operating modes tailored to specific applications. The High Flow and Pediatrics modes can be selected.

For especially sensitive applications, the Pediatrics mode is available. The mode offers control with low insufflation pressure and special safety limits in the pressure and flow area. At low flow rates, a tighter control of the flow rate is possible.

The ENDOFLATOR 40 is designed for a high flow rate of up to 40 l/min for quick compensation of heavy gas losses during complicated laparoscopic operations.

The ENDOFLATOR 40 is operated directly on the touchscreen or from an external control center via the SCB interface. All data during the current operation is displayed simultaneously on the touch screen.

4.2 Product overview



ENDOFLATOR 40 (UI400) – Front view

- 1 Power switch
- 2 Touch screen

- 3 Insufflation connection to the patient



ENDOFLATOR 40 (UI400) – Rear view

- | | |
|---|--------------------|
| 1 Gas connection (standard American type) | 4 Power socket |
| 2 SCB sockets | 5 Line fuse holder |
| 3 Potential equalization connector | |



Insufflation tube (UI004)



Holder for CO₂ cylinder, optional (UI005)

4.3 Possible combinations

It is recommended that the suitability of the products for the intended procedure be checked prior to use. Please note that the products listed here may not yet be available in all countries due to differences in approval requirements.

The product can be combined with gas filters or gas filters with connected insufflation tube. The product can also be combined with CO₂ supply tubes via the connection for high and low pressure tubes.

ENDOFLATOR 40	Gas Filters and Insufflation Tubes	High and Low-Pressure Tubes
UI400	031200-10 031122-25 UI004	UI020 UI021 UI022 UI023 UI024 UI025 UI026 UI027 UI028

Combinations with other products

Additional devices that support SCB can be connected via the SCB connection cable (20090170).

Expanded combinations

Can be combined with access systems with LUER-Lock connector.

Can be combined with access systems with HiCap connector.






4.4 Technical data













Designation	Value
Operating voltage (AC)	100 – 240 V
Operating frequency	50/60 Hz
Line fuse	2 x T 2.0 AH 250 V
Average power input during flow 40 l/min (input pressure 60 bar)	160 W
Maximum current consumption	1.6 A
Electrical protection class	I
Applied part type according to IEC 60601-1	CF
Degree of protection acc. to IEC 60529	IP 21
Gas input (maximum pressure)	100 bar
Dimensions (W x H x D)	305 x 164 x 315 mm
Weight	7.4 kg

Mode	Intracavitary pressure	Flow	Max. insufflation pressure
Pediatrics	1 – 15 mmHg Resolution: 1 mmHg	0.1 – 15 l/min Resolution: 0.1 – 2 l/min: 0.1 l/min 2 – 9.5 l/min: 0.5 l/min 10 – 15 l/min: 1 l/min	30 mmHg
High Flow	1 – 30 mmHg Resolution: 1 mmHg	1 – 40 l/min Resolution: 1 l/min	50 mmHg






4.5 Symbols employed

4.5.1 Symbols on the packaging








Symbol	Meaning
	Manufacturer
	Date of manufacture
	Medical device
	Article no.
	Serial number

Symbol	Meaning
	Batch code
	Number of products in the product packaging
	Unique Device Identifier
	Consult the printed or electronic instructions for use
	Unsterile
	Fragile, handle with care
	Keep dry
	Temperature limit
	Humidity limit
	Air pressure limit
	Federal (USA) law restricts this device to sale by or on the order of a physician.
	<p>CE marking</p> <p>With this marking, the manufacturer declares the conformity of the product with the applicable EU directives. A code number after the CE mark indicates the responsible notified body.</p> <p>The EU directives relevant to the product can be found in the EU Declaration of Conformity, which can be requested from KARL STORZ.</p>












4.5.2 Symbols on the product












Symbol	Meaning
	Follow the instructions for use. The color may differ on the product. The symbol is black/white on the packaging label.
	ON
	OFF
	Potential equalization connector
	Applied part of the type CF




4.5.3 Symbols on the type plate

Symbol	Meaning
	Manufacturer
	Date of manufacture
	Applied part of the type CF
	Federal (USA) law restricts this device to sale by or on the order of a physician.
	Prevention of pollution by electronic devices
	Separate collection of electrical and electronic devices. Do not dispose of in household refuse.
	CE marking With this marking, the manufacturer declares the conformity of the product with the applicable EU directives. A code number after the CE mark indicates the responsible notified body. The EU directives relevant to the product can be found in the EU Declaration of Conformity, which can be requested from KARL STORZ.

4.5.4 Symbols on the user interface

Symbol	Meaning
	Scroll back to the previous screen
	Call up settings
	Call up working area
	Increase value
	Decrease value
	Confirm entry
	Cancel entry
	Start/stop
	Pediatrics mode
	High Flow mode
	Gas bottle

Symbol	Meaning
	House connection
	Flow
	Pressure
	Alarm audio paused (30 s)
	Safety valve switched off
	House connection; input pressure OK
	House connection; input pressure too low or too high
	Gas bottle; input pressure > 30 bar
	Gas bottle; input pressure 20 – 30 bar
	Gas bottle; input pressure 10 – 20 bar
	Gas bottle; input pressure <10 bar

Symbol	Meaning
	Gas bottle; input pressure 1 bar
	Gas bottle; input pressure too high
	Gas consumption display (four-digit)

4.6 Ambient conditions

Storage and transport conditions	
Temperature	-10 °C ... +60°C (14°F ... 140°F)
Relative humidity (non-condensing)	5 – 95 %
Air pressure	500 – 1,080 hPa
Operating conditions	
Temperature	10 °C ... 35°C (50°F ... 95°F)
Relative humidity (non-condensing)	15 – 85 %
Max. operating altitude	3,000 m

5 Preparation

5.1 Unpacking the product

1. Carefully remove the product and accessories from the packaging.
2. Check the delivery for missing items and any possible damage.
3. In the case of damage, hidden defects, and short deliveries, document their nature and extent and contact the manufacturer or supplier immediately.
4. Keep packaging for further transport.

5.2 Setting up the product

When the product is installed, the position of the user must be taken into account. When operating the product, the user stands within a viewing cone with an angle of view of $\pm 45^\circ$ at a distance of approx. 30 to 70 cm from the front panel. For observation of the actual values during the application, a visual distance from the product of 2 m is assumed.

1. Set the product on a horizontal, flat surface.
2. Place the product so that it can be visually observed during use.
3. Install the product out of reach of patients.
4. To ensure that no liquid flows back into the product during use, place the product higher than the patient.
5. Ensure adequate air circulation.

5.3 Installing the bottle holder (option)

The CO₂ bottle can be fastened to the product.

1. Screw the bottle holder to the back of the product using the enclosed screws (6 mm).



2. Insert the CO₂ bottle into the holder vertically and secure it using the locking clip.

5.4 Connecting the product

1. Connect the potential equipotential cable.



2. Connect the power cord. Push the power plug fully into the power socket.



3. Connect the SCB cable.



4. Connect the other end of the SCB cable to a KARL STORZ control device or other SCB devices; see the instructions for use of the OR1 SCB CONTROL.
5. To remove the SCB cable, pull on the plug. The SCB cable is equipped with a protection device to prevent it from being pulled out accidentally.

6. Connect the product to the central gas supply with the low-pressure tube (3.3 to 7 bar).



7. Alternatively, connect the product to a gas bottle.

5.5 Connecting a gas bottle

⚠ WARNING

Wrong gas! Risk of injury!

If the wrong gas is used, patients, users, and third parties can be injured.

- ▶ Only use medical CO₂ gas.

⚠ WARNING

Wrong placement! Risk of injury!

If the CO₂ bottle is incorrectly placed, patients, users, and third parties can be injured. The product can be damaged and its functionality cannot be guaranteed.

- ▶ Ensure that the CO₂ bottle is in a vertical position.
- ▶ Secure the CO₂ bottle to prevent it from falling over.
- ▶ Observe the safety regulations when changing the CO₂ bottle.

- ① When using a CO₂ bottle, the use of a filter is recommended. The filter is integrated into high-pressure tube UI024, UI025, UI026, UI027 or UI028 and prevents particles from the bottle from entering the product. Contamination due to unclean gases can lead to product failure. Consult the high-pressure tube instructions for changing the filter.

1. Connect the CO₂ bottle to the product's KARL STORZ insufflator connector.



2. With a German connection to the CO₂ gas bottle W21.8x1/14 thread or ISO connection to the CO₂ gas bottle W27x2 thread, place the high-pressure tube on the outlet opening of the CO₂ bottle and screw tight.



3. Then open the valve stopcock of the CO₂ bottle with an approx. ½ turn counterclockwise.



4. Alternatively, with PIN-Index connection to the gas bottle, place the connecting piece of the high-pressure tube on the CO₂ bottle and screw tight.



5. Then open the CO₂ bottle with the wrench.



5.6 Putting the product into operation

1. Ensure that the gas is connected.


2. Turn on the product using the power switch.



⇒ The following start screen appears and the self-test is performed:




⇒ After a successful self-test, the information signal sounds and the mode selection appears:



⇒ The following messages are possible during the self-test:






3. If there is pressure on the device output, close the stopcock.
4. If the stopcock cannot be closed, remove the insufflation tube to allow the self-test to be performed without pressure.
5. If the self-test fails, switch the product off and on and check whether the product is connected correctly.


5.7 Selecting the mode

1. Start the product.
⇒ Mode selection appears:




2. Select the desired mode.

⇒ An availability signal sounds and the screen with the working area of the selected mode appears, e.g., **High Flow: CAUTION! Risk of injury! Only use the product if the availability signal was audible.**




⇒ If no gas is connected or the input pressure is too low, the following error message appears:



3. Check whether the gas bottle is open or has to be replaced.

⇒ If the gas supply has been provided late, the self-test has been delayed and the following message appears:



i The intended use is only possible after the self-test has been successfully completed.

The last used operating parameters of a mode are saved and called up again after a successful restart:

Mode	Operating parameter	Setting	Setting after restart
High Flow	Pressure	1 – 15 mmHg	1 – 15 mmHg

Mode	Operating parameter	Setting	Setting after restart
	Flow	16 – 30 mmHg	15 mmHg
		1 – 20 l/min	1 – 20 l/min
		≥ 21 l/min	20 l/min
Pediatrics	Pressure	1 – 12 mmHg	1 – 12 mmHg
		13 – 15 mmHg	12 mmHg
	Flow	0.1 – 1 l/min	0.1 – 1 l/min
		≥ 1.1 l/min	1 l/min

5.8 Connecting the insufflation filter and insufflation tube

⚠ WARNING

Expiry date passed! Risk of infection!

- ▶ Check the expiry date.
- ▶ Check the packaging for damage.
- ▶ Never use products that have passed the expired expiry date or have damaged or accidentally opened packaging but dispose of them properly.

⚠ WARNING

Disposable products! Risk of infection!

The reprocessing of disposable products can lead to infections in patients, users and third parties as well as damage to the product.

- ▶ Never reprocess disposable products.
- ▶ Dispose of disposable products in accordance with the applicable regulations.

⚠ WARNING

Contamination! Risk of infection!

Gas and body fluid that flows back can contaminate the product. Patients, users, and third parties can be infected and the product can be damaged.

- ▶ Place the product higher than the patient.
- ▶ Use a sterile insufflation filter between the insufflation connection and the insufflation tube.
- ▶ Replace the insufflation filter after every use.

1. Attach the sterile insufflation filter to the insufflation connection.



2. Attach the insufflation tube to the insufflation filter.
3. Connect the other end of the insufflation tube to the VERESS needle or the trocar.
- ① In the case of the single-use insufflation tube set, the insufflation filter is firmly connected to the tube.

5.9 Checking the functioning and sealing before use

⚠ CAUTION

Leakage! Risk of injury!

If the system is leaky, the intraabdominal pressure can increase uncontrollably.

- ▶ Check the system for leaks.
- ▶ Only work with a leak-tight system.


1. Select the desired mode, e.g., **Pediatrics**; See section *Selecting the mode* [p. 29].
2. Set the setpoint values 12 mmHg and 1 l/min.
3. Connect the instrument, e.g. trocar or Veress needle, to the insufflation tube and open the insufflation stopcock on the instrument.
4. Start the insufflation.
 - ⇒ After approx. 10 s, the flow display reaches stationary conditions with 1 l/min \pm 0.1 l/min.
 - ⇒ The cavity pressure remains at 0 mmHg.
5. If no flow is displayed, do not use the product and send it in for repair.
6. If the pressure is not 0 mmHg but another positive or negative value is shown, remove the insufflation tube and switch the product off and on again. Then start again with step 3.
7. Close the insufflation stopcock on the instrument.
 - ⇒ Pressure is built up, rises above the setpoint value, and is reduced after a few seconds via the safety valve.
 - ⇒ Depending on the mode and instrument, the message “301: Occlusion” or alarm “303: High pressure” appears.

8. Observe the instrument behavior for approx. 20 s.
 - ⇒ When the message “301: Occlusion” appears, a steady pressure between the setpoint value and the maximum adjustable pressure is established after a maximum of 4 insufflation blows.
 - ⇒ When the alarm “303: High pressure” appears, the pressure build-up and pressure reduction are repeated cyclically.
 - ⇒ The flow display may only have values > 0 l/min for a short time. If the target pressure is reached only slowly when the insufflation stopcock is closed or a stationary flow value of 0.2 l/min or higher occurs, the filter/patient tube/instrument system is leaky.
9. Check the plug connections and replace the tubing set if necessary.

6 Application

6.1 Working area

When a mode has been selected, the working area of the mode appears. In the working area, the setpoint values for pressure and flow are set and the insufflation can be started directly. The settings can be called up in the working area.




- | | |
|---------------------------------------|---|
| 1 Mode | 6 Set pressure values |
| 2 Gas bottle pressure display | 7 Set flow values |
| 3 Mode symbol (High Flow, Pediatrics) | 8 Start insufflation, stop insufflation |
| 4 Gas consumption display | 9 Settings |
| 5 Reset gas consumption display | 10 Browse back |

6.2 Setting operating parameters

1. Select the desired mode, e.g., **High Flow**, See section *Selecting the mode* [p. 29].

⇒ The last used operating parameters appear in the working area:



2. When the required cavity pressure is ≤ 15 mmHg and a maximum flow of 15 l/min is sufficient, select the mode **Pediatrics**.

3. Check the preset setpoint values for **Pressure** and **Flow** and adjust them with the **Plus** and **Minus** buttons. Use the lowest possible pressure and flow values for insufflation.

Recommended settings

Mode	Group	Weight	Flow range
Pediatrics	Children < 1 year	approx. 1 – 9 kg	0.1 – 0.5 l/min
	Children 1 – 3 years	approx. 10 – 15 kg	0.5 – 1.0 l/min
	Children 3 – 4 years	approx. 16 – 19 kg	1.0 – 2.0 l/min
	Children 4 – 14 years	≥ 20 kg	> 2.0 l/min
	All children	< 25 kg	≤ 14.0 l/min

The flow rates for laparoscopic procedures in newborns, infants, toddlers, and children are recommendations. The attending physician is solely responsible for selecting suitable flow and pressure values.

6.3 Performing insufflation

⚠ WARNING

Cross contamination! Risk of infection!

- ▶ To prevent cross contamination of patients, use a hydrophobic bacterial filter.

⚠ CAUTION

Increased intercavitary pressure! Risk of injury!

If external forces act on the cavity, the intercavitary pressure can increase and pressure fluctuations can result. This can cause injury to the patient.

- ▶ Make sure that no external forces are acting on the cavity.

⚠ CAUTION

Continuous operation! Risk of injury!

If the product is operated for longer than 24 hours, the safety and functionality of the product is no longer guaranteed. Patients, users, and third parties can be injured.


- ▶ After 24 hours of continuous operation, switch off the product and disconnect the patient tube.
- ▶ Switch the product on again.
 - ⇒ The self-test of the safety system is carried out.
 - ⇒ There must not be any pressure at the device output; otherwise no self-test can be performed and correct operation cannot be guaranteed.
- ▶ Before starting the operation, observe the warning messages on the product.

1. Before starting insufflation, rinse the entire system with 1 l CO₂.
2. Tap the **-0-** button to reset the volume display of the gas consumption.




3. In the case of single-use tubing sets, cut the tube off behind the LUER-Lock connector if a HICAP connector is used. In the case of reusable insufflation tubes, disconnect the male LUER-Lock.

4. Introduce the Veress needle and open the locking lever.
 5. Tap the **Start/Stop** button to start the insufflation.
- ⇒ The pressure, flow, and volume of the consumed gas are continuously displayed.



- ⇒ If the patient pressure rises above the preset value, an acoustic warning signal sounds and the pressure is reduced via the safety valve after 4 s delay time.



- ❗ To utilize the maximum flow in the High Flow mode, HiCap instruments fitted with a suitable connection must be used. The LUER lock connector at the end of the insufflation tube must be removed.

6.4 Ending the insufflation

1. Close the valve of the CO₂ bottle.
 2. Remove the insufflation tube and the insufflation filter from the product.
- ❗ The Secuvent safety valve is not suitable for reducing pressure after the intervention.

6.5 Settings

The settings can be called up in the working area of the respective mode.

1. Tap the **Settings** button.
- ⇒ The **Settings** screen appears with the following sub-menus:



Screen 1




Screen 2

6.5.1 Setting the language

The screen language is preset to English.

To change the language, proceed as follows:

1. In the **Settings** screen, tap on the **Language** sub-menu.
⇒ The **Language** screen appears:




2. Select the desired language and confirm.
3. Tap the **Working area** button to open the working area.

6.5.2 Adjusting the volume

The volume for warning signals and button tones can be adjusted.

1. In the **Settings** screen, tap the **Volume** sub-menu.
⇒ The **Volume** screen appears:



2. Select the desired volume for **Warning signals** and **Button tone**:
Adjustment range for warnings: 1 to 4
Adjustment range for button tones: 0 to 4
 ⓘ The volume setting for alarms and information signals is summarized under **Warning signals**. The volume setting for alarms and information signals is coupled. Since alarm signals always have to be louder than information signals, the information signal is set to mute at the lowest volume setting 1.
3. Confirm the settings.
4. Tap the **Working area** button to open the working area.

6.5.3 Selecting the source of gas supply

In the **Source of gas supply** screen, it is possible to specify whether the gas is supplied from a central house connection or from a bottle. If the bottle pressure falls below 10 bar, a warning can be given with a single or recurring audio signal.

1. In the **Settings** screen, tap the **Source of gas supply** sub-menu.
⇒ The **Source of gas supply** screen appears:



2. Select the *Central* entry for the house connection or the *bottle* entry with desired audio signal.
3. Confirm the selection.
4. Tap the **Working area** button to open the working area.

6.5.4 Activating the safety valve

When the safety valve is activated, excess pressure is reduced after 4 s. When the safety valve is deactivated, the following symbol appears in the working area:



A deactivated safety valve becomes active again after restarting.

1. In the **Settings** screen, tap the **Safety valve actuation** sub-menu.

⇒ The **Safety valve actuation** screen appears:



2. Activate or deactivate the safety valve and confirm.
3. Tap the **Working area** button to open the working area.

6.5.5 Device information

1. In the **Settings** screen, tap the **Device information** sub-menu.

⇒ The **Device information** screen appears.

2. Tap the **Working area** button to return to the working area.

6.5.6 System log

In the system log, alarms as well as user and service information are saved with their time of occurrence. Each entry occupies one row and consists of the following information:

- Consecutive number
- Time stamp
- ID
- Text message

A maximum of 50,000 entries can be saved; the newest entry has the number 000. When the maximum number has been exceeded, the oldest entries are overwritten.

The system log is backed up in the event of voltage drops and when the device is switched off, and it contains entries relating to switch-on and switch-off times. The entries can be exported in the administrator area.

1. In the **Settings** screen, tap the **System Log** sub-menu.
⇒ The **System Log** screen appears:




No	Timestamp	Id	Textmessage
001	2021-05-05T17:04:11	113	EVENT_PRESSURE_RECOMMENDATION_CONFIRMED
002	2021-05-05T17:04:10	100	Initial touch
003	2021-05-05T17:04:10	112	Please note
004	2021-05-05T17:04:09	041	Operating mode switched
005	2021-05-05T17:04:09	042	Set point changed
006	2021-05-05T17:04:09	042	Set point changed
007	2021-05-05T17:04:07	101	Heated patient tube connected / not connected
008	2021-05-05T17:04:05	101	Heated patient tube connected / not connected
009	2021-05-05T17:04:05	363	Change gas bottle (< 20 bar)

2. Tap the **Working area** button to open the working area.


6.5.7 Administration

The **Administration** screen is password-protected. The password on delivery is 2132. The operator is responsible for changing the password, if necessary.

1. In the **Settings** screen, tap the **Administration** sub-menu.
⇒ The **Password administration** screen appears:




2. Enter the password and confirm.
⇒ The **Administration** screen appears:



6.5.7.1 Setting pressure units

1. In the **Administration** screen, tap the **Unit pressure** sub-menu.

⇒ The **Unit pressure** screen appears:



2. Select the desired unit and confirm.
3. Tap the **Working area** button to open the working area.

6.5.7.2 Exporting the system log

For service purposes, data can be exported via the service interface using an USB stick.

1. Remove the cover on the back of the device with a Torx screwdriver (T10).



2. Insert a USB stick into the USB interface.



3. In the **Administration** screen, tap the **Export system protocol** sub-menu.
⇒ The export starts.
⇒ If the export was successful, the following message appears:




4. Tap the **Working area** button to open the working area.
5. Remove the USB stick and screw the cover onto the back of the device again.

6.5.7.3 Changing the password

1. In the **Administration** screen, tap the **Change administration password** sub-menu.
⇒ The **Change administration password** screen appears:



2. Enter the new password and confirm.
⇒ The **Change administration password** screen appears:



3. Enter the new password again and confirm.


4. Tap the **Working area** button to open the working area.

6.5.8 Auditor

The **Auditor** screen is password-protected. The password on delivery is: 1994. The operator is responsible for changing the password, if necessary.


1. In the **Settings** screen, tap the **Auditor** sub-menu.

⇒ The **Auditor password** screen appears:



2. Enter the password and confirm.

⇒ The **Auditor** screen appears:



6.5.8.1 Exporting the audit trail

In audit logs, activities in the system are recorded that concern network and system settings. The collected audit logs can be exported as an audit trail via USB stick.

1. Remove the cover on the back of the device with a Torx screwdriver (T10).



2. Insert a USB stick into the USB interface.



3. In the **Auditor** screen, tap the **USB export** sub-menu.
⇒ The export starts.
⇒ If the export was successful, the following message appears:



4. Tap the **Working area** button to open the working area.
5. Remove the USB stick and screw the cover onto the back of the device again.

6.5.8.2 Changing the password

1. In the **Auditor** screen, tap the **Change the auditor password** sub-menu.

⇒ The **Change the auditor password** screen appears:



2. Enter the new password and confirm.

⇒ The **Change the auditor password** screen appears:



3. Enter the new password again and confirm.
4. Tap the **Working area** button to open the working area.

6.5.9 Service

The service area is reserved for the service employees of KARL STORZ. The settings are described in the service manual.

6.6 Alarm and information signals



Alarm and information signals are self-explanatory messages that explain the behavior of the device and support the user with the operating functions. These signals improve the usability of the device and support service technicians in troubleshooting.

Information signals are continuously output when they indicate the cause of an inoperable device. All other alarm and information signals are output as long as the signal conditions are present. The minimum duration of the display is 5 s. The signals are ordered according to priority. Alarm signals have a higher priority than information signals.

Information signals are divided into five different priorities. Signals of a higher priority overwrite signals of lower priority, or signals of lower priority are suppressed as long as signals of higher priority are present. In the event of multiple signal conditions with the same priority, the most recently detected condition will appear in the title line.

6.6.1 Alarm specification

The product issues two alarms: "303: High pressure" and "373: CO₂ empty".



Alarm signals are not latching, i.e., the alarm signal is only issued for the duration that the signal condition is present. The alarms are displayed for a minimum of 5 s, and at least one tone sequence is output. The alarm thresholds and alarm delays are programmed as fixed settings. After a production stop or a power failure, the last set volume is adopted.

6.6.1.1 Excess pressure alarm

CAUTION

Excess pressure! Risk of injury!

- ▶ If an excess pressure warning occurs, observe the functioning of the product.
- ▶ If the insufflation or the safety valve is deactivated, take suitable measures to reduce the pressure.

The excess pressure alarm "303: High pressure" indicates that an excessive cavity pressure was measured at the device output. The excess pressure alarm is activated 3 s after the cavity pressure measurement.

The product has a safety valve that reduces the excess pressure after 4 s when the safety valve has not been deactivated and when the insufflation is active.

- ❗ The automatic pressure reduction for setpoint values near or at the maximum of the setting range does not take place at 16 mmHg (Pediatrics) or 31 mmHg (High Flow), but only at a particular excess pressure; see the table.

Mode	Setting	Excess pressure
Pediatrics	>12 mmHg	4 mmHg
High Flow	>26 mmHg	5 mmHg

When the insufflation is switched off, the pressure is not automatically reduced.

The excess pressure alarm is triggered at the following measured values:

Mode	Measured cavity pressure over setpoint value	Cavity pressure
Pediatrics	3 – 4 mmHg	≥16 mmHg
High Flow	4 – 5 mmHg	≥31 mmHg

The warning thresholds of the excess pressure are given with a resolution of one decimal place since the calculation depends on the setpoint value:

Mode	Pressure setpoint	Warning threshold
Pediatrics	1 – 15 mmHg	2.6 – 4 mmHg or pressure ≥16 mmHg
High Flow	1 – 30 mmHg	3.6 – 5 mmHg ≥31 mmHg

6.6.1.2 Gas empty alarm

⚠ CAUTION

Gas empty alarm! Risk of injury!

- ▶ Restore the gas supply.
- ▶ Restart the insufflation.

The gas empty alarm "373: CO₂ empty" indicates that the gas supply has failed during the application and insufflation has been deactivated.

The activation and deactivation thresholds of the alarm are the following:

Threshold	Input pressure
Activation	<1 bar
Deactivation	>1.5 bar

6.6.1.3 Visual alarm signal

The alarms are displayed in the title line as follows:

- Medium priority: blue font on yellow background
- Low priority: blue font on cyan background


6.6.1.4 Acoustic alarm signal

If a condition for an acoustic signal is present, at least one complete tone sequence will be output. The volume of the audio signals is set to the maximum at the factory and can be adjusted in the settings. The maximum volume is 74 dBA; the minimum volume is 49 dBA for the "freshfish" device version and 58 dBA for the "trumpetfish" device version.

The functioning of the alarm system is verified when the availability signal sounds after a mode has been selected.


- ▶ Tap the **Bell** button to deactivate the acoustic signals for 30 s. This affects all alarm and information signals.

⇒ The bell appears yellow during this time:



Medium priority alarm signal


The alarm signal is a burst of 3 tones lasting for 170 ms each, with a pause of 230 ms. The burst is repeated every 10 s and generates 4 harmonics:



Low priority alarm signal

The alarm signal is a burst of 2 tones lasting for 170 ms each, with a pause of 240 ms. The burst is repeated every 16 seconds and generates 4 harmonics:

- Tone 1: frequency 320 Hz
- Tone 2: frequency 254 Hz




Function confirmation of the alarm system

The functioning of the alarm system is confirmed when the availability signal sounds after a mode has been selected. For checking the conditions of the gas empty signal, See section *Checking the gas empty alarm* [p. 51].

6.6.2 Information signals

Information signals notify the user if the operating states of the insufflation system deviate from the expected value ranges, e.g., the gas supply pressure.

371: Input pressure too low




6.6.2.1 Visual information signal

The information signal is displayed with blue writing on a white background in the title line.

6.6.2.2 Acoustic information signal

The acoustic information signal is a double tone c-e (263 Hz – 330 Hz) lasting 300 ms.




Acoustic information signal

Depending on the priority of the message, the tone sequence will either be repeated or output once. If messages indicate an inoperable device, or if the message "371: Bottle pressure <10 bar" appears, the tone sequence is repeated every 20 s. For all other messages, the information signal sounds once.

Availability signal

The pitch of the availability signal is modulated with a frequency of 1.5 Hz by ± 2 Hz each time. 5 different harmonics are generated.



Harmonics of the availability signal

Button tones

When a button on the touch screen is tapped, a short beep is heard. The volume of this beep can be adjusted or turned off independently of the volume of all other information signals in the settings.

6.6.3 Checking the excess pressure alarm

The excess pressure alarm can be checked as follows:

1. Connect the stopcock of the instrument to the insufflation tube.
2. Start the insufflation.
3. Stop the insufflation.

4. When the insufflation is stopped, reduce the setpoint value by 6 mmHg so that the warning threshold is exceeded.

⇒ The visual and acoustic excess pressure alarm is output.

Since the automatic pressure reduction does not work when the insufflation is deactivated, the excess pressure condition is maintained when the system is sealed.

6.6.4 Checking the gas empty alarm

The gas empty alarm can be checked as follows:

1. Before switching on the product, briefly turn on the gas supply and close it again.
 2. Switch on the product.
 3. After a successful self-test, start the insufflation with a flow limit of >1 l/min.
- ⇒ When the inlet pressure has been reduced, the gas empty alarm sounds.

6.6.5 Checking information signals

Information signals can be checked as follows:

1. Before switching on the product, briefly turn on the gas supply and close it again.
2. Switch on the product.
3. Set "Bottle repetitive audio signal" in the settings for the gas supply system.
4. After a successful self-test, start the insufflation with a flow limit of >1 l/min.
5. Stop the insufflation immediately when the message "371: Input pressure too low" appears when the gas supply is closed.

The tightness of the system is sufficient to verify the characteristics of the information signals.

7 Maintenance, servicing, repairs, and disposal

7.1 Maintaining the product

If they are not described in more detail here, maintenance activities may only be performed by KARL STORZ or by a company authorized by KARL STORZ.

7.1.1 Maintenance

The following maintenance intervals are recommended:

Interval	Activity	To be performed by
annually	Safety test	KARL STORZ service technicians

7.2 Changing a fuse

⚠ WARNING

Undesired current flow! Risk of injury!

Live parts of the equipment can cause severe injuries due to electric shock.

- ▶ Do not open the housing.
 - ▶ Make sure that the connection to the power supply is disconnected.
 - ▶ Request a KARL STORZ service technician for service work.
- ✓ The product is switched off.
 - ✓ The power cord is disconnected from the product.
1. Use a screwdriver to remove the screw inserts on the line fuse holder.



2. Remove the defective fuse.

3. Insert the new fuse and tighten to 0.4 Nm. Only use fuses with the specified values, See section *Technical data* [p. 17].



4. Place the screw inserts back into the line fuse holder.
5. Connect to the power supply.



6. Switch on the product and test for proper operation.



7.3 Safety inspection in accordance with IEC 62353

⚠ WARNING

Risk of injury due to product deficiencies!

Patients, users, and third parties may be injured as a result of deficiencies with the product and accessories.

- ▶ Shut down the product.
- ▶ Have the deficiencies repaired by persons authorized by KARL STORZ.

Regardless of the national accident prevention regulations and testing intervals for medical devices, for this device safety checks must be performed as repeat inspections according to IEC 62353 and recorded by a qualified electrician at least once a year. Detailed specifications regarding the scope and execution of the safety inspection can be found in the service manual.

7.3.1 Visual inspection

1. Check the product and accessories for any mechanical damage.
2. Check labels for readability.

7.3.2 Electric measurements

i Limit values for electrical measurements can be found in the current IEC 62353.

1. Measure the protective ground resistance.
2. Measure the earth leakage current.
3. Measure the touch current.
4. Measure the patient leakage current.

7.3.3 Functional test

A complete functional test comprises the following checks:

7.3.3.1 Checking the function in the low flow range (pediatrics)

1. Set the Pediatrics mode with the following operating parameters: 10 mmHg, 0.5 l/min.
2. Connect the insufflation tube (031200) with the trocar LUER-Lock connector. Ensure that the gas can flow out.
3. Start the insufflation.
 - ⇒ The flow slowly increases to 0.5 l/min \pm 0.1 l/min.
4. When the stationary condition has been reached, reduce the flow setpoint to 0.1 l/min.
 - ⇒ The actual flow decreases to 0.1 l/min \pm 0.1 l/min. The cavity pressure display remains at 0 mmHg.
5. If an excessive flow is displayed when the tube is open, a fault is present: Send the product in for repair.
6. If the flow setpoint is 0.1 l/min, keep the LUER-Lock connector closed or close the stopcock.
7. After closing, set the nominal pressure with perhaps a slight excess pressure; the Secuvent and the excess pressure warnings must not be activated. Note that CO₂ diffuses through the tube and the pressure decreases over a long period.

7.3.3.2 Checking the Pediatrics function

1. Set the Pediatrics mode with the following operating parameters: 10 mmHg, 1 l/min.

2. Connect the insufflation tube (031200) with the trocar LUER-Lock connector.
3. Start the insufflation.
 - ⇒ When the tube is open, the flow display increases and reaches 1 l/min \pm 0.1 l/min after approx. 10 s at the end of the insufflation phases. The cavity pressure display remains at 0 mmHg.
4. When the stationary state is reached, keep the LUER-Lock connector closed or close the stopcock.
 - ⇒ After closing, the excess pressure is reduced via the Secuvent and the message "301: Occlusion" appears.
 - ⇒ After a maximum of 5 insufflation blasts, the pressure adjusts to 10 to 15 mmHg without the Secuvent reducing pressure.
 - ⇒ The message "301: Occlusion" is updated after each insufflation blast and is maintained until the LUER-Lock connector or the stopcock is opened again.

7.3.3.3 Checking the High Flow function

1. Set the High Flow mode with the following operating parameters: 16 mmHg, 1 l/min.
2. Connect the insufflation tube (031200) with the trocar LUER-Lock connector.
3. Start the insufflation.
 - ⇒ When the tube is open, the flow display increases and reaches 1 l/min \pm 0.1 l/min after approx. 10 s at the end of the insufflation phases. The cavity pressure display remains at 0 mmHg.
4. When the stationary state is reached, keep the LUER-Lock connector closed or close the stopcock.
 - ⇒ After closing, the excess pressure is reduced via the Secuvent and the message "301: Occlusion" appears.
 - ⇒ After a maximum of 5 insufflation blasts, the pressure adjusts to 16 to 30 mmHg without the Secuvent reducing pressure.
 - ⇒ The message "301: Occlusion" is updated after each insufflation blast and is maintained until the LUER-Lock connector or the stopcock is opened again.

7.4 Repairing the product

Repair work may only be performed by KARL STORZ or by a company authorized by KARL STORZ. The interventions described in this instruction manual are exempt from this rule.

- Please contact your local KARL STORZ subsidiary or authorized dealer (see the list of subsidiaries).

Contaminated devices may not be shipped. To prevent contact infections and airborne infections, products must first be decontaminated. KARL STORZ reserves the right to send back contaminated products.

7.5 Disposing of the product

The product meets the requirements of the Directive on Waste Electrical and Electronic Equipment (WEEE).

Within the scope of application of this directive, KARL STORZ SE & Co. KG is responsible for the proper disposal of this product.

1. The product must be disposed of in accordance with the applicable national laws and regulations at a suitable collection point for the reprocessing of electrical and electronic equipment.

2. Contact KARL STORZ SE & Co. KG, a KARL STORZ branch or an authorized dealer to find out the address of the collection point in your area.
3. Dispose of accessories in compliance with country-specific laws and regulations.

8 Accessories and spare parts

8.1 Accessories

Item	Order no.
Insufflation tubing set, with gas filter	031200-10
Filter, insufflation	031122-25
Insufflation tube	UI004
CO ₂ Low Pressure Tube US, 150 cm	UI020
CO ₂ Low Pressure Tube, 150 cm	UI021
CO ₂ Low Pressure Tube, 300 cm	UI022
CO ₂ Low Pressure Tube, 600 cm	UI023
CO ₂ high pressure tube, 55 cm	UI024
CO ₂ high pressure tube, 102 cm	UI025
CO ₂ High Pressure Tube, PIN Index, 55 cm	UI026
CO ₂ High Pressure Tube, PIN Index, 102 cm	UI027
CO ₂ High Pressure Tube ISO, 102 cm	UI028
Bottle stand, fold-away	UI005
Leakage tester	13242XL
Universal wrench	20400030
Power cord, length 300 cm	400A
Power cord, US version, 200 cm	400B

For additional accessories, insufflation tubes, Veress needles, and trocars, see www.karlstorz.com, KARL STORZ catalogs or on site with a KARL STORZ employee.

8.2 Spare parts

Item	Order no.
FUSE/ASM/T 2.0 AH/5 x 20 mm	ET15-1858390
O-rings 8 x 1 – NBR 40	ET15-1823090
Sintered Filter ISO	ET106-KS-SF-ISO-5
Sintered Filter PIN Index	ET106-KS-SF-PIN-5
Sintered Filter German	ET106-KS-SF-W21-5

9 Electromagnetic compatibility

9.1 General notes on the operating environment

The product is suitable for use in professional healthcare settings. Professional healthcare facilities include physician offices, dental offices, limited care facilities, freestanding surgical centers, freestanding birth centers, multiple treatment facilities, hospitals (emergency rooms, patient rooms, intensive care, surgical rooms, outside the HF-shielded room of an ME system for MRT).

- i** The emission characteristics of this product make it suitable for use in professional healthcare facilities as well as in a residential environment (CISPR 11 Class B). This product offers adequate protection to radio communication service. In the rare event of interference to the radio transmission operation, the user might need to take mitigation measures, such as relocating or re-orienting the product.

⚠ WARNING

Electromagnetic interferences! Malfunction!

Use of this equipment adjacent to or stacked with other equipment could result in improper operation.

- ▶ Avoid this situation.
- ▶ If such use is necessary: Ensure that this equipment and the other equipment are operating normally.

9.2 Accessories and cables

Accessories and cables for EMC compliance				
Type	Shield	Length [m]	Ferrite	Use
PA	No	>3	No	Potential equalization
SCB cable	Yes	≥ 1	No	Connection to the SCB network, other SCB devices
Power cord	No	3	No	Power connection

⚠ WARNING

Reduced immunity or increased emissions! Malfunction!

Use of the product with accessories, transducers and cables other than those specified in this manual may result in increased emissions or decreased immunity.

- ▶ Only use the accessories specified in the manual.

⚠ WARNING

Degradation of performance! Malfunction!

Portable RF communications equipment (including peripherals such as antenna cables and external antennas) could result in degradation of the performance of the product.

- ▶ Do not use portable communications equipment closer than 30 cm (12 inches) to any part of the product, including cables.

9.3 Table 1 – Compliance level for immunity tests

Guidelines and manufacturer's declaration – electromagnetic immunity

The product is intended for use in the electromagnetic environment specified below. The user of the product should make sure that it is used in such an environment.

Immunity tests	EN/IEC 60601 test level	Compliance level	Electromagnetic environment – guidelines
Electrostatic discharge (ESD) acc. to IEC 61000-4-2	± 8 kV contact discharge ± 15 kV air discharge	± 8 kV contact discharge ± 15 kV air discharge	Floors should be made of wood, concrete, or covered with ceramic tiles. If floors are covered with synthetic material, the relative humidity must be at least 30%.
Electrical fast transients/bursts acc. to IEC 61000-4-4	± 2 kV for power lines ± 1 kV for input and output lines 100 kHz repetition	± 2 kV for power lines ± 1 kV for input and output lines 100 kHz repetition	The power supply quality should be that of a typical commercial or hospital environment.
Surges acc. to IEC 61000-4-5	±1 kV voltage outer conductor – outer conductor ±2 kV voltage outer conductor – ground	±1 kV voltage outer conductor – outer conductor ±2 kV voltage outer conductor – ground	The power supply quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions, and voltage variations acc. to IEC 61000-4-11	<u>Voltage dip:</u> Dip to 0% for 1 cycle at 0° phase angle Dip to 70% for 25/30 cycles at 0° phase angle Dropout to 0% for 0.5 cycles @ 0°, 45°, 90°, 135°, 180°, 225°, 270°, and 315° phase angles <u>Voltage interruption:</u> 100% for 250/300 cycles	<u>Voltage dip:</u> Dip to 0% for 1 cycle at 0° phase angle Dip to 70% for 25/30 cycles at 0° phase angle Dropout to 0% for 0.5 cycles @ 0°, 45°, 90°, 135°, 180°, 225°, 270°, and 315° phase angles <u>Voltage interruption:</u> 100% for 250/300 cycles	The power supply quality should be that of a typical commercial or hospital environment. If the user of the product requires continued operation in the event of interruptions to the power supply network, it is recommended that the product be operated with an uninterruptible power supply or a battery.
Magnetic field at the power frequency (50/60 Hz) acc. to IEC 61000-4-8	30 A/m at 50 Hz / 60 Hz	30 A/m at 50 Hz / 60 Hz	If image distortion occurs, it may be necessary to install the product further from sources of electromagnetic fields or to install magnetic shielding. Before the product is installed, the electromagnetic field should be measured to ensure that it is sufficiently low.
Immunity test acc. to IEC 61000-4-3 for radiated, radio-frequency electromagnetic fields	3 V/m 80 MHz to 2.7 GHz	3 V/m 80 MHz to 2.7 GHz	

Immunity tests	EN/IEC 60601 test level	Compliance level	Electromagnetic environment – guidelines
	* Refer to Table 2 for wireless proximity RF field test levels		
Immunity to conducted disturbances, induced by radio-frequency fields acc. to IEC 61000-4-6	3 V _{rms} on 150 kHz to 80 MHz 1 kHz 80% AM modulation 6 V _{rms} in ISM band	3 V _{rms} on 150 kHz to 80 MHz 1 kHz 80% AM modulation 6 V _{rms} in ISM band	


9.4 Table 2 – Test levels for near fields from HF wireless communications equipment

Test frequency MHz	Frequency band MHz	Radio service	Modulation	Immunity test level V/m	Compliance level V/m
385	380 – 390	TETRA 400	Pulse modulation 18 Hz	27	27
450	430 – 470	GMRS 460, FRS 460	FM ± 5 kHz deviation 1 kHz sine wave	28	28
710	704 – 787	LTE band 13 and 17	Pulse modulation 217 Hz	9	9
745					
780					
810	800 – 960	GSM 800/900, TETRA 800, iDEN 820, CDMA 850, LTE band 5	Pulse modulation 18 Hz	28	28
870					
930					
1720	1700 – 1990	GSM 1800, CDMA 1900, GSM 1900, DECT, LTE band 1, 3, 4, 25, UMTS	Pulse modulation 217 Hz	28	28
1845					
1970					
2450	2400 – 2570	Bluetooth, WLAN 802.11 b/g/n, RFID 2450, LTE Band 7	Pulse modulation 217 Hz	28	28
5240	5100 – 5800	WLAN 802.11 a/n	Pulse modulation 217 Hz	9	9
5500					
5785					

9.5 Table 3 – Test levels for radiated and conducted immunity tests

Guidelines and manufacturer's declaration – electromagnetic immunity

The device is intended for use in the electromagnetic environment specified below. The user of the device should make sure that it is used in such an environment.

Interference immunity tests	EN/IEC 60601 test level	Compliance level	Electromagnetic environment – guidelines
Conducted RF disturbances acc. to IEC 61000-4-6	3 V _{rms} 150 kHz to 80 MHz	6 V _{rms}	<p>Portable and mobile RF communications equipment should be used no closer to any part of the device, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.</p> <p>Recommended separation distances:</p> $d = 1.2 \sqrt{P}$ <p>Where P is the rated power of the transmitter in watts [W] according to the information provided by the transmitter manufacturer and d is the recommended separation distance in meters [m].</p> <p>Field strengths from fixed RF transmitters as determined by an electromagnetic site survey ^a should be less than the compliance level in each frequency range ^b.</p> $d = 1.2 \sqrt{P}$ <p>80 MHz to 800 MHz</p> $d = 2.3 \sqrt{P}$ <p>800 MHz to 2.5 GHz</p> <p>Interferences may occur in the vicinity of equipment marked with the following symbol:</p> 
Radiated RF disturbances acc. to IEC 61000-4-3	3 V/m 80 MHz to 2.5 GHz	3 V/m	

Note: At 80 MHz and 800 MHz, the higher frequency range applies.

Note: These guidelines may not apply in all situations. The propagation of electromagnetic waves is affected by absorptions and reflections of buildings, objects, and people.

^a Field strengths from fixed transmitters, e. g., base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast, and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed trans-

Interference immunity tests	EN/IEC 60601 test level	Compliance level	Electromagnetic environment – guidelines
<p>mitters, an electromagnetic site survey should be considered. If the measured field strength at the location where the device is used exceeds the above compliance levels, the device should be monitored to ensure proper function. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the device.</p> <p>^b Over the frequency range from 150 kHz to 80 MHz, field strengths should be less than 3 V/m.</p>			

9.6 Table 4 – Emission class and group

Guidelines and manufacturer's declaration – electromagnetic emissions

The product is intended for use in such an environment as specified below. The customer or user of the device should ensure that it is used in such an environment.

Interference emission measurements	Conformity	Electromagnetic environment – Guidelines
RF emissions according to CISPR 11	Group 1	The product uses RF energy for its internal function only. Therefore, its RF emissions are very low and are not likely to cause any interference affecting nearby electronic equipment.
RF emissions according to CISPR 11	Class B	The product is suitable for use in all establishments including domestic establishments and those directly connected to the public low voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emissions acc. to IEC 61000-3-2	Class A	
Voltage fluctuations/flicker emissions acc. to IEC 61000-3-3	Compliant	

9.7 Table 5 – Recommended separation distances between portable and mobile HF communications devices and the product


The product is intended for use in an electromagnetic environment in which HF disturbances are controlled. The customer or user of the product can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile HF communications equipment (transmitters) and the product as recommended below, according to the output energy of the communications equipment.

Rated maximum output power of the transmitter [W]	Separation distance d [m] according to frequency of transmitter		
	150 kHz to 80 MHz $d = 1.2 \sqrt{P}$	80 MHz to 800 MHz $d = 1.2 \sqrt{P}$	800 MHz to 2.7 GHz $d = 2.3 \sqrt{P}$
0.01	0.12	0.12	0.23
0.1	0.38	0.38	0.73
1	1.2	1.2	2.3
10	3.8	3.8	7.3
100	12	12	23

For transmitters whose maximum rated power is not listed in the table above, the recommended separation distance d in meters (m) can be estimated using the equation from the respective column, whereby P is the maximum rated power of the transmitter in watts (W) according to the transmitter manufacturer.

Note: At 80 MHz and 800 MHz, the separation distance for higher frequency range applies.

Note: These guidelines may not apply in all situations. The propagation of electromagnetic waves is affected by absorptions and reflections of buildings, objects, and people.

-  The product was tested for compatibility with HF surgical devices in accordance with IEC 60601-2-2 Appendix BB.






10 Errors and messages

10.1 Troubleshooting

Fault	Possible causes	Actions
Product failed	Power supply failure	<ul style="list-style-type: none"> ▶ Have the power supply checked
	Defective fuse	<ul style="list-style-type: none"> ▶ Replace fuses as described in the instructions for use. Make sure to use the correct fuse type
Warning message "Control stopped (system error)"	Product defective	<ul style="list-style-type: none"> ▶ Switch the product on and off ▶ Contact Service ▶ Before doing so, open the system log in the settings and note the following from the last 5 entries: <ul style="list-style-type: none"> – Time stamp – ID – Text message – Detail ▶ Tap the respective line to call up the information on the detail
No gas flow		<ul style="list-style-type: none"> ▶ Observe warning messages ▶ Switch the product on and off ▶ Contact Service
No pressure build-up	Too much leakage	<ul style="list-style-type: none"> ▶ Increase flow setpoint
	Instrument resistance too high	<ul style="list-style-type: none"> ▶ Replace instrument
	Leaky insufflation system	<ul style="list-style-type: none"> ▶ Check the plug connections of the accessories and the tightness of the trocar ▶ Replace the tube or tubing set and perform a function test ▶ Replace the filter and perform a function test ▶ Contact Service

11 Overview of mitigating warnings

The original English warning text is as follows:

 WARNING	<p>To avoid the risk of electric shock, this equipment must only be connected to a supply mains with protective earth.</p>
 WARNING	<p>Portable RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the [ME EQUIPMENT or ME SYSTEM], including cables specified by the manufacturer. Otherwise, degradation of the performance of this equipment could result.</p>
 WARNING	<p>Use of this equipment adjacent to or stacked with other equipment should be avoided because it could result in improper operation. If such use is necessary, this equipment and the other equipment should be observed to verify that they are operating normally.</p>
 WARNING	<p>Use of accessories, transducers and cables other than those specified or provided by the manufacturer of this equipment could result in increased electromagnetic emissions or decreased electromagnetic immunity of this equipment and result in improper operation.</p>
 WARNING	<p>No modification of this equipment is allowed.</p>

12 Subsidiaries

KARL STORZ SE & Co. KG
Dr.-Karl-Storz-Straße 34, 78532 Tuttlingen/Germany
Postfach 230, 78503 Tuttlingen/Germany
Phone: +49 7461 708-0, Fax: +49 7461 708-105
E-mail: info@karlstorz.com

KARL STORZ Endoskope Berlin GmbH
Scharnhorststr. 3, 10115 Berlin/Germany
Phone: +49 30 30 69090, Fax: +49 30 30 19452

KARL STORZ Endoscopy Canada Ltd.
7171 Millcreek Drive, Mississauga, Ontario L5N 3R3 Canada
Phone: +1 905 816-4500, Fax: +1 905 816-4599
Toll free phone: 1-800-268-4880 (Canada only), Fax: 1-800-482-4198 (Canada only)
E-mail: info-canada@karlstorz.com

KARL STORZ Endoscopy America, Inc.
2151 East Grand Avenue, El Segundo, CA 90245-5017, USA
Phone: +1 424 218-8100, Fax: +1 424 218-8525
Toll free phone: 800 421-0837 (USA only), Fax: 800 321-1304 (USA only)
E-mail: communications@ksea.com

KARL STORZ Veterinary Endoscopy America, Inc.
1 South Los Cameros Road, Goleta, CA 93117, USA
Phone: +1 805 968-7776, Fax: +1 805 685-2588
E-mail: info@karlstorzvet.com

KARL STORZ Endoscopia Latino America, Inc.
815 N. W. 57th Avenue, Suite 480, Miami, FL 33126-2042, USA
Phone: +1 305 262-8980, Fax: +1 305 262-8986
E-mail: info@ksela.com

KARL STORZ Endoscopia México S.A. de C.V.
Av. Ejército Nacional No. 453 Piso 2, Colonia Granada, Alcaldía Miguel Hidalgo, C.P. 11520 Ciudad de México
Phone: +52 (55) 1101 1520
E-mail: mx-info@karlstorz.com

KARL STORZ Marketing América Do Sul Ltda.
Rua Joaquim Floriano, nº. 413, 20º andar – Itaim Bibi, CEP-04534-011 São Paulo, Brazil
Phone: +55 11 3526-4600, Fax: +55 11 3526-4680
E-mail: br-info@karlstorz.com

KARL STORZ Endoscopia Argentina S.A.
Zufriategui 627 6º Piso, B1638 CAA - Vicente Lopez, Provincia de Buenos Aires, Argentina
Phone: +54 11 4718 0919, Fax: +54 11 4718 2773
E-mail: info@karlstorz.com.ar

KARL STORZ Endoskopi Norge AS
Stamveien 1, 1483 Hagan, Norway
Phone: +47 6380 5600, Fax: +47 6380 5601
E-mail: post@karlstorz.no

KARL STORZ Endoskop Sverige AB
Storsätragränd 14, 127 39 Skärholmen, Sweden
Phone: +46 8 505 648 00
E-mail: kundservice@karlstorz.se

KARL STORZ Endoscopy Suomi OY
Taivaltie 5, 01610 Vantaa, Finland
Phone: +358 (0)96824774, Fax: +358 (0)968247755
E-mail: asiakaspalvelu@karlstorz.fi

KARL STORZ SE & Co. KG
Representative Office
Žalgirio St. 94, LT9300 Vilnius, Lithuania
Phone: +370 5 272 0448, Mobile: +370 685 67 000
E-mail: info-lt-lv@karlstorz.com

KARL STORZ Endoskopi Danmark A/S
Skovlytoften 33, 2840 Holte, Denmark
Phone: +45 45162600, Fax: +45 45162609
E-mail: marketing@karlstorz.dk

KARL STORZ Endoscopy (UK) Ltd.
415 Perth Avenue, Slough, Berkshire, SL1 4TQ, United Kingdom
Phone: +44 1753 503500, Fax: +44 1753 578124
E-mail: info-uk@karlstorz.com

KARL STORZ Endoscopie Nederland B.V.
Displayweg 2, 3821 BT Amersfoort, Netherlands
Phone: +31 (0)33 4545890
E-mail: info-nl@karlstorz.com

KARL STORZ Endoscopy Belgium N.V.
Phone: +31 (0)33 4545890
E-mail: info-be@karlstorz.com

KARL STORZ Endoscopie France S.A.S.
12, rue Georges Guynemer, Quartier de l'Europe, 78280 Guyancourt, France
Phone: +33 1 30484200, Fax: +33 1 30484201
E-mail: marketing-fr@karlstorz.com

KARL STORZ Endoskop Austria GmbH
Landstraßer Hauptstr. 148/1/G1, 1030 Vienna, Austria
Phone: +43 1 71 56 0470, Fax: +43 1 71 56 0479
E-mail: storz-austria@karlstorz.com

KARL STORZ Endoscopia Ibérica S.A.
Parque Empresarial San Fernando, Edificio Munich – Planta Baja, 28830 Madrid, Spain
Phone: +34 91 6771051, Fax: +34 91 6772981
E-mail: info-es@karlstorz.com

KARL STORZ Endoscopia Italia S.r.l.
Via dell'Artigianato, 3, 37135 Verona, Italy
Phone: +39 045 8222000, Fax: +39 045 8222001
E-mail: info-ita@karlstorz.com

 KARL STORZ Croatia d.o.o.
Capraška 6, 10000 Zagreb, Croatia
Phone: +385 1 6406 070, Fax: +385 1 6406 077
E-mail: info@karlstorz.hr

KARL STORZ Endoskopija d.o.o.
Cesta v Gorice 34b, 1000 Ljubljana, Slovenia
Phone: +386 1 620 5880, Fax: +386 1 620 5882
E-mail: pisarna@karlstorz.si

KARL STORZ Polska Sp. z o.o.
ul. Bojkowska 47, 44-100 Gliwice, Poland
Phone: +48 32 706 13 00, Fax: +48 32 706 13 07
E-mail: info-pl@karlstorz.com

KARL STORZ Endoszkóp Magyarország Kft.
Toberek utca 2. fsz. 17/b, HU-1112 Budapest, Hungary
Phone: +36 195 096 31, Fax: +36 195 096 31
E-mail: info-hu@karlstorz.com

KARL STORZ Endoscopia Romania srl
Str. Prof. Dr. Anton Colorian, nr. 74, Sector 4, 041393 Bucharest, Romania
Phone: +40 (0)31 4250800, Fax: +40 (0)31 4250801
E-mail: info-ro@karlstorz.com

KARL STORZ Endoskope Greece M.E.P.E.*
Patriarhou Grigoriou E' 34, 54248 Thessaloniki, Greece
Phone: +30 2310 304868, Fax: +30 2310 304862
E-mail: info-gr@karlstorz.com
*Repair & Service Subsidiary

KARL STORZ Industrial**
Gedik Is Merkezi B Blok, Kat 5, D 38-39, Bagdat Cad. No: 162, Maltepe Istanbul, Turkey
Phone: +90 216 442 9500, Fax: +90 216 442 9030
**Sales for Industrial Endoscopy

000 KARL STORZ Endoscopy — WOSTOK

Derbenyevskaya nab. 7, building 4, 115114 Moscow, Russia
Phone: +7 495 983 02 40, Fax: +7 495 983 02 41
E-mail: Info-ru@karlstorz.com

TOV LLC KARL STORZ Ukraine

Avenue Geroyiv Stalingrada Str. 2D, office 717 Kiev, 04210/Ukraine
Phone: +38 095 000-895-0, +38-097-000-895-0, +38 073 000-895-0
E-mail: marketing@karlstorz.com.ua

KARL STORZ SE & Co. KG Representation Office

Sabit Orudschow 1184, apt. 23, 1025 Baku, Azerbaijan
Phone: +99 450 613 30 60
E-mail: info-az@karlstorz.com

KARL STORZ ENDOSKOPE — East Mediterranean and Gulf (Offshore) S.A.L.

Spark Tower 1st floor Charles Helou St., Horch Tabet — Sin El Fil, Beirut, Lebanon
Phone: +961 1 501105, Fax: +961 1 501950
E-mail: info@karlstorz-emg.com

KARL STORZ Endoscopy (South Africa) (Pty) Ltd.

P.O. 6061, Roggebaai, 8012 Cape Town, South Africa
Phone: +27 21 417 2600, Fax: +27 21 421 5103
E-mail: info@karlstorz.co.za

TOO KARL STORZ Endoscopy Kasachstan

Saryarka, 6, BC "Arman", off. 910, 010000 Astana, Republic of Kazakhstan
Phone: +7 7172 552-549, 552-788, Fax: -444
E-mail: info@karlstorz.kz

KARL STORZ ENDOSKOPE East Mediterranean & Gulf (branch)

Building West Side 7A — Unit 7WA — 3008, Dubai Airport Free Zone, P.O. Box 54983, Dubai - United Arab Emirates
Phone: +971 (0)4 2958887, Fax: +971 (0)4 3205282
Service Hotline: +971 (0)4 3415882
E-mail: info-gne@karlstorz-emg.com

KARL STORZ Endoscopy India Private Limited

11th Floor, Dr. Gopal Das Bhawan, 28, Barakhamba Road, New Delhi 110001, India
Phone: +91 11 4374 3000, Fax: +91 11 4374 3010
E-mail: corporate@karlstorz.in

KARL STORZ SE & CO. KG

Interchange 21 Tower, Level 33, 399 Sukhumvit Road, North Klongtoey, Wattana, 10110 Bangkok, Thailand
Phone: +84 28 3823 8000 Fax: +84 28 3823 8039
E-mail: infovietnam@karlstorz.com

KARL STORZ SE & Co. KG

Resident Representative Office
14th Floor, MPlaza Saigon, 39 Le Duan, District 1, Ho Chi Minh City, Vietnam
Phone: +84 28 3823 8000, Fax: +84 28 3823 8039
E-mail: infovietnam@karlstorz.com

KARL STORZ Endoscopy China Ltd.

Room 2503-05, 25F AXA Tower, Landmark East, No. 100 How Ming Street, Kwun Tong, Kowloon, Hong Kong, People's Republic of China
Phone: +852 28 65 2411, Fax: +852 28 65 4114
E-mail: inquiry@karlstorz.com.hk

KARL STORZ Endoscopy (Shanghai) Ltd., Beijing Branch

Room 1805-1807, Building B, 18F Beijing IFC, No. 8, Jianguomenwai Street, Chaoyang District, 100022, Beijing, People's Republic of China
Phone: +86 10 5638188, Fax: +86 10 5638199
E-mail: info@karlstorz.com.cn

KARL STORZ Endoscopy (Shanghai) Ltd., Shanghai Branch

Room 701A Building 5 & Room 501 Building 7, No. 3000 Longdong Avenue, Pilot Free Trade Zone, 201203, Shanghai, People's Republic of China
Phone: +86 21 60339888, Fax: +86 21 60339808
E-mail: info@karlstorz.com.cn

KARL STORZ Endoscopy (Shanghai) Ltd., Chengdu Branch

Room 803-805, 8F Jin Jiang International Building, No. 1 West Linjiang Road, Wuhou District, 6100414, Chengdu, People's Republic of China
Phone: +86 28 86587977, Fax: +86 28 86587975
E-mail: info@karlstorz.com.cn

KARL STORZ Endoscopy (Shanghai) Ltd., Shenyang Branch

Room 2001-2005, 20F N-MEDIA International Center, No. 167 Youth Avenue, Shenhe District, 110014, Shenyang, People's Republic of China
Phone: +86 24 23181118, Fax: +86 24 23181119
E-mail: info@karlstorz.com.cn

KARL STORZ Endoscopy (Shanghai) Ltd., Guangzhou Branch

Room 02B & 03 & 04A, 35F Teem Tower, No. 208 Tianhe Road, Tianhe District, 510620, Guangzhou, People's Republic of China
Phone: +86 20 87321281, Fax: +86 20 87321286
E-mail: info@karlstorz.com.cn

KARL STORZ Endoscopy Asia Marketing Pte.

No. 8 Commonwealth Lane #03-02, Singapore 149555, Singapore
Phone: +65 69229150, Fax: +65 69229155
E-mail: infoasia@karlstorz.com

KARL STORZ Endoscopy Singapore Sales Pte Ltd

No. 8 Commonwealth Lane #03-02, Singapore 149555, Singapore
Phone: +65 69229150, Fax: +65 69229155
E-mail: infoasia@karlstorz.com

KARL STORZ SE & Co. KG Representative Office Indonesia

Sinarmas MSIG Tower Level 37, Jl. Jend. Surdirman No. Kav. 21, South Jakarta DKI Jakarta 12920
E-mail: infoindonesia@karlstorz.com

KARL STORZ Endoscopy Korea Co. Ltd.

9F Hyowon-Building, 97, Jungdae-ro, Songpa-gu, 05719 Seoul, Korea
Phone: +82-70-4350-7474, Fax: +82-70-8277-3299
E-mail: infokorea@karlstorz.com

KARL STORZ Endoscopy Taiwan Ltd.

12F, No. 192, Sec. 2, Chung Hsin Rd., Sindian District, New Taipei City, Taiwan
Phone: +886 933 014 160, Fax: +886 2 8672 6399
E-mail: info-tw@karlstorz.com

KARL STORZ SE & Co. KG Representative Office Philippines

1901 Picadilly Star Bldg., 4th Avenue, BGC, Taguig City 1636, Philippines
Phone: +63 2 317 45 00, Fax: +63 2 317 45 11
E-mail: philippines@karlstorz.com

KARL STORZ Endoscopy Japan K. K.

Stage Bldg. 8F, 2-7-2 Fujimi, Chiyoda-ku, Tokyo 102-0071, Japan
Phone: +81 3 6380-8622, Fax: +81 3 6380-8633
E-mail: info@karlstorz.co.jp

KARL STORZ Endoscopy New Zealand

Ltd, 31 Morningside Drive Mt Albert Auckland, 1025, New Zealand PO Box 56 511, Dominion Rd Auckland, 1446, New Zealand
Phone: +64 9 846 6044
Toll free: +64 508 84 84 84 (New Zealand only)
Email: sales-nz@karlstorz.com

KARL STORZ Endoscopy Australia Pty. Ltd .

68 Waterloo Road, Macquarie Park NSW 2113, P O Box 50 Lane Cove NSW 1595, Australia
Phone: +61 (0)2 9490 6700, Fax: +61 (0)2 9420 0695
Toll free: 1800 996 562 (Australia only)
E-mail: info@karlstorz.au

www.karlstorz.com



UIP400 • EN • V2.0 • 09-2022 • IFU • CE-MDR



KARL STORZ SE & Co. KG

Dr.-Karl-Storz-Straße 34
78532 Tuttlingen

Postfach 230
78503 Tuttlingen
Germany

Phone: +49 7461 708-0
Fax: +49 7461 708-105
E-mail: info@karlstorz.com
www.karlstorz.com

