

Certificate of Analysis



Cathode Buffer Container 3500 Series

Product No. 4408256
Lot No. 2402989
Date of Manufacture 06APR2024
Expiration Date 08JUN2025

TEST	SPECIFICATION	RESULT
VISUAL INSPECTION		
Clear, particulate free, viscous solution		Pass
PH TEST		
Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	7.960
CONDUCTIVITY		
Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	1785 µS/cm

For Research Use Only. Not for use in diagnostic procedures.

**Manufactured in compliance with our ISO 13485 certified quality management system.
Site: Bedford, MA USA**

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Janella Him
Quality Assurance
Issued 08APR2024