



DECLARATION OF CONFORMITY

ACCORDING TO (EU) 2017/745 MEDICAL DEVICE REGULATION

EU Representative

MedNet EC-REP GmbH
Borkstrasse 10, 48163 Münster,
Germany
SRN: DE-AR-000000002

Conformity Assessment

Conformity Assessment Procedure
Annex II+III of Regulation (EU) 2017/745

Applicable Standards

EN ISO 14971: 2019+A11: 2021, EN
455-1: 2020, EN 455-2: 2015, EN 455-3:
2015, EN 455-4: 2009, EN ISO 15223-1:
2021, EN ISO 20417: 2021, ISO 10993-1:
2018, EN ISO 10993-10: 2010,
EN ISO 10993-12: 2012, EN 62366-1:
2015+A1: 2020

Remark

*The declaration of conformity is valid in connection
with the release technical document.*

*All the supporting documentation is retained at the
premises of the manufacturer.*

*The Declaration of Conformity is exclusively under
the sole responsibility of the manufacturer.*

Manufacturer

Name: Lyncmed Medical Technology (Beijing) Co., Ltd.
Address: Room 1601, Building No.2, Zhubang 2000
Business Building, Balizhuangxili 99, Chaoyang District,
100022 Beijing, PEOPLE'S REPUBLIC OF CHINA
SRN: CN-MF-000014697

Product Information

Name: Patient Examination Gloves
(Nitrile Examination Gloves)

Model:

Model	Color	Size
LMNEGBE01	BLUE	XS, S, M, L, XL
LMNEGVT01	VIOLET	XS, S, M, L, XL
LMNEGBK01	BLACK	XS, S, M, L, XL
LMNEGBP01	BLUE PURPLE	XS, S, M, L, XL
LMNEGCB01	COBALT BLUE	XS, S, M, L, XL
LMNEGWE01	WHITE	XS, S, M, L, XL
LMNEGPK01	PINK	XS, S, M, L, XL

GMDN: 56286

Basic UDI-DI: 697160547A1X9

Classification: Class I, according to rule 1, Annex VIII
of MDR 2017/745

Declaration

We herewith declare that the above-mentioned
products meet the requirements of Medical Device
Regulation (EU) 2017/745 and the applicable standards
above.

Signature: 
Position: General Manager

Date: May 12, 2022

Place: Beijing