



Mast Group Ltd.
Mast House
Derby Road
Bootle
Merseyside L20 1EA
United Kingdom
Tel. +44 (0)151 933 7277
Fax +44 (0)151 944 1332
www.mastgrp.com

EC DECLARATION OF CONFORMITY

We hereby declare that the devices described below comply with those provisions which apply to them of the European Directive 98/79/EC 'on *in vitro* diagnostic medical devices', and are placed on the European market by Mast Group UK through our appointed EC Authorised Representative Mast Diagnostica GmbH, Feldstrasse 20, 23858 Reinfeld, Germany.

This declaration is valid for the IVD medical devices described below on or after the date hereof, and which bear the CE mark. It is also valid for all IVD medical devices described below which are manufactured by Mast and placed on the market on or after the date hereof by third parties with our consent and which bear the CE mark. All supporting documents relating to this declaration are retained at the manufacturer's premises.

Product code	Product description	IVD Directive classification	EDMS code
Various	MAST DISCS® AST – paper antimicrobial susceptibility test (AST) disc products. Each disc contains an antimicrobial agent of stated content in conformance to recognised AST standards such as European Committee on Antimicrobial Susceptibility Testing - EUCAST or US Clinical and Laboratory Standards Institute - CLSI.	Self-certification Annex III excluding section 6	1402050200 (Code registered - 02/12/2002) GMDN code - 63834

Standards applied: EN ISO 13485:2016, ISO 9001:2015, EN ISO 14971:2019, EN ISO 18113-1:2011, EN ISO 18113-2:2011, EN ISO 15223-1:2016, EN ISO 15223-2:2010.

Declaration made by 

Date: 18th May 2022

C Winstanley, Quality Assurance and Regulatory Affairs Manager – Mast Group Ltd.

Document valid till: 18th May 2025

Digitally signed by Monastîrschii Viorica
Date: 2023.02.07 13:25:09 EET
Reason: MoldSign Signature
Location: Moldova





European In Vitro Diagnostic Medical Directive 98/79/EC, Annex III – General IVD
 EC DECLARATION OF CONFORMITY- Self-Certification (Annex III excluding section 6) – General IVD
 Product Names: Mast® Disc AST products

Annex – list of Mast® Disc AST products

Serial No.	Device ID Number	Product Description	Method of use	Pack Size	Brand Name	Intended purpose of the medical device type
1	AK30C	Amikacin 30ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
2	AP2C	Ampicillin 2ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
3	AP10C	Ampicillin 10ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
4	SAM20C	Ampicillin 10ug/Sulbactam 10ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
5	AUG3C	Amoxicillin/clavulanic acid 2-1	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
6	AUG30C	Amoxicillin/clavulanic acid 20-10	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
7	ATH15C	Azithromycin 15ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
8	ATM30C	Aztreonam 30ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
9	BA10C	Bacitracin 10 units	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
10	PY100C	Carbenicillin 100ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Vials
11	CFC30C	Cefaclor 30ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
12	CFB30C	Ceftibuten 30ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
13	CMD30C	Cefamandole 30ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
14	CDX30C	Cefadroxil 30ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
15	BPR5C	Ceftobiprole 5ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
16	CPM30C	Cefepime 30ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
17	CFM5C	Cefixime 5ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
18	CPZ75C	Cefoperazone 75ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
19	CTX5C	Cefotaxime 5ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
20	CTX30C	Cefotaxime 30ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges



Serial No.	Device ID Number	Product Description	Method of use	Pack Size	Brand Name	Intended purpose of the medical device type
22	CTF30C	Cefotiam 30ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
22	FOX30C	Cefoxitin 30ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
23	CPD10C	Cefpodoxime 10ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
24	CZL30	Cefprozil 30ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
25	CPT5C	Ceftaroline 5ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
26	CPT30C	Ceftaroline 30ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
27	CZA14C	Ceftazidime/avibactam 10-4	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
28	CZA50C	Ceftazidime/avibactam 30-20	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
29	CAZ10C	Ceftazidime 10ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
30	CAZ30C	Ceftazidime 30ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
31	CRO5C	Ceftriaxone 5ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
32	CRO30C	Ceftriaxone 30ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
33	C/T40C	Ceftolozone/Tazobactam 30-10	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
34	CXM30C	Cefuroxime 30ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
35	CFX30C	Cephalexin 30ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
36	KF30C	Cephalothin 30ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
37	CZ30C	Cephazolin 30ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
38	C30C	Chloramphenicol 30ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
39	CIP5C	Ciprofloxacin 5ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
40	CLA15C	Clarithromycin 15ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
41	CD2C	Clindamycin 2ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
42	DLX5C	Delafloxacin 5ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
43	DOR10C	Doripenem 10ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges



Serial No.	Device ID Number	Product Description	Method of use	Pack Size	Brand Name	Intended purpose of the medical device type
44	DXT30C	Doxycycline 30ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
45	ERV20C	Eravacycline 20ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
46	ETP10C	Ertapenem 10ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
47	E15C	Erythromycin 15ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
48	FOT200C	Fosfomycin/Trometamol 200ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
49	FC10C	Fusidic Acid 10ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
50	FDC30C	Cefiderocol 30ug discs	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
51	GM10C	Gentamicin 10ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
52	GM30C	Gentamicin 30ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
53	GM120C	Gentamicin 120ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
54	IMI10C	Imipenem 10ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
55	IMR35C	Imipenem/relebactam 35µg discs	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
56	K30C	Kanamycin 30ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
57	LEV5C	Levofloxacin 5ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
58	LZD10C	Linezolid 10ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Vials
59	LZD30C	Linezolid 30ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
60	MEC10C	Mecillinam 10ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
61	MEM10C	Meropenem 10ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
62	MEM10	Meropenem 10ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
63	MEV30C	Meropenem/Vaborbactam 20-10ug	EUCAST/CLSI	100 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Vials
64	MZ5C	Metronidazole 5ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
65	MZ5	Metronidazole 5ug	CLSI	100 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Vials
66	MN30C	Minocycline 30ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
67	MFX5C	Moxifloxacin 5ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges



**Mast
Group**

Serial No.	Device ID Number	Product Description	Method of use	Pack Size	Brand Name	Intended purpose of the medical device type
68	MUP200C	Mupirocin 200ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
69	NA30C	Nalidixic Acid 30ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
70	NA30	Nalidixic Acid 30ug	EUCAST/CLSI	100 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Vials
71	NE10C	Neomycin 10ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
72	NET30C	Netilmicin 30ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
73	NI100C	Nitrofurantoin 100ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
74	NI300C	Nitrofurantoin 300ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
75	NIB30C	Nitroxoline 30ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
76	NOR10C	Norfloxacin 10ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
77	NO5C	Novobiocin 5ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
78	OFX5C	Ofloxacin 5ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
79	OX1C	Oxacillin 1ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
80	OX5C	Oxacillin 5ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
81	OX1	Oxacillin 1ug	EUCAST/CLSI	100 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Vials
82	PEF5C	Pefloxacin 5ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
83	PG1C	Penicillin G 1 unit	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
84	PG1	Penicillin G 1 unit	EUCAST	100 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Vials
85	PG10C	Penicillin G 10 unit	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
86	PG10	Penicillin G 10 unit	CLSI	100 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Vials
87	PRL30C	Piperacillin 30ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
88	PRL100C	Piperacillin 100ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
89	PTZ36C	Piperacillin/Tazobactam 30-6	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
90	PTZ110C	Piperacillin/Tazobactam 100-10	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
91	SYN15C	Quinupristin/dalfopristin 15ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
92	RP5C	Rifampicin 5ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges



Serial No.	Device ID Number	Product Description	Method of use	Pack Size	Brand Name	Intended purpose of the medical device type
93	RP5	Rifampicin 5ug	EUCAST/CLSI	100 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Vials
94	SPC100C	Spectinomycin 100ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
95	S10C	Streptomycin 10ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
96	S300C	Streptomycin 300ug Discs	EUCAST	5x50 discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
97	TEC30C	Teicoplanin 30ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
98	TEL15C	Telithromycin 15ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
99	TEM30C	Temocillin 30ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
100	T30C	Tetracycline 30ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
101	TC75C	Ticarcillin 75ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
102	TGC15C	Tigecycline 15ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
103	TIM85C	Timentin 85ug (Ticarcillin 75ug/Clavulanic acid 10ug)	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
104	TN10C	Tobramycin 10ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
105	TM5C	Trimethoprim 5ug	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
106	TS25C	Cotrimoxazole 25ug (Trimethoprim 1.25ug/Sulphamethosazole 23.75ug)	EUCAST/CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
107	TS25	Cotrimoxazole 25ug (Trimethoprim 1.25ug/Sulphamethosazole 23.75ug)	EUCAST/CLSI	100 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Vials
108	VA5C	Vancomycin 5ug	EUCAST	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges
109	VA5	Vancomycin 5ug	EUCAST	100 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Vials
110	VA30C	Vancomycin 30ug	CLSI	5x50 Discs	MASTDISCS	Antibiotic Susceptibility Testing Discs in Cartridges



Mast Group Ltd.
Mast House
Derby Road
Bootle
Merseyside L20 1EA
United Kingdom

Tel. +44 (0)151 933 7277
Fax +44 (0)151 944 1332
www.mastgrp.com

EC DECLARATION OF CONFORMITY

We hereby declare that the devices described below comply with those provisions which apply to them of the European Directive 98/79/EC 'on *in vitro* diagnostic medical devices', and are placed on the European market by Mast Group UK through our appointed EC Authorised Representative Mast Diagnostica GmbH, Feldstrasse 20, 23858 Reinfeld, Germany.

This declaration is valid for the IVD medical devices described below on or after the date hereof, and which bear the CE mark. It is also valid for all IVD medical devices described below which are manufactured by Mast and placed on the market on or after the date hereof by third parties with our consent and which bear the CE mark. All supporting documents relating to this declaration are retained at the manufacturer's premises.

Product code	Product description	IVD Directive classification	EDMS code
Various (DM---)	MAST® Dehydrated culture media – Powdered product formulations for use in microbial culture media.	Self certification Annex III excluding section 6	1401010100, 1401010200, 1403010100 (Codes registered - 24/03/2003) GMDN codes 30630, 30660

Standards applied: EN ISO 13485:2016, ISO 9001:2015, EN ISO 14971:2019, EN ISO 18113-1:2011, EN ISO 18113-2:2011, EN ISO 15223-1:2016, EN ISO 15223-2:2010.

Declaration made by 

Date: 26 May 2022

Christine Winstanley, Quality Assurance and Regulatory Affairs Manager – Mast Group Ltd.

Document valid till: 26th May 2025



European In Vitro Diagnostic Medical Directive 98/79/EC, Annex III – General IVD

EC DECLARATION OF CONFORMITY- Self-Certification (Annex III excluding section 6) – General IVD

Product Names: Mast Dehydrated Culture Media

Serial No.	Device ID Number	Product Description	Pack Size	EDMS/GMDN code	Brand Name	Intended purpose of the medical device type
1	DM095D	Baird-Parker Medium	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
2	DM494D	Buffered Peptone Water	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
3	DM494E	Buffered Peptone Water	5Kg	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
4	DM269D	Buffered Rappaport Vassiliadis Broth	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
5	DM253D	Burkholderia cepacia Medium	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
6	DM110A	C.L.E.D. Medium	10x200ml sachets	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
7	DM110B	C.L.E.D. Medium	10x500ml sachets	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
8	DM110D	C.L.E.D. Medium	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
9	DM111D	C.L.E.D. with Andrade's Indicator	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
10	DM111E	C.L.E.D. with Andrade's Indicator	5Kg	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
11	DM119E	Cooked Meat Granules	5Kg	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media



Serial No.	Device ID Number	Product Description	Pack Size	EMDS/GMDN code	Brand Name	Intended purpose of the medical device type
12	DM130D	D.C.A. (Hynes)	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
13	DM215B	D.S.T. Agar (Sulphonamide Antagonist Free)	10x 500ml	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
14	DM215D	D.S.T. Agar (Sulphonamide Antagonist Free)	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
15	DM132D	DNase Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
16	DM133D	Eosin Methylene Blue Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
17	DM136D	G.C. Agar Base	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
18	DM134D	Hektoen Enteric Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
19	DM134E	Hektoen Enteric Agar	5Kg	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
20	DM146D	Kimmig Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
21	DM137D	Kligler's Iron Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
22	DM138-1D	Kohn's No 1 Medium	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
23	DM258D	Legionella BCYE Agar Base	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
24	DM440D	M.S.R.V. (Salmonella) Medium	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
25	DM140D	MacConkey Agar (without salt)	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
26	DM140E	MacConkey Agar (without salt)	5Kg	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media



Serial No.	Device ID Number	Product Description	Pack Size	EMDS/GMDN code	Brand Name	Intended purpose of the medical device type
27	DM141D	MacConkey Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
28	DM141E	MacConkey Agar	5Kg	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
29	DM142D	MacConkey Agar No.2	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
30	DM143D	MacConkey Agar No.3	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
31	DM143E	MacConkey Agar No.3	5Kg	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
32	DM150D	MacConkey Broth	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
33	DM160D	Mannitol Salt Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
34	DM170B	Mueller Hinton Agar	10x200ml sachets	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
35	DM170D	Mueller Hinton Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
36	DM170E	Mueller Hinton Agar	5Kg	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
37	DM171D	Mueller Hinton Broth	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
38	DM171N	Mueller Hinton Broth	20Kg	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
39	DM195D	Plate Count Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
40	DM251D	Preston Blood Free Campylobacter Agar Base	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
41	DM205D	S.S. Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media



Serial No.	Device ID Number	Product Description	Pack Size	EMDS/GMDN Code	Brand Name	Intended purpose of the medical device type
42	DM205E	S.S. Agar	5Kg	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
43	DM211D	Simmons Citrate Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
44	DM491D	Sorbitol MacConkey Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
45	DM218D	T.C.B.S. Cholera Medium	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
46	D401D	T.C.B.S. Cholera Medium	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
47	DM219S	Tetrathionate Broth	460g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
48	DM221D	Thioglycollate Broth USP	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
49	DM224D	Triple Sugar Iron Agar (TSI)	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
50	DM228D	Urea Agar Base	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
51	DM228S	Urea Solution (40% v/v)	10 x10ml	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
52	DM480D	Violet Red Bile (Lactose) Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
53	DM235D	Wilkins Chalgren Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
54	DM230D	X.L.D. Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
55	DM230E	X.L.D. Agar	5Kg	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media
56	DM496D	Yeast Extract Agar	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media



**Mast
Group**

Serial No.	Device ID Number	Product Description	Pack Size	EMDS/GMDN Code	Brand Name	Intended purpose of the medical device type
57	DM252D	Yersinia Agar Base	500g	EDMS - 1401010100 GMDN - 30630	Dehydrated culture media and accessories	Cultural Media & Supplements - Clinical & Industrial Culture Media



Mast Group Ltd.
Mast House
Derby Road
Bootle
Merseyside L20 1EA
United Kingdom

Tel. +44 (0)151 933 7277
Fax +44 (0)151 944 1332

www.mastgrp.com

EC DECLARATION OF CONFORMITY

We hereby declare that the devices described below comply with those provisions which apply to them of the European Directive 98/79/EC 'on *in vitro* diagnostic medical devices', and are placed on the European market by Mast Group UK through our appointed EC Authorised Representative Mast Diagnostica GmbH, Feldstrasse 20, 23858 Reinfeld, Germany.

This declaration is valid for the IVD medical devices described below on or after the date hereof, and which bear the CE mark. It is also valid for all IVD medical devices described below which are manufactured by Mast and placed on the market on or after the date hereof by third parties with our consent and which bear the CE mark. All supporting documents relating to this declaration are retained at the manufacturer's premises.

Product code	Product description	IVD Directive classification	EDMS code/ GMDN code
Various (M-----)	MAST® ASSURE Salmonella Antisera – for serological identification and epidemiological typing of Salmonella bacteria.	Self certification Annex III excluding section 6	EDMS code - 1501000000 (code registered - 24/03/2003) GMDN code - 39521

Standards applied: EN ISO 13485:2016, ISO 9001:2015, EN ISO 14971:2012, EN ISO 18113-1:2011, EN ISO 18113-2:2011, EN ISO 15223-1:2016, EN ISO 15223-2:2010.

Declaration made by C Winstanley Date: 26th May 2022

C Winstanley, Quality Assurance and Regulatory Affairs Manager – Mast Group Ltd.

Document valid till: 26th May 2025



**Mast
Group**

European In Vitro Diagnostic Medical Directive 98/79/EC, Annex III – General IVD
EC DECLARATION OF CONFORMITY - Self-Certification (Annex III excluding section 6) – General IVD
Product Names: MAST® ASSURE Salmonella Antisera

Serial No.	Product code	Pack Size	Brand/Trade Name	Indented Purpose of the Medical Device Type
1	M10308	2ml	MAST® ASSURE Salmonella O - POLY O Factor O2, O4, O7, O8, O9, O46, O3, 10, O1,3,19	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
2	M10309	2ml	MAST® ASSURE Salmonella O - POLY O1 Factor O11, O13, O6, 14, O16, O18, O21, O35	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
3	M14294	2ml	MAST® ASSURE Salmonella O - POLY O A-G	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
4	M14300	2ml	MAST® ASSURE Salmonella O - POLY O A-S	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
5	M92537	2ml	MAST® ASSURE Salmonella O - Omnivalent (Kauffmann-White group A-067)	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
6	M14317	2ml	MAST® ASSURE Salmonella H - POLYVALENT PHASE 1&2 (a-z29)	Liquid stable polyvalent antiserum for the determination of H antigens for the serological identification of salmonellae by slide agglutination.
7	M10339	2ml	MAST® ASSURE Salmonella H - POLYVALENT PHASE 2 (H-1)	Liquid stable polyvalent antiserum for the determination of H antigens for the serological identification of salmonellae by slide agglutination.
8	M14324	2ml	MAST® ASSURE Salmonella H - RAPID DIAGNOSTIC 1 Factors b, d, E, r	Liquid stable polyvalent antiserum for the determination of H antigens for the serological identification of salmonellae by slide agglutination.
9	M14331	2ml	MAST® ASSURE Salmonella H - RAPID DIAGNOSTIC 2 Factors b, E, k, l	Liquid stable polyvalent antiserum for the determination of H antigens for the serological identification of salmonellae by slide agglutination.
10	M14348	2ml	MAST® ASSURE Salmonella H - RAPID DIAGNOSTIC 3 Factors d, E, G, k	Liquid stable polyvalent antiserum for the determination of H antigens for the serological identification of salmonellae by slide agglutination.



**Mast
Group**

Serial No.	Product code	Pack Size	Brand/Trade Name	Indented Purpose of the Medical Device Type
11	M10310	2ml	MAST® ASSURE Salmonella O - Mono Factor O2	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
12	M10311	2ml	MAST® ASSURE Salmonella O - Mono Factor O4	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
13	M10312	2ml	MAST® ASSURE Salmonella O - Mono Factor O7	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
14	M10313	2ml	MAST® ASSURE Salmonella O - Mono Factor O8	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
15	M10314	2ml	MAST® ASSURE Salmonella O - Mono Factor O9	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
16	M10315	2ml	MAST® ASSURE Salmonella O - Mono Factor O9.46	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
17	M10316	2ml	MAST® ASSURE Salmonella O - Mono Factor O3,10	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
18	M10318	2ml	MAST® ASSURE Salmonella O - Mono Factor O1,3,19	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
19	M10319	2ml	MAST® ASSURE Salmonella O - Mono Factor O11	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
20	M10320	2ml	MAST® ASSURE Salmonella O - Mono Factor O13	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
21	M10321	2ml	MAST® ASSURE Salmonella O - Mono Factor O6,14	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
22	M10322	2ml	MAST® ASSURE Salmonella O - Mono Factor O16	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
23	M10323	2ml	MAST® ASSURE Salmonella O - Mono Factor O18	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
24	M10324	2ml	MAST® ASSURE Salmonella O - Mono Factor O21	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
25	M10325	2ml	MAST® ASSURE Salmonella O - Mono Factor O35	Liquid stable antiserum for the determination of O antigens for the serological identification of salmonellae by slide agglutination.
26	M10326	2ml	MAST® ASSURE Salmonella O - Mono Factor Vi	Liquid stable antiserum for the determination of Vi antigens for the serological identification of salmonellae by slide agglutination.



**Mast
Group**

Serial No.	Product code	Pack Size	Brand/Trade Name	Indented Purpose of the Medical Device Type
27	M10327	2ml	MAST® ASSURE Salmonella H - Mono Factor a	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
28	M10328	2ml	MAST® ASSURE Salmonella H - Mono Factor b	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
29	M10329	2ml	MAST® ASSURE Salmonella H - Mono Factor c	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
30	M10330	2ml	MAST® ASSURE Salmonella H - Mono Factor d	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
31	M10331	2ml	MAST® ASSURE Salmonella H - Mono Factor e, h	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
32	M14335	2ml	MAST® ASSURE Salmonella H - Mono Factor E	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
33	M10332	2ml	MAST® ASSURE Salmonella H - Mono Factor G	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
34	M10333	2ml	MAST® ASSURE Salmonella H - Mono Factor i	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
35	M10334	2ml	MAST® ASSURE Salmonella H - Mono Factor k	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
36	M10336	2ml	MAST® ASSURE Salmonella H - Mono Factor r	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
37	M10344	2ml	MAST® ASSURE Salmonella H - Mono Factor 2	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
38	M10345	2ml	MAST® ASSURE Salmonella H - Mono Factor 5	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
39	M10346	2ml	MAST® ASSURE Salmonella H - Mono Factor 6	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
40	M10364	2ml	MAST® ASSURE Salmonella H - Mono Factor f	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
41	M10365	2ml	MAST® ASSURE Salmonella H - Mono Factor m	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
42	M10366	2ml	MAST® ASSURE Salmonella H - Mono Factor p	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.



**Mast
Group**

Serial No.	Product code	Pack Size	Brand/Trade Name	Indented Purpose of the Medical Device Type
43	M10367	2ml	MAST® ASSURE Salmonella H - Mono Factor q	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
44	M10368	2ml	MAST® ASSURE Salmonella H - Mono Factor s	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
45	M10369	2ml	MAST® ASSURE Salmonella H - Mono Factor t	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
46	M10376	2ml	MAST® ASSURE Salmonella H - Mono Factor x	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
47	M10335	2ml	MAST® ASSURE Salmonella H - Mono Factor L	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
48	M10337	2ml	MAST® ASSURE Salmonella H - Mono Factor y	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
49	M10338	2ml	MAST® ASSURE Salmonella H - Mono Factor e,n	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
50	M10340	2ml	MAST® ASSURE Salmonella H - Mono Factor v	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
51	M10341	2ml	MAST® ASSURE Salmonella H - Mono Factor w	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
52	M10342	2ml	MAST® ASSURE Salmonella H - Mono Factor z13	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
53	M10343	2ml	MAST® ASSURE Salmonella H - Mono Factor z28	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
54	M10347	2ml	MAST® ASSURE Salmonella H - Mono Factor 7	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
55	M10348	2ml	MAST® ASSURE Salmonella H - Mono Factor z6	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
56	M10370	2ml	MAST® ASSURE Salmonella H - Mono Factor u	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
57	M10372	2ml	MAST® ASSURE Salmonella H - Mono Factor z23	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
58	M10373	2ml	MAST® ASSURE Salmonella H - Mono Factor z24	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.



**Mast
Group**

Serial No.	Product code	Pack Size	Brand/Trade Name	Indented Purpose of the Medical Device Type
59	M10377	2ml	MAST® ASSURE Salmonella H - Mono Factor z15	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
60	M10374	2ml	MAST® ASSURE Salmonella H - Mono Factor z32	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
61	M10378	2ml	MAST® ASSURE Salmonella H - Mono Factor z	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
62	M10379	2ml	MAST® ASSURE Salmonella H - Mono Factor z4	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
63	M10380	2ml	MAST® ASSURE Salmonella H - Mono Factor z10	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.
64	M10381	2ml	MAST® ASSURE Salmonella H - Mono Factor z29	Liquid stable antiserum for the determination of H antigens for the serological identification of salmonellae by slide and tube agglutination.

Certificate of Registration

QUALITY MANAGEMENT SYSTEM - ISO 9001:2015

This is to certify that:

Mast Group Ltd
Mast House, Derby Road
Bootle
Liverpool
L20 1EA
United Kingdom

Holds Certificate Number:

FM 724380

and operates a Quality Management System which complies with the requirements of ISO 9001:2015 for the following scope:

Design, manufacture and supply of in-vitro diagnostic devices and associated services.

For and on behalf of BSI:



Matt Page, Managing Director Assurance - UK & Ireland

Original Registration Date: 1994-06-14

Latest Revision Date: 2021-12-03

Effective Date: 2021-11-11

Expiry Date: 2024-05-31



Page: 1 of 2

...making excellence a habit.™

Certificate No: FM 724380

Location	Registered Activities
Mast Group Ltd Mast House, Derby Road Bootle Liverpool L20 1EA United Kingdom	Design, manufacture and supply of in-vitro diagnostic devices and associated services.
Mast Group Ltd Atlantic House, Derby Road Bootle Liverpool L20 1EA United Kingdom	Manufacturing, QC and warehousing of IVD kits and reagents, bacteriological media and other kits and reagents for the life sciences industry.



Original Registration Date: 1994-06-14
Latest Revision Date: 2021-12-03

Effective Date: 2021-11-11
Expiry Date: 2024-05-31

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract.
An electronic certificate can be authenticated [online](#).
Printed copies can be validated at www.bsigroup.com/ClientDirectory

Information and Contact: BSI, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP. Tel: + 44 345 080 9000
BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A Member of the BSI Group of Companies.

Certificate of Registration

QUALITY MANAGEMENT SYSTEM - ISO 13485:2016 & EN ISO 13485:2016

This is to certify that:

Mast Group Ltd
Mast House, Derby Road
Bootle
Liverpool
L20 1EA
United Kingdom

Holds Certificate Number:

MD 724379

and operates a Quality Management System which complies with the requirements of ISO 13485:2016 & EN ISO 13485:2016 for the following scope:

Design, manufacture and supply of in-vitro diagnostic devices for clinical microbiology and molecular biology and bacteriological media.



For and on behalf of BSI:

Gary E Slack, Senior Vice President - Medical Devices

Original Registration Date: 2020-10-07

Latest Revision Date: 2021-11-28

Effective Date: 2021-06-01

Expiry Date: 2024-05-31

Page: 1 of 2



...making excellence a habit.™

Certificate No: **MD 724379**

Location	Registered Activities
Mast Group Ltd Mast House, Derby Road Bootle Liverpool L20 1EA United Kingdom	Design, manufacture and supply of in-vitro diagnostic devices for clinical microbiology and molecular biology and bacteriological media
Mast Group Ltd Atlantic House, Derby Road Bootle Liverpool L20 1EA United Kingdom	Manufacture of IVD kits and reagents for clinical microbiology and molecular biology, bacteriological media.



Original Registration Date: 2020-10-07

Latest Revision Date: 2021-11-28

Effective Date: 2021-06-01

Expiry Date: 2024-05-31



SRL SANMEDICO
A. Corobceanu Street 7A, apt. 9,
Chişinău MD-2012,
Moldova

Mast Group Ltd.
Mast House
Derby Road
Bootle
Merseyside L20 1EA
United Kingdom

Tel: +44 (0) 151 933 7277

www.mast-group.com

TO WHOM IT MAY CONCERN

LETTER OF AUTHORISATION

This letter is to confirm, SRL SANMEDICO of address A. Corobceanu Street 7A, apt. 9, Chişinău MD-2012, Moldova is authorised to register products, present proposals, offer quotations and accept orders and to participate in tenders as required for Mast Group Limited products.

The undersigned herewith states that the above is true and correct.

Yours faithfully,

Susan Thomson

Bootle UK

December 9th 2021

International business Manager, on behalf of Mast Group



Mueller Hinton Agar

DM170

Intended Use

A standardised medium for susceptibility testing.

Contents

See pack label.

Formulation*

Material:	Concentration in medium:
Casein hydrolysate	17.5g/litre
Beef infusion from 300g	2.0g/litre
Starch	1.5g/litre
Agar	17.0g/litre
Final pH at 25°C: 7.3 ± 0.1	

Storage and shelf life

All dehydrated culture media containers should be kept tightly closed and stored in a dry place at 10 to 25°C until the expiry date shown on the pack label.

Precautions

For *in vitro* diagnostic use only. Observe approved hazard precautions and aseptic techniques. To be used only by adequately trained and qualified laboratory personnel. Sterilise all biohazard waste before disposal. Refer to Product Safety Data sheet (available on request or via MAST® website).

Materials required but not provided

Standard microbiological supplies and equipment such as loops, MAST® selective supplements, swabs, applicator sticks, incinerators and incubators, etc., as well as serological and biochemical reagents and additives such as blood.

Procedure

1. Refer to pack label for quantities and volumes required. Prepare MAST® Mueller Hinton Agar (DM170D) by suspending the powder in distilled or deionised water. For sachet packs, dissolve the entire contents of the sachet in the volume shown on the label.
2. Autoclave at 121°C (15 p.s.i.) for 15 minutes.
3. If required cool to 50 to 55°C and hold at this temperature in a water bath. If required add 5 to 7% sterile defibrinated blood to enhance the growth of fastidious organisms or antibiotics (MAST® ADATAB) for dilution susceptibility test methods.
4. Alternative growth supplements can be used.
5. Pour culture plates (25ml per plate) and allow to set.
6. Prepared culture plates may be used immediately or stored in plastic bags at 2 to 8°C for up to one week before use.

7. Antimicrobial Susceptibility Testing should be performed in accordance with standards set down by regulatory bodies such as CLSI® (Clinical and Laboratory Standards Institute) and EUCAST (The European Committee on Antimicrobial Susceptibility Testing).

Interpretation of results

After incubation record diameter of zones of inhibition or Minimum Inhibitory Concentration (MIC). Interpret results as sensitive, intermediate or resistant according to the criteria laid down in the method of use.

Quality control

Check for signs of deterioration. Quality control must be performed with at least one organism to demonstrate expected performance. Do not use the product if the result with the control organism is incorrect. The list below illustrates a range of performance control strains which the end user can easily obtain.

Test Organisms	
<i>Enterococcus faecalis</i> ATCC® 29212	Growth and correct susceptibility pattern
<i>Escherichia coli</i> ATCC® 25922	Growth and correct susceptibility pattern
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Growth and correct susceptibility pattern
<i>Staphylococcus aureus</i> ATCC® 25923	Growth and correct susceptibility pattern

References

Bibliography available on request.

E. coli
Salmonella
Haemophilus

mastassure™

Bacterial Agglutinating
Antisera Guide

Shigella

Vibrio

Campylobacter

Bordetella

Clostridium

Legionella

Pseudomonas

Staphylococcus

Listeria

Streptococcus

Yersinia

Proteus

Brucella



Contents

Section 1	Page No
Salmonella agglutination methods	3-15
Salmonella stock antisera	16
Section 2	
Pathogenic Escherichia coli agglutination methods	17-25
E.coli Stock antisera	26
Section 3	
Shigella sp agglutination methods	27-32
Shigella sp stock antisera	32
Section 4	
Vibrio cholera agglutination methods	33-35
Vibrio chloerae stock antisera	35

This document does not replace the IFU's intended for use with these products. It is primarily aimed at being a resource for education, training and a guide for a laboratory writing Standard Operating Procedures looking at the more traditional methods associated with bacterial agglutinating sera.

References available upon request.

A scanning electron micrograph (SEM) showing a Salmonella bacterium in the foreground. The bacterium is rod-shaped with a textured, ribbed surface and is covered in fine, hair-like flagella. It is positioned diagonally across the frame. The background consists of a red, textured surface, possibly a biological membrane, with several green, rounded structures in the upper portion of the image.

mastassure™

SALMONELLA ANTISERA

A. MAST ASSURE™ Salmonella Antisera

1. Introduction

a. Salmonella and Disease

Salmonella are Gram negative organisms belonging to the family *Enterobacteriaceae*. They are non-spore forming rods, and most of them are motile possessing flagella.

Salmonellae are widely distributed in nature, with all vertebrates and some invertebrates capable of harbouring salmonella in their gut. Most animal infections seem to be symptomless or to cause a self-limiting gastroenteritis of variable severity. Many serotypes e.g. *S. typhimurium* show a wide host range, while others e.g. *S. typhi* and *S. paratyphi* A, B and C are primarily human pathogens and are rarely isolated from animals other than humans. Others are particularly adapted to animal hosts include *S. cholera-suis* (pigs), *S. dublin* (cattle), *S. gallinarum-pullorum* (poultry), *S. abortus-equi* (horses) and *S. abortus-ovis* (sheep).

Salmonella infection in man can cause a wide spectrum of clinical illness however it mainly manifests itself in four syndromes. These are enteric fever, gastroenteritis, bacteraemia with or without metastatic infection, and the asymptomatic carrier state.

- Enteric fever is most usually caused by *S. typhi* and *S. paratyphi* A, B and C but can be caused by other Salmonella serotypes. The clinical features tend to be more severe with *S. typhi* (typhoid fever). Enteric fever is a systemic disease whose major symptoms are fever and headache and may be fatal if untreated.
- Acute gastroenteritis is characterised by vomiting, abdominal pain, diarrhoea and fever and is commonly seen as a result of food poisoning from infected poultry.
- Bacteraemia is a constant feature of enteric illness and may occur as a rare complication of any salmonella infection. Transient bacteraemia may occur in gastroenteritis but in most cases organisms are cleared from the blood stream without ill effect.
- Carrier State. Most sufferers from a salmonella infection continue to excrete the organism in their stools for days or weeks after complete clinical recovery, but eventual clearance from the body is usual. A few patients continue to excrete salmonellae for prolonged periods. Chronic carriers may excrete salmonellae for a year or more and may have no symptoms of disease.

Laboratory diagnosis of salmonella infection usually depends upon the isolation and identification of the causal salmonella from a specimen of the patient's blood or faeces. Further antigenic analysis of the organism is often needed to identify the species or strain for epidemiological purposes.

b. Antigenic characterisation of Salmonellae

Salmonella possess two main types of antigen, the O (somatic or body) and the H (flagellar) antigens. The O antigens are heat stable and upon which grouping of the organisms are based. The H antigens are heat labile and are used for confirming and identifying the serotypes within the groups. The H antigen of an organism of a known O group may occur as two different serological entities, known as phases. Diphasic organisms may be separated into individual phases by growing the culture in the presence of a small amount of the specific phase antiserum of the phase already recognised, such that the only motile organisms isolated will be of the alternative phase. Occasionally a third antigen called the Vi antigen is also present in certain Salmonella strains. When present the Vi antigen may block the activity of the O antigens and must be inactivated before proceeding with the serological grouping of the organism.

As an example *Salmonella typhimurium* possess the O antigens 1, 4, [5], 12 and may possess either phase 1 antigen i or phase 2 antigen 1, 2. This may be written as 1, 4, [5], 12: i : 1, 2.

In the early 1920s White recognised the antigenic variation in *Salmonella* species and its importance in differentiating *Salmonella* on the basis of serotyping. Kauffmann later confirmed White's observations and greatly extended them to form the current Kauffmann-White Schema for classification of types within the genus *Salmonella*. Today more than 2000 serotypes of *Salmonella* have been identified.

c. MAST ASSURE™ *Salmonella* Antisera: preparation and presentation.

MAST ASSURE™ *Salmonella* Antisera are prepared from rabbit's hyperimmunised with standard strains of *Salmonella* organisms possessing defined antigenic factors. All sera are heat inactivated at 56°C for 30 minutes, absorbed to remove cross-reacting agglutinins and filter sterilised. The MAST ASSURE™ *Salmonella* Antisera provide a comprehensive range of O, H and Vi antisera and in addition *Salmonella* Phase Induction Antisera, for the induction of a hidden flagellar phase. Antigens are identified normally by qualitative slide agglutination or by quantitative tube agglutination tests or in the case of *Salmonella* Phase Induction Antisera using culture tubes (semi-solid medium) or agar plates (bridging method).

All MAST ASSURE™ *Salmonella* Antisera except the Phase Induction Antisera are provided as 2ml (or 5ml) amounts in vials with dropper attachments and contain 0.1% sodium azide as preservative. Supplied ready to use. This is sufficient for 50 (125) slide agglutination tests or 20 (50) tube agglutination tests. The MAST ASSURE™ *Salmonella* Phase Induction Antisera are provided sterile in 5 ml volumes and sealed in injection vials. Supplied ready to use with no preservative is added.

2. Culture of *Salmonella* - Preparation for Serology

Salmonella belong to the family *Enterobacteriaceae* and there is much cross-reaction and antigenic relationships between *Salmonella* and other genera within this family. Hence it is important that organisms undergoing serological classification should be correctly identified as *Salmonella* by morphological and biochemical features first.

3. *Salmonella* O - Grouping Antisera

Classically *Salmonella* somatic O antigenic factors have been grouped together according to alphabetical designations as detailed in the Kaufmann-White Scheme. However these have become outdated as new antigenic factors were realised. Mast prefers to use the numbered designations as recommended by the WHO Collaborative Centre for Reference and Research on *Salmonella* in their MAST ASSURE™ Antisera range.

Table 1 below indicates the correlation between the alphabetical and numerical systems.

Table 1 - Correlation between the Alphabetical Grouping and Numerical O Factor Designation Systems for *Salmonella* O Grouping.

Alphabetical group designation	Numerical O factor designation	Alphabetical group designation	Numerical O factor designation	Alphabetical group designation	Numerical O factor designation
A	O2	F	O11	Q	O39
B	O4	G ₁	O13, 22	R	O40
C ₁ + C ₄	O6, 7	G ₂	O13, 23	S	O41
C ₂	O6, 8	H	O6, 14	T	O42
C ₃	O8, [20]	I	O16	U	O43
D ₁	O9, 12	J	O17	V	O44
D ₂	O9, 46	K	O18	W	O45
D ₃	O9, 46, 27	L	O21	X	O47
E ₁	O3, 10 *	M	O28	Y	O48
E ₂	O3, 15 *	N	O30	Z	O50
E ₃	O3, 15, 34 *	O	O35	no code	O51 - O67
E ₄	O3, 19	P	O38		

NOTE * = Group E₁ was originally classified as O3, 10, Group E₂ as O3,15 and Group E₃ as O3, 15, 34. It was later noted that Group E₁ strains could be lysogenised by phage ε15 (O3, 10 → O3, 15) and then by phage ε34 (O3, 15 → O3, 15, 34). Because of this these strains are now classified under one common group.

Table 2 lists all the MAST ASSURE™ Salmonella O Grouping Antisera (and Vi Antiserum) available to date together with the product codes, the Kauffmann-White alphabetical group equivalent and the actual antigenic agglutinins present. It is impractical to provide single factor sera for all known Salmonella antigens, however a wide range of sera is available in the MAST ASSURE™ range which is sufficient to identify with a reasonable degree of probability the majority of Salmonella types isolated.


To determine the O-group of a Salmonella strain, polyvalent O antisera should be used initially to narrow down the range before specific grouping sera are used. This is illustrated schematically in Figure 1 and according to methods as detailed in section 4.

Table 2 - MAST ASSURE™ Salmonella O-Grouping Antisera Details

Mast Product Code	MAST ASSURE™ Salmonella O Antiserum	Kauffmann-White alphabetical grouping equivalent	Agglutinins present
M10308	Polyvalent O	A - E	O2, 3, 4, 5, 7, 8, 9, 10, 12, 15, 19, 34, 46.
M10309	Polyvalent O1	F - O excluding J, M and N.	O11, 13, 14, 16, 18, 21, 22, 23, 24, 35.
M14294	Polyvalent A-G	A -G	O2, 3, 4, 5, 7, 8, 9, 10, 11, 12, 13, 15, 19, 22, 34, 46.
M14300	Polyvalent A-S	A -S	O2, 3, 4, 5, 6, 7, 8, 9, 10, 11, 12, 13, 14, 15, 16, 17, 18, 19, 21, 22, 23, 24, 28, 30, 34, 35, 38, 39, 40, 41, 46.
M10310	O2	A	O2
M10311	O4	B	O4, 5
M10312	O7	C1 + C4	O7
M10313	O8	C2 + C3	O8
M10314	O9	D	O9
M10315*	O9, 46	D2	O46
M10316	O3, 10	E1 + E2 + E3 *	O10, 15, 34
M10318*	O1, 3, 19	E4	O19
M10319*	O11	F	O11
M10320	O13	G1 + G2	O13, 22, 23
M10321	O6, 14	H	O14, 24, 25
M10322*	O16	I	O16
M10323*	O18	K	O18
M10324*	O21	L	O21
M10325*	O35	O	O35
M10326	Vi	-	Vi

* Denotes an 8 week lead time

Figure 1 - Summary of Salmonella O-Grouping Procedures



This document does not replace the IFU's intended for use with these products. It is primarily aimed at being a resource for education, training and a guide for a laboratory writing Standard Operating Procedures looking at the more traditional methods associated with bacterial agglutinating sera

References available upon request.



4. Procedures for O serotyping and Interpretation of Results

MAST ASSURE™ Salmonella O Grouping Antisera are intended for use in the identification of O antigens by qualitative slide agglutination, although they may be used in quantitative tube agglutination tests for confirmatory purposes.


Cultures of organisms identified as *Salmonella* by their morphological and biochemical features may be serotyped by the following procedures. Also refer to figure 2 for a summary of the Salmonella O grouping procedure.

a. Slide Agglutination for O-antigen and Vi grouping.

1. Place two drops of sterile 0.85% saline solution (saline) onto a carefully cleaned microscope slide. The slide may be partitioned into several parts using a chinograph or glass pencil Picture a ,b. With an inoculation loop or wire emulsify into each drop of saline a live cell colony from a fresh agar plate or slope culture to produce a distinct and uniform turbidity Picture c, and d.



2. Place a drop of polyvalent antiserum onto one of the drops of emulsified isolate and to the other a drop of saline as a control.



Note: allow the antiserum to freefall from the dropper provided with the bottle. Do not contaminate the antiserum with organism.

3. Mix the reagents by tilting the slide back and forth for 60 seconds while viewing under indirect light against a dark background.

Distinct clumping or agglutination within this period, without clumping in the saline control (auto- agglutination) should be regarded as a positive result.



An isolate producing a distinct positive reaction with a polyvalent antiserum is assumed to be *Salmonella* bearing one or more of the O antigenic factors represented by that antiserum. Using this information, further testing of the isolate should be conducted, as described in steps 1 - 3, with specific O antisera to reveal the full O antigenic grouping of the isolate.

5. If no agglutination is found with any of the polyvalent sera and saline, repeat steps 1 - 3 above using the Vi antiserum. If a positive reaction is found with the Vi antiserum, prepare a dense cell suspension of the organism in 0.85% saline and heat the suspension to 100°C for 60 minutes or autoclave at 121°C for 15 minutes, then repeat the agglutination test using polyvalent and Vi sera on the heated cell suspension.

If the live cell isolate produces a negative result with the polyvalent antisera and a positive result with the Vi antiserum, whilst the heated cell isolate produces a positive result with the polyvalent antisera and a negative result with the Vi antiserum the isolate should probably be regarded as *Salmonella typhi* (O9, 12, Vi). A few other organisms e.g. *Salmonella paratyphi* C (O6, 7, Vi) also contain the Vi antigen but may easily be distinguished by their O antigenic specificities.

b. Tube Agglutination method.

This method may be used for confirmatory testing only.

Live cell suspensions may be used as antigens, but care must be taken to handle cultures with care to avoid laboratory infection. Killed O antigens may be prepared by heating a saline suspension of organisms to 100°C for 10 minutes, centrifuging and resuspending the deposit in saline. The use of phenol or formalin in O antigen preparations should be avoided since these substances inhibit O agglutination in the presence of H antigens.

1. Remove colonies from a suitable agar or broth culture and prepare a fairly light suspension of bacteria of approximately 10^9 organisms per ml in sterile saline (0.85% saline).
2. Make serial dilutions of antiserum in 0.5ml volumes of saline from 1:10 to 1:640 or 1:1280. Round bottomed glass tubes approximately 9 x 85mm are most suitable.
3. To each tube add 0.5ml of antigen suspension.

Note: this doubles the dilution of the antiserum.

4. A control tube should be additionally set up containing 0.5ml of antigen suspension and 0.5ml of saline.
5. Shake the tubes thoroughly and incubate O factor tubes at 50°C for 4 hours or Vi factor tubes at 37°C for 2 hours then at 4°C for 18 hours.

Note: Vi factor tubes should be allowed to warm to room temperature before reading.

6. Examine the tubes for agglutination. Positive agglutination will be observed as obvious granular agglutination. In a negative reaction and in the saline control the appearance of the suspension should be unchanged, with a cloudy appearance, and show a typical swirl on agitation. The titre value is the dilution of the last tube showing agglutination. Titres at or near the stated value (available on request) indicate that the antigen is of the same serotype as the antiserum.

If agglutination is observed in the saline control, the test is invalid. A new antigen solution should be prepared from a fresh culture preparation of the organism and retested.

5. Salmonella H - Typing Antisera

Salmonella flagellar (H) antigens may occur as two different serological entities, known as phases. Antigens labelled alphabetically may appear in phases 1 or 2, but antigens labelled numerically only appear in phase 2. The alphabetically labelled antigens are designated as:- a-z, and z_1 - z_{68} and the numerically labelled antigens are designated as:- 1, 2, 5, 6 and 7. For the alphabetically labelled antigens certain factors are always seen in association with others. These are listed below:-

E complex - contains e and one or more of n, h, x, z₁₅

G complex - contains g and one or more of f, m, p, q, s, t, u, z₅₁, (z₅₂, z₆₂, z₆₃)

L complex - contains l and one or more of v, w, z₁₃, z₂₈, (z₄₀)

z₄ complex - contains z₄ and one or more of z₂₃, z₂₄, z₃₂

For the numerically labelled antigens factors 2, 5, 6 and 7 are always seen in association with factor 1, e.g. 1, 2 or 1,7. Antigen z₆ is also linked with this group.

Once the O - group of a Salmonella strain has been determined the organism may be identified further using MAST ASSURE™ Salmonella H-Typing Antisera. A list of the MAST ASSURE™ Salmonella H-Typing Antisera available is given in Table 3.


To determine the H-type of a Salmonella strain, polyvalent H antisera should be used initially to narrow down the range before specific typing sera are used. This is illustrated schematically in Figure 2 and according to methods as detailed in section 6.

Table 3 - MAST ASSURE™ Salmonella H-Typing Antisera Details

Mast Product Code	MAST ASSURE™ Salmonella H Antiserum	Agglutinins present	Comments
M14317	Polyvalent Phase 1 & 2	a-z ₂₉	
M10339	Polyvalent Phase 2 (H-1)	1, 2, 5, 6, 7, z ₆	
M14324	Rapid Diagnostic 1	b; d; E (n, x, h); r	
M14331	Rapid Diagnostic 2	d; E (n, x, h); k; L (l, v, w)	
M14348	Rapid Diagnostic 3	d; E (n, x, h); G (g, m, s, t); k	
M10327	H - a	a	
M10328	H - b	b	
M10329	H - c	c	
M10330	H - d	d	
M14335	H - E	n, x, h	
M10331	H - e, h	h	A factor of H - E
M10338*	H - e, n	n, x.	A factor of H - E
M10376*	H - x	x	A factor of H - E and H - e, n
M10377*	H - z ₁₅	z ₁₅	A factor of H - E and H - e, n
M10332	H - G	g, m, s, t	
M10364	H - f	f	A factor of H - G
M10365	H - m	m	A factor of H - G
M10366	H - p	p	A factor of H - G
M10367	H - q	q	A factor of H - G
M10368	H - s	s	A factor of H - G
M10369	H - t	t	A factor of H - G
M10370*	H - u	u	A factor of H - G
M10333	H - i	i	
M10334	H - k	k	
M10335	H - L	v, w,	
M10336	H - r	r	
M10337*	H - y	y	
M10378*	H - z	z	
M10379*	H - z ₄	z ₄ , z ₂₃ , z ₂₄	
M10372*	H - z ₂₃	z ₂₃	A factor of H - z ₄
M10373*	H - z ₂₄	z ₂₄	A factor of H - z ₄
M10374*	H - z ₃₂	z ₃₂	A factor of H - z ₄
M10380*	H - z ₁₀	z ₁₀	
M10381*	H - z ₂₉	z ₂₉	
M10344	H - 2	2	A factor of H - 1
M10345	H - 5	5	A factor of H - 1
M10346	H - 6	6	A factor of H - 1
M10348*	H - z ₆	z ₆	A factor of H - 1

* Denotes an 8 week lead time

Figure 2 - Summary of Salmonella H-Typing Procedures



This document does not replace the IFU's intended for use with these products. It is primarily aimed at being a resource for education, training and a guide for a laboratory writing Standard Operating Procedures looking at the more traditional methods associated with bacterial agglutinating sera

References available upon request.

6. Procedures for H serotyping and Interpretation of Results

MAST ASSURE™ Salmonella H Typing Antisera are intended for use in the identification of H (flagellar) antigens. Polyvalent antisera should preferably be used in slide agglutination tests. Specific H-typing sera may also be used in initial qualitative slide agglutination, although they may be used in quantitative tube agglutination tests for confirmatory purposes.

Normally an isolate is tested initially against the Polyvalent H phase 1 & 2 serum (specific and non-specific antigens). If this shows agglutination it should be tested against the Polyvalent H phase 2 serum (non-specific antigens i.e. 1, 2, 5, 6, 7, z₆). If it is not agglutinated by this serum the isolate is assumed to be in the specific phase and attempts should be made to determine the phase 1 antigens using the Rapid antisera, Polyvalent H - G, L or E complex antisera or monovalent H-antisera, as described below.

The commonly occurring Phase 1 antigen, with the exception of factor i, may be identified by slide agglutination tests using the Rapid Diagnostic Antisera, as illustrated below:-

Antigenic factor(s)	Rapid Diagnostic 1	Rapid Diagnostic 2	Rapid Diagnostic 3
B	+	+	-
d	+	-	+
E	+	+	+
G	-	-	+
k	-	+	+
L	-	+	-
r	+	-	-

For H-serotypes possessing common determinants (for example the G group: f,g; g,m; g,p; g,s,t) specific antisera should be used for characterising the antigenic group further e.g. f, m, p, s or t.

Sometimes only one of the two phases possessed by a diphasic organism are detected in this test. To determine the antigens in the other phase, secure the culture in the other phase by phase induction (see section 8 - Phase Induction Procedures).

a. Slide Agglutination for H-antigen typing.

Cultures of organisms identified as *Salmonella* by their morphological and biochemical features and O-grouped as detailed in section 4 may be H-serotyped by the following procedures.

1. Place two drops of sterile 0.85% saline solution (saline) onto a carefully cleaned microscope slide. The slide may be partitioned into several parts using a chimograph or glass pencil. With an inoculation loop or wire emulsify into each drop of saline a live cell colony from a fresh agar plate or slope culture to produce a distinct and uniform turbidity.
2. Place a drop of polyvalent antiserum onto one of the drops of emulsified isolate and to the other a drop of saline as a control.

Note: allow the antiserum to freefall from the dropper provided with the bottle. Do not contaminate the antiserum with organism.

3. Mix the reagents by tilting the slide back and forth for 60 seconds while viewing under indirect light against a dark background.
4. Distinct clumping or agglutination within this period, without clumping in the saline control (auto-agglutination) should be regarded as a positive result.

An isolate producing a distinct positive reaction with a polyvalent antiserum is assumed to be a *Salmonella* bearing one or more of the H antigenic factors represented by that antiserum. Using this information, further testing of the isolate should be conducted, as described in steps 1 - 3, with specific H antisera to reveal the full H antigenic grouping of the isolate.

b. Tube Agglutination method.

1. Prepare a suspension of bacteria in 0.5% (v/v) formal saline or formalinised broth cultures may be used. Colonies taken from primary isolation media may be unsatisfactory for use in determining the H serotype due to poor motility of organisms. This may be improved by using organisms grown at 37°C for 6-8 hours in a broth culture, by subculture onto moist agar slopes, using 0.5% agar in a Petri dish or in 0.2% agar in a Craigie tube and picking the leading edge of the culture after incubation.

2. Make serial dilutions of antiserum in 0.5ml volumes of saline from 1:10 to 1:640 or 1:1280. Round bottomed glass tubes approximately 9 x 85mm are most suitable.

3. To each tube add 0.5ml of antigen suspension.

Note: this doubles the dilution of the antiserum.

4. A control tube should be additionally set up containing 0.5ml of antigen suspension and 0.5ml of saline.

5. Shake the tubes thoroughly and incubate tubes at 50-52°C for 1-2 hours.

6. Examine the tubes for agglutination. Positive agglutination will be observed as characteristic floccular agglutination. In a negative reaction and in the saline control the appearance of the suspension should be unchanged, with a cloudy appearance, and show a typical swirl on agitation. The titre value is the dilution of the last tube showing agglutination. Titres at or near the stated value (available on request) indicate that the antigen is of the same serotype as the antiserum.

If agglutination is observed in the saline control, the test is invalid. A new antigen solution should be prepared from a fresh culture preparation of the organism and retested.

7. Salmonella H Antigen Phase Induction

Sometimes it is necessary to isolate the second phase of a diphasic organism for complete serological identification. A list of the MAST ASSURE™ *Salmonella* H Phase Induction Antisera available is given in Table 4.

These antisera are provided sterile in sealed injection vials and contain no preservative. For use, the required volume may be withdrawn under sterile conditions using a syringe and needle.

8. Salmonella H Antigen Phase Induction Procedures

Phase induction may be done according to the following procedures:-

A. Craigie Tube Method

1. Add 0.1ml of the H antiserum by which the organism is agglutinated, to about 3ml of semi-solid nutrient agar held at 50°C in a water bath. Mix the contents taking care to avoid frothing.
2. After mixing the serum and agar, aseptically place a previously sterilised Craigie tube vertically into the medium, with the upper end projecting well above the agar surface.
3. After the medium has cooled and solidified, inoculate the organism using a straight wire into the agar inside the Craigie tube. Incubate the culture at 37°C overnight (16-18 hours).

4. After incubation, remove organisms from the agar outside the Craigie tube, place into glucose broth and incubate at 37°C for 6-8 hours. By this time there should be enough growth to assess the H antigens of the induced phase.
5. Determine the H antigens of the induced phase using MAST ASSURE™ - SALMONELLA Antisera.
For method see separate instructions.

If the organism fails to appear outside the Craigie tube at 37°C overnight (16-18 hours), the cultures should be left a further 24 hours, or reduce the volume of serum used for phase induction by half to 0.05ml. If such procedures fail to work, it should be assumed that the organism has flagella of only one phase.

B. Bridging Method

1. Cut a 50 × 20mm ditch in a well dried nutrient agar plate.
2. Soak a strip of previously sterilised filter paper (approximately 36 × 7mm) in the H antiserum by which the organism is agglutinated and place this strip across the ditch at right angles. At one end of the filter paper strip place a sterilised filter paper disc (approximately 7mm in diameter) so that half of it is on the serum strip and half is on the agar.
3. Inoculate the agar at the opposite end of the paper strip to the disc with organisms from a 6-8 hour nutrient broth culture of the organism and incubate overnight (16-18 hours) at 37°C.
4. After incubation, remove the paper disc with sterile forceps and place it into glucose broth and incubate at 37°C for 4 hours. By this time there should be enough growth to assess the H antigens of the induced phase.
5. Determine the H antigens of the induced phase using MAST ASSURE™ - SALMONELLA Antisera.
For method see separate instructions.

If the organism fails to appear on the paper disc after incubation at 37°C overnight (16-18 hours), repeat the test. If the procedures fail to work again, it should be assumed that the organism has flagella of only one phase.

Note:- ensure that the surface of the agar plate is dry before use and that the filter paper strip is not over saturated with antiserum. If moisture is present on the surface of the plate organisms may swarm round the side of the ditch and be recovered on the filter paper disc giving erroneous results.

Table 4 - MAST ASSURE™ Salmonella Phase Induction Antisera Details Non –CE marked, special order only, please check lead times

Mast Product Code	MAST ASSURE™ Salmonella H Antiserum	Agglutinins present
M10349/NCE	H - a	a
M10350/NCE	H - b	b
M10351/NCE	H - c	c
M10352/NCE	H - d	d
M10353/NCE	H - e, h	h
M10354/NCE	H - G	g, m, s, t
M10355/NCE	H - i	i
M10356/NCE	H - k	k
M10357/NCE	H - L	l, v, w
M10358/NCE	H - r	r
M10359/NCE	H - y	y
M10360/NCE	H - e, n	n, x
M10361/NCE	H - 1	1, 2, 5, z ₆
M10382/NCE	H - z	z
M10383/NCE	H - z ₄	z ₄
M10384/NCE	H - z ₁₀	z ₁₀
M10385/NCE	H - z ₂₉	z ₂₉

9. Antisera Sets

MAST ASSURE™ Salmonella Antisera are also available as convenient sets as listed in Table 5 and comprise of individual bottles of stated antisera.

Table 5 - MAST ASSURE™ Salmonella Antisera Sets

Mast Product Code	MAST ASSURE™ Salmonella Antisera Set Components	Comments
M10303*	Salmonella Antisera Set 1 - This set consists of 17 O-Grouping + Vi antisera. i.e. Polyvalent O and O1, and O-Grouping sera O2; O4; O7; O8; O9; O ₉ , 46; O ₃ , 10; O ₁ , 3, 19; O11; O13; O ₆ , 14; O16; O18; O21; O35; and Vi	2ml × 18
M10304*	Salmonella Antisera Set 2 - This set consists of 17 H-Typing antisera. i.e. a; b; c; d; e, h; G; i; k; L; r; y; e, n; 1; z; z ₄ ; z ₁₀ ; z ₂₉	5ml × 17
M10386*	Salmonella Antisera Set 3 - This set consists of 4 H-L complex antisera. i.e. v; w; z ₁₃ ; z ₂₈	5ml × 4
M10387*	Salmonella Antisera Set 4 - This set consists of 5 H-1 complex antisera. i.e. 2; 5; 6; 7; z ₆	5ml × 5
M10388*	Salmonella Antisera Set 5 - This set consists of 7 H-G complex antisera. i.e. f; m; p; q; s; t; u	5ml × 7
M10389*	Salmonella Antisera Set 6 - This set consists of 3 H-z ₄ complex antisera and 2 H-e,n complex antisera. i.e. z ₂₃ ; z ₂₄ ; z ₃₂ ; x; z ₁₅	5ml × 5
M10390*	Salmonella Antisera Set 7 - This set consists of O-Grouping O2 and O9 antisera, H-Typing antisera a and d, and Vi antiserum	2ml × 3 + 5ml × 2
M10391/NCE*	Salmonella H-sera for Phase Induction - This set consists of 17 H-Phase Induction antisera. i.e. a; b; c; d; e, h; G; i; k; L; r; y; e, n; 1; z; z ₄ ; z ₁₀ ; z ₂₉	5ml × 17

* Non CE marked, special order only, please check lead times

* Denotes an eight week lead time.

Bacterial Agglutinating Antisera

Order No

Product

Packsize

mastassure™

MAST ASSURE™ ANTISERA AVAILABLE FROM STOCK

SALMONELLA O ANTISERA - MONOVALENT

M10310	Salmonella O Factor O2	2ml
M10311	Salmonella O Factor O4	2ml
M10312	Salmonella O Factor O7	2ml
M10313	Salmonella O Factor O8	2ml
M10314	Salmonella O Factor O9	2ml
M10316	Salmonella O Factor O3,10	2ml
M10321	Salmonella O Factor O6,14	2ml
M10326	Salmonella O Factor Vi	2ml

SALMONELLA O ANTISERA - POLYVALENT

M10308	POLY O Factor O2, O4, O7, O8, O9, O9, 46, O3, 10 and O1,3,19	2ml
M10309	POLY O1 Factor O11, O13, O6, 14, O16, O18, O21, O35	2ml
M14294	POLY O A-G	2ml
M14300	POLY O A-S	2ml
M92537	Omnivalent (Kauffmann-White group A-067)	2ml

SALMONELLA H ANTISERA - MONOVALENT

M10327	Salmonella H Factor a	2ml
M10328	Salmonella H Factor b	2ml
M10329	Salmonella H Factor c	2ml
M10330	Salmonella H Factor d	2ml
M10331	Salmonella H Factor e, h	2ml
M14335	Salmonella H Factor G	2ml
M10332	Salmonella H Factor E	2ml
M10333	Salmonella H Factor i	2ml
M10344	Salmonella H Factor 2	2ml
M10345	Salmonella H Factor 5	2ml
M10346	Salmonella H Factor 6	2ml
M10364	Salmonella H Factor f	2ml
M10365	Salmonella H Factor m	2ml
M10366	Salmonella H Factor p	2ml
M10367	Salmonella H Factor q	2ml
M10368	Salmonella H Factor s	2ml
M10369	Salmonella H Factor t	2ml

SALMONELLA H ANTISERA - POLYVALENT

M14317	POLYVALENT PHASE 1&2 (a-z29)	2ml
M10339	POLYVALENT PHASE 2 (H-1)	2ml
M14324	RAPID DIAGNOSTIC 1 Factors b, d, E, r	2ml
M14331	RAPID DIAGNOSTIC 2 Factors b, E, k, l	2ml
M14348	RAPID DIAGNOSTIC 3 Factors d, E, G, k	2ml

mastassure™

**PATHOGENIC
ESCHERICHIA COLI ANTISERA**

A. MAST ASSURE™ Pathogenic Escherichia coli Antisera

1. Introduction

a. Escherichia coli and Disease

Escherichia coli (*E. coli*) is a group of Gram negative organisms belonging to the family *Enterobacteriaceae*. They are non-spore forming rods, and most are motile possessing flagella.

Many strains of *E. coli* occur naturally in human and animal intestine where they are the predominant organism in the aerobic commensal flora. Some strains however may be associated with human infection, causing urinary tract infections (the most common cause of acute, uncomplicated urinary tract infection), diarrhoea and gastroenteritis, suppurative lesions, abscesses in a variety of wounds, neonatal septicaemia and meningitis, or with animal infections causing mastitis, pyometria in bitches, coli granulomata in fowls and white scours in calves.

There are four groups of pathogenic *E. coli* that cause diarrhoea, acute gastritis or colitis based on their pathogenic mechanisms:- Enteropathogenic (EPEC), Enterotoxigenic (ETEC), Enteroinvasive (EIEC) and Enterohaemorrhagic (EHEC) or now referred to as Verotoxigenic (VTEC).

- The Enteropathogenic *E. coli* (EPEC) cause infantile enteritis especially in tropical countries, where outbreaks often occur in hospitals resulting in high mortality rates. Such outbreaks in industrialised countries have become uncommon. Symptoms include diarrhoea, stomach ache, fever and vomiting.
- The Enterotoxigenic *E. coli* (ETEC) produce a heat-labile (ST) or a heat-stable (LT) enterotoxin or both. In addition they possess colonisation factors that are specific for the host animal species and which enables the organisms to adhere to the epithelium of the small intestine. Symptoms include diarrhoea, stomach ache, fever and vomiting.
- The Enteroinvasive *E. coli* (EIEC) cause an illness identical to *Shigella* dysentery in patients of all ages. Symptoms include diarrhoea (with mucus and blood), stomach ache and fever.
- The Verotoxigenic *E. coli* (VTEC) produce one or two Vero cytotoxins (VT1 and VT2). VT1 is closely related to the so-called Shiga toxin produced by strains of *Shigella dysenteriae* 1 and is sometimes called Shiga-like toxin. VTEC organisms cause a range of symptoms from mild, watery diarrhoea to a severe diarrhoea with large amounts of fresh blood in the stool (haemorrhagic colitis). An important complication in children is the Haemolytic uraemic syndrome. Symptoms include diarrhoea (with blood), stomach ache and fever.

The most common groups of diarrhoea causing *E. coli* are given below:-

Pathogenic <i>E. coli</i> group	Common <i>E. coli</i> Serogroup
Enteropathogenic <i>E. coli</i> (EPEC)	O26, O44, O55, O86, O111, O114, O119, O125, O126, O127, O128, O142, O158
Enterotoxigenic <i>E. coli</i> (ETEC)	O6, O8, O15, O20, O25, O27, O63, O78, O85, O115, O148, O153, O159, O167
Enteroinvasive <i>E. coli</i> (EIEC)	O28, O29, O32, O42, O112, O124, O136, O143, O144, O152, O164
Verotoxigenic <i>E. coli</i> (VTEC)	O157

b. Antigenic characterisation of Escherichia coli

Members of the species *E. coli* possess three main types of surface antigen, the O (somatic), the K (capsular) and the H (flagellar) antigens. The O antigens are heat stable and upon which grouping of the organisms are based. The H and K antigens are heat labile and normally only the H antigens are used for identifying the serotype of an organism further. The term K antigen was originally used collectively for surface or capsular antigens and were divided into three classes L, A and B according to the effect of heat on the agglutinability, antigenicity and antibody binding power. The designations L, A and B are now no longer used and indeed the determination of K antigens is no longer deemed necessary for identification of *E. coli* serotype. Many of the antigens thought to be K antigens are now known not to be K antigens. K antigens are only used diagnostically for example with certain toxigenic *E. coli* of veterinary importance. These toxigenic *E. coli* have pili antigens which are antigenically different from those of other *E. coli*. The pili are important not only for attachment and infection of the organism to the intestinal mucous epithelial cells but also

associated with the organism's ability to produce enterotoxin. The K88 and 987P antigens occur in strains of *E. coli* associated with diarrhoeal disease in swine, and the K99 antigen with diarrhoeal disease in lambs and calves. Enterotoxigenic *E. coli* (ETEC) in humans also have colonisation factor antigens (CFA I, II, and E8775) and there may be others, but they are not routinely used for clinical diagnosis².

In 1947 Kauffmann first proposed a scheme for classifying *E. coli* on the basis of their O antigens. Currently more than 160 O, 90 K and 50 H antigens have been described, and this forms the basis for current serological testing of pathogenic *E. coli*.

c. MAST ASSURE™ Pathogenic *E. coli* Antisera: preparation and presentation.

MAST ASSURE™ Pathogenic *E. coli* Antisera are prepared from rabbits hyperimmunised with standard strains of *E. coli* organisms possessing defined antigenic factors. All sera are heat inactivated at 56°C for 30 minutes, absorbed to remove cross-reacting agglutinins and filter sterilised. The MAST ASSURE™ Pathogenic *E. coli* Antisera provide a comprehensive range of O and H antisera for the determination of O and H antigens of pathogenic *E. coli*. Antigens are identified normally by qualitative slide agglutination or by quantitative tube agglutination tests.

MAST ASSURE™ Pathogenic *E. coli* Antisera are provided as 2ml (or 5ml) amounts in vials with dropper attachments and contain 0.1% sodium azide as preservative. Supplied ready to use. This is sufficient for 50 (125) slide agglutination tests or 20 (50) tube agglutination tests.

2. Culture of *E. coli* - Preparation for Serology

E. coli belong to the family Enterobacteriaceae and there is much cross-reaction and antigenic relationships between *E. coli* and other genera within this family particularly *Shigellae*. Hence it is important that organisms undergoing serological classification should be correctly identified as *E. coli* by morphological and biochemical features first.

3. Pathogenic *E. coli* O -Grouping and H -Typing Antisera Products

Table 1 lists the MAST ASSURE™ Polyvalent Pathogenic *E. coli* O-Grouping Antisera available together with the product codes, Table 2 lists the MAST ASSURE™ single factor Pathogenic *E. coli* O-Grouping Antisera available and Table 3 lists the MAST ASSURE™ Pathogenic *E. coli* H-Typing Antisera available. It is impractical to provide single factor sera for all known *E. coli* antigens, however a wide range of sera is available in the MAST ASSURE™ range which is sufficient to identify with a reasonable degree of probability the majority of *E. coli* types isolated.

To determine the O-group of a pathogenic *E. coli* isolate, polyvalent O-Grouping antisera should be used initially to narrow down the range before specific grouping sera are used. This is illustrated schematically in Figure 1 and according to methods as detailed in section 4.

MAST ASSURE™ Pathogenic *E. coli* Antisera are also available as convenient sets as listed in Table 4 and comprise of individual bottles of stated antisera.

4. Procedures for Serogrouping and Typing, and Interpretation of Results

MAST ASSURE™ Pathogenic *E. coli* Antisera are intended for use in the identification of O antigens by qualitative slide agglutination, although they may be used in quantitative tube agglutination tests for confirmatory purposes.

Cultures of organisms identified as *E. coli* by their morphological and biochemical features may be serotyped by the following procedures. Also refer to figure 6 for a summary of the Pathogenic *E. coli* O-grouping and H-Typing procedures.

A. O-Antigen Grouping

a. Slide Agglutination of live organisms

1. Place two drops of sterile 0.85% saline solution (saline) onto a carefully cleaned microscope slide. The slide may be partitioned into several parts using a chinograph or glass pencil. With an inoculation loop or wire emulsify into each drop of saline a live cell colony from a fresh agar plate or slope culture to produce a distinct and uniform turbidity.
2. Place a drop of polyvalent antiserum onto one of the drops of emulsified isolate and to the other a drop of saline as a control.

Note: allow the antiserum to freefall from the dropper provided with the bottle. Do not contaminate the antiserum with organism.

3. Mix the reagents by tilting the slide back and forth for 60 seconds while viewing under indirect light against a dark background.
4. Distinct clumping or agglutination within this period, without clumping in the saline control (auto-agglutination) should be regarded as a positive result.
An isolate producing a distinct positive reaction with a polyvalent antiserum is assumed to be an *E. coli* bearing one or more of the O antigenic factors represented by that antiserum. Using this information, further testing of the isolate should be conducted, as described in steps 1 - 3, with single factor O antisera to reveal the full O antigenic grouping of the isolate.
Always confirm the O grouping by slide agglutination on heat killed organisms (see below).

In general the serogroups covered by polyvalent O antisera D1, D2 and D3 are considered to be enteropathogenic, serogroups covered by polyvalent sera D4, D5 and D6 are considered to be enterotoxigenic, and serogroups covered by polyvalent sera D7 and D8 are considered to be enteroinvasive. These represent the most common O antigenic groups encountered.

b. Slide Agglutination for heat killed organisms

If the live organisms give a positive reaction with a particular single factor O antiserum, prepare a dense cell suspension of the organism in 0.85% saline and heat the suspension to 100°C for 60 minutes or autoclave at 121°C for 15 minutes then repeat the agglutination test as above using single factor O antiserum on the heated cell suspension. This should be done to identify the O antigen type as distinct from the K antigen.

c. Interpretation of Results

If both live and heat killed cells of the bacterial isolate, which has been previously morphologically characterised as an *E. coli*, positively agglutinate with a single factor antiserum the organism should be regarded as a pathogenic *E. coli* belonging to the sero-group represented by single factor antiserum. However if the live cells give a positive agglutination result while the heat killed cells do not, the O sero-group cannot be determined as represented by that single factor antiserum.

d. Tube Agglutination

Quantitative tube agglutination tests may be carried out according to the following method:-

1. The cell suspension should be prepared by removing colonies from a suitable pure agar culture or from a centrifuged and washed broth culture, and resuspending them in sterile saline or 0.5% formal saline to give a fairly light suspension (about 7.5×10^8 organisms per ml).

2. Heat the suspension for 1 hour at 100°C to obtain an O antigen suspension. Alternatively a heat treated 4-6 hour broth culture may be used.
3. Make serial dilutions of antisera in 0.5ml volumes of saline, from 1:10 to 1:1280. Round bottomed glass tubes approximately 9 × 85mm are most suitable.
4. To each tube add 0.5ml of the antigen suspension.

Note: this doubles the dilution of the antiserum.

5. A control tube should additionally be set up containing 0.5ml of antigen suspension and 0.5ml of saline.
6. Shake the tubes thoroughly and incubate the tubes at 50°C for 16-24 hours.
7. Examine the tubes for agglutination. Positive agglutination will appear as a clearing of the fluid with sediment that rises in a granular mass when the tube is flicked with the finger. The control tube should be cloudy and any sediment should resuspend on flicking. Agglutination titres of 1:20 are not counted as significant. Titres at or near that stated for the antiserum indicate that the antigen is of the O group stated by the antiserum that caused the agglutination.

B. H - Antigen Serotyping

The majority of *E. coli* strains are poorly motile when first isolated; hence it is normally necessary to passage them serially through several tubes of semi-solid agar motility medium to enhance motility and H antigen development. The number of passages required may vary with individual cultures.

Tube Agglutination Test

1. Allow the organism to pass through Craigie tubes with semi-solid nutrient medium 3-5 times. Then inoculate the passaged organism into a suitable nutrient broth tubes and incubate at 37°C for 6-8 hours.
2. After incubation make a 1:2 dilution by adding an equal volume of 0.85% saline containing 1% (v/v) formalin to the culture.
3. Add 2 drops of the required type specific H serum into a small test tube, and add 0.45 - 0.5ml of the treated organism suspension it. As a control prepare a similar tube which contains only the antigen suspension.
4. Shake the contents of the test tube thoroughly and allow the tubes to stand in a water bath at 50-52°C for 1 hour.
5. Observe the tubes for spontaneous and distinct agglutination seen easily with the naked eye. Do not shake them as this will disturb the agglutination pattern. An isolate producing a distinct positive reaction is assumed to be an *E. coli* bearing the H antigenic factors represented by that antiserum.

Quantitative tube agglutination tests may be performed as described in section d above (O - Antigen Grouping, Tube Agglutination Procedure).

Limitations of the Test

1. The test results obtained when using these sera cannot determine whether a particular organism is pathogenic. To determine pathogenicity it is necessary to test for factors associated with pathogenicity e.g. toxins.
2. In general the serotypes covered by polyvalent sera 1, 2 and 3 are considered to be enteropathogenic, serotypes covered by polyvalent sera 4, 5 and 6 are considered to be enterotoxigenic, and serotypes covered by polyvalent sera 7 and 8 are considered to be enteroinvasive. However, it must be noted that there are other pathogenic *E. coli* that are not covered by these antisera.
3. Organisms must be morphologically and biochemically characterised as being an *E. coli* before testing with the antisera as cross-reactions with closely related members of the Enterobacteriaceae may occur.

Table 1 - MAST ASSURE™ Pathogenic *E. coli* Polyvalent O Antisera Details

Mast Product Code	MAST ASSURE™ Pathogenic <i>E. coli</i> O Antisera	O serotypes detected
M12005*	<i>E. coli</i> O - Polyvalent D1	O1; O26; O86a; O111; O119; O127a, O128;
M12006*	<i>E. coli</i> O - Polyvalent D2	O44; O55; O125; O126; O146; O166
M12007*	<i>E. coli</i> O - Polyvalent D3	O18; O114; O142; O151; O157; O158
M12008*	<i>E. coli</i> O - Polyvalent D4	O6; O27; O78; O148; O159; O168
M12009*	<i>E. coli</i> O - Polyvalent D5	O20; O25; O63; O153; O167
M12010*	<i>E. coli</i> O - Polyvalent D6	O8; O15; O115; O169
M12011*	<i>E. coli</i> O - Polyvalent D7	O28ac; O112ac; O124; O136; O144
M12012*	<i>E. coli</i> O - Polyvalent D8	O29; O143; O152; O164
M14263	<i>E. coli</i> O - Polyvalent 2	O26; O55; O111; O119; O126
M14270	<i>E. coli</i> O - Polyvalent 3	O86; O114; O125; O127; O128
M14287	<i>E. coli</i> O - Polyvalent 4	O44; O112; O124; O142

* Denotes an eight week lead time.

Table 2 - MAST ASSURE™ Pathogenic *E. coli* Single Factor O-Grouping Antisera

Mast Product Code	<i>E. coli</i> O Serotype
M12013*	O1
M12014	O26
M12015*	O86a
M12016	O111
M12017	O119
M12018	O127a
M12019*	O128
M12020*	O44
M12021*	O55
M12022*	O125
M12023*	O126
M12024*	O146
M12025*	O166
M12026*	O18
M12027*	O114
M12028*	O142
M12029*	O151
M12030	O157
M12031*	O158
M12032*	O6
M12033*	O27
M12034*	O78
M12035*	O148
M12036*	O159
M12037*	O168
M12038*	O20
M12039*	O25
M12040*	O63
M12041*	O153
M12042*	O167
M12043*	O8
M12044*	O15
M12045*	O115
M12046*	O169
M12047*	O28ac
M12048*	O112ac
M12049*	O124
M12050*	O136
M12051*	O144
M12052*	O29
M12053*	O143
M12054*	O152
M12055*	O164

* Denotes an eight week lead time.

Table 3 - MAST ASSURE™ Pathogenic *E. coli* Single Factor H -Typing Antisera

Mast Product Code	<i>E. coli</i> H Serotype
M12056*	H-2
M12057*	H-4
M12058*	H-5
M12059*	H-6
M12060*	H-7
M12061*	H-9
M12062*	H-10
M12063*	H-11
M12064*	H-12
M12065*	H-16
M12066*	H-18
M12067*	H-19
M12068*	H-20
M12069*	H-21
M12070*	H-27
M12071*	H-28
M12072*	H-34
M12073*	H-40
M12074*	H-41
M12075*	H-42
M12076*	H-45
M12077*	H-51


* Denotes an eight week lead time.

Table 4 - MAST ASSURE™ Pathogenic *E. coli* Polyvalent O-Grouping and H-Typing Antisera Sets

Mast Product Code	MAST ASSURE™ Pathogenic <i>E. coli</i> Antisera Set components	Comments
M12001*	<i>E. coli</i> O Antisera Set 1 - This set consists of 8 O-Grouping Polyvalent Antisera and 43 single factor O-Grouping antisera. i.e. Polyvalent. D1 - D8, and single factor sera O1; O26; O86a; O111; O119; O127a, O128; O44; O55; O125; O126; O146; O166; O18; O114; O142; O151; O157; O158; O6; O27; O78; O148; O159; O168; O20; O25; O63; O153; O167; O8; O15; O115; O169; O28ac; O112ac; O124; O136; O144; O29; O143; O152; O164.	2ml × 51
M12002*	<i>E. coli</i> O Antisera Set 2 - This set consists of 22 H-Typing Antisera. i.e. 2; 4; 5; 6; 7; 8; 9; 10; 11; 12; 16; 18; 19; 20; 21; 27; 28; 34; 40; 42; 45; 51	5ml × 22

* Denotes an eight week lead time.

Figure 1- Summary of Pathogenic *E. coli* O-Grouping and H-Serotyping Procedures



This document does not replace the IFU's intended for use with these products. It is primarily aimed at being a resource for education, training and a guide for a laboratory writing Standard Operating Procedures looking at the more traditional methods associated with bacterial agglutinating sera

References available upon request.

Bacterial Agglutinating Antisera

Order No

Product

Packsize

mastassure™

MAST ASSURE™ ANTISERA AVAILABLE FROM STOCK

E.COLI O ANTISERA - MONOVALENT

M12014	<i>Escherichia coli</i> Factor O26	2ml
M12016	<i>Escherichia coli</i> Factor O111	2ml
M12017	<i>Escherichia coli</i> Factor O119	2ml
M12018	<i>Escherichia coli</i> Factor O127a	2ml
M12030	<i>Escherichia coli</i> Factor O157	2ml

E.COLI O ANTISERA - POLYVALENT

AVAILABLE FROM STOCK

M14263	<i>E. coli</i> POLY 2 Factors O26, O55, O111, O119, O126	2ml
M14270	<i>E. coli</i> POLY 3 Factors O86, O114, O125, O127, O128	2ml
M14287	<i>E. coli</i> POLY 4 Factors O44, O112, O124, O142	2ml

A 3D digital illustration of Shigella bacteria, depicted as orange, rod-shaped organisms with flagella, and numerous purple, spherical particles. The scene is set against a dark, gradient background with soft lighting and shadows, giving it a scientific and microscopic appearance.

mastassure™

SHIGELLA ANTISERA

A. MAST ASSURE™ Shigella Antisera

1. Introduction

a. Shigella and Disease

The genus *Shigella* is a group of Gram negative organisms belonging to the family *Enterobacteriaceae* and is closely related to the genus *Escherichia*. They are non-spore forming rods that possess no flagella and hence are non-motile.

Shigellae are the causative agents of classic bacillary dysentery. The disease is usually spread via the faecal-oral route, though occasional epidemics have been traced to unchlorinated water supplies. The disease is endemic in most warm-climate countries and is also prevalent in many countries with a temperate climate. In the UK and other countries outbreaks are common in institutions and schools where more lax sanitation allows spread of organism.

The disease varies in severity according to species and serotype of the causal organism, but is characterised by severe blood stained, mucopurulent diarrhoea accompanied by fever and malaise. A brief episode of watery diarrhoea often precedes the severe bloody flux. In the more virulent forms such as that caused by *Shigella dysenteriae* symptoms such as severe water and electrolyte loss from the small intestine are caused by toxin production, and indeed the infectious dose of organisms required to produce disease is small. After reaching the large intestine the organisms multiply in the gut lumen where many adhere to and invade epithelial cells of the gut mucosa. Organisms then multiply within the epithelial cells and spread to adjacent cells and patches of necrotic epithelial cells slough off and ulcers form. A cellular immune response occurs resulting in the presence of large numbers of polymorphonuclear phagocytes, which can be readily observed on microscopic examination of the stool.

Laboratory diagnosis of Shigella infection usually depends upon the isolation and identification of the causal organism from a specimen of the patient's blood or faeces. Microscopic examination of the specimen is normally conducted to be able to distinguish bacterial dysentery from amoebic dysentery and to note the cellular exudate. Further antigenic analysis of the organism is often needed to identify the species or strain for epidemiological purposes.

b. Antigenic characterisation of Shigellae

Modern classification of Shigellae owe much to Ewing who built on the foundation of Murray, Andrews and Inman and Boyd whose recognition of type specific and group specific antigens introduced some system into the serological analysis of Shigellae. Shigellae are currently differentiated into four subgroups on the basis of their O (somatic) antigens and further differentiated into serotypes as detailed below:-

- | | |
|----------------|---|
| Group A | - <i>Sh. dysenteriae</i> , contains 12 distinct antigenic serotypes. |
| Group B | - <i>Sh. flexneri</i> , contains 6 serotypes (I-VI) that can be divided further into sub serotypes according to their possession of group factors designated 3,4; 4; 6; 7; and 7, 8 (see Table1). |
| Group C | - <i>Sh. boydii</i> , contains 18 distinct antigenic serotypes. |
| Group D | - <i>Sh. sonnei</i> , contains only one distinct serotype that may occur in two forms, phase I (smooth) and phase II (rough). |

Table 1 - Antigenic Structure of *Sh. flexneri*

Serotype	Sub serotype	Type antigen	Group antigen*
1	1a	I	4
1	1b	I	4,6
2	2a	II	3,4
2	2b	II	7,8
3	3a	III	(3,4),6,7,8
3	3b	III	(3,4),6
4	4a	IV	3,4
4	4b	IV	6
5	5a	V	3,4
5	5b	V	7,8
6		VI	(4)
X variant		-	7,8
Y variant		-	3,4

* Not all group antigens are listed.

c. MAST ASSURE™ Shigella Antisera: preparation and presentation.

MAST ASSURE™ Shigella Antisera are prepared from rabbit's hyperimmunised with standard strains of Shigella organisms possessing defined antigenic factors. All sera are heat inactivated at 56°C for 30 minutes, absorbed to remove cross-reacting agglutinins and filter sterilised. The MAST ASSURE™ Shigella Antisera are a comprehensive range of polyvalent and single specificity antisera for the agglutination of specific O serotype and group antigens of Shigellae. Antigens are identified normally by qualitative slide agglutination. Positive results may be confirmed by tube agglutination tests.

All MAST ASSURE™ Shigella Antisera are provided as 2ml amounts in vials with dropper attachments and contain 0.1% sodium azide as preservative. Supplied ready to use. This is sufficient for 50 slide agglutination tests or 20 tube agglutination tests. Supplied ready to use with no preservative is added.

2. Culture of Shigella - Preparation for Serology

Shigella belongs to the family *Enterobacteriaceae* and there is much cross-reaction and antigenic relationships between *Shigella* and other genera within this family especially between the genus *Escherichia*. Hence it is important that organisms undergoing serological classification should be correctly identified as Shigella by morphological and biochemical features first.

3. MAST ASSURE™ Shigella Typing and Grouping Antisera Products

Table 2 lists the MAST ASSURE™ Shigella Typing and Grouping Antisera available, including monovalent and polyvalent antisera, together with the product codes. Antisera are also available as convenient sets and comprise of individual bottles of antisera as stated.

Table 2 - Mast Assure™ Shigella Grouping and Typing Antisera

Order Code	Product	Serotype
M10109	<i>Sh. dysenteriae</i> Polyvalent A	Types 1, 2, 3, 4, 5, 6 and 7
M10110	<i>Sh. dysenteriae</i> Polyvalent A1	Types 8, 9, 10, 11 and 12
M10116	<i>Sh. dysenteriae</i> type 1	Type 1
M10117*	<i>Sh. dysenteriae</i> type 2	Type 2
M10118*	<i>Sh. dysenteriae</i> type 3	Type 3
M10119*	<i>Sh. dysenteriae</i> type 4	Type 4
M10120*	<i>Sh. dysenteriae</i> type 5	Type 5
M10121*	<i>Sh. dysenteriae</i> type 6	Type 6
M10122*	<i>Sh. dysenteriae</i> type 7	Type 7
M10123*	<i>Sh. dysenteriae</i> type 8	Type 8
M10124*	<i>Sh. dysenteriae</i> type 9	Type 9
M10125*	<i>Sh. dysenteriae</i> type 10	Type 10
M10152*	<i>Sh. dysenteriae</i> type 11	Type 11
M10153*	<i>Sh. dysenteriae</i> type 12	Type 12
M10111	<i>Sh. flexneri</i> Polyvalent B	Types I, II, III, IV, V, and VI and groups (3) 4, 6 and 7 (8)
M10126	<i>Sh. flexneri</i> type I	Type I
M10127	<i>Sh. flexneri</i> type II	Type II
M10128*	<i>Sh. flexneri</i> type III	Type III
M10129*	<i>Sh. flexneri</i> type IV	Type IV
M10130	<i>Sh. flexneri</i> type V	Type V
M10131	<i>Sh. flexneri</i> type VI	Type VI
M10132*	<i>Sh. flexneri</i> group (3)4	Group (3) 4
M10133*	<i>Sh. flexneri</i> group 6	Group 6
M10134*	<i>Sh. flexneri</i> group 7(8)	Group 7 (8)
M10112	<i>Sh. boydii</i> Polyvalent C	Types 1, 2, 3, 4, 5, 6, and 7
M10113	<i>Sh. boydii</i> Polyvalent C1	Types 8, 9, 10 and 11
M10114	<i>Sh. boydii</i> Polyvalent C2	Types 12, 13, 14 and 15
M10154	<i>Sh. boydii</i> Polyvalent C3	Types 16, 17 and 18
M10135*	<i>Sh. boydii</i> type 1	Type 1
M10136	<i>Sh. boydii</i> type 2	Type 2
M10137*	<i>Sh. boydii</i> type 3	Type 3
M10138*	<i>Sh. boydii</i> type 4	Type 4
M10139*	<i>Sh. boydii</i> type 5	Type 5
M10140*	<i>Sh. boydii</i> type 6	Type 6
M10141*	<i>Sh. boydii</i> type 7	Type 7
M10142*	<i>Sh. boydii</i> type 8	Type 8
M10143*	<i>Sh. boydii</i> type 9	Type 9
M10144*	<i>Sh. boydii</i> type 10	Type 10
M10145*	<i>Sh. boydii</i> type 11	Type 11
M10146*	<i>Sh. boydii</i> type 12	Type 12
M10147*	<i>Sh. boydii</i> type 13	Type 13
M10148*	<i>Sh. boydii</i> type 14	Type 14
M10149*	<i>Sh. boydii</i> type 15	Type 15
M10155*	<i>Sh. boydii</i> type 16	Type 16
M10156*	<i>Sh. boydii</i> type 17	Type 17
M10157*	<i>Sh. boydii</i> type 18	Type 18
M10115	<i>Sh. sonnei</i> Polyvalent D	Phase I and II

* Denotes an eight week lead time.

Table 2 - Mast Assure™ Shigella Grouping and Typing Antisera (CONTINUED)


Order Code	Product	Serotype
M10150	<i>Sh. sonnei</i> phase I	Phase I
M10151	<i>Sh. sonnei</i> phase II	Phase II
M10104*	Shigella Antisera Set 1	8 Polyvalent - A, A1, B, C, C1, C2, C3, and D, and 41 monovalent sera constituting the above polyvalent sera.
M10105*	Shigella Antisera Set 2	8 Polyvalent - A, A1, B, C, C1, C2, C3, and D, and 11 monovalent sera covering all <i>Sh. flexneri</i> types and groups and <i>Sh. sonnei</i> phase I and II
M10106*	Shigella Antisera Set 3	8 Polyvalent sera - A, A1, B, C, C1, C2, C3, and D.

* Denotes an eight week lead time.

MAST ASSURE™ Shigella Typing and Grouping Antisera are intended for use in the identification of O antigens by qualitative slide agglutination.

Cultures of organisms identified as *Shigella* by their morphological and biochemical features may be serotyped by the following procedures. To determine the Type, Group or Phase of Shigella species isolate, polyvalent antisera should be used initially to narrow down the range before specific sera are used. This is illustrated schematically in Figure 1.

Figure 1 - Summary of Shigella Sero-grouping and Typing Procedures



1. Place two drops of sterile 0.85% saline solution (saline) onto a carefully cleaned microscope slide. The slide may be partitioned into several parts using a chimograph or glass pencil. With an inoculation loop or wire emulsify into each drop of saline a live cell colony from a fresh agar plate or slope culture to produce a distinct and uniform turbidity.
2. Place a drop of polyvalent antiserum onto one of the drops of emulsified isolate and to the other a drop of saline as a control.

Note: allow the antiserum to freefall from the dropper provided with the bottle. Do not contaminate the antiserum with organism.

3. Mix the reagents by tilting the slide back and forth for 60 seconds and while viewing under indirect light against a dark background.
4. Distinct clumping or agglutination within this period, without clumping in the saline control (auto-agglutination) should be regarded as a positive result. Weak agglutination should be counted as negative.

An isolate producing a distinct positive reaction with a polyvalent antiserum is assumed to be a *Shigella* from the letters (A-D) represented by the antiserum. Using this information, further testing of the isolate should be conducted, as described in steps 1 - 4, with monovalent antisera. If the organism is identified as *Sh. flexneri* it should be typed and grouped separately.

5. If the live cells do not produce positive agglutination, prepare a dense cell suspension of the organism in 0.85% saline and heat the suspension to 100°C for 60 minutes or autoclave at 121°C for 15 minutes then repeat the agglutination tests as described in steps 1 - 3. Some *Shigellae* strains possess heat labile capsular (K) antigens, which mask the presence of the heat stable somatic (O) antigens.

Bacterial Agglutinating Antisera

Order No	Product	Packsize
----------	---------	----------

mastassure™


MAST ASSURE™ ANTISERA AVAILABLE FROM STOCK

SHIGELLA ANTISERA - MONOVALENT

M10116	<i>Shigella dysenteriae</i> Type 1	2ml
M10126	<i>Shigella flexneri</i> Type I	2ml
M10127	<i>Shigella flexneri</i> Type II	2ml
M10130	<i>Shigella flexneri</i> Type V	2ml
M10131	<i>Shigella flexneri</i> Type VI	2ml
M10136	<i>Shigella boydii</i> Type 2	2ml
M10150	<i>Shigella sonnei</i> Phase I	2ml
M10151	<i>Shigella sonnei</i> Phase II	2ml

SHIGELLA ANTISERA - POLYVALENT

M10109	<i>S.dysenteriae</i> POLY A Types 1, 2, 3, 4, 5, 6, 7	2ml
M10110	<i>S.dysenteriae</i> POLY A1 Types 8, 9, 10, 11, 12	2ml
M10111	<i>S.flexneri</i> POLY B Types I, II, III, IV, V, VI, (Groups (3)4, 6 & 7 (8)	2ml
M10112	<i>S.boydii</i> POLY C Types 1, 2, 3, 4, 5, 6, 7	2ml
M10113	<i>S.boydii</i> POLY C1 Types 8, 9, 10, 11	2ml
M10114	<i>S.boydii</i> POLY C2 Types 12, 13, 14, 15	2ml
M10154	<i>S.boydii</i> POLY C3 Types 16, 17, 18	2ml
M10115	<i>S.sonnei</i> POLY D Phase I & II	2ml



mastassure™

**VIBRIO CHOLERAЕ
ANTISERA**

A. MAST ASSURE™ *Vibrio cholerae* Antisera

1. Introduction

a. *Vibrio cholerae* and Disease

Vibrio cholerae belongs to the genus *Vibrio* which are Gram negative aerobic or facultative anaerobic comma shaped rods, they are fermentative and nearly all are oxidase positive. They are motile possessing a single polar flagellum.

Cholera is typically characterised by the sudden onset of effortless vomiting and profuse watery diarrhoea. Although vomiting is a common feature the rapid onset of dehydration and hypovolaemic shock, which may cause death in 12 - 24 hours, are related mainly to the profuse, watery, colourless stools with flecks of mucus and a distinctive fishy smell ('rice-water' stools) which contain little protein and are very different from the mucopurulent blood-stained stools of bacillary dysentery. Anuria develops, muscle cramps occur and the patient quickly becomes weak with loss of skin turgor, low blood pressure and an absent or thready pulse. Symptoms vary in severity and milder cases cannot be easily distinguished from other secretory diarrhoeas. Symptomless infections are also common.

Laboratory diagnosis involves culturing the vibrios from stool specimens in alkaline peptone water followed by surface inoculation onto a suitable solid medium such as thiosulphate-citrate-bile salts-sucrose agar (TCBS). On this medium *Vibrio cholerae* strains appear as yellow sucrose-fermenting colonies. Colonies may subsequently be assessed for oxidase production, biochemical confirmatory tests and agglutination by antisera.

b. Antigenic characterisation of *Vibrio cholerae*

Taxonomically *Vibrio cholerae* is a homogeneous species comprising of organisms that are similar to each other biochemically, share a common H (flagellar) antigen and are closely related genetically. The serology of *Vibrio cholerae* is based on the Somatic O antigen scheme of Sakazaki et al. Now more than 80 Serogroup have been identified within the species. Serogroup O1 organisms comprise the causal agents of most epidemic and pandemic cholera outbreaks. A series of pandemic outbreaks of O1 cholera originating in the Bengal basin ravaged the world in the 19th and early 20th centuries. The strain responsible for the seventh outbreak was labelled the El Tor biotype as it was isolated from pilgrims at the quarantine station known as El Tor. Other outbreaks have been attributed to the classical biotype strains.

Strains of *Vibrio cholerae* serotype O1 may be subdivided with absorbed antisera into serovars called Ogawa, Inaba and Hikojima. These serovars share three somatic antigens - a, b and c. Absorption of O1 antisera with Ogawa organisms produces a serum which agglutinates Inaba and Hikojima strains. Similarly, absorption of O1 antisera with Inaba organisms produces a serum which agglutinates Ogawa and Hikojima strains. Unabsorbed serum (containing a, b and c) called 'polyclonal *Vibrio cholerae* antiserum agglutinates all three O1 variants.

Non-O1 *Vibrio cholerae* cause mild, sometimes bloody diarrhoea, often accompanied by abdominal cramps. Symptoms may occasionally be severe in which case the disease resembles cholera. Also some non-O1 strains produce virulence factors including toxins. Hence it may be important for epidemiological purposes to serotype the strain in an isolate.

In the early 1990's a new serotype of *Vibrio cholerae* was found to be the causative agent of a pandemic in India and Bangladesh, and was named serotype O139. It gives rise to an infection as serious as those caused by serotype O1.

c. MAST ASSURE™ *Vibrio cholerae* Antisera: preparation and presentation.

MAST ASSURE™ *Vibrio cholerae* antisera are prepared from rabbits hyperimmunised with standard strains of *Vibrio cholerae* O1 Inaba type and Ogawa type or O139 Bengal type organisms. All sera are heat inactivated at 56°C for 30 minutes, absorbed to remove cross-reacting agglutinins and filter sterilised.

MAST ASSURE™ - VIBRIO CHOLERAEE Antisera are a set of antisera for the agglutination of specific *Vibrio cholerae* O1 antigens and the O139 (Bengal) antigen. Antigens are identified normally by qualitative slide agglutination. Positive results may be confirmed by tube agglutination tests.

All MAST ASSURE™ - VIBRIO CHOLERAEE Antisera are provided as 2ml amounts in vials with dropper attachments and contain 0.1% sodium azide as preservative. Supplied ready to use. This is sufficient for 50 slide agglutination tests or 20 tube agglutination tests. Supplied ready to use with no preservative is added.

2. Culture of *Vibrio cholerae* - Preparation for Serology

It is important that organisms undergoing serological classification should be correctly identified as *Vibrio cholerae* by morphological and biochemical features first.

3. MAST ASSURE™ *Vibrio cholerae* Antisera Products

Table 1 lists the MAST ASSURE™ *Vibrio cholerae* Antisera available , including monovalent and polyvalent antisera, together with the product codes.

Table 1 - Mast Assure™ *Vibrio cholerae* Antisera

Order Code	Product
M11002	<i>Vibrio cholerae</i> Polyvalent (O1) - Serovar Inaba and Ogawa
M11003	<i>Vibrio cholerae</i> (O1) - Serovar Inaba
M11004	<i>Vibrio cholerae</i> (O1) - Serovar Ogawa
M15001	<i>Vibrio cholerae</i> O139 "Bengal"
M11001*	<i>Vibrio cholerae</i> (O1) Antisera Set consisting of 1X Polyvalent antiserum - Serovar Inaba and Ogawa 2X Monovalent antisera - Serovar Inaba and serovar Ogawa

* Denotes an eight week lead time.

4. Procedures for O Antigen Typing *Vibrio Cholerae* Isolates and Interpretation of Results

Cultures of organisms identified as *Vibrio cholerae* by their morphological and biochemical features may be serotyped by the following procedures.

1. Place two drops of sterile 0.85% saline solution (saline) onto a carefully cleaned microscope slide. The slide may be partitioned into several parts using a chimograph or glass pencil. With an inoculation loop or wire emulsify into each drop of saline a live cell colony from a fresh agar plate or slope culture to produce a distinct and uniform turbidity.
2. Place a drop of polyvalent antiserum onto one of the drops of emulsified isolate and to the other a drop of saline as a control.

Note: allow the antiserum to freefall from the dropper provided with the bottle. Do not contaminate the antiserum with organism.

3. Mix the reagents by tilting the slide back and forth for 60 seconds while viewing under indirect light against a dark background.
4. Distinct clumping or agglutination within this period, without clumping in the saline control (auto-agglutination) should be regarded as a positive result.
5. Specimens that show agglutination only with Inaba-type serum should be reported as *Vibrio cholerae* O1 serovar Inaba and specimens that show agglutination only with Ogawa-type serum should be reported as *Vibrio cholerae* O1 serovar Ogawa. Specimens that show agglutination both types of serum should be reported as *Vibrio cholerae* O1 serovar Hikojima. Specimens that show agglutination only with O139 Bengal serum should be reported as *Vibrio cholerae* O139 Bengal.

Note:- it should be remembered that El Tor biotype of *Vibrio cholerae* O1 cannot be distinguished from the classical biotype by serological means.



V 1.0 SJW 10/16

United Kingdom
 Mast Group Ltd.
 Mast House
 Derby Road, Bootle
 Merseyside L20 1EA

Tel: + 44 (0) 151 933 7277
 Fax: + 44 (0) 151 944 1332
 e-mail: sales@mastgrp.com

Germany
 Mast Diagnostica GmbH
 Feldstrasse 20
 DE-23858 Reinfeld

Tel: + 49 (0) 4533 2007 0
 Fax: + 49 (0) 4533 2007 68
 e-mail: mast@mast-diagnostica.de

France
 Mast Diagnostic
 12 Rue Jean-Jacques Mention
 CS 91106
 80011 Amiens CEDEX 1

Tél. + 33 (0) 322 80 80 67
 Fax + 33 (0) 322 80 99 22
 e-mail: info@mast-diagnostic.fr

www.mastgrp.com



MASTDISCS[®]

Leading the field with a complete solution
for AST and Identification disc testing



- Comprehensive range
- Premium quality products
- Compatible with EUCAST and CLSI standards
- Bespoke service available

Mast Group Ltd., has been a manufacturer of antibiotic susceptibility test products since 1957, and continues to be at the forefront of developments in this field. With an ever-expanding portfolio of antibiotic susceptibility, combination and identification discs, Mast Group Ltd. supplies the most comprehensive range available worldwide. Mast Group Ltd. also offers a range of services including **mastpharma**® development and **mastpharma**® stability for the evaluation of novel antimicrobial compounds for the pharmaceutical industry.

Disc diffusion is still the most popular method for determining bacterial antibiotic susceptibility in the treatment of infectious disease, as highlighted by the variety of international standardised procedures including The European Committee on Antimicrobial Susceptibility Testing (EUCAST) and The Clinical and Laboratory Standards Institute (CLSI). From January 2016 The British Society for Antimicrobial Chemotherapy (BSAC) guidelines were superseded by EUCAST. Mast Group Ltd. continues to support laboratories with the supply of BSAC concentrations until January 2020.

MASTDISCS® are available as an extensive range of test discs presented in dispensing cartridges or vials, with quality assurance guaranteed in accordance to standardised test protocols as issued by the above national and international organisations.

Disc Diffusion Susceptibility Testing

Order No	Product	Packsize
MASTDISCS[®] AST - ANTIBIOTIC SUSCEPTIBILITY DISCS IN CARTRIDGES AND VIALS		
MASTDISCS[®] in Cartridges (5 × 50 discs per pack)		
STOCKCART	Stock Susceptibility Cartridge Discs of a single type	 1 pack
FUNGCART	Stock Antifungal Cartridge Discs of a single type	1 pack
SPECIALCART	Cartridge Discs made to special order*	Min 18 packs
VOLCART	Cartridge Discs made to special order*	 Min 98 packs
MASTDISCS[®] (100 discs per vial)		
STOCKDISC	Stock Susceptibility Discs of a single type	1 vial
FUNGDISC	Stock Antifungal Discs of a single type	1 pack
SPECIALDISC	Discs made to special order*	Min 22 packs
VOLDISC	Discs made to special order*	Min 97 packs
TOOL/C	*Set up charge for Special Discs in vials or cartridges A one off charge for customisation of each new specification	per new specification

Disc Diffusion Susceptibility Testing

MASTDISCS[®]AST

ANTIBIOTIC SUSCEPTIBILITY DISCS IN CARTRIDGES - STOCK RANGE

Description Antibiotic & Content µg per disc (unless otherwise stated)	STANDARD			Order Code Cartridges 5 × 50 discs
	BSAC	EUCAST	CLSI	
Amikacin 30	✓	✓	✓	AK30C
Amoxicillin 2	✓	-	-	A2C
Amoxicillin 10	✓	-	-	A10C
Amoxicillin 25	✓	-	-	A25C
Amoxicillin/clavulanic acid 2-1	✓	✓	-	AUG3C
Amoxicillin/clavulanic acid 20-10	✓	✓	✓	AUG30C
Ampicillin 2	✓	✓	-	AP2C
Ampicillin 10	✓	✓	✓	AP10C
Ampicillin 25	✓	-	-	AP25C
Ampicillin/Sulbactam 10-10	-	-	✓	SAM20C
Azithromycin 15	✓	✓	✓	ATH15C
Aztreonam 30	✓	✓	✓	ATM30C
Bacitracin 10 units	-	-	-	BA10C
Carbenicillin 100 ♦	✓	-	✓	PY100C
Cefachlor 30	✓	✓	✓	CFC30C
Cefadroxil 30	✓	✓	-	CDX30C
Cefalexin 30	✓	-	-	CFX30C
Cefalothin 30	✓	-	✓	KF30C
Cefamandole 30	✓	-	✓	CMD30C
Cefazolin 30	-	-	✓	CZ30C
Cefepime 30	✓	✓	✓	CPM30C
Cefixime 5	✓	✓	✓	CFM5C
Cefoperazone 30	✓	-	-	CPZ30C
Cefotaxime 5	✓	✓	-	CTX5C
Cefotaxime 30	✓	-	✓	CTX30C
Cefoxitin 10	✓	-	-	FOX10C
Cefoxitin 30	✓	✓	✓	FOX30C
Cefpodoxime 10	✓	✓	✓	CPD10C
Cefradine 30	✓	-	-	CRD30C
Ceftaroline 5	-	✓	-	CPT5C
Ceftaroline 30	-	-	✓	CPT30C
Ceftazidime 10	-	✓	-	CAZ10C
Ceftazidime 30	✓	✓	✓	CAZ30C
Ceftazidime/avibactam 10-4	-	✓	-	CZA14C
Ceftazidime/avibactam 30-20	-	-	✓	CZA50C
Ceftibuten 30	-	✓	✓	CFB30C
Ceftolozane/tazobactam 30-10	-	✓	✓	CT40C
Ceftriaxone 5	✓	-	✓	CRO5C
Ceftriaxone 30	✓	✓	✓	CRO30C
Cefuroxime 5	✓	-	-	CXM5C
Cefuroxime 30	✓	✓	✓	CXM30C
Chloramphenicol 10	✓	-	-	C10C
Chloramphenicol 30	✓	✓	✓	C30C
Ciprofloxacin 1	✓	-	-	CIP1C
Ciprofloxacin 5	✓	✓	✓	CIP5C
Clarithromycin 2	✓	-	-	CLA2C
Clarithromycin 5	✓	-	-	CLA5C
Clarithromycin 15	-	✓	✓	CLA15C
Clindamycin 2	✓	✓	✓	CD2C
Delafloxacin 5 NEW	-	-	✓	DLX5C
Doripenem 10	-	✓	✓	DOR10C
Doxycycline 30	✓	-	✓	DXT30C
Eravacycline 20 NEW	-	✓	✓	ERV20C
Ertapenem 10	✓	✓	✓	ETP10C
Erythromycin 5	✓	-	-	E5C
Erythromycin 15	-	✓	✓	E15C
Florfenicol 30	-	-	-	FFC30

♦ Selected individually priced discs

Disc Diffusion Susceptibility Testing

MASTDISCS[®]AST

ANTIBIOTIC SUSCEPTIBILITY DISCS IN CARTRIDGES - STOCK RANGE

Description Antibiotic & Content µg per disc (unless otherwise stated)	STANDARD			Order Code Cartridges 5 × 50 discs
	BSAC	EUCAST	CLSI	
Fosfomycin 200/Glucose-6-Phosphate 50	✓	-	✓	FOT200C
Fusidic Acid 10	✓	✓	-	FC10C
Gentamicin 10	✓	✓	✓	GM10C
Gentamicin 30	-	✓	-	GM30C
Gentamicin 120	-	-	✓	GM120C
Gentamicin 200	✓	-	-	GM200C
Imipenem 10	✓	✓	✓	IMI10C
Imipenem/Relebactam 10-25 Coming Soon	-	-	✓	IMR35C
Kanamycin 30	-	-	✓	K30C
Levofloxacin 1	✓	-	-	LEV1C
Levofloxacin 5	✓	✓	✓	LEV5C
Linezolid 10	✓	✓	-	LZD10C
Linezolid 30	-	-	✓	LZD30C
Mecillinam 10	✓	✓	✓	MEC10C
Meropenem 10	✓	✓	✓	MEM10C
Meropenem/Vaborbactam 20-10 NEW	-	✓	✓	MEV30C
Metronidazole 5	✓	-	-	MZ5C
Minocycline 30	✓	-	✓	MN30C
Moxifloxacin 1	✓	-	-	MXF1C
Moxifloxacin 5	✓	✓	✓	MXF5C
Mupirocin 5	✓	-	-	MUP5C
Mupirocin 20	✓	-	-	MUP20C
Mupirocin 200	-	✓	-	MUP200C
Nalidixic Acid 30	✓	✓	✓	NA30C
Neomycin 10	✓	-	-	NE10C
Neomycin 30	-	-	-	NE30C
Netilmicin 10	-	✓	-	NET10C
Netilmicin 30	✓	✓	-	NET30C
Nitrofurantoin 100	-	✓	-	NI100C
Nitrofurantoin 200	✓	-	-	NI200C
Nitrofurantoin 300	-	-	✓	NI300C
Nitroxoline 30	-	-	✓	NIB30C
Norfloxacin 2	✓	-	-	NOR2C
Norfloxacin 10	-	✓	✓	NOR10C
Novobiocin 5	-	-	-	NO5C
Ofloxacin 5	✓	✓	✓	OFX5C
Oxacillin 1	✓	✓	✓	OX1C
Pefloxacin 5	-	✓	-	PEF5C
Penicillin G 1 unit	✓	✓	-	PG1C
Penicillin G 10 units	-	-	✓	PG10C
Piperacillin 30	-	-	-	PRL30C
Piperacillin 75	✓	-	-	PRL75C
Piperacillin 100	-	-	✓	PRL100C
Piperacillin/tazobactam 30-6	-	✓	-	PTZ36C
Piperacillin/tazobactam 75-10	✓	-	-	PTZ85C
Piperacillin/tazobactam 100-10	-	-	✓	PTZ110C
Polymyxin B 300 units	-	-	-	PB300C
Quinupristin/dalfopristin 15	✓	-	✓	SYN15C
Rifampicin 2	✓	-	-	RP2C
Rifampicin 5	✓	✓	✓	RP5C
Spectinomycin 25	✓	-	-	SPC25C
Streptomycin 10	✓	-	✓	S10C
Streptomycin 300	-	✓	✓	S300C
Teicoplanin 30	✓	-	✓	TEC30C
Temocillin 30 *	-	-	-	TEM30C
Tetracycline 10	✓	-	-	T10C
Tetracycline 30	-	✓	✓	T30C

* Temocillin 30µl is included in CA-SFM standard

Disc Diffusion Susceptibility Testing

MASTDISCS[®]AST

ANTIBIOTIC SUSCEPTIBILITY DISCS IN CARTRIDGES - STOCK RANGE

Description Antibiotic & Content µg per disc (unless otherwise stated)	STANDARD			Order Code Cartridges 5 × 50 discs
	BSAC	EUCAST	CLSI	
Ticarcillin 75	✓	-	✓	TC75C
Tigecycline 15	✓	✓	✓	TGC15C
Ticarcillin/clavulanic acid 75-10	✓	-	✓	TIM85C
Tobramycin 10	✓	✓	✓	TN10C
Trimethoprim 2.5	✓	-	-	TM2.5C
Trimethoprim 5	✓	✓	✓	TM5C
Trimethoprim/sulfamethoxazole 1.25-23.75	✓	✓	✓	TS25C
Vancomycin 30	-	-	✓	VA30C
Blank discs	-	-	-	BD0680W/C

MASTDISCS[®]AST

NON CLINICAL ANTIBIOTIC SUSCEPTIBILITY DISCS IN CARTRIDGES - STOCK RANGE

Description Antibiotic & Content µg per disc (unless otherwise stated)	Order Code Cartridges 5 × 50 discs
Cefperazone/sulbactam 75-30	SPZ105C/NCE
Colistin Sulphate 10	CO10C/NCE
Colistin Sulphate 25	CO25C/NCE
Enrofloxacin 5	ENF5C/NCE

MASTDISCS[®]AST

SPECIALIST VETERINARY SUSCEPTIBILITY DISCS - STOCK RANGE

Description Antibiotic & Content µg per disc (unless otherwise stated)	Order Code	
	Cartridges 5 × 50 discs	Vials 100 discs
Cefquinone 30	CEQ30C/NCE	-
Gamithromycin 15	GAM15C/NCE	-
Pradofloxacin 5	PRA5/NCE	-
Marbofloxacin 5	MAR5/NCE	-
Tildipirison 60	TIP60/NCE	-
Tylosin 30	TY30C/NCE	TY30/NCE

Disc Diffusion Susceptibility Testing

Order No

Product

Mechanism

Packsize (tests)

MASTDISCS[®]AST

ANTIBIOTIC SUSCEPTIBILITY DISCS IN VIALS - STOCK RANGE

Description Antibiotic & Content µg per disc (unless otherwise stated)	STANDARD			Order Code Cartridges 5 × 50 discs
	BSAC	EUCAST	CLSI	
Amikacin 30	✓	✓	✓	AK30
Bacitracin 8 units	-	-	-	BA8
Bacitracin 10 units	-	-	-	BA10
Cefpodoxime 10	✓	✓	✓	CPD10
Chloramphenicol 10	✓	-	-	C10
Chloramphenicol 30	✓	✓	✓	C30
Gentamicin 10	✓	✓	✓	GM10
Meropenem 10	✓	✓	✓	MEM10
Metronidazole 2.5	-	-	-	MZ2.5
Metronidazole 5	✓	-	-	MZ5
Nalidixic Acid 30	✓	✓	✓	NA30
Novobiocin 5	-	-	-	NO5
Oxacillin 1	✓	✓	✓	OX1
Penicillin G 1 unit	✓	✓	-	PG1
Penicillin G 2 units	-	-	-	PG2
Rifampicin 5	✓	✓	✓	RP5
Tylosin 30	-	-	-	TY30
Vancomycin 5	✓	✓	-	VA5
Blank discs	-	-	-	BD0638W

MASTDISCS[®]AST

ANTIFUNGAL SUSCEPTIBILITY DISCS IN CARTRIDGES AND VIALS - STOCK RANGE

Description Antibiotic & Content µg per disc (unless otherwise stated)	Order Code	
	Cartridges 5 × 50 discs	Vials 100 discs
Amphotericin B 20	AMB20C	-
Clotrimazole 10	CTM10C	-
Econazole 10	ECN10C	-
Fluconazole 10	FCN10C	-
Fluconazole 25	FCN25C	FCN25
Flucytosine 1	FY1C	-
Ketoconazole 10	KCA10C	-
Miconazole 10	MCL10C	-
Nystatin 100	NY100C	-

Disc Diffusion Susceptibility Testing

Order No	Product	Mechanism	Packsizes (tests)
----------	---------	-----------	-------------------

MASTDISCS[®] Combi

COMBINATION DISC SETS FOR THE DETECTION OF ANTIBIOTIC RESISTANCE

D52C	Extended Spectrum β Lactamase Set	ES β L	50 tests
D62C	Cefotaxime 30 & Cefotaxime 30/Clavulanic Acid 10	ES β L	150 tests
D63C	Cefepime 30 & Cefepime 30/Clavulanic Acid 10	ES β L	150 tests
D64C	Ceftazidime 30 & Ceftazidime 30/Clavulanic Acid 10	ES β L	150 tests
D66C	Cefpodoxime 10 & Cefpodoxime 10/Clavulanic Acid 1	ES β L	150 tests
D67C	Extended Spectrum β Lactamase Set (CPD10)	ES β L	50 tests
D68C	AmpC & ESBL Detection Set	AmpC/ES?L	50 tests
D69C	AmpC Detection Set	AmpC	50 tests
D72C	AmpC, ESBL & Carbapenemase Detection Set	AmpC/ESBL/Carbapenemase	50 tests
D73C	MAST [®] Carba plus	MBL/KPC/OXA	50 tests

Additional carbapenemase screening and identification tests

PACE-ID 	Colorimetric test for the rapid detection of carbapenemase producing <i>Pseudomonas</i> spp., <i>Acinetobacter</i> spp. and Enterobacterales.	48 tests
D71C	MAST [®] CAT-ID - For presumptive identification of carbapenemase production	250 tests
D74	MAST [®] ICT - screening test for the detection of carbapenemase production in Enterobacterales, <i>Pseudomonas</i> and <i>Acinetobacter</i> spp.	25 tests
TEM30C	To aid presumptive identification of OXA-48 when used with D70C	5 x 50 discs

MASTDISCS[®] AST

PENICILLIN DISCS

For detecting the emergence of penicillin resistance. (vials of 100 discs or 5 x 50 discs in cartridges)

PG1	Penicillin G	1 unit (vials)	1 vial
PG1C	Penicillin G	1 unit (cartridges)	1 pack

HIGH CONTENT AMINOGLYCOSIDE DISCS

The following discs are available to special order and subject to 6 weeks lead time.

For detecting high level aminoglycoside resistance. (vials of 100 discs or 5 x 50 discs in cartridges)

GM500C/NCE	500 μ g	Min order
K1000C	1000 μ g	18 packs
K1000	1000 μ g	Min order
		22 vials

Disc Diffusion Susceptibility Testing

Order No

Product

Packsize

DISCMASTER DISPENSER

Robust and reliable antimicrobial cartridge disc dispensers designed for use with MASTDISCS® Antimicrobial Susceptibility Test Cartridges.

MDD65	MAST® DISCMASTER Dispenser - 6 place	1
SILICA63	Silica Gel Capsule for MAST® DISCMASTER	4
SHD5	Single Cartridge Hand Dispenser	5
CANISTER	Canister for MDD64 models and below	1
CANISTER63	Canister for MDD65 MAST® DISCMASTER Dispenser	1



ZT147 v6.0 CEA 07/19

United Kingdom
Mast Group Ltd.
Mast House
Derby Road, Bootle
Merseyside L20 1EA

Tel: + 44 (0) 151 933 7277
Fax: + 44 (0) 151 944 1332
e-mail: sales@mast-group.com

Germany
Mast Diagnostica GmbH
Feldstrasse 20
DE-23858 Reinfeld

Tel: + 49 (0) 4533 2007 0
Fax: + 49 (0) 4533 2007 68
e-mail: mast@mast-diagnostica.de

France
Mast Diagnostic
12 Rue Jean-Jacques Mention
CS 91106
80011 Amiens CEDEX 1

Tél. + 33 (0) 322 80 80 67
Fax + 33 (0) 322 80 99 22
e-mail: info@mast-diagnostic.fr