

CERTIFICATE

of Conformity

Electrical Equipment (Safety) Regulations 2016

(UK SI 2016 No. 1101)

Registration Nr.: AQ 50539846 0001
Report Nr.: CN21HLH0 001

Holder: DLAB SCIENTIFIC CO.,LTD.
Yu An Road 31,
Airport Economic core zone,
Shun yi District,
Beijing, 101318
P.R. China

Product: Laboratory Equipment (*PCR Thermocycler*)

Test standard: BS EN 61010-1:2010+A1
BS EN IEC 61010-2-010:2020

Identification: TC1000-S, SCI1000-S, 50X4101300
TC1000-G, SCI1000-G, 50X4102300
(X=1-9, X stands for different client)

This certificate of conformity is based on an evaluation of a sample of the above mentioned product, technical report and documentation. This certificate does not imply the assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

Certification Body



Date: 2022-04-24

Xuhua Liu



The UKCA mark may be used, if all relevant UK Legislations and the UK Declaration of Conformity has been completed.



TÜV Rheinland UK Ltd.

Friars Gate (Third Floor), 1011 Stratford Road, Shirley, Solihull, B90 4BN