



# EC Declaration of Conformity

We, TaiDoc Technology Corporation

B1-7F, No.127, Wugong 2nd Road, 24888 Wugu Dist, New Taipei City, TAIWAN

declare under our sole responsibility that the product

**Product Name** : Blood Glucose Monitoring System

**Product Model** : TD-4116

**Classification** : 98/79/EC (IVDD), Annex II, List B

**Conformity Assessment Route** : 98/79/EC (IVD), Annex IV excluding section 4&6

**EC Certificate Number** : V1 052126 0069 Rev.02

**European Representative** : MedNet EC-REP GmbH  
Borkstraße 10, 48163 Münster , Germany

**Notified Body (CE0123)** : TÜV SÜD Product Service GmbH  
Ridlerstraße 65, 80339 München, Germany

**GMDN code** : 62537

to which this declaration relates is in conformity with the following standard(s) or other normative document(s) :

ISO 13485:2016	Medical devices - Quality management systems - Requirements for regulatory purposes
EN ISO 14971:2012	Medical devices - Application of risk management to medical devices
EN ISO 15197:2015	In vitro diagnostic test systems —Requirements for blood-glucose monitoring systems for self-testing in managing diabetes mellitus
EN 15223-1:2016	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied. Part I: General requirements
EN ISO 18113-1:2011	In vitro diagnostic medical devices. Information supplied by the manufacturer (labelling). In vitro diagnostic instruments for professional use
EN ISO 18113-2:2011	In vitro diagnostic medical Devices. Information supplied by the manufacturer (labelling). In vitro diagnostic reagents for professional use
EN ISO 18113-3:2011	In vitro diagnostic medical devices. Information supplied by the manufacturer (labelling). In vitro diagnostic instruments for professional use
EN ISO 18113-4:2011	In vitro diagnostic medical Devices. Information supplied by the manufacturer (labelling). In vitro diagnostic reagents for self-testing



www.taidoc.com

泰博科技股份有限公司  
TaiDoc Technology Corp.

新北市24888五股區五工二路127號B1 7樓  
B1-7F, No.127, Wugong 2nd Rd., Wugu Dist.,  
New Taipei City 24888, Taiwan

Tel : +886-2-6625-8188  
Fax : +886-2-6625-0288

EN ISO 18113-5:2011	In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 5: In vitro diagnostic instruments for self-testing
EN ISO 23640:2015	In vitro diagnostic medical devices. Evaluation of stability of in vitro diagnostic reagents
EN ISO 20417:2021	Medical device – Information to be supplied by the manufacturer.
EN 13532:2002	General requirements for in vitro diagnostic medical devices for self-testing
EN 13612:2002	Performance evaluation of in vitro diagnostic medical devices
EN 61010-1:2010	Safety requirements for electrical equipment for measurement, control and laboratory use. General requirements.
EN 61010-2-101:2015	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment
EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements
EN 61326-2-6 :2013	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-6: Particular requirements - In vitro diagnostic (IVD) medical equipment
EN 62304:2006+A1:2015	Medical device software -- Software life cycle processes
ISO/IEC 12207:2008	Systems and software engineering- Software life cycle processes
EN 62366-1:2015	Medical devices -- Application of usability engineering to medical

2022. 1. 3.

Date of Issue

**Jim Jan**  
Management Representative