

MEDİ C TEKS	Plaster of Paris Bandage Analysis Report	DATE
		20/05/2021

PRODUCT NAME: Plaster of Paris Bandage

SIZE : 15CM X 2,70M

LOT NUMBER :130521106

TESTS PERFORMED	TS 4434	CONCLUSION	CONDITIONS OF CONFORMITY
OF THE WRAP SHEET MASS PER UNIT AREA(g/m ²)	MIN 26 g	30,62 gr	A
THE WRAP AND WEFT WIRE WINDING NUMBER OF CLOTH	Cm2 WIRE MUST BE AT LEAST 18 PER	18 wire	A
FEATURE OF THE WRAP SHEET	100% COTTON AND SHOULD BE HYDROPHILE	100% COTTON AND HYDROPHIL	A
PLASTERY MASS PER UNIT AREA OF THE WRAP SHEET	AT LEAST 340 g	437 gr	A
WETTABILITY	WHEN THROWN IN TO THE WATER SHOULD NOT DRY SURFACE	FULL WETTING OCCURRED AFTER 6 SECONDS 30°C WATER	A
SETTING TIME	1-8 MINUTES	2,40 MINUTES	A

CONDITIONS OF CONFORMITY : A:Available NA: None/Available

CONCLUSION: Test results conform to the relevant standard

QUALITY CONTROL OFFICER

GÖZDE DURGİT

MEDİTEKS SAĞLIK HİZMETLERİ
TIBBİ MALZEME TEKSTİL ÜRÜNLERİ
SAN.DİŞ TİC.A.Ş EDİRNE SÜLOĞLU ŞUBESİ
Organize San. Bölgesi No:5 Süloğlu-Edirne
Davutpaşa V. 0284 316 012 4314
Mersis No: 083 012 4314 00016

FR-075/ 00//Release Date: 09.05.2018 REF.: PR-009

Edirne Organize Sanayi Bölgesi 4.Cadde No:5 Süloğlu / Edirne

Tel : 0284 316 23 81-82 Fax : 0284 316 23 84 www.mediteks.com