

EC DECLARATION OF CONFORMITY

Lorne Laboratories Ltd declares that the following in vitro diagnostic reagent:

Product name	Catalogue number
ASO Latex kit	031100A

has been classified as non List A, non List B (Directive 98/79/EC, Annex II) and complies with the essential requirements and provisions of Directive 98/79/EC of the European Parliament and of the Council (also SI 2002 No.618 which transposes the requirements of Directive 98/79/EC).

and is in conformity with the national standards transposing harmonised standards:

- BS EN 980:2008
- BS EN ISO 13485:2012
- BS EN 13612:2002
- BS EN 13640:2002
- BS EN 13641:2002
- BS EN ISO 14971:2012
- BS EN ISO 18113, parts 1&2

The conformity assessment procedure performed was in accordance with Annex III of Directive 98/79/EC.

This declaration of conformity is issued under the sole responsibility of Lorne Laboratories Ltd and is valid from 13 April 2016.



Eddy Velthuis
Technical Director

EC DECLARATION OF CONFORMITY

Lorne Laboratories Ltd declares that the following in vitro diagnostic reagent:

Product name	Catalogue number
RF Latex kit	830100A

has been classified as non List A, non List B (Directive 98/79/EC, Annex II) and complies with the essential requirements and provisions of Directive 98/79/EC of the European Parliament and of the Council (also SI 2002 No.618 which transposes the requirements of Directive 98/79/EC).

and is in conformity with the national standards transposing harmonised standards:

- BS EN 980:2008
- BS EN ISO 13485:2012
- BS EN 13612:2002
- BS EN 13640:2002
- BS EN 13641:2002
- BS EN ISO 14971:2012
- BS EN ISO 18113, parts 1&2

The conformity assessment procedure performed was in accordance with Annex III of Directive 98/79/EC.

This declaration of conformity is issued under the sole responsibility of Lorne Laboratories Ltd and is valid from 13 April 2016.



Eddy Velthuis
Technical Director

EC DECLARATION OF CONFORMITY

Lorne Laboratories Ltd declares that the following in vitro diagnostic reagent:

Product name	Catalogue number
CRP Latex kit	850100A

has been classified as non List A, non List B (Directive 98/79/EC, Annex II) and complies with the essential requirements and provisions of Directive 98/79/EC of the European Parliament and of the Council (also SI 2002 No.618 which transposes the requirements of Directive 98/79/EC).

and is in conformity with the national standards transposing harmonised standards:

- BS EN 980:2008
- BS EN ISO 13485:2012
- BS EN 13612:2002
- BS EN 13640:2002
- BS EN 13641:2002
- BS EN ISO 14971:2012
- BS EN ISO 18113, parts 1&2

The conformity assessment procedure performed was in accordance with Annex III of Directive 98/79/EC.

This declaration of conformity is issued under the sole responsibility of Lorne Laboratories Ltd and is valid from 13 April 2016.



Eddy Velthuis
Technical Director