

Anexa 92,

Bio RS-24 Mini-rotator cu platforma PRS-22

Specificarea tehnică deplină solicitată, Standarde de referință	Specificarea tehnică deplină oferită, Standarde de referință
Se acceptă la limita 16 ml $\pm$ 1 ml Unghi de rotație - 1 până la 90° Interval de viteză - 1 până la 100 rpm Voltaj - 120 V Interval de temperatură (metric) - 4°C până la 40°C Amperaj - 0,5 A Tip de produs - Mixer de mostre Certificari/Conformitate - Certificat de analiză (COA) Interval de temperatură - 39,2 °F până la 104 °F	Platforma acceptă la limita 16 ml $\pm$ 1 ml, diapason 1,5 - 15 ml 22 tuburi, 10-16 mm diameter (1.5 ml -15 ml tubes). Unghi de rotație - 1 până la 360° Interval de viteză - 5 până la 30 rpm Voltaj - 100–240V; 50/60 Hz; Output DC 12V Interval de temperatură (metric) + 4°C până la 40°C Amperaj - 0,011 A, 12V, /1.3 W Tip de produs - Mixer de mostre Certificari/Conformitate - Certificat de analiză (COA) Interval de temperatură - 39,2 °F până la 104 °F

Bio RS-24, Mini-rotator

DESCRIPTION

Mini-rotator **Bio RS-24** provides vertical rotation of the platform. The rotator is an ideal instrument for preventing blood coagulation in tubes and for fulfilment of procedures of biological components extraction.

The device is simple to operate; it is designed as a low cost solution.

Rotator can be used in cold rooms or incubators, operating at ambient temperature range +4°C to +40°C.

PRS - series platforms are equipped with universal rubber clamps for different size tube fixation;

PRSC - series platforms have metal clamps able to hold heavier solutions (e.g. soil, sand)



SPECIFICATIONS

Speed control range	5-30 RPM
Digital time setting	1 min–24 hrs / non–stop (increment 1 min)
Vertical rotation movement	overhead, 360°
Maximum continuous operation time	8 h or more - with longer pauses
Maximum load	0.375 kg
Overall dimensions (W×D×H)	325×190×155 mm
Weight	1.4 kg
Input current/power consumption	12 V, 110 mA/1.3 W
External power supply	Input AC 100–240 V; 50/60 Hz; Output DC 12 V

CAT. NUMBER

Including PRS-22 platform	Including PRS-22 platform
BS-010133-AAG	230VAC 50/60Hz Euro plug
BS-010133-AAK	100-240VAC 50/60Hz Multi plug (EU, UK, AU, US)
BS-010133-BK	IQ OQ document
BS-010133-CK	PQ document

ACCESSORIES



PRS-22
BS-010117-FK
platform

Can accommodate 22 tubes, 10-16 mm diameter (1.5 ml - 15 ml tubes).



PRSC-18
BS-010117-EK
platform

Can accommodate 18 tubes, 16 mm diameter (15 ml).



PRS-4/12
BS-010117-AK
platform

Can accommodate 4 tubes, 20-30 mm diameter (up to 50ml tubes) and 12 tubes, 10-16 mm diameter (1.5 ml - 15 ml tubes).

EU Declaration of Conformity

Unit type Rockers, shakers, rotators, vortexes, spinners

Models MR-1, MR-12;
3D, Multi Bio 3D, PSU-10i, PSU-20i, MPS-1, PSU-2T;
Bio RS-24, Multi Bio RS-24, Multi RS-60;
V-1 plus, V-32, MSV-3500;
UIS-360

Serial number 14 digits styled XXXXXXYYMMZZZZ, where XXXXXX is model code, YY and MM – year and month of production, ZZZZ – unit number.


Manufacturer SIA BIOSAN
Latvia, LV-1067, Riga, Ratsupites 7 k-2

The objects of the declaration described above is in conformity with the following relevant Union harmonization legislations:

LVD 2014/35/EU	LVS EN 61010-1:2011 + A1:2019 Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements. LVS EN 61010-2-051:2021 + A11:2021 Particular requirements for laboratory equipment for mixing and stirring.
EMC 2014/30/EU	LVS EN 61326-1:2021 Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements.
Directive 2011/65/EU RoHS3 2015/863/EU	Directive on the restriction of the use of certain hazardous substances in electrical and electronic equipment.
WEEE 2012/19/EU	Directive on waste electrical and electronic equipment.

I declare that the Declaration of Conformity is issued under sole responsibility of the manufacturer and belongs to the above-mentioned objects of the declaration.

Deniss Beleckis
Manufacturing director


Signature

31.03.2025
Date

EC DECLARATION OF CONFORMITY

We, NÜVE Sanayi Malzemeleri İmalat ve Ticaret A.Ş.
Saracalar Mah. Saracalar Kümeevleri No: 4/2
Akyurt 06750 Ankara / TURKEY

Herewith declare that the following described product meets the provisions of the following directives. All supporting documentation is retained under the premises of the manufacturer.

PRODUCT:

WATER DISTILLER

MODEL:

ND 4 – ND 8 – ND 12

APPLICABLE EC DIRECTIVES

LVD : 93/68/EEC & 2006/95/EC

EMC :92/31/EEC & 2004/108/EC

**APPLICABLE HARMONISED
STANDARDS:**

**EN 61010-1
EN 61000-6-3
EN 61326-1**

DATE OF ISSUE:

March 01, 2014

SIGNATURE:



**Beril İZGİN
General Manager**

CE