

# Certificate of Analysis

**ThermoFisher**  
SCIENTIFIC


## Cathode Buffer Container 3500 Series


Product No. 4408256  
Lot No. 2203807  
Date of Manufacture 31MAR2022  
Expiration Date 04DEC2022

TEST	SPECIFICATION	RESULT
<b>VISUAL INSPECTION</b>		
Clear, particulate free, viscous solution		Pass
<b>PH TEST</b>		
Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	7.874
<b>CONDUCTIVITY</b>		
Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	1794 µS/cm

**For Research Use Only. Not for use in diagnostic procedures.**

**Manufactured in compliance with our ISO 13485 certified quality management system. Site:  
Bedford, MA, USA**

 Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

  
Luciana Torcasio (Apr 1, 2022 11:36 EDT)

**Quality Assurance  
Issued 01APR2022**