

Certificate of Registration

QUALITY MANAGEMENT SYSTEM - ISO 13485:2016

This is to certify that:

BIOTRONIK AG
Ackerstrasse 6
8180 Bülach
Switzerland

Facility ID Number: F000099

Holds Certificate No:

MDSAP 688646

Statement of Conformity: The company listed on this certificate has been audited to and found to conform with the following criteria: ISO 13485:2016 and Australia - Therapeutic Goods (Medical Devices) Regulations, 2002, Schedule 3 Part 1 (excluding Part 1.6) - Full Quality Assurance Procedure; Brasil - RDC ANVISA n. 16/2013, RDC ANVISA n. 23/2012, RDC ANVISA n. 67/2009; Canada - Medical Devices Regulations - Part 1 - SOR 98/282; Japan - MHLW Ministerial Ordinance 169, Article 4 to Article 68, PMD Act; USA - 21 CFR 820, 21 CFR 803, 21 CFR 806, 21 CFR 807 - Subparts A to D

Design, development, manufacture, and distribution of the following sterile devices: PTCA balloon catheters, PTA balloon catheters, drugreleasing PTCA balloon catheters, drug-releasing PTA balloon catheters, coronary stents and stent systems, peripheral stents and stent systems, drugeluting coronary stents and stent systems, coronary guidewires, peripheral guidewires, drug-eluting resorbable coronary scaffolds and scaffold systems.



For and on behalf of BSI:

Gary E Slack, Senior Vice President - Medical Devices

Original Registration Date: 2018-10-11

Effective Date: 2021-10-11

Expiry Date: 2024-10-10



BSI Group America Inc. is an MDSAP authorized auditing organization

Page: 1 of 1

...making excellence a habit.™

EC Certificate - Full Quality Assurance System

Directive 93/42/EEC on Medical Devices, Annex II excluding Section 4

No.

CE 608280

Issued To:

**BIOTRONIK AG
Ackerstrasse 6
8180 Bülach
Switzerland**

In respect of:

Design and manufacture of PTCA balloon catheters, PTA balloon catheters, drug-releasing PTCA balloon catheters, drug-releasing PTA balloon catheters, coronary stent systems, peripheral vascular stent systems, drug-eluting coronary stent systems, drug-eluting resorbable coronary scaffold systems, coronary guidewires and peripheral guidewires

on the basis of our examination of the quality assurance system under the requirements of Council Directive 93/42/EEC, Annex II excluding section 4. The quality assurance system meets the requirements of the directive. For the placing on the market of class III products an Annex II section 4 certificate is required.

For and on behalf of BSI, a Notified Body for the above Directive (Notified Body Number 2797):



Gary E Slack, Senior Vice President Medical Devices

First Issued: **2014-04-01**

Date: **2019-10-30**

Expiry Date: **2024-05-26**

...making excellence a habit.™

Page 1 of 4

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body. This approval excludes all products designed and/or manufactured by a third party on behalf of the company named on this certificate, unless specifically agreed with BSI.

This certificate was issued electronically and is bound by the conditions of the contract.

EC Certificate - Full Quality Assurance System

Supplementary Information to CE 608280

Issued To:

BIOTRONIK AG
Ackerstrasse 6
8180 Bülach
Switzerland

Number	Device Name	Intended purpose per IFU
Class III		
---	Magmaris Sirolimus-Eluting Resorbable Coronary Magnesium Scaffold System	See CE 608221
	PRO-Kinetic Energy Coronary Stent System	See CE 608282
	Pantera LEO Fast-Exchange PTCA catheter	See CE 608283
	Orsiro Sirolimus-Eluting Coronary Stent System	See CE 608284
	Pantera Lux Paclitaxel releasing PTCA Balloon Catheter	See CE 608285
	PK Papyrus Covered Coronary Stent System	See CE 608286
	Synsiro Sirolimus-Eluting Coronary Stent System	See CE 608289
	Passeo-18 Lux Paclitaxel releasing PTA Balloon Catheter	See CE 610590
	Cruiser and Cruiser Hydro coronary and peripheral artery guidewires	See CE 619676
	Pantera Pro Coronary Dilatation Catheter	See CE 620197
	Orsiro Mission Sirolimus Eluting Coronary Stent System	See CE 704680
	Synsiro Pro Sirolimus Eluting Coronary Stent System	See CE 708283

First Issued: **2014-04-01**

Date: **2019-10-30**

Expiry Date: **2024-05-26**

...making excellence a habit.™

Page 2 of 4

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body. This approval excludes all products designed and/or manufactured by a third party on behalf of the company named on this certificate, unless specifically agreed with BSI.

This certificate was issued electronically and is bound by the conditions of the contract.

Information and Contact: BSI, Say Building, John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands Tel: + 31 20 346 0780

BSI Group The Netherlands B.V. registered in The Netherlands under 33264284.

A member of BSI Group of Companies.

EC Certificate - Full Quality Assurance System

Supplementary Information to CE 608280

Issued To:

BIOTRONIK AG
Ackerstrasse 6
8180 Bülach
Switzerland

Number	Device Name	Intended purpose per IFU
Class IIb		
47932	Self-expanding NiTi peripheral stents	For use in patients with atherosclerotic disease of the iliac arteries and for the treatment of insufficient results after percutaneous transluminal angioplasty (PTA), e.g. residual stenosis and dissection.
		For use in patients with atherosclerotic disease of the femoral and infrapopliteal arteries and for the treatment of insufficient results after percutaneous transluminal angioplasty (PTA), e.g. residual stenosis and dissection.
		For use in patients with atherosclerotic disease of the superficial femoral, proximal popliteal and infrapopliteal arteries and for the treatment of insufficient results after percutaneous transluminal angioplasty (PTA), e.g. residual stenosis and dissection.

First Issued: **2014-04-01**Date: **2019-10-30**Expiry Date: **2024-05-26**

...making excellence a habit.™

Page 3 of 4

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body. This approval excludes all products designed and/or manufactured by a third party on behalf of the company named on this certificate, unless specifically agreed with BSI.

This certificate was issued electronically and is bound by the conditions of the contract.

Information and Contact: BSI, Say Building, John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands Tel: + 31 20 346 0780

BSI Group The Netherlands B.V. registered in The Netherlands under 33264284.

A member of BSI Group of Companies.

EC Certificate - Full Quality Assurance System

Supplementary Information to CE 608280

Issued To:

BIOTRONIK AG
Ackerstrasse 6
8180 Bülach
Switzerland

Number	Device Name	Intended purpose per IFU
Class IIb		
47932	Balloon-expandable Cobalt Chromium peripheral stents	To improve sub-optimal angiographic results ($\geq 50\%$ residual stenosis) and/or flow-limiting dissections after PTA of atherosclerotic lesions in the infrapopliteal arteries.
44279	Iliac artery stents	For the treatment of de novo or restenotic atherosclerotic lesions in iliac arteries.
45852	Renal artery stents	For improving arterial luminal diameter in patients with clinical symptoms attributable to atherosclerotic stenosis of the renal arteries.
Class IIa		
MD 0106	PTA balloon catheters	---

First Issued: **2014-04-01**Date: **2019-10-30**Expiry Date: **2024-05-26**

...making excellence a habit.™

Page 4 of 4

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body. This approval excludes all products designed and/or manufactured by a third party on behalf of the company named on this certificate, unless specifically agreed with BSI.

This certificate was issued electronically and is bound by the conditions of the contract.

Information and Contact: BSI, Say Building, John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands Tel: + 31 20 346 0780

BSI Group The Netherlands B.V. registered in The Netherlands under 33264284.

A member of BSI Group of Companies.

EC-Declaration of Conformity



DOC No. 11-02-01

Issue: 18

Manufacturer: Biotronik AG
Ackerstrasse 6
8180 Bülach
Switzerland

Authorised Representative: BIOTRONIK SE & Co. KG
Woermannkehe 1
12359 Berlin
Germany

Product Category: Drug-eluting stents for vascular intervention

Product Name: Orsiro Sirolimus Eluting Coronary Stent System

Class: III, according to Council Directive 93/42/EEC, Annex IX, rule 8 and 13

Conformity Assessment Route: Council Directive 93/42/EEC, Annex II, Section 3 and 4

Scope: 54 different variants. See list on next pages

We hereby declare that the above-mentioned products meet the provisions of Council Directive 93/42/EEC for medical devices. All supporting documentation is retained under the premises of the manufacturer.

For these products the following EC-Design Examination Certificate has been issued:

Certificate Number:	CE 608284
Notified Body:	BSI Group The Netherlands B.V.
EEC No:	2797
Expiry date:	26.May.2024

To these products our approved Full Quality Assurance System according to Annex II of the Directive 93/42/EEC is applied. For this Quality Assurance System, the following certificate has been issued:

Certificate Number:	CE 608280
Notified Body:	BSI Group The Netherlands B.V.
EEC No:	2797
Expiry date:	26.MAY.2024

Date of first CE-marking: 25.FEB.2011

Place, Date of issue: Bülach, 21.MAY.2021

Signature:

Marcel Schäfer, Ph.D.
Senior Director Regulatory Affairs and Post Market Surveillance

A11 REG 151631 EN 07

Scope of DoC No. 11-02-01

Pos.	Designation	Catalogue number (REF)	Stent diameter [mm]	Stent length [mm]	Nominal Total Drug Load TDL [µg]
1	Orsiro 2.25/9	364469	2.25	9	55
2	Orsiro 2.5/9	364470	2.5	9	55
3	Orsiro 2.75/9	364471	2.75	9	55
4	Orsiro 3.0/9	364472	3.0	9	55
5	Orsiro 3.5/9	364473	3.5	9	70
6	Orsiro 4.0/9	364474	4.0	9	70
7	Orsiro 2.25/13	364475	2.25	13	80
8	Orsiro 2.5/13	364476	2.5	13	80
9	Orsiro 2.75/13	364477	2.75	13	80
10	Orsiro 3.0/13	364478	3.0	13	80
11	Orsiro 3.5/13	364479	3.5	13	95
12	Orsiro 4.0/13	364480	4.0	13	95
13	Orsiro 2.25/15	364481	2.25	15	93
14	Orsiro 2.5/15	364482	2.5	15	93
15	Orsiro 2.75/15	364483	2.75	15	93
16	Orsiro 3.0/15	364484	3.0	15	93
17	Orsiro 3.5/15	364485	3.5	15	113
18	Orsiro 4.0/15	364486	4.0	15	113
19	Orsiro 2.25/18	364487	2.25	18	109
20	Orsiro 2.5/18	364488	2.5	18	109
21	Orsiro 2.75/18	364489	2.75	18	109
22	Orsiro 3.0/18	364490	3.0	18	109
23	Orsiro 3.5/18	364491	3.5	18	131
24	Orsiro 4.0/18	364492	4.0	18	131
25	Orsiro 2.25/22	364499	2.25	22	134
26	Orsiro 2.5/22	364500	2.5	22	134
27	Orsiro 2.75/22	364501	2.75	22	134
28	Orsiro 3.0/22	364502	3.0	22	134
29	Orsiro 3.5/22	364503	3.5	22	162
30	Orsiro 4.0/22	364504	4.0	22	162
31	Orsiro 2.25/26	364505	2.25	26	159
32	Orsiro 2.5/26	364506	2.5	26	159
33	Orsiro 2.75/26	364507	2.75	26	159
34	Orsiro 3.0/26	364508	3.0	26	159
35	Orsiro 3.5/26	364509	3.5	26	193
36	Orsiro 4.0/26	364510	4.0	26	193
37	Orsiro 2.25/30	364511	2.25	30	184
38	Orsiro 2.5/30	364512	2.5	30	184
39	Orsiro 2.75/30	364513	2.75	30	184
40	Orsiro 3.0/30	364514	3.0	30	184
41	Orsiro 3.5/30	364515	3.5	30	224
42	Orsiro 4.0/30	364516	4.0	30	224
43	Orsiro 2.25/35	391234	2.25	35	213

44	Orsiro 2.5/35	391235	2.5	35	213
45	Orsiro 2.75/35	391236	2.75	35	213
46	Orsiro 3.0/35	391237	3.0	35	213
47	Orsiro 3.5/35	391018	3.5	35	261
48	Orsiro 4.0/35	391019	4.0	35	261
49	Orsiro 2.25/40	391238	2.25	40	247
50	Orsiro 2.5/40	391239	2.5	40	247
51	Orsiro 2.75/40	391240	2.75	40	247
52	Orsiro 3.0/40	391241	3.0	40	247
53	Orsiro 3.5/40	391020	3.5	40	298
54	Orsiro 4.0/40	391021	4.0	40	298

Change History

Check version index is up to date prior to use.

Version of SAP Document	Main changes from previous release to current release
01	Transfer from previous template to new template TMP 111387. New issue due to transfer of Notified Body to BSI Group The Netherlands B.V.
02	New issue due to sterilizer addition
03	Revised for the introduction of the electronic IFU in compliance with regulation 207/2012.
04	Declaration of Conformity updated with the new expiry date of the EC Full Quality Assurance System Certificate
05	Implementation of the recertification
06	New issue due to changes affecting Sirolimus at supplier Biocon Limited (heavy metals in specification; test methods for particle size and residual solvent; and re-test period).
07	Designation of Authorised (EU) Representative. Addition of name and address.

EC-Declaration of Conformity

DOC No. 13-06-02

Issue: 11

Manufacturer: Biotronik AG
Ackerstrasse 6
8180 Bülach
Switzerland

Authorised Representative: BIOTRONIK SE & Co. KG
Woermannkehre 1
12359 Berlin
Germany

Product Category: Coronary stent system

Product Name: PK Papyrus Covered Coronary Stent System

Class: III, according to Council Directive 93/42/EEC, Annex IX, rule 8

Conformity Assessment Route: Council Directive 93/42/EEC, Annex II, Section 3 and 4

Scope: 17 different variants. See list on next page

We hereby declare that the above-mentioned products meet the provisions of Council Directive 93/42/EEC for medical devices. All supporting documentation is retained under the premises of the manufacturer.

For these products the following EC-Design Examination Certificate has been issued:

Certificate Number:	CE 608286
Notified Body:	BSI Group The Netherlands B.V.
EEC No:	2797
Expiry date:	10.Jun.2023

To these products our approved Full Quality Assurance System according to Annex II of the Directive 93/42/EEC is applied. For this Quality Assurance System the following certificate has been issued:

Certificate Number:	CE 608280
Notified Body:	BSI Group The Netherlands B.V.
EEC No:	2797
Expiry date:	26.May.2024

Date of first CE-marking: 12.Jun.2013

Place, Date of issue: Bülach, 21.MAY.201

Signature:



Marcel Schäfer, Ph.D.

Senior Director Regulatory Affairs and Post Market Surveillance

A11 REG 146249 EN 06


Scope of DoC No. 13-06-02

Pos.	Designation	Catalogue number (REF)	Stent diameter [mm]	Stent length [mm]
1	PK Papyrus 2.5/15	369380	2.5	15
2	PK Papyrus 3.0/15	369381	3.0	15
3	PK Papyrus 3.5/15	369382	3.5	15
4	PK Papyrus 4.0/15	369383	4.0	15
5	PK Papyrus 4.5/15	369384	4.5	15
6	PK Papyrus 5.0/15	369385	5.0	15
7	PK Papyrus 2.5/20	369386	2.5	20
8	PK Papyrus 3.0/20	369387	3.0	20
9	PK Papyrus 3.5/20	369388	3.5	20
10	PK Papyrus 4.0/20	369389	4.0	20
11	PK Papyrus 4.5/20	369390	4.5	20
12	PK Papyrus 5.0/20	369391	5.0	20
13	PK Papyrus 3.0/26	381789	3.0	26
14	PK Papyrus 3.5/26	381790	3.5	26
15	PK Papyrus 4.0/26	381791	4.0	26
16	PK Papyrus 4.5/26	369392	4.5	26
17	PK Papyrus 5.0/26	369393	5.0	26


Change History

Check version index is up to date prior to use.

Version of SAP Document	Main changes from previous release to current release
01	New Document using current template. Replaces "PK Papyrus 130602 Issue 5".
02	Declaration of Conformity updated with the new expiry date of the EC Design Examination Certificate.
03	New issue due to transfer of Notified Body to BSI Group The Netherlands B.V.
04	New issue due to sterilizer addition.
05	Declaration of Conformity updated with the new expiry date of the EC Full Quality Assurance System Certificate. As-
06	Designation of Authorised (EU) Representative. Addition of name and address.



Covered single
stent design



Low crossing profile



Exceptional
deliverability




Technical data /
ordering info

Vascular Intervention // **Coronary**
Covered Coronary Stent System

BIO **BIOTRONIK**
excellence for life

PK Papyrus[®]




PK Papyrus


Designed to save lives when seconds count.

Covered single stent design

With its covered single stent design, PK Papyrus achieves greater bending flexibility and a smaller crossing profile compared to the traditional sandwich design stent, allowing you to seal perforations with confidence.



Traditional sandwich design stent



PK Papyrus
Covered single stent design

Low crossing profile¹






Illustration depicts crimped devices prior to inflation

Innovative polyurethane membrane

Electrostatic forces spin polyurethane fibers onto the stent surface, creating a thin and highly elastic membrane.




Polyurethane fibers
(2000x) \rightarrow 10 μ m




Exceptional deliverability

58% greater flexibility²



Shorter median time to deliver

Single center, retrospective investigation of 61 patients treated with covered coronary stents.^{3,4}



“In rare cases of a coronary perforation, time is the enemy.”

Dr. Dean Kereiakes, Interventional Cardiologist and Medical Director of The Christ Hospital and Vascular Center, Cincinnati, Ohio, USA.

5F Compatibility

For main sizes - no need for guide catheter upgrade (ø 2.5-4.0 mm).

5F





PK Papyrus

Indicated for acute coronary artery perforations.*

Vascular
Intervention
Coronary



Technical Data

Stent

Stent cover material	Non-woven, electrospun polyurethane
Stent cover thickness	90 µm
Stent strut thickness	ø 2.5 - 3.0 mm: 60 µm (0.0024"); ø 3.5 - 4.0 mm: 80 µm (0.0031"); ø 4.5 - 5.0 mm: 120 µm (0.0047")
Stent material	Cobalt chromium (L-605) with proBIO (Amorphous Silicon Carbide) coating
Maximum stent expansion diameter	ø 2.5 - 3.0 mm: 3.50 mm; ø 3.5 - 4.0 mm: 4.65 mm; ø 4.5 - 5.0 mm: 5.63 mm

Delivery system

Guide wire diameter	0.014"
Usable catheter length	140 cm
Recommended guide catheter	ø 2.5 - 4.0 mm: 5F (min. I.D.** 0.056"); ø 4.5 - 5.0 mm: 6F (min. I.D.** 0.070")
Nominal pressure (NP)	ø 2.5 - 3.5 mm: 8 atm; ø 4.0 - 5.0 mm: 7 atm
Rated burst pressure (RBP)	ø 2.5 - 4.0 mm: 16 atm; ø 4.5 - 5.0 mm: 14 atm

**I.D. = Inner Diameter

Ordering Information

Stent ø (mm) Catheter length 140 cm Stent length (mm)

		15	20	26
5F	2.5	369380	369386	-
	3.0	369381	369387	381789
	3.5	369382	369388	381790
	4.0	369383	369389	381791
6F	4.5	369384	369390	369392
	5.0	369385	369391	369393

1. Compared to Graftmaster 2.8/16 (BIOTRONIK data on file); 2. Compared to Jostent Graftmaster 3.0/16 (BIOTRONIK data on file); 3. Hernández-Enríquez M, Lairez O, Campelo-Parada F, et al. Outcomes after use of covered stents to treat coronary artery perforations. Comparison of old and new-generation covered stents. J Interv Cardiol. 2018;1-7. doi: 10.1111/joic.12525; 4. Population is representative of real world interventional practice and was not a randomized prospective clinical trial.

PK Papyrus and **proBIO** are trademarks or registered trademarks of the BIOTRONIK Group of Companies. Jostent and Graftmaster are trademarks or registered trademarks of the Abbott Group of Companies.

*Indication as per IFU.

394771/F/Jul_2019_DV

BIOTRONIK AG
Ackerstrasse 6
8180 Bülach, Switzerland
Tel +41 (0) 44 8645111
Fax +41 (0) 44 8645005
info.vi@biotronik.com
www.biotronik.com

© 2019 BIOTRONIK AG – All rights reserved.
Specifications are subject to modification,
revision and improvement.

 **BIOTRONIK**
excellence for life






Outstanding
patient outcomes



Highly deliverable



Ultrathin 60* μm
struts




Technical data /
ordering info

Vascular Intervention // **Coronary**
Drug-Eluting Stent System

BIOTRONIK
excellence for life

Orsiro[®]



* \varnothing 2.25 – 3.0 mm

Orsiro

Ultrathin struts[§]. Outstanding patient outcomes[◇].


Outstanding patient outcomes

Improving patient outcomes, year after year*

BIOFLOW-V (n = 1,334) the FDA pivotal trial

Significant differences in TLF observed at year 1 and 2 were maintained and further increased at year 3 (8.6% vs. 14.4%, p = 0.003), driven by significant differences in TV-MI (5.5% vs. 10.1%, p = 0.004) and Ischemia-driven TLR (3.4% vs. 6.9%, p = 0.008) that favor Orsiro over Xience.^{1,2,3}

TLF and components at 12, 24 and 36 Months



40%

lower
TLF rate³ ^ϕ

(p=0.003)

46%

lower
TV-MI rate³ ^ϕ

(p=0.004)

52%

lower
Ischemia-driven
TLR rate³ ^ϕ

(p= 0.008)

TLF – Target Lesion Failure; TV-MI – Target Vessel Myocardial Infarction; TLR – Target Lesion Revascularization.

[§]As characterized with respect to strut thickness in Bangalore et al. Meta-analysis.

[◇]Based on investigator's interpretation of BIOFLOW-V primary endpoint results.

*Compared to Xience, based on three consecutive years.


[‡]p-values for 36-m frequentist analysis.

^ϕvs. Xience, based on 36-m frequentist analysis.

Superiority in STEMI⁴

BIOSTEMI (n=1,300) is the first RCT demonstrating superiority between two contemporary DES


Orsiro is superior to Xience in STEMI patients undergoing primary PCI with respect to Target Lesion Failure (TLF) rate at 12 months.



Long-term safety

In the randomized, all-comers BIOSCIENCE trial (n= 2,119)⁵

Orsiro shows numerically equal or lower Stent Thrombosis (ST) in complex patients in comparison to Xience.



*Compared to Xience, BIOTRONIK data on file based on the Rate Ratio of 0.59.


**BCI: Bayesian Credibility Interval.

[‡]n= 1,300 newly enrolled STEMI patients including 407 patients from the BIOSCIENCE STEMI subgroup used as prior information.

Highly deliverable


Better push

Transmits up to 72% more force from hub to tip.¹³




Easier cross

Up to 79% less force needed to successfully cross demanding anatomies.¹³



Lower crossing profile

Improved acute performance – up to 7% lower crossing profile.¹³




79%
easier to cross
vs. Synergy¹³




Ultrathin 60 µm struts


Improved outcomes start in the early phase



48 hours
Thinner struts mean
less vessel injury⁶



30 days^Δ
80.4% strut coverage⁷



90 days^Δ
98.7% strut coverage⁷

Thinner struts make the difference

Ultrathin vs. second generation DES in a large scale meta-analysis including more than 11,000 patients^{8,9}

16%

Relative risk reduction in TLF at 12 months
RR (95% CI) 0.84 (0.72, 0.99)

‡ Driven by peri-procedural MI events (<48 hours). In-hospital rate may include events > 48 hours.

Δ Images: Secco G et al. Time-related changes in neointimal tissue coverage following a new generation SES implantation: an OCT observational study. Presented at: euro PCR, May 20, 2014; Paris, France.

Strut thickness
in perspective¹⁰

Orsiro
BIOTRONIK
CoCr-SES



60 µm*

Synergy
Boston Scientific
PtCr-EES



74 µm

Ultimaster
Terumo
CoCr-SES



80 µm

Resolute Onyx^{11,12}
Medtronic
CoNi-ZES



81 µm

Xience Family
Abbott
CoCr-EES




81 µm

Promus
Boston Scientific
PtCr-EES



81 µm


BioMatrix
Biosensors
316L-BES



120 µm

* ø 2.25 – 3.0 mm





Indicated for discrete de novo stenotic lesions and in-stent restenotic lesions.*

Technical Data		Stent
		Stent material
		Cobalt chromium, L-605
		Passive coating
		proBIO (Amorphous Silicon Carbide)
		Active coating
		BIOlute bioabsorbable Poly-L-Lactide (PLLA) eluting a limus drug
		Drug dose
		1.4 µg / mm ²
		Strut thickness
		ø 2.25 - 3.0 mm: 60 µm (0.0024"); ø 3.50 - 4.0 mm: 80 µm (0.0031")
Delivery system		
		Catheter type
		Rapid exchange
		Recommended guide catheter
		5F (min. I.D. 0.056")
		Lesion entry profile
		0.017"
		Guide wire diameter
		0.014"
		Usable catheter length
		140 cm
		Balloon material
		Semi crystalline polymer material
		Coating (distal shaft)
		Hydrophilic coating
		Marker bands
		Two swaged platinum-iridium markers
		Proximal shaft diameter
		2.0F
		Distal shaft diameter
		2.6F: ø 2.25 - 3.5 mm; 2.8F: ø 4.0 mm
		Nominal pressure (NP)
		8 atm
		Rated burst pressure (RBP)
		16 atm

Compliance Chart		Balloon diameter x length (mm)					
		ø 2.25 x 9-40	ø 2.50 x 9-40	ø 2.75 x 9-40	ø 3.00 x 9-40	ø 3.50 x 9-40	ø 4.00 x 9-40
Nominal Pressure (NP)	atm**	8	8	8	8	8	8
	ø (mm)	2.25	2.50	2.75	3.00	3.50	4.00
Rated Burst Pressure (RBP)	atm**	16	16	16	16	16	16
	ø (mm)	2.50	2.77	3.05	3.33	3.88	4.44

**1 atm = 1.013 bar

Ordering Information	Stent ø (mm)	Catheter length 140 cm Stent length (mm)								
		9	13	15	18	22	26	30	35	40
	2.25	364469	364475	364481	364487	364499	364505	364511	391234	391238
	2.50	364470	364476	364482	364488	364500	364506	364512	391235	391239
	2.75	364471	364477	364483	364489	364501	364507	364513	391236	391240
	3.00	364472	364478	364484	364490	364502	364508	364514	391237	391241
	3.50	364473	364479	364485	364491	364503	364509	364515	391018	391020
	4.00	364474	364480	364486	364492	364504	364510	364516	391019	391021

1. Kandzari D et al. Ultrathin, bioresorbable polymer sirolimus-eluting stents versus thin, durable polymer everolimus-eluting stents in patients undergoing coronary revascularisation (BIOFLOW V): a randomised trial. Lancet. 2017 Oct 21; 390(10105):1843-1852; 2. Kandzari D et al. Ultrathin bioresorbable polymer sirolimus-eluting stents versus thin durable polymer everolimus-eluting stents. Journal of the American College of Cardiology. 2018 Dec 17;72(25):3287-97; 3. Kandzari D et al. J Am Coll Cardiol. Cardiovasc Interven. 2020, doi: 10.1016/j.jcin.2020.02.019. 4. Iglesias JF et al. Biodegradable polymer sirolimus-eluting stents versus durable polymer everolimus-eluting stents in patients with ST-segment elevation myocardial infarction (BIOSTEMI): a single-blind, prospective, randomised superiority trial; Lancet, September, 2019; 5. Pilgrim T et al. 5-year outcomes of the BIOSCIENCE randomised trial. Supplementary appendix; Lancet 2018; published online Aug 28. [http://dx.doi.org/10.1016/S0140-6736\(18\)31715-X](http://dx.doi.org/10.1016/S0140-6736(18)31715-X); 6. Foin et al. Impact of stent strut design in metallic stents and biodegradable scaffolds. Int J Cardiol.2014 Dec 20;177(3):800-8; 7. Secco G et al. Time-related changes in neointimal tissue coverage of a novel Sirolimus eluting stent: Serial observations with optical coherence tomography. Cardiovascular Revascularization Medicine 17.1 (2016): 38-43; 8. Bangalore S et al. Newer-generation ultrathin strut drug-eluting stents versus older second-generation thicker strut drug-eluting stents for coronary artery disease: meta-analysis of randomized trials. Circulation. 2018 Nov 13;138(20):2216-26; 9. Bangalore S, et al. Newer-generation ultrathin strut drug-eluting stents versus older second-generation thicker strut drug-eluting stents for coronary artery disease: meta-analysis of randomized trials. Circulation. 2018 Jul. 24: 2216-2226; 10. Stefanini GG et al. Coronary stents: novel developments. Heart. 2014 Jul 1;100(13):1051-61; 11. Low AF. Stent platform for procedural success: Introducing the Continuous Sinusoidal & Core Wire Technologies. Presented at: AsiaPCR; 22-24 January, 2015; Singapore, Singapore; 12. Tolentino A. Evolving DES Strategy: Biodegradable Polymer vs. Bioabsorbable Scaffold. Presented at: Cardiovascular Nurse/TechnologistSymposium; June 17, 2016; New York, USA; 13. BIOTRONIK data on file.

Target Lesion Failure (TLF), Target Lesion Revascularization (TLR), Target Vessel Myocardial Infarction (TV-MI), Stent Thrombosis (ST).

*Indication as per IFU.

Orsiro, proBIO and BIOlute are trademarks or registered trademarks of the BIOTRONIK Group of Companies.

Synergy and Promus are trademarks or registered trademarks of the Boston Scientific group of companies. Resolute and Resolute Onyx are trademarks or registered trademarks of the Medtronic group of companies. Xience, Xience Prime and Xience Sierra are trademarks or registered trademarks of the Abbott group of companies. Ultimaster is a trademark or registered trademark of the Terumo group of companies. BioMatrix is a trademark or registered trademark of the Biosensors International Group.