

**CERTIFICATE OF ANALYSIS**

Certificado de Análisis / Certificat de contrôle / Chargen-Prüfprotokoll

Pourquier\* Brucellosis Positive Control [Brucellosis Positive Control]

---

<b>Product code :</b>	<b>P00130</b>	<b>Expiration date :</b>	<b>14-JUL-2024</b>
<b>Batch :</b>	<b>439</b>		
<b>Tests per kit :</b>	<b>/</b>		
<b>Insert version :</b>	<b>/</b>	<b>Manufacturing date :</b>	<b>JUL-2019</b>

---

Reconstitute\* with 1ml distilled water

\* After reconstitution, the control serum must be stored at a temperature  $\leq -16^{\circ}\text{C}$  for up to 1 year

He can be frozen and unfrozen up to three times

---

<b>Control Test</b>	<b>Spécifications</b>	<b>Results</b>
<b>EAT :</b>	Positive : Agglutination	Positive
<b>SAW :</b>	Positive ( $\geq 30$ UI/ml)	60 IU/ml [30-120]
<b>FC :</b>	Positive ( $\geq 20$ ICFTU/ml)	267 ICFTU/ml [133 - 533]
<b>ELISA :</b>	Positive	Positive

---

This product was performance tested by French Regulatory Agency (ANSES) and aligned with our results required for release.

This information is released by :

Title : **Quality & EHS Assistant, Mauran Rachel**

Signature :



Manufacturer: IDEXX Montpellier SAS - 326, rue de la Galéra – Parc Euromédecine - 34090 Montpellier - France

Tél. : 33(0) 4 99 23 24 25 - Fax : 33(0) 4 67 04 20 25 - [info-montpellier@idexx.com](mailto:info-montpellier@idexx.com)

EN 05-01-17 - Vers : C

Document printed on : 20-Aug-2019