



# Certificate of Analysis

## Capillary Array 50cm 3500 Series

Product No. 4404685  
Lot No. L517F65  
Date of Manufacture 14-Jul-2017

TEST	SPECIFICATION	RESULT
<b>CAPILLARY LEAK TEST</b> Each array is tested to assure absence of leaks.		
Leak Test	No Leaks	PASS
<b>CAPILLARY PLUG TEST</b> Each array is tested to assure absence of plugged capillaries.		
Plug Test	No Plugged Capillaries	PASS
<b>FLUORESCENCE BACKGROUND TEST</b> Each array is tested to assure background fluorescence is within specification.		
Fluorescence Background Test	Conforms to Specification	PASS
<b>CAPILLARY MAPPING TEST</b> Each array is tested to verify that capillary inlet and outlet spatial positioning correspond.		
Mapping Test	Conforms to Specification	PASS

*For Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.*

**ISO 9001  
REGISTERED**

Foster City, CA

*Agnes Teubner*  
Quality Assurance

To obtain a Certificate of Analysis on-line go to [www.appliedbiosystems.com](http://www.appliedbiosystems.com) or email us at [cofarequest@lifetech.com](mailto:cofarequest@lifetech.com)

4442856 C