

Declaration of Conformity

- for EU and UK -

Manufacturer's Name: SHIMADZU CORPORATION
Analytical & Measuring Instruments Division
Address : 1, Nishinokyo-Kuwabara-cho, Nakagyo-ku,
 Kyoto 604-8511, Japan

declares in sole responsibility that the following product
Product Name Total Organic Carbon Analyzer
Model Name TOC-L series
P/N Depend on configuration. See Appendix 1.

referred to in this declaration conforms with following directives and standards

EMC Directive 2014/30/EU	EMC Regulation S.I. 2016/1091
EN 61326-1:2013	EN 61326-1:2013
Low Voltage Directive 2014/35/EU	Electrical Equipment Regulation S.I. 2016/1101
EN 61010-1:2010	EN 61010-1:2010
RoHS Directive 2011/65/EU	RoHS Regulation S.I. 2012/3032

Note 1) This declaration becomes invalid if technical or operational modifications are introduced without manufacturer's consent.

Note 2) This declaration is valid if this product is used alone or in combination with the accessories of this product which are mentioned in attached Appendix 1 or other instruments which fulfill with the requirement of mentioned directive.

Note 3) Authorised Representative in EU is as follows:

SHIMADZU EUROPA GmbH

Address :Albert-Hahn-Strasse 6-10,47269 Duisburg, F.R.Germany

Note4) Authorised Representative in UK is as follows:

SHIMADZU UK Limited

Address : Unit 1A, Mill Court, Featherstone Road Wolverton Mill South,
 Milton Keynes MK12 5RD, U.K.

Kyoto, JAPAN 9 June 2021

Place and date of issue



Signature

Toshiyuki Kawano

Name

General Manager of Quality Assurance Department

Analytical & Measuring Instruments Division

Title



Appendix 1

Product Name : Total Organic Carbon Analyzer

Model No.	P/N
TOC-LCSN	638-91109-58、 638-91149-59
TOC-LCPN	638-91110-58、 638-91150-58
TOC-LCSH	638-91111-58、 638-91151-58
TOC-LCPH	638-91112-58、 638-91152-58







Product Options

Model No.	P/N
ASI-L (Auto Sampler,24ML)	638-93199-58
ASI-L (Auto Sampler,9-40ML)	638-93200-58
OCT-L (8 Port Sampler,OCT-L-1)	638-93201-58
OCT-L (8 Port Sampler,OCT-L-2)	638-93202-58
TNM-L(Total Nitrogen Measuring unit)	638-91108-58
CARRIER GAS PURIFY KIT L E	638-41447-04
EXTERNAL SPARGE ASSY	638-77183-40、 41
POC KIT,L	638-42101-01、 02
OPTION BOARD (LAN) ASSY,L ROHS	638-79070-41
OPTION BOARD (DIO) ASSY,L ROHS	638-79069-41、 42
MAINTENANCE SET,PCBA FLOWCON	638-79041-49
STIRRER ASSY,24A L	638-67099-41
STIRRER ASSY,40A L	638-67100-41
LabSolutions TOC DB	638-64172-91
LabSolutions TOC DB Standard License	638-64165-94
LabSolutions TOC CS	638-64173-91
LabSolutions TOC CS Standard License	638-64165-91
LabSolutions TOC CS Additional License	638-64165-95
LabSolutions TOC DB Additional License	638-64165-92



新規・改訂内容一覧表

担当部署：品質保証部

改訂記号	改訂理由	改訂内容	作成者 日付	検討者 日付	承認者 日付
	—	新規作成	矢幡 2014-3-14	石田 2014-3-14	谷水 2014-3-14
A	誤記訂正	Appendix 1 P/N 638-91112-58 Model No.訂正 TOC-L CSN→TOC-L CPH	神吉 2015-9-8	石田 2015-9-8	谷水 2015-9-8
B	追記訂正	Appendix 1 オプション追加 (Carrier gas, External Sparge, POC, Option board LAN, DIO, PCBA Flowcon)	中森 2015-9-14	石田 2015-9-15	谷水 2015-9-15
C	追記訂正 誤記訂正	Appendix 1 オプション追加 スターラー24A,40A P.1 EMC 2004/108→2014/30 EN 61326-1:2006→:2013	中森 2015-10-6	石田 2015-10-6	谷水 2015-10-6
D	最新規格対応	p.1 Low Voltage Directive 2006/95/EC→2014/35/EU The last two digits...以下2行削除			
E	・UK対応 ・新製品追加	・UKの規格番号の追記 ・Appendix 1 LabSolution TOC 関連の製品追加			

配布先

名称 TITLE

Declaration of Conformity TOC-L

番号 No.

ZEAE-2118²

SHEET

3/3

改定 REV

D