

Detalii specifice și standarde tehnice

LOT „Procurare de Termociclor”:

№	Name of product	Minimal technical specifications	Offered specifications; model, manufacturer.
1	Termociclor	<p>The thermal cycler for PCR with 96 positions for 0.2 ml tubes. Power configurations: VAC 220-240V, 50 Hz.</p> <ul style="list-style-type: none"> - Memory for minimum 5 PCR programs, that can be edited manually. - Temperature range: 04-100°C; - Real-time graphical display of actual and set block temperatures, - Real-time digital display of all sensor values; - Temperature accuracy: +/-0.5°C; - Heating rate: up to 6°C/s; - Cooling rate: Up to 3°C/s; - Calibrated to traceable international standards. 	<p style="text-align: center;">T100 Thermal Cycler Cod ref:1861096 Producer Bio Rad Laboratories Inc.</p> <p>The thermal cycler for PCR with 96 positions for 0.2 ml tubes, 96 x 0.2 ml strips or 96-well plate. Power configurations: VAC 220-240V, 50-60 Hz.</p> <ul style="list-style-type: none"> - Memory onboard for minimum 500 PCR programs, that can be edited manually, unlimited with USB flash. - Temperature range: 04-100°C; - Real-time graphical display of actual (block) and calculated temperatures, - Real-time digital display of all sensor values; - Temperature accuracy: +/-0.5°C; - Maximum ramp rate (Heating & Cooling): 4°C/s(A number from 0.1 to 4°C per sec. Applies only to a temperature step. Instructs the thermal cycler to ramp to the target temperature at the specified ramp rate in that step. If not specified, the T100 thermal cycler runs at a maximum ramp rate of 4°C/sec); - Calibrated to traceable international standards. - Warranty period 12 months.
	Mandatory requirements	<ul style="list-style-type: none"> - Original manufacturer packing for all items; - Quality Certificates (CE and/or DC) or any other equivalent; - ISO 9001-2015 	<ul style="list-style-type: none"> - Original manufacturer packing for all items; - Quality Certificates (Declaration of conformity); - ISO 9001-2015

- **Toate bunurile vor fi testate de beneficiar. Acceptarea bunurilor și semnarea actelor de recepție va avea loc după testarea pozitivă efectuată de beneficiar.**

NUMELE FURNIZORULUI SRL”**Dialab Solutions**”

Semnătura și ștampila autorizate _____

Locul:mun. Chișinău

Data:23.08.2023