

Certificate of Analysis

Certificat d'Analyse

applied
biosystems
by Thermo Fisher Scientific

PrioCHECK™ African Swine Fever Virus Ab Kit



Tests

LOT

A230901L

REF

A56981



2023-09-11



2024-12-31

KIT CONFIGURATION CONFIGURATION DU KIT

COMPONENT DESCRIPTION	REF	UNIT	LOT	
1: Test Plate	7611011	5 units	A230901L201	2024-12-31
2: Positive Control (PC)	7611014	600 µL	A230901L220	2024-12-31
3: Negative Control (NC)	7611015	600 µL	A230901L224	2024-12-31
4: ASFV Sample diluent	7611016	150 mL	A230901L204	2024-12-31
5: ASFV Conjugate	7611013	60 mL	A230901L212	2024-12-31
6: Chromogen (TMB) Substrate	7588925	60 mL	A230901L214	2024-12-31
7: Stop Solution	7610302	60 mL	A230901L216	2024-12-31
8: Washing fluid (20x)	7611006	2x 60 mL	A230901L206	2024-12-31
Additional kit contents	-	10 plate sealers	-	-

INSTRUCTIONS FOR USE

Notices d'utilisation



English

MAN0019591 Rev. A.0

QUALITY CONTROL RESULTS RESULTATS DE CONTROLE QUALITE

Repeatability: CV < 15 %			
Components	Specification	Result	Conclusion
Composants	Spécification	Résultat	Conclusion
OD ₄₅₀ Negative Control	≤ 0.250	0.167	PASSED
OD ₄₅₀ Positive Control	≥ 0.400	1.578	PASSED
OD ₄₅₀ Positive Control / OD ₄₅₀ Negative Control	≥ 5	9	PASSED

SUPPLEMENTAL INFORMATION - Available upon request to thermofisher.com/support

Informations complémentaires - Disponibles sur demande auprès de thermofisher.com/support

Additional quality control data can be provided upon request.

The aforementioned batch meets all specifications.

Le lot susmentionné est conforme aux spécifications



Prionics Lelystad B.V.
Platinastraat 33
8211 AR Lelystad
The Netherlands

www.thermofisher.com/support

M. Birkhuijsen

Quality Assurance

Date: 06 Oct 2023