

EU Declaration of Conformity

as per Annex IV of the Regulation EU 2017/746 on in-vitro diagnostic medical devices

Manufacturer: Roche Diagnostics GmbH

Address: Sandhofer Strasse 116
68305 Mannheim
Germany

Single Registration Number: DE-MF-000006260

Roche Diagnostics GmbH declares, under the sole responsibility, that the product/the product line

Product Name	Cat. No.	Basic UDI-DI
Elecsys CA 19-9	11776193122	761333600730AH
Elecsys CA 19-9	11776193214	761333602081AD
Elecsys CA 19-9	07027028190	761333600799BM
Elecsys CA 19-9	07027028214	761333602050A2

Intended Use:

Immunoassay for the in vitro quantitative determination of CA 19-9 in human serum and plasma. The electrochemiluminescence immunoassay “ECLIA” is intended for use on cobas e immunoassay analyzers.

Product Name	Cat. No.	Basic UDI-DI
CA 19-9 CalSet	11776215122	761333600732AM

Intended Use:

CA 19-9 CalSet is used for calibrating the quantitative Elecsys CA 19-9 assay on cobas e immunoassay analyzers.

Risk Class: ☐ A ☐ B ☒ C ☐ D

Conformity Route: ☐ Self-Declaration of Conformity (Class A)
☐ Self-Declaration of Conformity after Notified Body involvement for sterile manufacturing conditions acc. Art. 48 (10) (Class A sterile)
☒ Technical Documentation Assessment Class B/C – Annex IX
☐ Technical Documentation Assessment Class D – Annex IX
☐ Technical Documentation Assessment Class B/C/D for Self-Testing – Annex IX
☐ Technical Documentation Assessment Class B/C/D for Near-Patient Testing – Annex IX

☐ Technical Documentation Assessment Class C/D for Companion Diagnostics – Annex IX

Certificates:

☒ EU QM Certificate No.: V12 010283 0639

☐ EU Technical Documentation Assessment Certificate No. (Class D, Near-Patient Testing, Self-Testing and Companion Diagnostics):

Other:

☐ Common Specifications:

Notified Body (NB) Name: TÜV Süd Product Service GmbH

NB Address: Ridlerstraße 65
80339 Munich
Germany

NB Ident. No.: 0123

to which this declaration relates fulfils the requirements of Regulation EU 2017/746 on in-vitro diagnostic medical devices.

Mannheim, 12 August 2024

Roche Diagnostics GmbH

ppa./on behalf of the company

DocuSigned by:

485913ABEB04408...

Dr. Peer Lorenz
Site Quality Head / Network Lead, Mannheim

ppa./on behalf of the company

DocuSigned by:

FC5EDEC1054B44C...

Dr. Stefan Scheib
Global Head of Regulatory Affairs, Core Lab

Contact address: Roche Diagnostics GmbH
Abt./Dept. Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim