

MAGLUMI[®] IgE (CLIA)

INTENDED USE

The kit is an *in vitro* chemiluminescence immunoassay for the quantitative determination of total Immunoglobulin E (IgE) in human serum and plasma using the MAGLUMI series Fully-auto chemiluminescence immunoassay analyzer and Biolumi series Integrated System, and the assay is used for an aid in the diagnosis of individuals with suspected or confirmed allergic disorders.

SUMMARY

Immunoglobulin E (IgE) is the antibody isotype that contains the ϵ heavy chain and it is a monomer with five domains in the immunoglobulin structure¹. IgE is well known for its role in allergic disease, the manifestations of which are mediated through its two Fc receptors, Fc ϵ RI and CD23 (Fc ϵ RII)².

IgE is a marker of allergic diseases; high serum levels are present in patients with hay fever, allergic rhinitis, extrinsic asthma, atopic dermatitis, etc³. Patients with chronic spontaneous urticaria (csU), the most frequent type of non-acute urticaria, have been described to exhibit increased levels of IgE⁴. A protective role of IgE antibodies in infections with certain parasites in humans, as the levels of parasite-specific IgE and resistance to infection correlate positively⁵. Parasite infections are initially characterized by production of large amounts of nonspecific IgE while parasite-specific IgE is detectable only at later stages of primary infections or after multiple infections⁶. Several classical autoantigens which are known as targets in rheumatoid arthritis, systemic lupus erythematosus (SLE), in addition to autoimmune diseases of the skin, gut, eye and endocrine glands, were also reported to react with IgE antibodies⁶. High IgE levels in HIV-1-infected children and adults have been associated with progression of the disease⁷. IgE levels could be a marker of poor prognosis in some patients in the early or late stages of HIV -1 infection⁷. Aspergillosis is the name given to all diseases caused by the fungus in the genus aspergillus and includes allergic, superficial, saprophytic and invasive disease⁸. Total IgE and Aspergillus-Specific IgE levels are often elevated. Therefore, close long-term follow-up with serial assessments of total serum IgE is advised for patients⁹.

Immediate GI hypersensitivity is an IgE-mediated GI reaction that develops within minutes to 2 hours after exposure of the GI immune system to the offending food allergen and presents clinically with nausea, vomiting, abdominal pain, and diarrhea¹⁰. Elevated IgE in chronic graft-versus-host disease (cGVHD) was associated with increased levels of multiple immune cell subsets and immunoglobulins suggesting that IgE may be a marker of more robust post-transplant immune reconstitution¹¹. IgE synthesis may be accelerated during the early stages following burn injury¹². Multiple myeloma (MM) is a neoplastic condition whose hallmark is the proliferation of malignant plasma cells in the bone marrow, resulting in an increase in serum and/or urine monoclonal-(M) protein and end-organ damage. IgE MM is a rare disease, accounting for only 0.01% of all patients with MM¹³. Serum IgE levels are particularly increased in patients with alcoholic liver disease¹⁴. For viral hepatitis, increased IgE concentrations have been observed during acute hepatitis A and B. Chronic hepatitis B carriers may also have high IgE levels¹⁴. Furthermore, low or undetectable serum IgE levels have been frequently observed in patients with primary biliary cirrhosis¹⁵.

TEST PRINCIPLE

Sandwich chemiluminescence immunoassay.

The prediluted sample, ABEI labeled with anti-IgE monoclonal antibody, buffer, magnetic microbeads coated with another anti-IgE monoclonal antibody are mixed thoroughly and incubated, reacting to form sandwich complexes. After precipitation in a magnetic field, the supernatant is decanted and then a wash cycle is performed. Subsequently, the Starter 1+2 are added to initiate a chemiluminescent reaction. The light signal is measured by a photomultiplier as relative light units (RLUs), which is proportional to the concentration of IgE present in the sample.

REAGENTS

Kit Contents

| Component | Description | 100 tests/kit | 50 tests/kit | 30 tests/kit |
|----------------------------|--|---------------|--------------|--------------|
| Magnetic Microbeads | Magnetic microbeads coated with anti-IgE monoclonal antibody (~6.67 μ g/mL) in PBS buffer, NaN ₃ (<0.1%). | 2.5 mL | 1.5 mL | 1.0 mL |
| Calibrator Low | A low concentration of IgE in PBS buffer, NaN ₃ (<0.1%). | 1.0 mL | 1.0 mL | 1.0 mL |
| Calibrator High | A high concentration of IgE in PBS buffer, NaN ₃ (<0.1%). | 1.0 mL | 1.0 mL | 1.0 mL |
| Buffer | Tris-HCl buffer, NaN ₃ (<0.1%). | 6.5 mL | 4.0 mL | 3.0 mL |
| ABEI Label | ABEI labeled with anti-IgE monoclonal antibody (~0.100 μ g/mL) in Tris-HCl buffer, NaN ₃ (<0.1%). | 7.5 mL | 4.5 mL | 3.3 mL |
| Diluent | 0.9% NaCl. | 25.0 mL | 13.5 mL | 9.0 mL |
| Control 1 | A low concentration of IgE (45.0 IU/mL) in PBS buffer, NaN ₃ (<0.1%). | 1.0 mL | 1.0 mL | 1.0 mL |
| Control 2 | A high concentration of IgE (180 IU/mL) in PBS buffer, NaN ₃ (<0.1%). | 1.0 mL | 1.0 mL | 1.0 mL |

All reagents are provided ready-to-use.

Warnings and Precautions

- For *in vitro* diagnostic use.
- For professional use only.
- Exercise the normal precautions required for handling all laboratory reagents.
- Personal protective measures should be taken to prevent any part of the human body from contacting samples, reagents, and controls, and should comply with local operating requirements for the assay.
- A skillful technique and strict adherence to the package insert are necessary to obtain reliable results.
- Do not use kit beyond the expiration date indicated on the label.
- Do not interchange reagent components from different reagents or lots.
- Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).
- All waste associated with biological samples, biological reagents and disposable materials used for the assay should be considered potentially infectious and should be disposed of in accordance with local guidelines.
- This product contains sodium azide. Sodium azide may react with lead or copper plumbing to form highly explosive metal azides. Immediately after disposal, flush with a large volume of water to prevent azide build-up. For additional information, see Safety Data Sheets available for professional user on request.

Note: If any serious incident has occurred in relation to the device, please report to Shenzhen New Industries Biomedical Engineering Co., Ltd. (Snibe) or our authorized representative and the competent authority of the Member State in which you are established.

Reagent Handling

- To avoid contamination, wear clean gloves when operating with a reagent kit and sample. When handling reagent kit, replace the gloves that have been in contact with samples, since introduction of samples will result in unreliable results.
- Do not use kit in malfunction conditions; e.g., the kit leaking at the sealing film or elsewhere, obviously turbid or precipitation is found in reagents (except for Magnetic Microbeads) or control value is out of the specified range repeatedly. When kit in malfunction conditions, please contact Snibe or our authorized distributor.
- To avoid evaporation of the liquid in the opened reagent kits in refrigerator, it is recommended that the opened reagent kits to be sealed with reagent seals contained within the packaging. The reagent seals are single use, and if more seals are needed, please contact Snibe or our authorized distributor.
- Over time, residual liquids may dry on the septum surface. These are typically dried salts and have no effect on assay efficacy.
- Use always the same analyzer for an opened reagent integral.
- For magnetic microbeads mixing instructions, refer to the Preparation of the Reagent section of this package insert.
- For further information about the reagent handling during system operation, please refer to Analyzer Operating Instructions.

Storage and Stability

- Do not freeze the integral reagents.

- Store the reagent kit upright to ensure complete availability of the magnetic microbeads.
- Protect from direct sunlight.

| Stability of the Reagents | |
|---------------------------|----------------------------------|
| Unopened at 2-8°C | until the stated expiration date |
| Opened at 2-8°C | 6 weeks |
| On-board | 4 weeks |

| Stability of Controls | |
|--------------------------|----------------------------------|
| Unopened at 2-8°C | until the stated expiration date |
| Opened at 10-30°C | 6 hours |
| Opened at 2-8°C | 6 weeks |
| Frozen at -20°C | 3 months |
| Frozen and thawed cycles | no more than 3 times |

SPECIMEN COLLECTION AND PREPARATION

Specimen Types

Only the specimens listed below were tested and found acceptable.

| Specimen Types | Collection Tubes |
|----------------|--|
| Serum | Tubes without additive/accessory, or tubes containing clot activator or clot activator with gel. |
| Plasma | K2-EDTA or Na-heparin |

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. Follow tube manufacturers' instructions carefully when using collection tubes.

Specimen Conditions

- Do not use heat-inactivated samples or grossly hemolyzed/hyperlipidaemia specimens and specimens with obvious microbial contamination.
- Ensure that complete clot formation in serum specimens has taken place prior to centrifugation. Some serum specimens, especially those from patients receiving anticoagulant or thrombolytic therapy, may exhibit increased clotting time. If the serum specimen is centrifuged before a complete clotting, the presence of fibrin may cause erroneous results.
- Samples must be free of fibrin and other particulate matter.
- To prevent cross contamination, use of disposable pipettes or pipette tips are recommended.

Preparation for Analysis

- Inspect all specimens for foam. Remove foam with an applicator stick before analysis. Use a new applicator stick for each specimen to prevent cross contamination.
- Frozen specimens must be completely thawed before mixing. Mix thawed specimens thoroughly by low speed vortexing or by gently inverting. Visually inspect the specimens. If layering or stratification is observed, mix until specimens are visibly homogeneous. If specimens are not mixed thoroughly, inconsistent results may be obtained.
- Specimens should be free of fibrin, red blood cells, or other particulate matter. Such specimens may give reliable results and must be centrifuged prior to testing. Transfer clarified specimen to a sample cup or secondary tube for testing. For centrifuged specimens with a lipid layer, transfer only the clarified specimen and not the lipemic material.
- The sample volume required for a single determination of this assay is 20 μ L.

Specimen Storage

Specimens removed from the separator, red blood cells or clot may be stored up to 8 hours at 10-30°C or 7 days at 2-8°C, or 6 months frozen at -20°C. Frozen specimens subjected to up to 2 freeze/thaw cycles have been evaluated.

Specimen Shipping

- Package and label specimens in compliance with applicable local regulations covering the transport of clinical specimens and infectious substances.
- Do not exceed the storage limitations listed above.

Specimen Dilution

- Samples, with IgE concentrations above the analytical measuring interval, can be diluted with Diluent either by following automated dilution protocol or manual dilution procedure. The recommended dilution ratio is 1:3. The concentration of the diluted sample must be >1067 IU/mL.
- For manual dilution, multiply the result by the dilution factor. For dilution by the analyzers, the analyzer software automatically takes the dilution into account when calculating the sample concentration.

PROCEDURE

Materials Provided

IgE (CLIA) assay, control barcode labels.

Materials Required (But Not Provided)

- General laboratory equipment.
- Fully-auto chemiluminescence immunoassay analyzer Maglumi 600, Maglumi 800, Maglumi 1000, Maglumi 2000, Maglumi 2000 Plus, Maglumi 4000, Maglumi 4000 Plus, MAGLUMI X8, MAGLUMI X3, MAGLUMI X6 or Integrated System Biolumi 8000, Biolumi CX8.
- Additional accessories of test required for the above analyzers include Reaction Module, Starter 1+2, Wash Concentrate, Light Check, Tip, and Reaction Cup. Specific accessories and accessories' specification for each model refer to corresponding Analyzer Operating Instructions.
- Please use accessories specified by Snibe to ensure the reliability of the test results.

Assay Procedure

Preparation of the Reagent

- Take the reagent kit out of the box and visually inspect the integral vials for leaking at the sealing film or elsewhere. If there is no leakage, please tear off the sealing film carefully.
- Open the reagent area door; hold the reagent handle to get the RFID label close to the RFID reader (for about 2s); the buzzer will beep; one beep sound indicates successful sensing.
- Keeping the reagent straight insert to the bottom along the blank reagent track.
- Observe whether the reagent information is displayed successfully in the software interface, otherwise repeat the above two steps.
- Resuspension of the magnetic microbeads takes place automatically when the kit is loaded successfully, ensuring the magnetic microbeads are totally resuspended homogenous prior to use.

Assay Calibration

- Select the assay to be calibrated and execute calibration operation in reagent area interface. For specific information on ordering calibrations, refer to the calibration section of Analyzer Operating Instructions.
- Execute recalibration according to the calibration interval required in this package insert.

Quality Control

- When new lot used, check or edit the quality control information.
- Scan the control barcode, choose corresponding quality control information and execute testing. For specific information on ordering quality controls, refer to the quality control section of the Analyzer Operating Instructions.

Sample Testing

- After successfully loading the sample, select the sample in interface and edit the assay for the sample to be tested and execute testing. For specific information on ordering patient specimens, refer to the sample ordering section of the Analyzer Operating Instructions.

To ensure proper test performance, strictly adhere to Analyzer Operating Instructions.

Calibration

Traceability: This method has been standardized against the WHO 3rd International Standard 11/234.

Test of assay specific calibrators allows the detected relative light unit (RLU) values to adjust the master curve.

Recalibration is recommended as follows:

- Whenever a new lot of Reagent or Starter 1+2 is used.
- Every 28 days.
- The analyzer has been serviced.
- Control values lie outside the specified range.

Quality Control

Controls are recommended for the determination of quality control requirements for this assay and should be run in singlicate to monitor the assay performance. Refer to published guidelines for general quality control recommendations, for example Clinical and Laboratory Standards Institute (CLSI) Guideline C24 or other published guidelines¹⁶.

Quality control is recommended once per day of use, or in accordance with local regulations or accreditation requirements and your laboratory's quality control procedures, quality control could be performed by running the IgE assay:

- Whenever the kit is calibrated.
- Whenever a new lot of Starter 1+2 or Wash Concentrate is used.

Controls are only applicable with MAGLUMI and Biolumi systems and only used matching with the same top seven LOT numbers of corresponding reagents. For each target value and range refer to the label.

The performance of other controls should be evaluated for compatibility with this assay before they are used. Appropriate value ranges should be established for all quality control materials used.

Control values must lie within the specified range, whenever one of the controls lies outside the specified range, calibration should be repeated and controls retested. If control values lie repeatedly outside the predefined ranges after successful calibration, patient results must not be reported and take the following actions:

- Verify that the materials are not expired.
- Verify that required maintenance was performed.
- Verify that the assay was performed according to the package insert.
- If necessary, contact Snibe or our authorized distributors for assistance.

If the controls in kit are not enough for use, please order IgE (CLIA) Controls (REF: 160201498MT) from Snibe or our authorized distributors for more.

RESULTS

Calculation

The analyzer automatically calculates the IgE concentration in each sample by means of a calibration curve which is generated by a 2-point calibration master curve procedure. The results are expressed in IU/mL. For further information please refer to the Analyzer Operating Instructions.

Conversion factors: IU/mL x 2.40 = ng/mL.

Interpretation of Results

The expected range for the IgE assay was obtained by testing 1455 apparently healthy individuals in China, gave the following expected value:

| Age (years) | 95 th percentile (IU/mL) |
|-------------|-------------------------------------|
| <1 | 15 |
| 1-5 | 60 |
| 6-9 | 90 |
| 10-17 | 200 |
| Adults | 100 |

Results may differ between laboratories due to variations in population and test method. It is recommended that each laboratory establish its own reference interval.

LIMITATIONS

- Results should be used in conjunction with patient's medical history, clinical examination and other findings.
- If the IgE results are inconsistent with clinical evidence, additional testing is needed to confirm the result.
- Specimens from patients who have received preparations of mouse monoclonal antibodies for diagnosis or therapy may contain human anti-mouse antibodies (HAMA). Such specimens may show either falsely elevated or depressed values when tested with assay kits which employ mouse monoclonal antibodies^{17,18}. Additional information may be required for diagnosis.
- Heterophilic antibodies in human serum can react with reagent immunoglobulins, interfering with *in vitro* immunoassays. Patients routinely exposed to animals or animal serum products can be prone to this interference and anomalous values may be observed¹⁹.
- Bacterial contamination or heat inactivation of the specimens may affect the test results.
- Do not use samples from patients under treatment with omalizumab or similar drugs containing anti-IgE antibodies, to avoid misleadingly low IgE results.
- Increases in total serum IgE levels can be observed in many diverse conditions, from infection to atopy to primary immunodeficiency. The differentiation of atopy from immunodeficiency most often is made on a clinical basis, after taking findings from history, physical examination, and laboratory studies into consideration²⁰.

SPECIFIC PERFORMANCE CHARACTERISTICS

Representative performance data are provided in this section. Results obtained in individual laboratories may vary.

Precision

Precision was determined using the assay, samples and controls in a protocol (EP05-A3) of the CLSI (Clinical and Laboratory Standards Institute): duplicates at two independent runs per day for 5 days at three different sites using three lots of reagent kits (n = 180). The following results were obtained:

| Sample | Mean (IU/mL) (n=180) | Within-Run | | Between-Run | | Reproducibility | |
|---------------|-------------------------|------------|------|-------------|------|-----------------|------|
| | | SD (IU/mL) | %CV | SD (IU/mL) | %CV | SD (IU/mL) | %CV |
| Serum Pool 1 | 60.637 | 1.910 | 3.15 | 1.070 | 1.76 | 2.669 | 4.40 |
| Serum Pool 2 | 100.045 | 2.673 | 2.67 | 1.696 | 1.70 | 4.719 | 4.72 |
| Serum Pool 3 | 201.039 | 2.746 | 1.37 | 1.356 | 0.67 | 4.593 | 2.28 |
| Plasma Pool 1 | 59.857 | 1.911 | 3.19 | 0.979 | 1.64 | 2.435 | 4.07 |
| Plasma Pool 2 | 99.705 | 2.781 | 2.79 | 0.984 | 0.99 | 4.086 | 4.10 |
| Plasma Pool 3 | 200.055 | 3.031 | 1.52 | 1.685 | 0.84 | 6.879 | 3.44 |
| Control 1 | 44.729 | 1.412 | 3.16 | 0.908 | 2.03 | 1.977 | 4.42 |
| Control 2 | 183.802 | 4.138 | 2.25 | 1.300 | 0.71 | 6.720 | 3.66 |

Linear Range

0.500-3200 IU/mL (defined by the Limit of Quantitation and the maximum of the master curve).

Reportable Interval

0.300-9600 IU/mL (defined by the Limit of Detection and the maximum of the master curve×Recommended Dilution Ratio).

Analytical Sensitivity

Limit of Blank (LoB) =0.100 IU/mL.

Limit of Detection (LoD) =0.300 IU/mL.

Limit of Quantitation (LoQ) =0.500 IU/mL.

Analytical Specificity

Interference

Interference was determined using the assay, three samples containing different concentrations of analyte were spiked with potential endogenous and exogenous interferents in a protocol (EP7-A2) of the CLSI. The measurement deviation of the interference substance is within ±10%. The following results were obtained:

| Interference | No interference up to | Interference | No interference up to |
|--------------|-----------------------|---------------------|-----------------------|
| Hemoglobin | 1000 mg/dL | Rheumatoid factor | 2000 IU/mL |
| Intralipid | 5000 mg/dL | Human albumin | 12 g/dL |
| Bilirubin | 200 mg/dL | K2-EDTA | 22.75 µmol/mL |
| HAMA | 40 ng/mL | Heparin sodium salt | 80 IU/mL |
| ANA | 398 AU/mL | Biotin | 0.5 mg/dL |

Cross-Reactivity

Cross-reactivity was determined using the assay, three samples containing different concentrations of analyte were spiked with potential cross-reactant in a protocol (EP7-A2) of the CLSI. The measurement deviation of the interference substance is within ±10%. The following results were obtained:

| Cross-reactant | No interference up to | Cross-reactant | No interference up to |
|----------------|-----------------------|----------------|-----------------------|
| IgA | 4800 mg/dL | IgM | 2500 mg/dL |
| IgG | 8000 mg/dL | IgD | 1100 mg/dL |

High-Dose Hook

No high-dose hook effect was seen for IgE concentrations up to 12000 IU/mL.

Method Comparison

A comparison of the IgE assay with a commercially available immunoassay, gave the following correlations (IU/mL):

Number of samples measured: 135

Passing-Bablok: y=0.9950x+0.0998, r=0.976.

The clinical specimen concentrations were between 0.562 and 3192 IU/mL.

REFERENCES

1. Amarasekera M. Immunoglobulin E in health and disease[J]. Asia Pacific Allergy, 2011, 1(1): 12–15.
2. Sutton B J, Davies A M. Structure and dynamics of IgE–receptor interactions: FcεRI and CD23/FcεRII[J]. Immunological Reviews, 2015, 268(1): 222–235.
3. Wüthrich B, Schindler C, Medici T C, et al. IgE Levels, Atopy Markers and Hay Fever in Relation to Age, Sex and Smoking Status in a Normal Adult Swiss Population[J]. International Archives of Allergy and Immunology, Karger Publishers, 1996, 111(4): 396–402.
4. Kessel A, Helou W, Bamberger E, et al. Elevated Serum Total IgE – A Potential Marker for Severe Chronic Urticaria[J]. International Archives of Allergy and Immunology, Karger Publishers, 2010, 153(3): 288–293.
5. Mukai K, Tsai M, Starkl P, et al. IgE and mast cells in host defense against parasites and venoms[J]. Seminars in Immunopathology, 2016, 38(5): 581–603.
6. Valenta R, Mittermann I, Werfel T, et al. Linking allergy to autoimmune disease[J]. Trends in Immunology, 2009, 30(3): 109–116.
7. Marone G, Florio G, Triggiani M, et al. Mechanisms of IgE Elevation in HIV-1 Infection[J]. Critical Reviews & Trade; in Immunology, Begel House Inc., 2000, 20(6).
8. Denning D W. Chronic forms of pulmonary aspergillosis[J]. Clinical Microbiology and Infection, Elsevier, 2001, 7: 25–31.
9. Patterson K C, Strek M E. Diagnosis and Treatment of Pulmonary Aspergillosis Syndromes[J]. Chest, 2014, 146(5): 1358–1368.
10. Sabra A, Bellanti J A, Rais J M, et al. IgE and non-IgE food allergy[J]. Annals of Allergy, Asthma & Immunology, 2003, 90(6, Supplement): 71–76.
11. Goklemez S, Pirsil F, Curtis L M, et al. Clinical Significance of IgE in a Large Cohort of Patients with Moderate or Severe Chronic Graft-versus-Host Disease[J]. American journal of hematology, 2017, 92(8): E162–E164.
12. Polacek V, Jira M, Fara M, et al. Immunoglobulin E (IgE) in patients with severe burns[J]. Burns, 1987, 13(6): 458–461.
13. Pandey S, Kyle R A. Unusual myelomas: a review of IgD and IgE variants[J]. Oncology (Williston Park, N.Y.), 2013, 27(8): 798–803.
14. González-Quintela A, Alende M R, Lojo S, et al. Total Serum IgE Levels in Chronic Hepatitis C: Influence of Interferon Alpha Therapy[J]. International Archives of Allergy and Immunology, Karger Publishers, 2001, 125(2): 176–181.
15. González-Quintela A, Vidal C, Gude F. Alcohol-induced alterations in serum immunoglobulin e (IgE) levels in human subjects[J]. Frontiers in Bioscience: A Journal and Virtual Library, 2002, 7: e234-244.
16. CLSI. Statistical Quality Control for Quantitative Measurement Procedures: Principles and Definitions. 4th ed. CLSI guideline C24. Wayne, PA: Clinical and Laboratory Standards Institute; 2016.
17. Robert W. Schroff, Kenneth A. Foon, Shannon M. Beatty, et al. Human Anti-Murine Immunoglobulin Responses in Patients Receiving Monoclonal Antibody Therapy[J]. Cancer Research, 1985, 45(2):879-85.
18. Primus F J, Kelley E A, Hansen H J, et al. "Sandwich"-type immunoassay of carcinoembryonic antigen in patients receiving murine monoclonal antibodies for diagnosis and therapy[J]. Clinical Chemistry, 1988, 34(2):261-264.
19. Boscato L M, Stuart M C. Heterophilic antibodies: a problem for all immunoassays. Clin Chem 1988;34(1):27-33.
20. Pien G C, Orange J S. Evaluation and clinical interpretation of hypergammaglobulinemia E: differentiating atopy from immunodeficiency[J]. Annals of Allergy, Asthma & Immunology: Official Publication of the American College of Allergy, Asthma, & Immunology, 2008, 100(4): 392–395.

SYMBOLS EXPLANATIONS

| | | | |
|--|---|--|---|
| | Consult instructions for use | | Manufacturer |
| | Temperature limit (Store at 2-8°C) | | Use-by date |
| | Contains sufficient for <n> tests | | Keep away from sunlight |
| | This way up | | Authorized representative in the European Community |
| | <i>In vitro</i> diagnostic medical device | | Kit component |
| | Catalogue number | | Batch code |
| | CE marking with notified body ID number | | |

MAGLUMI® and Biolumi® are trademarks of Snibe. All other product names and trademarks are the property of their respective owners.

Shenzhen New Industries Biomedical Engineering Co., Ltd.
No.23, Jinxiu East Road, Pingshan District, 518122 Shenzhen, P.R. China
Tel: +86-755-21536601 Fax: +86-755-28292740

Shanghai International Holding Corp. GmbH (Europe)
Eiffestrasse 80, 20537 Hamburg, Germany
Tel: +49-40-2513175 Fax: +49-40-255726