



BAYMED



DECLARATION OF CONFORMITY (UYGUNLUK BEYANI)

MANUFACTURER: BAYTEKS TEKNİK TEKSTİL SAN. VE TİC. A.Ş

Organize Sanayi Bölgesi 19 Nolu Cad. No:11/2 MERKEZ/KİLİS

Tel: 0342 337 30 30

Fax: 0342 337 30 35

This declaration of conformity has been prepared in MDR 2017/745 Part II, Article 19, Annex IV; The document is arranged in the home page program for all users.

UDI -DI	: See the Sterile Products table.
PRODUCT	: Sterile Disposable Surgical Gowns, Sterile Disposable Surgical Drapes, Sterile Disposable Surgical Packs.
NOTIFIED BODY	: UDEM Adriatic d.o.o R3 Radnička cesta 54/ R3, Green Gold Centar, IV. kat, 10000, Zagreb, Hırvatistan
ID NO	2696
CND(EMDN) CODE	: See the Sterile Products table.
SRN NUMBER	: TR-MF-000024488
CLASSIFICATION	: Sınıf I Steril -(MDR 2017/745 - EKVIII/Bölüm III - Madde 4 - Kural 1)
SCOPE	: Sterile Disposable Surgical Gowns, Sterile Disposable Surgical Drapes, Sterile Disposable Surgical Packs.
ADDITIONAL	: EK XI A
APPLIED STANDARDS	: TS EN ISO 13485:2016, ISO 14971:2020, EN ISO 11135:2014, TS EN ISO 15223-1:2021, TS EN ISO 11737-1:2018, EN ISO 11737-2:2020, TS EN ISO 14644-1:2021, TS EN ISO 14644-2:2021, TS EN ISO 10993-1:2021, TS EN ISO 11607-1:2020, TS EN ISO 11607-2:2020, TS EN ISO 11138-1:2017, TS EN ISO 11140-1:2015

The directive for our product is the Council Directive MDR 2017/745 ECC for all versions of medical devices. The Manufacturer of the product, Bayteks Teknik Tekstil San. ve Tic. A.Ş, is responsible for the requirements of this council directive. Our products are not medical devices that contains human blood derivatives, animal products, animal skin, tissues, or blood derivatives or phthalates.

STERILE PRODUCT			
PRODUCT NAME	BASIC UDI	MODEL NO	EMDN
INSTRUMENT TABLE DRAPE 50X70 CM	86817441ST30100000976BN	SD-00200-38	T020199
TUR DRAPE 200X250 CM	86817441ST30100003663BK	SD-04803-09	T020102
ANGIOGRAPHY PACK	86817441ST30100001938BM	SP-05006-104	T0202
LAPAROSCOPY DRAPE	86817441ST30100004140AM	SD-04205-24	T0202
FOOT DRAPE 175X175 CM	86817441ST30100004111AE	SD-04202-02	T030199
ANGIOGRAPHY PACK	86817441ST30100004047AW	SP-05004-89	T0202
ENDOVASCULAR PACK	86817441ST30100001665B9	SP-05006-161	T0202
TOWEL – 40X40 CM	86817441ST30100003625BB	SU-40004-01	T030199
RADIAL ANGIOGRAPHY PACK	86817441ST30100002752BC	SP-05004-56	T0202
CRANIOTOMY PACK	86817441ST30100002329AV	SP-01005-35	T0202
SPINAL DRAPE 90X75 CM	86817441ST30100003647BM	SD-05211-55	T020102



BAYMED

CE

The products listed in the list above and their contents are classified Class 1 Sterile products. These products, their content, and their accessories do not take part in any other class. We herewith declare that the above mentioned products conforms general requirements of the Council Directive MDR 2017/745 ECC for all versions of Medical Device Directive .

Applied Directives

Medical Device Directive MDR 2017/745 ECC ANNEX XI A ALL VERSIONS.

DATE OF ISSUE : 25.08.2024

REV. NO. : 1

Name and Surname : Yaşar ÇALIŞKAN

Position : Sales Chief

Date of signature : 20.01.2025

Place of Signature : Kilis - Türkiye

