



Critical Care Blood Gas Analyser

A technology evolution in critical care testing




New Technologies Simplify Use and Offer Additional Tests

Stat Profile Prime Plus is a comprehensive, whole blood critical care analyser that combines blood gases, electrolytes, metabolites, co-oximetry, and 32 calculated results in a simple, compact device. Prime Plus combines maintenance-free, replaceable cartridge technology for sensors and reagents with patented, new, maintenance-free, and non-lysing whole blood co-oximetry technology.

Prime Plus results are produced very rapidly—a complete test menu panel in about one minute—and are combined with bidirectional connectivity and a powerful onboard data management system.

Nova MicroSensor Card™ Technology

Most comprehensive critical care menu



PO₂ PCO₂ pH Hct tHb Na Cl K TCO₂
iCa iMg Glu Lac SO₂ O₂Hb COHb
MetHb HHb tBil HbF

- All Prime Plus biosensors use proven Nova technology in a miniaturised, maintenance-free sensor card format.
- Nova MicroSensor cards combine all 22 whole blood assays including co-oximetry.



Important New Assays

Urea (BUN), Creatinine and eGFR

Over 50% of patients admitted to the intensive care unit (ICU) will develop some stage of acute kidney injury (AKI).¹ Prime Plus is the only blood gas analyser to provide optional

whole blood urea (BUN) and creatinine (plus eGFR) tests for rapid assessment of kidney function.

Ionised magnesium (iMg)

Disruptions in the balance of iMg, Na, K, and iCa can cause cardiac arrhythmias, reduced cardiac contraction, and cardiac arrest. Prime Plus is the only blood gas analyser to provide a comprehensive profile of electrolytes including iMg.



1. Mandelbaum T et al. Outcome of critically ill patients with acute kidney injury using the AKIN criteria. *Crit Care Med* 2011;39:2259-2264.

New Disposable CO-Oximeter Eliminates Maintenance

CO-Oximetry test menu

O₂Hb COHb MetHb HHb tHb HbF tBil

Prime Plus incorporates a new, patented* multi-wavelength optical system that scans a continuous spectrum of optical wavelengths to produce a comprehensive co-oximetry panel result without lysing the sample. The optical components in contact with blood are contained in the disposable sensor card, which is replaced every 16 days.


- Cleaning and deproteinising are completely eliminated.
- Lysing and all its required mechanical components are eliminated, along with lysing and deproteinising reagents. This new technology reduces maintenance costs and improves reliability.

Fast Stat Results

Prime Plus's exceptional throughput easily handles the high sample workload of a busy critical care setting. Prime Plus delivers a 22-test critical care profile in about one minute. Other analysers can require up to four minutes, even with fewer tests reported.

Clot Protection


Prime Plus's unique Clot Block™ sample flow path protects sensor cards from blood clot blockages.



* Patent numbers: 95350531, 9933411B2

Individual Sensors and Calibrators Maximise Uptime

Individual sensor cards and calibrator cartridges offer a significant benefit in analyser uptime compared to combined sensor/calibrator cartridge systems.



MicroSensor cards have fastest replacement time

MicroSensor cards can be replaced and calibrated in 45 minutes. Other combined systems usually take one hour to calibrate and remain unstable with drift and frequent re-calibrations for two hours or longer.

Calibrator cartridges replaced in seconds

Calibrator and quality control (QC) cartridges are immediately ready to use and easily replaced in seconds. Replacing only a calibrator cartridge significantly reduces analyser downtime because it has no warm-up time, compared to the over two-hour wait for other combined systems.

Individual Sensors and Calibrators Lower Costs

Individual sensor cards and calibrator cartridges are a low cost alternative to the inflexibility and waste of combined sensor/calibrator cartridge systems. For example, an analyser in a high patient workload setting requires fewer sensor cards than calibrators, and a low volume workload setting requires the reverse. In both cases, Prime Plus eliminates waste and reduces overall consumable costs by using the full life of each card and cartridge.

Bidirectional Connectivity and Point-of-Care Management

NovaNet bidirectional middleware for all Nova-connected devices

NovaNet is a robust, economical solution for bidirectional interface of all Nova point-of-care (POC) devices to the LIS/HIS/EMR. NovaNet ensures timely, accurate capture of Nova analyser POC test results for clinicians and managers to retrieve wherever and whenever needed.




Prime Plus

Test Orders, Demographics,
Admissions, Discharges,
Transfers

Patient Results

NovaNet



LIS/HIS/EMR

- NovaNet provides bidirectional connectivity to transfer patient test orders, demographics, admissions, discharges, and data to Prime Plus analysers.
- POC data is captured seamlessly for medical record review, retention, and billing.
- POC patient and QC results transmissions are confirmed with acknowledgements. NovaNet flags and reports any results that fail to transmit.
- NovaNet's industry standard HL7, ASTM, or POCT01-A2 formats are easily implemented with LIS/HIS/EMR systems.

No third party middleware connectivity costs

NovaNet eliminates the cost of third party middleware to connect Nova analysers to the LIS/HIS/EMR. For hospitals that already have third party middleware connectivity, NovaNet provides supplemental remote review and remote control capabilities for connected Nova analysers.


Management reports for patient and QC data, devices, and operators

NovaNet is specifically designed to meet POC programme management and regulatory requirements by capturing patient testing, QC compliance, and operator records. A large library of reports is available including:

- Patient abnormal/critical results
- Patient report exceptions
- Daily QC
- QC cumulative statistics
- Sample comments
- Operator certifications
- Corrective actions
- Calibrator and sensor replacements

Remote Review and Remote Control

NovaNet provides information on analyser connectivity, calibration, QC, reagent, and sensor status. The dashboard allows POC coordinators to review the status of remote analysers and correct for calibration or QC needs.




Dashboard review

Individuals with password privileges can view a dashboard of all connected devices from anywhere on the network.

Remote control

Key operators can remotely perform essential analyser functions such as:

- Initiate calibration and QC cycles
- Upload or edit set-up parameters
- Assign, certify, or remove operators and privilege levels



High Level Data Encryption and Network Security

As part of Nova's cybersecurity and protected health information (PHI) risk protection, Prime Plus analysers and NovaNet middleware comply with U.S. Homeland Security and U.S. Food and Drug Administration (FDA) cybersecurity risk mitigation measures, and U.S. HIPAA PHI security measures.¹ Utilising high level proprietary and SSL encryption, the following capabilities can be enabled for Prime Plus analysers and NovaNet middleware:

- Encryption of the entire hard drive and all PHI data held in Prime Plus and NovaNet databases
- Encryption of all PHI traveling between Prime Plus, NovaNet, and the LIS or middleware
- Lockdown on access to Windows, protecting the Prime Plus and NovaNet operating systems and the hospital network from malware intrusion

These features provide the highest level of analyser, PHI, and network security of any blood gas analyser.



1. Health Insurance Portability and Accountability Act

Automated, True Liquid QC


Liquid QC provides the only reliable test of an analyser

U.S. federal government and many international government regulations have eliminated electronic equivalent QC and are requiring true liquid QC.¹

Automated QC complies with U.S. CLIA, German RiLiBAK, and other international QC requirements

QC cartridges contain a 30-day supply of liquid QC material. Controls run automatically at user-selected intervals. Prime Plus quality controls:

- Comprise a matrix similar to patient samples.
- Are analysed as patient samples.
- Follow the same sample pathway as patient samples, from sample probe to waste container.
- Challenge all analytical phases of testing.
- Challenge multiple levels of each analyte.




QC statistics and reports are automatically maintained and easily accessed.

Saves time and labour

Maintaining QC is one of the most time consuming aspects of critical care testing. Prime Plus's fully automated, onboard liquid QC saves hours of time each week compared to individualized QC plans (IQCPs) and manually running controls.

Supplemental Quality Monitoring (SQM)

Prime Plus provides an automated electronic quality monitoring supplement to liquid QC. SQM continuously monitors the status and performance of all analytical components (including sensors, reagents, calibrators, sample integrity, software, and electronics), providing real-time, sample-to-sample assurance of correct performance.




1. Centers for Medicare and Medicaid Services, Center for Clinical Standards and Quality/Survey and Certification Group. Policy clarification on acceptable control materials used when quality control (QC) is performed in laboratories. Baltimore, MD: CMS, April 8, 2016.

Simple, Fast Operation

25 cm wide, high definition, colour touchscreen operation

The large colour touchscreen is easy to read and operate with intuitive prompts.



Three simple steps to initiate a full 22-test profile


1. Press 
2. Scan or enter patient ID
3. Press 

Integrated barcode scanner


The 1D/2D barcode scanner, conveniently located within the sample port allows for fast, error-free entry of operator and patient IDs. The optional, wireless, external barcode scanner also allows for positive patient ID, further eliminating pre-analytical error.

Safe Operation

The unique safety design of the sample port protects the user from accidental contact with the analyser probe.



Syringes can be docked and sampled with hands-free operation.



Hands-free capillary sampling can be performed without adapters.



Samples can be aspirated directly from tubes. Sample transfer to a syringe or capillary is eliminated.



QC proficiency ampoules can be sampled without adapters.

Stat Profile Prime Plus™ Specifications

Critical Care Tests

pH	Direct ISE
PCO ₂	Severinghaus
PO ₂	Amperometric
SO ₂ %	Optical, reflectance
Haematocrit	Conductivity/Na ⁺ correction
Na ⁺	Direct ISE
K ⁺	Direct ISE
Cl ⁻	Direct ISE
TCO ₂	Direct ISE
Ca ⁺⁺	Direct ISE
Mg ⁺⁺	Direct ISE
Glucose	Enzyme/Amperometric
Lactate	Enzyme/Amperometric
Urea (BUN)	Enzyme/Amperometric
Creatinine	Enzyme/Amperometric

Calculated Tests

eGFR	A-aDO ₂	Ca ⁺⁺ /Mg ⁺⁺ Ratio
HCO ₃ ⁻	a/A	Normalized Ca ⁺⁺
TCO ₂	PO ₂ /FIO ₂	Normalized Mg ⁺⁺
BE-ecf	Anion Gap	Osmolality
BE-b	SBC	Haemoglobin
A	Base Excess	O ₂ Saturation
pH/PCO ₂ /PO ₂	Corrected to Patient Temperature	
Respiratory Index	(If % FIO ₂ value entered)	
Actual Bicarbonate		
Standard Bicarbonate		

CO-Oximetry Tests

HHb, deoxyhaemoglobin	O ₂ Hb, oxyhaemoglobin
MetHb, methaemoglobin	COHb, carboxyhaemoglobin
tHb, total haemoglobin	SO ₂ %, oxygen saturation
tBil, total bilirubin	HbF, fetal haemoglobin

Special Calculated Tests (CO-Oximetry Required)

Tests	Resolution
A-v DO ₂	0.1 mmHg (0.01 kPa)
CaO ₂	0.1 mL/dL (0.01 kPa)
CcO ₂	0.1 mL/dL (0.01 kPa)
P50	0.1 mmHg (0.01 kPa)
C(a-v)O ₂	0.1 mmHg (0.001 kPa)
CvO ₂	0.1 mmHg (0.001 kPa)
Qsp/Qt	0.1 mmHg (0.001 kPa)
O ₂ Ct	0.1 mL/dL (0.01 mL/L)
O ₂ Cap	0.1 mL/dL (0.01 mL/L)

Methodology

Complete Management Reports

- Calibration Report
- Cartridge Log Report
- Daily Sample Log Report
- Edit Log Report
- Error Log Report
- Maintenance Log Report
- Operator Setup Report
- Patient Report
- Levey-Jennings QC Report
- QC Corrective Actions Report
- QC Data Report
- QC Statistics Report
- QC Setup Report
- Sample Audit Log Report

Monitored Interferences

sHb, sulphaemoglobin (Measured; user alerted if abnormal, > 1.5%)

Measurement Ranges

pH	6.5 - 8 (H+: 316.2 - 10 nmol/L)
PCO ₂	3 - 200 mmHg (0.4 - 26.7 kPa)
TCO ₂	5 - 70 mmol/L (90 - 1260 mg/dL)
PO ₂	5 - 765 mmHg (0.66 - 102 kPa)
Hct	12 - 70%
Na ⁺	80 - 200 mmol/L
K ⁺	1 - 20 mmol/L
Cl ⁻	50 - 200 mmol/L
Ca ⁺⁺	0.1 - 2.7 mmol/L (0.4 - 10.8 mg/dL)
Mg ⁺⁺	0.1 - 1.5 mmol/L (.24 - 3.65 mg/dL)
Glucose	0.8 - 28 mmol/L (15 - 500 mg/dL)
Lactate	0.3 - 20 mmol/L (2.7 - 180.2 mg/dL)
Urea (BUN)	0.17 - 5.5 mmol/L (3 - 100 mg/dL)
Creatinine	10 - 660 µmol/L (0.2 - 12 mg/dL)
HHb	0 - 33% (0 - 0.33)
O ₂ Hb	0 - 100% (0 - 1)
MetHb	0 - 80% (0 - 0.8)
COHb	0 - 60% (0 - 0.6)
SO ₂ %	30 to 100%
O ₂ Ct	495.04 - 2952.56 µmol/L (5.6 - 33.4 mg/dL)
O ₂ Cap	495.04 - 2952.56 µmol/L (5.6 - 33.4 mg/dL)
tBil	44.2 - 3094 µmol/L (0.5 - 35 mg/dL)
HbF	0 - 92%
tHb	5 - 25 g/dL (50 - 250 g/L)
sHb	Alert > 1.5%
BarP	400 - 800 mmHg (53.3 - 106.7 kPa)

Other Features

Full colour, 25-cm touchscreen, multilingual, QC statistics, onboard data management, automatic sampler, integrated capillary adapter, optional barcode scanner, QC data storage, optional mobile cart with UPS

Sample Volume

MicroSensor Card
60 µL

Operating Temperature Range

15°C-32°C

Physical Specifications

Height: 45.7 cm (18.2 in) Width: 35.6 cm (14.2 in)
Depth: 39.1 cm (15.5 in) Weight: 15.88 kg (35 lb) without reagent packs

Electrical Power Requirement

< 90 Watts

Printer

Onboard thermal printer

CE Marked, FDA Cleared

Calibration

Fully automatic two-point calibration every 2 hours; user-selectable single-point calibration every 45 minutes or with each sample. Manual calibration initiated at any time.

Acceptable Samples

Whole blood (heparinized), arterial, venous, mixed venous, capillary, dialysate. Sample draw requirement is 135 µL.

Communication Protocols

ASTM, HL7, or POCT01-A2 connectivity formats

POC Monitoring Systems

Nova's whole blood meters and test strips provide accurate results by utilising Multi-Well™ biosensor technology, which measures and corrects for interferences such as haematocrit, paracetamol, ascorbic acid, and uric acid, which can cause erroneous results on other handheld, whole blood meters.

Other features include:

- Easy, handheld operation
- Samples as small as 0.6 microlitres
- Results as fast as 6 seconds
- No calibration coding
- Single connectivity solution
- Choice of hospital connectivity or Xpress meter



StatStrip® Glucose/
Ketone Meter



StatStrip® Xpress2
Glucose/Ketone
Meter



StatSensor®
Creatinine Meter



StatSensor® Xpress
Creatinine Meter



StatStrip®
Lactate Meter



StatStrip® Xpress
Lactate Meter

Compact Size for POC Use

Dimensions for Prime Plus, including onboard co-oximetry and built-in bidirectional connectivity:



Weight:

15.88 kg (35 lb) without calibrator cartridges
19.30 kg (42.5 lb) with calibrator cartridges

nova
biomedical

novabiomedical.com

Specifications subject to change without notice.

Nova Biomedical Headquarters: 200 Prospect St., Waltham, MA 02453-3465 U.S.A. +1-781-894-0800 800-458-5813 FAX: +1-781-894-5915 Int'l FAX: +1-781-899-0417 e-mail: info@novabio.com
Nova Biomedical Brasil: Rua Massena, 107, Jardim Canadá, Nova Lima - MG, CEP: 34007-746 Brasil, TEL: +55-31-3360-2500, email: sac@novabiomedical.com.br
Nova Biomedical Canada: Ltd. 17 - 2900 Argente Road, Mississauga, Ontario L5N 7X9 Canada TEL: +1-905-567-7700 800-263-5999 FAX: +1-905-567-5496 e-mail: info@novabio.ca
Nova Biomedical France: Parc Technopolis - Bât. Sigma 3 Avenue du Canada 91940 Les Ulis Courtaboeuf, France TEL: +33-1-64 86 11 74 FAX: +33-1-64 46 24 03 e-mail: info@novabiomedical.fr
Nova Biomedical GmbH, Deutschland: Hessestr. 13 A, Geb. G, 64546 Mörfelden-Walldorf, Germany TEL: +49-6105 4505-0 FAX: +49-6105 4505-37 e-mail: info@novabiomedical.de
Nova Biomedical Iberia: Av. Corts Catalanes 9-11 08173 Sant Cugat Barcelona, Spain: +1-781-894-0800 FAX: +1-781-894-5915 Int'l FAX: +1-781-899-0417 e-mail: info@novabio.com
Nova Biomedical Italia Srl: Via Coma 19 20020 LAINATE (MI) TEL: +39-02-87070041 FAX: +39-02-87071482 e-mail: info@novabiomedical.it
Nova Biomedical K.K.: 43MT Building-7F, 13-16 Nishi 3-chome, Minato-ku, Tokyo 108-0073, Japan TEL: +81-3-5418-4141 FAX: +81-3-5418-4676 e-mail: novabio@novabiomedical.co.jp
Nova Biomedical Switzerland GmbH: Turmstrasse 18, 6312 Steinhausen, Switzerland TEL: +41-41-521-6655 FAX: +41-41-521-6656 e-mail: info@novabiomedical.ch
Nova Biomedical U.K.: Innovation House, Aston Lane South, Runcorn, Cheshire WA7 3FY United Kingdom TEL: +44-1928 704040 FAX: +44-1928 796792 e-mail: sales@novabiomedical.co.uk

REF 57845

R_X Only IVD 20°C 

Stat Profile Prime Plus® Chemistry Controls 4 and 5

Controles de química 4 y 5 Stat Profile Prime Plus®, Contrôles 4 et 5 chimie Stat Profile Prime Plus®, Stat Profile Prime Plus® Blutchemiekontrollen 4 und 5, Υλικό ελέγχου Χημείας 4 και 5 Stat Profile Prime Plus®,
Controlli chimici 4 e 5 Stat Profile Prime Plus®, Controles de Química Stat Profile Prime Plus® 4 e 5, Stat Profile Prime Plus® Kimya Kontrolleri 4 ve 5, Stat Profile Prime Plus® 4-es és 5-ös kémiai kontroll,
5 - 1 4 בקרת כימיה Stat Profile Prime Plus®, Stat Profile Prime Plus® 生化学検査用コントロール 4 および 5, Stat Profile Prime Plus® 화학 조절제 4 및 5, Stat Profile Prime Plus® 化学对照溶液 4 和 5

CONTROL 4 5

LOT 25260035 2026-09-15



EN

Product Description
Aspirator quality control material for monitoring the performance of Na⁺, K⁺, Cl⁻, iCa, iMg, Glucose, Lactate, BUN (urea), and Creatinine for use with Stat Profile Prime Plus Analyzers ONLY.
Formulation: 2 Levels

Intended Use
Intended for in vitro diagnostic use by healthcare professionals for monitoring the performance of the Stat Profile Prime Plus Analyzers.

Methodology
Refer to Stat Profile Prime Plus Analyzer Instruction for Use Manual for Methodology and Principles.

Composition
Controls are buffered solutions containing known concentrations of Na⁺, K⁺, Cl⁻, iCa, iMg, Glucose, Lactate, BUN (urea), Creatinine, and creatinins. Each ampoule contains a minimum volume of 1.7 mL. Controls contain no constituents of human origin. However, good laboratory practices should be followed during handling of these materials. (REF: NCCLS DOCUMENT M29-T21)

Warnings and Precautions
Controls must be stored at 2-8°C (37-47°F). Once opened, analyze within one minute and discard the unused portion in accordance with local guidelines. Refer to Stat Profile Prime Plus Analyzers Instructions for Use Manual for complete information. Intended for in vitro diagnostic use. Follow standard practices for handling laboratory reagents.

Storage
Store at 2-8°C (37-47°F). DO NOT FREEZE. Each ampoule has a Lot Number and Expiration Date printed on the label.

Directions for Use
Store controls at 2-8°C (37-47°F) until ready for use. Ensure control is at room temperature prior to use. Shake ampoule well before opening. After open ampoule (product is liquid with glass in cap), close ampoule, invert control within one minute. Discard the unused portion in accordance with local guidelines. Verify that the Lot Number on the Expected Ranges Table corresponds to the Lot Number on the ampoule. Refer to Stat Profile Prime Plus Analyzer Instructions for Use Manual for complete instructions.

Limitations
The Expected Range values are specific for instruments and controls manufactured by Nova Biomedical.

Traceability of Standards
Analyses are traced to NIST Standard reference materials.

Reference Intervals
Concentrations are formulated at normal and abnormal expected values in patient blood. The expected clinical range of these values in patient blood is referenced to Tietz, MD et al. 1993 Textbook of Clinical Chemistry, W.B. Saunders Co. Users may wish to determine Mean Values and Expected Ranges in their own laboratory.

Expected Ranges
The expected range for each parameter was determined at Nova Biomedical using replicate determinations on Nova analyzers. The expected range includes the maximum deviations from the Mean Value that may be expected under laboratory conditions for instruments operating within specifications. Refer to Expected Ranges Table.

NCCLS Document M29-T2
How to Define and Determine Reference Intervals in the clinical laboratory; approved guidelines-second edition, NCCLS C28-A2, Volume 20, Number 12.

LOT



CONTROL 4 25258033 2026-09-15
CONTROL 5 25258034 2026-09-15

Expected Ranges, Rangos esperados, Plages attendues, Erwartungsbereiche, Αναμενόμενα εύρος, Intervalli previsti, Intervalos previstos, Beklenen Aralıklar, Várt tartományok, הצפויים הטווחים, 予測範囲, 예상 범위, 预期范围值

		CONTROL 4		CONTROL 5	
		min - x - max		min - x - max	
Na ⁺	mmol/L	134.3 - 138.3 - 142.3		107.6 - 111.6 - 115.6	
K ⁺	mmol/L	3.73 - 3.98 - 4.23		6.04 - 6.34 - 6.64	
Cl ⁻	mmol/L	118.9 - 123.4 - 127.9		92.1 - 96.6 - 101.1	
iCa	mmol/L	1.00 - 1.08 - 1.16		1.46 - 1.58 - 1.70	
iCa	mg/dL	4.0 - 4.3 - 4.6		5.9 - 6.3 - 6.8	
iMg	mmol/L	0.54 - 0.61 - 0.68		0.93 - 1.08 - 1.23	
iMg	mg/dL	1.3 - 1.5 - 1.7		2.3 - 2.6 - 3.0	
Glu	mg/dL	71 - 79 - 87		267 - 292 - 317	
Glu	mmol/L	3.9 - 4.4 - 4.8		14.8 - 16.2 - 17.6	
Lac	mmol/L	1.4 - 1.7 - 2.0		5.8 - 6.5 - 7.2	
Lac	mg/dL	12.5 - 15.1 - 17.8		51.7 - 57.9 - 64.1	
BUN	mg/dL	11 - 16 - 21		42 - 52 - 62	
BUN	mmol/L	3.9 - 5.7 - 7.5		15.0 - 18.6 - 22.1	
Urea	mg/dL	23.6 - 34.3 - 45.0		90.1 - 111.5 - 133.0	
Urea	mmol/L	3.9 - 5.7 - 7.5		15.0 - 18.6 - 22.1	
Creatinine	mg/dL	0.80 - 1.10 - 1.40		6.30 - 7.30 - 8.30	
Creatinine	mmol/L	0.07 - 0.10 - 0.12		0.56 - 0.65 - 0.73	
Creatinine	μmol/L	71 - 97 - 124		557 - 645 - 734	

ES

Descripción del producto
Material accesorio de control de calidad para supervisar el desempeño de Na⁺, K⁺, Cl⁻, iCa, iMg, glucosa, lactato, BUN (urea) y creatinina, para uso UNICAMENTE con analizadores Stat Profile Prime Plus.
Formulación en 2 niveles

Indicaciones de uso
Destinado al uso diagnóstico in vitro por parte de profesionales de la salud para supervisar el desempeño de los analizadores Stat Profile Prime Plus.

Metodología
Para conocer la metodología y los principios de prueba, consulte el Manual de instrucciones de uso del analizador Stat Profile Prime Plus.

Composición
Los controles son soluciones tamponadas que contienen concentraciones conocidas de Na⁺, K⁺, Cl⁻, iCa, iMg, glucosa, lactato, BUN (urea), creatinina y creatininas. Cada ampolla contiene un volumen mínimo de 1.7 mL. Los controles no contienen ninguna sustancia de origen humano. Sin embargo, se deben seguir las buenas prácticas de laboratorio al manejar estos materiales. (REF: NCCLS DOCUMENT M29-T21)

Advertencias y precauciones
Los controles deben almacenarse a 2-8°C (37-47°F). Una vez abiertos, revisar el análisis de inmediato y desechar la parte que no cubra de inmediato con los tapones locales. Después de la parte no utilizada de acuerdo con las normas locales. Verifique que el número de lote que figura en la etiqueta de rangos esperados coincida con el número de lote impreso en la ampolla. Para conocer las instrucciones completas, consulte el Manual de instrucciones de uso del analizador Stat Profile Prime Plus.

Almacenamiento
Conservar a 2-8°C (37-47°F). NO CONGELAR. Cada ampolla tiene un número de lote y la fecha de vencimiento impresa en la etiqueta.

Instrucciones de uso
Conservar los controles a 2-8°C (37-47°F) hasta que estén listos para usar. Asegurarse de que el control esté a temperatura ambiente antes de usar. Agitar bien la ampolla y dejarla lista para su uso antes de romper los sellos. Una vez abierta, analizar la muestra en el control en un minuto. Después de la parte no utilizada de acuerdo con las normas locales. Verifique que el número de lote que figura en la etiqueta de rangos esperados coincida con el número de lote impreso en la ampolla. Para conocer las instrucciones completas, consulte el Manual de instrucciones de uso del analizador Stat Profile Prime Plus.

Limitaciones
Los valores de rangos esperados son específicos para los instrumentos y controles fabricados por Nova Biomedical.

Compatibilidad de reactivos
Analice únicamente a los materiales de referencia estándar del NIST.

Intervalos de referencia
Las concentraciones están formuladas como valores esperados normales y anormales en la sangre del paciente. Se puede consultar el rango clínico esperado de estos valores en la sangre del paciente en Tietz, MD et al. 1993 Textbook of Clinical Chemistry, W.B. Saunders Co. Los usuarios pueden desear determinar valores medios y rangos esperados en su propio laboratorio.

Alcance esperado
El tiempo esperado para cada parámetro ha sido determinado en Nova Biomedical usando determinaciones replicadas en analizadores Nova. El rango esperado indica las desviaciones máximas del valor medio que pueden esperarse bajo condiciones de laboratorio idénticas para instrumentos que funcionen dentro de las especificaciones. Consulte la Tabla de Rangos Esperados.

NCCLS Document M29-T2
How to Define and Determine Reference Intervals in the clinical laboratory; approved guidelines-second edition, NCCLS C28-A2, Volume 20, Number 12.

FR

Description du produit
Matériau de contrôle de la qualité destiné pour surveiller la performance des éléments suivants: Na⁺, K⁺, Cl⁻, iCa, iMg, glucose, lactate, BUN (urée) et créatinine, pour utilisation avec les analyseurs Stat Profile Prime Plus UNICAIEMENT.

Formulation à 2 niveaux
Destiné à être utilisé par les professionnels de la santé pour le diagnostic in vitro afin de surveiller la performance des analyseurs Stat Profile Prime Plus.

Méthodologie
Voir le manuel d'utilisation de l'analyseur Stat Profile Prime Plus pour la méthodologie et les principes.

Composition
Les contrôles sont des solutions tamponées contenant des concentrations connues de Na⁺, K⁺, Cl⁻, iCa, iMg, glucose, lactate, BUN (urée), créatinine et de créatinines. Chaque ampoule contient un volume minimum de 1,7 mL. Les contrôles ne contiennent aucun constituant d'origine humaine. Cependant, de bonnes pratiques de laboratoire doivent être suivies lors de la manipulation de ces matériaux. (REF: NCCLS DOCUMENT M29-T21)

Avertissements et mises en garde
Les contrôles doivent être stockés entre 2 et 8 °C. Une fois ouverts, analyser dans la minute qui suit et jeter la partie non utilisée conformément aux directives locales. Voir le manuel d'utilisation de l'analyseur Stat Profile Prime Plus pour obtenir des informations complètes. Pour usage diagnostique in vitro. Suivre les pratiques standard pour la manipulation des réactifs de laboratoire.

Stockage
Stockez entre 2 et 8 °C. NE PAS CONGELER. Un numéro de lot et une date d'expiration sont imprimés sur l'étiquette de chaque ampoule.

Mise d'échantillon
Garder les contrôles entre 2 et 8 °C jusqu'à l'utilisation. S'assurer que le contrôle est à température ambiante avant utilisation. Bien agiter l'ampoule avant de l'ouvrir, puis casser l'ampoule (en protégeant les doigts avec la gaine ou un gant). Une fois ouvert, analyser le contrôle dans la minute qui suit. Jeter la partie non utilisée conformément aux directives locales. Vérifier que le numéro de lot sur le tableau des plages attendues correspond au numéro de lot sur l'ampoule. Voir le manuel d'utilisation de l'analyseur Stat Profile Prime Plus pour obtenir des instructions complètes.

Limites d'utilisation
Les valeurs de plages attendues sont spécifiques aux instruments et contrôles fabriqués par Nova Biomedical.

Fiabilité des étalons
Les substances à analyser sont tracées selon les matériaux de référence de NIST.

Intervalles de référence
Les concentrations sont formulées à des valeurs normales et anormales attendues dans le sang du patient. La plage clinique attendue de ces valeurs dans le sang du patient est résumée dans Tietz, MD et al. 1993 Textbook of Clinical Chemistry, W.B. Saunders Co. Les utilisateurs peuvent souhaiter déterminer les valeurs moyennes et les plages attendues dans leur propre laboratoire.

Plages attendues
La plage attendue pour chaque paramètre a été déterminée en Nova Biomedical en utilisant des essais répétés sur des analyseurs Nova. La plage attendue indique les écarts maximaux du résultat moyen par rapport aux conditions de laboratoire identiques pour les instruments fonctionnant selon les spécifications. Voir la table Plages attendues.

NCCLS Document M29-T2
How to Define and Determine Reference Intervals in the clinical laboratory; approved guidelines-second edition, NCCLS C28-A2, Volume 20, Number 12.

DE

Produktdescription
Wässriges Qualitätskontrollmaterial zur Überwachung von Na⁺, K⁺, Cl⁻, iCa, iMg, Glucose, Laktat, BUN (Harnstoff) und Kreatinin. NUR zur Verwendung mit Stat Profile Prime Plus Analysegeräten.

Verwendungszweck
Für die in vitro-Diagnose durch medizinisches Fachpersonal zur Überwachung der Leistung der Stat Profile Prime Plus Analyser.

Verfahrensweise
Die Verfahrensweise und -prinzipien sind dem Anleithandbuch für das Stat Profile Prime Plus Analysengerät zu entnehmen.

Zusammensetzung
Die Kontrollen sind gepufferte Lösungen, die bekannte Konzentrationen von Na⁺, K⁺, Cl⁻, iCa, iMg, Glucose, Laktat, BUN (Harnstoff), Kreatinin und Kreatininsubstanzen enthalten. Jede Ampulle enthält mindestens 1,7 ml. Die Kontrollen enthalten keine Bestandteile menschlichen Ursprungs, jedoch sind bei der Handhabung dieser Kontrollen gute Laborpraktiken zu befolgen. (REF: NCCLS DOCUMENT M29-T21)

Warnhinweise und Vorkehrungsmaßnahmen
Die Kontrollen sind bei 2-8 °C zu lagern. Nach Öffnen ist die Analyse innerhalb von einer Minute durchzuführen, und Reagenzien sind entsprechend dem vor Ort geltenden Richtlinien zu entsorgen. Vollständige Informationen sind dem Anleithandbuch für das Stat Profile Prime Plus Analysengerät zu entnehmen. Zum Gebrauch bei der in-vitro-Diagnose bestimmt. Die übliche Vorgehensweise für die Handhabung von Labormaterialien ist zu befolgen.

Lagerung
Bei 2-8 °C lagern. NICHT EINFRIEREN. Auf dem Etikett jeder Ampulle ist die Chargennummer und das Verfallsdatum aufgedruckt.

Gebrauchsanweisung
Bei der Verwendung bei 2-8 °C lagern. Vor der Verwendung abkühlen lassen, dass die Kontrolle Raumtemperatur hat. Vor dem Öffnen die Ampulle gut schütteln und dann aufschneiden (mit dem Finger mit Gaze oder Handschuhen schützen). Nach dem Öffnen die Kontrolle innerhalb einer Minute analysieren. Reagenzien entsprechend der vor Ort geltenden Richtlinien entsorgen. Sichern stellen, dass die Chargennummer in der Tabelle mit den erwarteten Bereichen mit der Chargennummer auf der Ampulle übereinstimmt. Vollständige Gebrauchsanweisungen sind dem Anleithandbuch für das Stat Profile Prime Plus Analysengerät zu entnehmen.

Einschränkungen
Die Werte für die erwarteten Bereiche gelten nur für von Nova Biomedical hergestellte Instrumente und Kontrollen.

Rückführbarkeit der Standards
Die Analyse wird auf Standard Referenzmaterialien des NIST rückführbar.

Referenzintervalle
Die Konzentrationen sind so formuliert, dass sie normalen und anormalen erwarteten Werten im Patientenblut entsprechen. Der erwartete klinische Bereich dieser Werte im Patientenblut ist bei Tietz, R. W. (Hg., 1993), Textbook of Clinical Chemistry, W.B. Saunders Co. aufgeführt. Sie sollten keine der Standard-Maßnahmen und die erwarteten Bereiche im eigenen Labor ändern.

Erwartete Bereiche
Der Erwartungsbereich für jedes Parameter wurde von Nova Biomedical anhand von Wiederbestimmungen an Nova Analysengeräten ermittelt. Der erwartete Bereich zeigt die maximalen Abweichungen vom Mittelwert an, die unter identischen Laborbedingungen für Instrumente der festgelegten Qualitätswerte innerhalb der Spezifikationen zu erwarten sind. Holten Sie eine Tabelle mit den Erwartungsbereichen.

NCCLS Document M29-T2
How to Define and Determine Reference Intervals in the clinical laboratory; approved guidelines-second edition, NCCLS C28-A2, Volume 20, Number 12.

EC REP Nova Biomedical GmbH Hessenring 13A, Geb. G 64546 Mörfelden-Walldorf Germany



200 Prospect Street
Walham, MA 02454 U.S.A.

LPN 59009F 2024-11

