



TURKISH ACCREDITATION AGENCY

ACCREDITATION CERTIFICATE

As a "Type C" Inspection Body

BAYLAN ÖLÇÜ ALETLERİ SANAYİ VE TİCARET ANONİM ŞİRKETİ

Central Address: AOSB MAH. 10032 SK. NO:16 ÇIĞLI İzmir / Türkiye

**The list of the branches operating under the same accreditation depending on the central address and the scope of these branches are given in the annexes.*

is accredited in accordance with TS EN ISO/IEC 17020:2012 standard within the scope given in Annex following the assessment conducted by TURKAK.

Accreditation Number : AB-0068-M

Accreditation Date : 15.09.2009

Revision Date / Number : 19.04.2023 / 10


This certificate shall remain in force until **10.01.2026**, subject to continuing compliance with the standard **TS EN ISO/IEC 17020:2012**, related regulations and requirements.

Gülden Banu Müderrisoğlu
Secretary General



Turkish Accreditation Agency (TURKAK) is a signatory to the European co-operation for Accreditation (EA) Multilateral Agreement (MLA) and International Laboratory Accreditation Cooperation (ILAC) Mutual Recognition Agreement (MRA) in the scope of ISO/IEC 17020.

This document has been signed by Gülden Banu Müderrisoğlu on {1} with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.

 Inspection TS EN ISO/IEC 17020 AB-0068-M	BAYLAN ÖLÇÜ ALETLERİ SANAYİ VE TİCARET ANONİM ŞİRKETİ	
	Accreditation Nr: AB-0068-M Revision Nr: 10 Date: 19.04.2023	
"Type C" Inspection Body		
Address : AOSB MAH. 10032 SK. NO:16 ÇIĞLI İzmir / Türkiye		Phone : +90 232 497 9700 Fax : - Email : g.yigit@baylanwatermeters.com Website : www.baylanwatermeters.com / https://www.baylanelektrik.com/

Measuring Instruments		
Inspection Field	Inspection Type	Standard/Specification
23.1.Water Meters	23.A. Initial Inspection Static Pressure Test Cold and Hot Water Meters (DN15 - 200) (Q3: 1,6 - 400 m ³ /h)	OIML R49-2, Article 7.3, Article 10 EN ISO 4064-2, Article 7.3, Article 10 EN 14154-3, Article 5.2, (Although it was abolished on 04.06.2014, it is included in the scope since the organization still exports to countries where the mentioned standard is in force.)
23.4. Thermal Energy Meters	23.A. Initial Inspection (For cases where a type approval change is not needed after repair) 23.B. Periodical Inspection 23.C. Complaint Inspection(DN15-50) (Qp:1,5-15 m ³ /h; 3-95°C)	TS EN 1434-5 OIML R75-2 Heat meters Directive Annex- Article 5.1 (Official newspaper numbered 30470)
23.1.Water Meters	23.A. Initial Inspection Determination of Intrinsic Errors of Indications & Sealing Cold and Hot Water Meters (DN15 - 200) (Q3: 1,6 - 400 m ³ /h)	OIML R49-2, Article 7.4, Article 10 EN ISO 4064-2, Article 7.4, Article 10 Measuring and measuring Instruments Stamp Directive (Published in the official newspaper no 26400 at the date of 11.01.2007) EN 14154-3, Article 5.3 (Although it was abolished on 04.06.2014, it is included in the scope since the organization still exports to countries where the mentioned standard is in force.)

This document has been signed by Gülden Banu Müderrisoğlu on {1} with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.

