



Installation manual

User manual

**ENDOSCOPES WASHER
DISINFECTOR**

EW 2/1

EW 2/2

Serial N°:



Miele Group
Member



Via Balegante, 27
31039 Riese Pio X (TV)
ITALY

Manufacturer: STEELCO S.p.A. Via Balegante, 27 - 31039 Riese Pio X (TV) - ITALY

CONTENTS

1. GENERAL RULES	8
1.1 LIMITS OF MANUFACTURER’S LIABILITY	8
1.2 MANUAL VALIDITY, CONTENTS AND CONSERVATION	9
1.3 REGULATIONS	10
1.4 ENDOSCOPE REQUIREMENTS	11
2. SAFETY INFORMATION	13
2.1 INTENDED USE, IMPROPER USE.....	13
2.2 IMPORTANT WARNINGS AND SUGGESTIONS.....	14
2.3 SAFETY RECOMMENDATIONS	15
2.4 ATTENTION.....	17
2.5 RESIDUAL RISKS	19
2.6 SAFETY SIGNALS USED.....	21
2.7 TRAINING	22
2.7.1 STAFF QUALIFICATION.....	22
2.8 INDICATION OF SOUND LEVEL.....	23
3. UNIT CHARACTERISTICS.....	24
3.1 MACHINE COMPONENTS.....	24
3.2 CONSTRUCTION CHARACTERISTICS	25
3.3 ENVIRONMENTAL EMISSIONS	26
3.4 INSTALLATION	27
3.5 RE-INSTALLATION	27
3.6 TECHNICAL CHARACTERISTICS.....	28
3.6.1 MEASUREMENTS	28
3.6.2 AMBIENT INSTALLATION AND USE CONDITIONS	28
3.6.3 NOISE	28
3.6.4 VIBRATIONS.....	29

3.6.5	UNIT WEIGHT	29
3.6.6	HEAT LOSS AND AMBIENT VENTILATION	29
3.6.7	SIGNALS AND WARNINGS	29
4.	INSTALLATION.....	30
4.1	SHIPPING AND UNPACKING	30
4.2	ASSEMBLY	30
4.2.1	MAXIMUM FLOOR LOAD	30
4.2.2	POSITIONING OF THE MACHINE	30
4.3	POSITIONING	31
4.3.1	UNIT WITH ONE DOOR	31
4.3.2	UNIT WITH TWO DOORS.....	31
4.4	CONNECTION	32
4.4.1	ELECTRICAL CONNECTION	32
4.4.2	FUSES	36
4.4.3	COMPRESSED AIR CONNECTION.....	36
4.4.4	PLUMBING CONNECTION.....	37
4.4.5	CONNECTION TO THE DRAIN	38
4.4.6	CONNECTING THE CHEMICALS	39
4.4.6.1	REPLACEMENT OF CHEMICAL PRODUCT CONTAINER	40
4.4.6.2	WATER INTAKE FILTER INSTALLATION	42
4.4.6.3	CALIBRATION PROCEDURE	44
4.5	CONNECTIVITY.....	49
5.	CHECKING THE PARAMETERS AND SCREEN PAGE FUNCTIONS ON THE MONITOR.....	50
5.1	MAIN MAINTENANCE AREA PARAMETERS.....	53
5.2	CLOCK AND DATE UPDATE	54
5.3	RESTORING THE PARAMETERS SAVED	56
5.4	MANUAL CONTROL OF THE MECHANICS	57
5.5	CYCLES HISTORICAL VISUALIZATION	68
5.6	PARAMETER CONFIGURATION	73
5.7	USER DATA CONFIGURATION	74

5.8	THERMALDISINFECTION CYCLE.....	84
5.9	SELF-DISINFECTION PROCEDURE WITH A 12% SODIUM HYPOCHLORITE SOLUTION.....	86
5.10	TABLE MANAGEMENT.....	87
5.10.1	ENDOSCOPE TYPE TABLE.....	89
5.10.2	MANUFACTURERS TABLE.....	91
5.11	OPERATOR SETUP.....	93
5.12	ENDOSCOPE SETUP.....	96
5.12.1	ENDOSCOPE CONFIGURATION PARAMETER SETTING.....	98
5.13	CREATING A BACKUP OF THE SETTINGS.....	106
5.14	UTILITY.....	107
5.15	TECHNICAL AREA.....	114
5.15	MAINTENANCE AREA PARAMETERS.....	115
5.16	STORING A TECHNICAL INTERVENTION.....	116
5.17	I/O STATISTICS.....	118
5.18	CALIBRATIONS.....	119
5.19	RECALLING A TECHNICAL INTERVENTION FROM MAINTENANCE FROM HISTORY.....	125
5.20	CONSULT THE ALARMS HISTORY.....	128
5.21	CHANGING MAINTENANCE PARAMETERS.....	131
5.22	CHANGING CYCLE PARAMETERS.....	141
6.	WASHING CYCLE.....	149
6.1	PREPARATION OF THE ENDOSCOPES.....	149
6.2	CHECK OF ENDOSCOPE INSTRUMENT CONNECTORS.....	150
6.3	START CYCLE.....	151
6.4	FREE PROGRAMS SETTING.....	157
7.	PROGRAMMED MAINTENANCE.....	183
7.1	OPERATOR MAINTENANCE.....	183
7.2	TECHNICIAN MAINTENANCE.....	186
7.3	NECESSARY MATERIALS.....	199
8.	MACHINE ALARMS.....	200
8.1	TROUBLE SHOOTING.....	200

8.2	MACHINE ALARMS CONCERNING MECHANICAL PARTS	246
9.	DEMOLITION AND DISPOSAL	251
10.	QUALIFICATION.....	252
10.1	INSTALLATION QUALIFICATION (IQ).....	252
10.2	OPERATIVE QUALIFICATION (OQ)	252
10.3	PERFORMANCE QUALIFICATION (PQ).....	252
ANNEX A – TRAINING CERTIFICATE		255
ANNEX B – WATER SAMPLING DURING THE FINAL RINSING PHASE.....		256
ANNEX C – TEST PROCEDURE WHEN INSTALLING.....		267
ANNEX D – BARCODE READER SETTING		268
ANNEX E: SPECIFICATION OF ROUTER NAT PROGRAMMING.....		275

ATTENTION!!

THE PRESENT HANDBOOK HAS BEEN WRITTEN FOR TECHNICIANS AUTHORIZED

Thank you for purchasing this appliance.

The installation, maintenance and operating instructions given in the following pages have been prepared to ensure the long life and good performance of the appliance.

Following the instructions carefully.

The appliance was designed and constructed using the latest technological innovations available.
Please take good care of it.

Your satisfaction is our best reward.

WARNING:

NON OBSERVANCE, EVEN IN PART, OF THE RULES INDICATED IN THIS MANUAL WILL CAUSE THE PRODUCT GUARANTEE TO BECOME INVALID AND RELIEVES THE MANUFACTURER OF ANY RESPONSIBILITY.

1. GENERAL RULES

1.1 Limits of manufacturer's liability

The manufacturer shall not be held liable for failures or problems which arise due to tampering and/or incorrect applications and/or improper use of the machine.

The purchaser must comply with all instructions set forth in the user's manual, and he must in particular:

- Always work within the allowable limits for the use of the machine;
- Always carry out constant and diligent maintenance;
- Allow use of the machine by persons with proper skills and abilities for their role and purpose who have been properly trained and instructed;
- Use only manufacturer original spare parts.

Any modifications, adaptation or the like which may be made to machines which are subsequently placed on the market do not oblige the manufacturer to intervene on previously supplied machines, nor to consider the machine and the related user's manual lacking and inadequate.

The installation, maintenance and operating instructions given in the following pages have been prepared to ensure the long life and outstanding performance of the appliance.

For some especially demanding programming or maintenance operations, this manual serves as a memorandum of the main operations to be carried out. Education on these topics can be obtained by attending training course held by the manufacturer.

The instructions in this manual do not replace but rather are in addition to employer requirements to adhere to current legislation on standards of prevention and safety.

1.2 Manual validity, contents and conservation

This manual reflects the state of the art at the moment of manufacture and delivery of the appliance and is valid for its entire life cycle.

The manufacturer is at clients' disposal for further information or to receive suggestions for making the manual more compliant with the needs for which it was prepared.

The translation of the contents into the client's language has been carefully prepared.

In order to prevent possible accidents to persons or property due to incorrect translation of the instructions, the client must:

- Not perform operations or manoeuvres with the machine if there are any doubts or uncertainties about the operation to be performed;
- Ask technical service for clarification of the instruction.
- If lost, ask for a new copy from the manufacturer.

It is important to keep this instruction manual with the machine for future reference.

If the machine is sold or transferred, the manual must be handed over to the new owners or user in order for them to become acquainted with its functioning and the relative warnings.

Read the warnings carefully before installing and using the machine.

This is a translation of the Italian text, which prevails in case of doubts.

1.3 Regulations

The purpose of the warnings is to safeguard the user in compliance with following Regulations and “Technical Product Standards”:

EUROPE:

- 93/42/EEC and s.m.i (Medical Devices Directive);
- 2014/35/UE (Low Voltage Directive);
- 2014/30/UE (Electromagnetic compatibility directive);
- 2014/68/UE (Pressure Equipment Directive);
- EN 61010-1 (Safety);
- EN 61010-2-040 (Safety);
- 2011/65/EC (RoHS II);
- 2012/19/EC (WEEE);


and recognized international standards:

- IEC 61000 (Electromagnetic compatibility);
- IEC 61326-1 (Electromagnetic compatibility);
- ISO 14971 (Medical devices risk analysis);
- ISO 15883-1 (General requirements, terms, definitions and tests);
- ISO 15883-4 (Requirements and tests for medical washer and disinfectors that carry out the chemical disinfection of flexible endoscopes);
- ISO/TS 15883-5 (Soil test – A soil test method to prove the effectiveness of cleaning activities);
- IEC 60529 (IP Grade);

1.4 Endoscope requirements

The flexible endoscopes treated in the EW 2 system must fulfil the following requirements:

- Resistance to a temperature between +10°C and +60°C;
- Resistance of the endoscope channels to a maximum pressure of 1.5 bar;
- Resistance to a pressure of max. 300 mbar during the leak test;
- Resistance to the detergents and disinfectants used;
- Presence of connections where to attach each of the endoscope channels;
- Waterproof with protection against the effects of immersion and, if applicable, equipped with parts that protect the endoscope from humidity.



The colored code connection suggested in the figure above refers - but is not limited – to the majority of the endoscopes in the market.

During endoscopes calibration and setting, please check carefully each channel pressure and flow.

When the values detected do not allow the detection of obstructions or disconnections, a customized connection is needed.

When this happens, the color code identification (color – channel function) is not valid anymore, but the connection should be done based on the flowmeters' diameters and booster pumps. (See picture below).

Such a connection must be able to detect any obstruction or disconnection.
 Moreover, every colored hose which is not needed, shall be connected to a flow reducer.
 In case of need / doubts please contact the manufacturer.

			Ø mm Inner	Ø mm Range
1	Brown	variable	5,6	1 – 5,6
2	Green	variable	5,6	1 – 5,6
3	Blue	Water	1	0,5 – 1
4	Yellow	Air	1	0,5 – 1
5	Transparent	Leak test	---	
6	Red	variable	1	0,5 – 1
7	Black	variable	1,2	0,6 – 1,2

2. SAFETY INFORMATION

Compliance with safety standards allow the operator to work productively and calmly, without the danger of harming himself or others. Before starting work, the worker must be completely familiar with the functions and proper operation of the machine. He must know the precise function of all command and control devices of the machine.

2.1 Intended use, improper use

INTENDED USE:

The use of this device is allowed only and exclusively for the reprocessing of medical devices such as:

- **Flexible endoscopes**
- **Rigid endoscopes**
- **Video bronchoscopes or others**
- **Transesophageal probes**

Improper use of this unit may be hazardous to the operator and may seriously damage the machine itself.

WARNING: If the appliance is used in a manner not specified by the manufacturer, protection of the appliance may be compromised.

IMPROPER USE:

ANY USE OTHER THAN THAT FOR WHICH THE MACHINE WAS INTENDED IS FORBIDDEN.

2.2 Important warnings and suggestions

For proper use of the machine, and in order to safeguard employed staff, carefully comply with the following general and specific standards.

THE OPERATOR MUST:

- **Carefully adhere to the provisions and instructions** provided by the employer, managers and supervisors for individual and group safety.
- **Use safety devices appropriately and with care**, as well as group and individual safety gear provided by the employer.
- **Immediately inform the employer**, the manager and the supervisor of deficiencies in the aforementioned devices and means, as well as any hazardous conditions which he may become aware of, taking action directly in urgent cases within their scope of responsibilities and abilities to eliminate or reduce the deficiencies or hazards.

THE OPERATOR MUST NEVER:

- **Remove or modify, without authorization, the safety devices**, nor those for signalling and measuring, nor the individual and group safety gear.
- **Undertake on his own initiative operations or manoeuvres which are not his responsibility which** may compromise safety.
- **Insert foreign objects into the electrical parts.**
Do not insert foreign bodies into the covers of the electrical motors or into the moving parts of the machine.
- **Provide power to the machine by tampering with the main switch and the safety devices.**

2.3 Safety recommendations

- If the new machine seems damaged, contact the retailer before using it.
- Any modification of electrical and hydraulic systems necessary to install the machine must be carried out by qualified, authorised persons only.
- This machine must be operated by trained persons only;
- This machine has been designed for the reprocessing of flexible and rigid endoscopes, transesophageal probes and the thermal-disinfection of washing chambers;
- Any use other than that for which the machine was intended is forbidden.
- The user is forbidden to carry out any work or repairs on the machine.
- Technical Assistance for this machine should be carried out by qualified and authorised operators only.
- The equipment should be installed by authorised persons only.
- The electrical safety of machine is only guaranteed if it is connected to an efficient earth system.
- Take great care when handling detergents and additives: avoid contact, wear gloves, goggles and mask and act in compliance with the safety recommendations indicated by the manufacturer of the chemical products.
- Do not inhale the fumes produced by chemical products.

WARNING: The chemical products are an irritant for the eyes, in case of contact rinse thoroughly with plenty of water and consult a doctor.

If these products come into contact to the skin, rinse with plenty of water.

- The water in the tank is not drinking water.
- Do not lean on the door and do not use it as a step.
- Do not install the equipment in rooms where there is the risk of explosion. (ATEX)
- Do not expose the equipment to intense cold.
- Do not wash the machine using high-pressure jets of water.
- The machine reaches a temperature of 80/90°C during the self-thermaldisinfection: take great care to avoid burns.
- Disconnect the machine from the electrical supply before carrying out maintenance work.
- The acoustic pressure of the machine is < 40 dB(A).

IMPORTANT WARNING FOR THE MAINTENANCE TECHNICIAN

LEAVE THE "Min.lim.water inconsistency" I.= 0.0 PARAMETER




WARNING ??	
BOUND VARIOUS1	LIMIT VARIOUS2
TIMES VARIOUS	
Max.channel press.mb:	2450
Max.channel flow ml/' :	0900
Leak contr.press.lim.mb:	0400
Chann.null press.lim.mb:	10
Tank level ON=litres:	08.0
Min.lim.incong.water l.:	0.0
Min.lim.incong.chem.ml:	25
Nr of cycles protected:	050

SAVE DATA PRINT DATA



2.4 Attention

- The user must oversee the machine during the cycle.
- The injection tube for washing water must always be connected to the appropriated basket.
- When the machine is running do not interrupt the cycle since this jeopardises disinfection.
- Use only detergents and chemical additives that have been tested and approved by the equipment manufacturer;
The use of other products may damage the machine and the correct reprocessing of the medical device and the cycle cannot be guaranteed or validated in accordance with standard ISO 15883.
- The use of opportune PPEs is compulsory in order to avoid contact with infected material and to prevent contamination during the handling procedures of medical devices to be reprocessed.
- The chemical products recommended by the manufacturer are those that have been certified and validated by ISO 15883:4 and 15883:5-TS standard.
- Check that type of chemical product is suitable for the specific washing program used.
- Comply with the instructions provided by the chemical product manufacturer.
- The machine was designed for use with water and chemical additives.
Do not use organic or other types of solvent as this may result in the risk of explosion or the rapid deterioration of certain machine parts.
- Residues of solvents or acids, particularly “hydrochloric acid”, can damage steel.
- Use original accessories only.
- Do never use soap powder.
- Do never use foaming detergent.
- The machine is to be used only with the accessories included by the manufacturer.
- Accessories which are not approved by the manufacturer may compromise the results achieved as well as user safety.
- Do never use chemical products based on chlorides (bleaches, sodium hypochlorite, hydrochloric acid and so on). These kinds of chemical detergents irreparably damage the machine and jeopardise the integrity of materials and objects treated.
- Check at every cycle the integrity of the connexions used to connect the endoscopes.
- The operator always has to verify before starting of the cycle the presence of the filters water in the sump and their correct positioning.

	ATTENTION
	IT IS NECESSARY TO START A THERMODISINFECTION CYCLE TO DECONTAMINATE THE MACHINE AT THE FIRST START OF MACHINE – SEE ANNEX C.

ATTENTION:

The taps of the water must be always turned off, as the safety and diagnosis system will be deactivated, in the following situations:

- If the machine is left unused;
- If the machine is disconnected from the electrical connection.

The Manufacturer cannot be held responsible for damage or injury caused by failure to observe the above rules.

The non-observance of these rules produces the total and prompt cancellation of the guarantee.

2.5 Residual risks

The appliance includes a series of fixed guards to prevent access to hazardous internal parts or zones.


It is however considered that the **EW 2** includes some residual risks.

Hereunder for each phase or significant work intervention are useful measures to be taken:

PHASE	BASKET LOADING
RISK	Contusions and cuts to the upper limbs , due to accidental contact with due to falling or striking against tools, objects and instruments, mainly while loading and handling the basket.
MEASURE	Assign staff that is instructed and equipped with work equipment (e.g. basket with protections, transport carts) and appropriate clothing and individual protection gear (e.g. shirts and protective gloves).

PHASE	MAINTENANCE OF INTERNAL EQUIPMENT
RISK	Burns of body parts by hot parts of the appliance.
MEASURE	Allow maintenance to be performed only by trained personnel, equipped with appropriate clothing and individual protection gear. Wear suitable clothing and protective gloves.

PHASE	EMISSION OF HAZARDOUS GAS
RISK	Inhalation of vapours of hazardous gas.
MEASURE	With a correct installation, concurring with the manufacturer prescription, using the authorized chemical product and concurring with the rules in force in your country, the machine doesn't generate hazardous gas. However, the machine is supplied with vapours discharge, that have to be connected concurring with the instruction on chapter 3.

PHASE	OBTAINING DETERGENTS/CHEMICAL ADDITIVES
RISK	Contact with body parts with chemical washing products.
MEASURE	Assign staff that is instructed and equipped with appropriate clothing and individual protection gear. Wear clothing, gloves, goggles and mask and act in compliance with the safety recommendations indicated by the manufacturer of the chemical products.
FIRST AID MEASURE	<ul style="list-style-type: none"> • Immediately take off/remove clothing which has been contaminated or soaked by the product. • If the substances come into contact with the skin, wash off affected skin areas immediately and rinse with water.
RISK	Inhalation of vapours of chemical wash products.
MEASURE	Assign staff that is instructed and equipped with appropriate clothing and individual protection gear. Comply with the safety instructions provided by the manufacturer of the chemical products and if there are none, wear a mask for the protection of the respiratory airways.
RISK	Accidental release of chemical wash product
MEASURE	Do not flush concentrate into drains, surface or ground waters. Collect spillage with adsorbent material (e.g. sand, earth, vermiculite, diatomaceous earth). Flush away minor amounts with plenty of water.
	IN CASE OF CONTACT WITH BODY OR RELEASE OF CHEMICAL PRODUCT LOOK ALWAYS AT THE SAFETY MEASURES INDICATED IN THE CHEMICAL TECHNICAL DATASHEET.

2.6 Safety signals used

To inform personnel operating on the machines of obligations of behaviour and residual risks, adequate safety signals (as set forth by 92/58 EEC) are applied to the machine and near the workplace.

GENERIC SAFETY SIGNALS:

In particular, labels with signals of obligation, prohibition and danger contained in this manual, and pertinent to this machine and most commonly used are:



Electrical risk



Warning!
See annex documentation



CAUTION
HOT SURFACE
Caution hot surface

INDIVIDUAL SAFETY WEAR:

The evaluation of risks for the health and safety of workers carried out in the workplace and on any equipment used, as well as the evaluation of residual risks as indicated, allow the employer to evaluate the need to adopt the individual protection gear which is most suitable and appropriate to be provided to workers. Considering the type of machine, it is felt that the individual protection gear should be provided to staff.

2.7 Training

Instructions for use of the machine will be provided by the **STEELCO INSTALLATION TECHNICIAN** during the start-up phase to **MACHINE OPERATORS** and **MAINTENANCE TECHNICIANS** for their areas of responsibility, which will be thus instructed and trained. Moreover, an appropriate course certificate is issued (see Annex A).

It will be the duty of the **EMPLOYER** to check that the degree of staff training is suitable for assigned duties.

2.7.1 Staff qualification

Depending on the difficulty of certain installation operations, and of the operation and maintenance of the system, professional profiles are identified as follows:

Is INSTALLATION and REPAIR TECHNICIAN:

Specialized installation and maintenance staff capable of carrying out all machine positioning and installation operations, connection of various systems and machine start-up at the client's place of business, as well as all routine and special maintenance operations.

This operator is responsible for training staff for machine operation and for testing the machine.

As RESPONSIBLE AUTHORITY FOR THE MACHINE IN THE WORKPLACE:

Specialized staff assigned to the verification of safety devices and procedures for proper use of the machine in complete absence or hazards.

The responsible authority is personally responsible for training courses for staff assigned to machine operation and maintenance.

He must ensure that staff assigned to operation has acquired all information required for use and routine maintenance of the machine, registering attendance and documenting comprehension tests.

The responsible authority must have a perfect understanding of all command, control and safety devices of the machine.

He must inform all personnel assigned to machine operation and maintenance of the instructions concerning safety standards, the actions to be avoided and the first aid interventions connected with use of the machine and the chemical wash agents it contains.

The responsible authority must be aware of all correct procedures for carrying out in absolute absence of danger all operation and maintenance of the machine, as well as all procedures for disposal of any residual pollutants and manufacturing wastes.

He must always be present during extraordinary or routine maintenance and give his approval to proceed to staff assigned to operation or to personnel assigned to routine or special maintenance.

The responsible authority will be responsible for operation of all command, control and safety devices in the machines of the system.

He shall carry out scheduled verification of those devices in order to ensure their continued operation over time.

Ac MACHINE OPERATOR:

Skilled personnel assigned to machine operation.

The machine operator must be perfectly aware of all of the machine command and control devices.

Only after approval by the safety supervisor, the machine operator must be capable of using the assigned commands to do the following:

- Commissioning and start-up of the machine;
- Loading and unloading of material to be washed in the baskets;
- Operation of the machine in the various possible working modes, such as the start of various programmed wash cycles.
- Programming and setting data from the operator panel, adjustment of single control devices during working phases, starting or resetting of work functions.
- In addition, the machine operator must, by making use of all required individual protection gear and following adequate safety measures, be capable of performing some routine maintenance such as cleaning inside the machine, cleaning clogged filters, and disposing of pollutant waste materials produced during working.

2.8 Indication of sound level

The value shown refers to the measurement obtained on a machine of the same type as that covered herein and measured with an instrument at a height of 1,5 m at a distance of 1 m from the machine.


AVERAGE SOUND PRESSURE LEVEL: < 40 dB (A)

3. UNIT CHARACTERISTICS

3.1 Machine components



Wash cart for the simultaneous re-processing of 20 rigid endoscopes equipped with luer lock connections.



Wash cart for the simultaneous re-processing of 2 transesophageal probes inside the SAFE CASES.



Rotating basket for 2 flexible endoscopes.



Wash cart with baskets for 2 flexible endoscopes.



Wash cart for simultaneous re-processing with an independent monitoring system of the bioptic/suction channel and leak test.

3.2 Construction characteristics

The typical construction characteristics of the unit are:

- pressed washing chamber with ample curves to facilitate the flow of chemicals;
- watertight washing chamber, thermally and acoustically insulated, with rounded edges, made of AISI 316L stainless steel. External body in AISI 304 stainless steel;
- hung access door in HTS glass, washing chamber, wash rotors, filters and internal circuits made of AISI 316L stainless steel;
- system to lock the door during the cycle and unlock it at the end of the cycle;
- water non-return system to prevent contamination of mains water (**UNI EN 1717**);
- 1 air pump (for testing endoscopes for leaks);
- possibility of treating endoscopes with 8 channels each (one channel for the leak test and 7 functional channels);
- possibility of treating indifferently 7 channels with a diameter greater or smaller than 1 mm for each endoscope;
- two stainless steel wash rotors in the chamber (plus a third in the middle of the basket);
- chemical disinfection (using peracetic acid diluted to 0.15%) at a temperature of 35°C for rigid and flexible endoscopes;
- chemical disinfection carried out at 55°C with peracetic acid diluted by 0.20%;
- **thermal self-disinfection (according to the EN ISO 15883-4 standard) at 80°C for 10 minutes or at 90°C for 3 minutes;**
- temperature control with 3 PT 1000 probes (plus one for drying version);
- control of pressure and flow for each channel of each endoscope;
- equipped with 3 adjustable Teflon dosing pumps for chemicals;
- checking for the presence of chemicals with RFID;
- double check of the quantity of chemical/disinfectant introduced;
- double check of the quantity of water introduced;
- control of the sealing phase before starting and during the entire treatment cycle;
- touch-screen control system with display of the single phases of the wash cycle, showing the time left;
- complete self-diagnosis system to facilitate technical service;
- 200 programs can be stored;
- 200 endoscopes can be stored for each type;
- 10 different endoscope Manufacturers can be stored;
- 20 different endoscope types can be stored;
- stores the last 800 cycles completely (all data of all the machine sensors are collected);
- total control of the cycle phases, time, temperature, chemical dosing, pressure range on the channels, quantity of water, drying;
- acoustic alarms for malfunctions and end of cycle;
- cycles stored with the events recorded (made available via an RS 232 connection, network port or USB port);

- set up for connection to the local Ethernet;
 - 3 Password levels to access programming;
-
- touch-screen display with complete info on machine functions;
 - dual filter system for inlet water (the first is 0.45 micron and the second 0.1 micron, absolute filter, medical device, 0.1 micron, the filter stops bacteria entering up to 10^{12} . The filter is positively charged for removing the endotoxins);
 - compressed air filtering system at 0.2 micron, able to guarantee absolute sterility of the process (optional – connection to medical air required);
 - 0.2 micron filter leak test;
 - possibility of setting a time for the self-disinfection cycle to be done automatically;
 - 0.3 micron chamber filter (pressure balancing);
 - automatic traceability system (optional software – Steelco Data or ARES);
 - Independent Double control PLC;
 - possibility of computerising the treatment process by means of operator and/or endoscope identification using barcode scanners and/or transponder scanners.

3.3 Environmental emissions

The closed-circuit System has no significant environmental emissions and any emissions there may be when changing the tanks are not toxic or harmful to man. However, the System has a fitting for the forced evacuation of the fumes from the chemical tanks and from the high disinfection process. This fitting is at the back in the case of a single-door unit and on top in the case of the through model and is supplied with a hose about 30 cm long (on request) for connecting to the extraction hood (if applicable).

3.4 Installation

- The endoscope washer is delivered after several tests that guarantee the correct functioning of the machine. However, before to use it, it is necessary to record the various flexible endoscopes type used and their features.
- Each endoscope will be codified (manufacturer, identification number, number and type of the channels, type of associated cycle, etc.) and the appropriate pressure and flow rate values will be associated to each one.
- These values can be insert manually or using a self-learning function; another way is to use the standard parameters of an instrument, copying from one of the instruments on the list, copying it in a new position and deselecting the channels don't used.
- During the next treatment cycles execution, if any set value is different from the real value the machine shows an anomaly (leak form the channels, disconnection from the endoscope circuit, etc).
- The external panels have always to remain closed.
- The data of the pressure and flow rate values of the different channels have to be inserted before the first use of the endoscope and have to be checked with a test cycle after the endoscope standard/special maintenance.
- The installation qualification, the operative qualification, the performance qualification and the periodical following qualification are assigned to the user. **An annual Performance Qualification (PQ) is recommended.**
- All the qualification tests have to be documented, approved and preserved from an appointed person before the endoscope washer use.
- All the procedures for the routine test execution, included the verifications, check and the materials to be used have to be documented.
- All the procedures for the standard maintenance of the endoscope washer and for the check operations of the validated cycles performance reproducibility, has to be documented.
- PQ and OQ protocols are available only upon request.

3.5 Re-installation

The machine can be moved from the installation site and be installed again in a new one only if this operation will be done following the same method and watchfulness of the first installation and from qualified personnel, including the documentation as on the first installation (installation test).

The new installation site has to be communicated to the Manufacturer/Distributor or other in accord to the prescription of the Directive 93/42 CEE.

ATTENTION: PLEASE CONSULT ANNEX C IN THIS MANUAL.

3.6 Technical characteristics

3.6.1 Measurements

- EW2-2 external measurements mm. 860x710x1690h
- EW2-1 external measurements mm. 860x660x1690h
- internal dimensions of the wash chamber mm. 560x590x600h (equal to a volume of 200l);
- maximum load capacity: 2 flexible endoscopes, 8 fibro bronchoscopes (optional basket) or 4 video bronchoscopes (optional basket), 2 transesophageal probes;
- washing pump power W 550;
- triphase electrical voltage 400Vac - 50Hz (Optional - 230Vac 50Hz) + monophasic 230 V 50 Hz ~1.;
- fully compliant with standards: UNI EN ISO 15883-1 and UNI EN ISO 15883-4.
- compliant where it is possible with standard UNI EN ISO 15883-5.

3.6.2 Ambient installation and use conditions

The ambient requirements of the storage facilities, transport, assembly, installation and use are given in Table.

REQUIREMENT	VALUE
Machine chamber operating temperature	35/45°C (in self-disinfection 55°C + chemical or max. 80/90°C water only)
Maximum ambient operating pressure	0,2 bar
Use	Reprocessing endoscope accessories and rigid and flexible endoscopes
Relative humidity from	20Rh at 90Rh
Temperature	from +5 at +40°C

Table: Ambient installation and use conditions

3.6.3 Noise

Sound pressure is < 40 dB. The people responsible for the instrument washer must ensure that the sound pressure level of the unit, once installed, does not exceed a hazardous limit.

3.6.4 Vibrations

If the machine is installed correctly vibrations are virtually non-existent.

3.6.5 Unit weight

The unit has a gross weight of 260 kg. Make sure it is installed on a surface that can support this weight.

3.6.6 Heat loss and ambient ventilation

The door is made of HTS glass.

The ventilation system of the premises must be in conformity with what is provided for by specific legislation and it must comply with the values indicated in the installation plan.

3.6.7 Signals and warnings

Signals and warnings necessary for a correct installation of the unit are specified on the CE label (included in the user manual and on the machine) and in this manual. It is strictly forbidden to remove and/or damage the labels on the EW 2 unit. If they are no longer legible ask the Manufacturer for replacements.

4. INSTALLATION

This unit was designed, built and checked in accordance with EC Directives and harmonised standards in order to ensure the utmost safety and effectiveness.

4.1 Shipping and unpacking

The unit can be transported by the usual carriers. It is packed on a pallet for easier handling with a transpallet. Handle the unit with care anytime and do not turn it over. The unit must be unpacked carefully to avoid damaging it. Unpack the unit following these instructions:

- Cut the strap or open the box and remove the expanded polystyrene corner guards;
- Remove the box followed by the nylon bag.
- **CAUTION: the bag represents a serious hazard for children and should be disposed of immediately.**
- All the packaging materials can be recycled.
- The lifting capability of the forklift must be greater than the total weight of the machine to be moved;
- The machine must be kept as close as possible to the ground during movement.
- The belts, ropes or chains and lifting hook used to the machine lifting, must be suitable to total weight of machine to be moved.
- **The forklift operator must perform movement only when there are no persons or objects in the movement area.**

4.2 Assembly

The unit does not have to be disassembled prior to installation – it is delivered ready to position and install.

4.2.1 Maximum floor load

For machine installation, the floor must be rated for a minimum load of:

300 daN/m²

4.2.2 Positioning of the machine

In normal conditions, the minimum dimensions are suggested for the use of the machine in a single installation or with the coil nearby. For different installation ask for the distributors.

Minimum room ceiling height: Machine height (in m) + 0,3 m

4.3 Positioning

The unit is delivered ready to install. It only needs connecting to the compressed air and to the hydraulic and electrical power sources. In the case of special installations, please refer to the Manufacturer's installation diagrams.

When positioning the unit in the place of installation, it must be levelled using the adjustable feet.

4.3.1 Unit with one door

To be able to use and service the unit easily a clearance of at least 1 mt must be left at the front (so the loading door can be opened and loading is easier), at least 30 cm on the left (looking at the loading side) and at least 10 cm from the wall it is against.

4.3.2 Unit with two doors

The two-door unit is used to distinctly separate the flows of soiled material from those of the clean material. To respect this constraint, the unit is placed astride the wall separating the soiled area from the clean area. To be able to use and service the unit easily a clearance of at least 1 mt must be left at the front and back (so the loading door can be opened and to make loading easier), at least 30 cm on the left (looking at the loading side).

4.4 Connection

4.4.1 Electrical connection

The Retailer and/or Importer and/or Installer are obliged to provide for a suitable insulation class of the power cord based on the working environment and in compliance with the Technical Standards in force in the country where the unit is installed.

- The electrical connection must be carried out in compliance with current technical regulations.
- Make sure that the mains voltage measured corresponds to the voltage indicated on the machine's rating plate.
- A single-pole magnetothermal switch suitable for the current input and with a contact opening of at least 3 mm must be installed.
- Make sure the electrical systems are efficiently earthed.
- The machine is equipped with a terminal for the equipotential connection between the various devices (see standards on electrical systems).
- The unit's power supply line must include a four-pole circuit breaker with a calibration threshold of 30 mA in compliance with current standards. For the amperage of the unit please see the CE labels on the unit itself. This device must be installed as near as possible to the unit and must be easy to reach for the operator using it. The switch must be marked "UNIT DISCONNECTION DEVICE" and must indicate precisely the unit in question. The cross sections of the cables must be suitable for the power of the unit installed (see the CEI/IEC standards).

THE MACHINES ARE NORMALLY EQUIPPED WITH A THREE-PHASE 220/230/240 VAC - 60HZ POWER SUPPLY AND THREE-PHASE WITH NEUTRAL 380/400/415 VAC - 60HZ POWER SUPPLY.

FOR THE SPECIFIC VOLTAGE, PLEASE REFER TO THE LABEL ON THE MACHINE.

DESCRIPTION	REQUIREMENT
Installation conditions	Fixed
Protection against electrical charges	Class I
Degree of pollution	2
Operation	Non-stop
Hook-up connection	Permanent
External measurements	860 x 660 x 1690h [mm] Single Door


Table: Electrical specifications

DESCRIPTION 400VAC STANDARD	230VAC ON REQUEST
Voltage 400Vac 3N+E or 230Vac 3N+E	50 Hz 230Vac single-phase - 50 Hz
Watt [W] 5800 or 9800	2500 for the 220V
Ampere [A] 14.5 or 24.5	11.5 for the 220V

Table: Power

- The unit was designed and built also taking into consideration standard CEI EN 61326-1 of 1998: 'Electromagnetic compatibility (EMC) requirements.
 - The unit is connected permanently. The disconnection device must be installed in the main power supply panel.
-
- The wiring diagrams provided with the unit must be complied with in full, paying particular attention to the earth connection. The earth terminal is next to the electrical power supply terminals, inside the electric box for connecting the unit to the mains.
 - This machine is electrically safe only if connected to an efficient earth system that complies with the laws/standards in force in the country of installation.
 - Check that the power supply voltage does not differ by more than 10% from its nominal value.
 - The frequency of the power supply voltage must not differ by more than 1% of its value.
 - Connection of the machine to the mains must be provided with an earth connection and an equipotential circuit as set forth by current standards.
 - Make sure that the electrical systems are efficiently earthed.
 - The earth conductor is to be connected to the earth terminal identified by the standard symbol.
 - The machine is equipped with a terminal identified by the relative symbol for equipotential connections between appliances (see rules for electrical plants).
 - In case of prolonged unused of the machine is recommended that you execute the disconnection procedure of the electrical connection by placing the main-switch in "OFF" state.
 - The magnetothermal switch must be positioned in an accessible place, free and not covered from other machines or anything that could obstruct the switch control.
 - The magneto-thermal switch must be provided with quality markings and must be indicated as an electrical shut-off device for the machine.
 - Near the magneto-thermal switch, a sign must be placed which reads:

“DISCONNECTION DEVICE FOR MACHINE”



4.4.2 Fuses

The fuses are used to protect the electrical circuits of machine from possible failure as overload or short circuits.

If a fuse takes action the downstream connections and their function are no longer available.

The fuses must respect the characteristics (size, dimensions and tripping characteristic) indicated in the wiring diagram.

Replacement of fuse:

ATTENTION

The replacement of fuse must be done from authorized operators only.

Before making the replacement procedure of fuse, establish and remove the cause of the fault.

If necessary, contact our technical assistance service.

Replacement procedure of fuse:

- Log off the machine in safety condition by magnetothermal switch.
- Access at the electrical panel.
- Identify the fuse subjects to replacement, based on the wiring diagram.
- Remove the related fuse from electrical panel.
- Replace failure fuse with another fuse with same characteristics. The correct value of fuses is in the wiring diagram.

If at the reactivation of electrical devices the new fuse intervenes, repeat the diagnosis and replacement procedure as described previously.

ATTENTION

Use only fuses with the amperage and characteristics indicated in the wiring diagram.

The use of fuses other than those specified in the wiring diagram, void the warranty and can cause the risk of damage the machine.

4.4.3 Compressed air connection

In order to the equipment to work properly, it requires a connection to the medical air compressed system and another to the oil and water residue-free technical compressed air system.

The characteristics of the air must be in conformity with the indications specified in the installation plan.

NB. If technical air is used, an 0.2 micron bacteriological filter must be installed as well as a device to eliminate humidity from the air.

4.4.4 Plumbing connection

The installation diagrams must be strictly complied with, paying particular attention to the diameter of the pipes which must be no less than that of the corresponding fittings. Be sure to use hoses suitable for the supply fluid characteristics and of an appropriate length to facilitate their installation.

The water introduced in the chamber is controlled by two flow meters (in order to have better control of the process).

Comply strictly with the following instructions:

- this machine must be connected to the water mains in compliance with current legislation;
- the water non-return system is installed inside the machine;
- for the connection use cocks with a male $\frac{3}{4}$ " fitting (ND 20 mm) installed where they are easy to reach (behind the machine in the one-door version, inside the machine in the 2-door version);
- ensure that mains water pressure is between 150 and 500 kPa;
- if mains pressure is higher than 5 bar (500 kPa) a pressure reducer must be installed (optional);
- if pressure is less than 1.5 bar (150 kPa) a pressure increase pump must be installed;
- each machine is delivered with hoses for filling the water with a $\frac{3}{4}$ " F threaded fitting;
- the unit is set up to use cold and mixed water (mixer optional) 35°C/50°C depends the type of chemicals;
- do not shorten or damage the hoses supplied with the machine;
- use only the pipes included with the machine;

When the unit is not working always close the supply cocks.

Comply with the prescribed pressures and temperatures. Water supply pipes with stainless steel meshed reinforcements are to be used.

4.4.5 Connection to the drain

- The drain pipes must be able to withstand temperatures of up to 100°C as well as chemical and biological agents.
- Since the machine drainage system is hermetically closed at the end of each cycle (pneumatic valve) it requires no other particular device for a floor or wall drain. If a siphon is used, only for model equipped with draining pump, it must be possible to remove it for cleaning or it must be equipped with an accessible opening to do the cleaning.
- The unit is pre-set for a 1" female connection on the bottom.
- The drain must have a minimum flow rate of 25 l/min.

The drain must be made in compliance with international directives. No liability shall be accepted should improper use be the source of pollution. If the drain is clogged, be very careful when treating the water and avoid contact with hands, eyes, etc. In the case of contact wash the parts with plenty of water. Use the PPE.

4.4.6 Connecting the chemicals

The system uses specific chemical compounds designed to obtain properly reprocessed devices, therefore, any other type of chemical compound will not guarantee these results or its compatibility with the machine.

Standard machine accessories are 3 pumps for dosing the chemicals and 6 [flowmeters](#) (two for each chemical in order to have better control of the process). A dosing unit is dedicated to the detergent, another to the chemical disinfectant and a third one to a different type of chemical, for example [using a bicomponent disinfectant or rinse aid](#). The fourth chemical is available as an optional. The chemicals dispensing system consists of tanks of product (those supplied by the Manufacturer), suction devices, pumps and flow meters.


- Check the technical data sheets of the chemicals used and do not exceed the recommended concentrations.
- Comply with the instructions given on the technical data and safety sheets for the environment and on storage methods.
- The unit checks that there are chemicals at each cycle (the presence of detergent is checked by a level sensor and RFID control while the peracetic acid is checked by a [scale](#) and RFID control). A warning is given if no chemicals are found.
- Access to the chemicals is only allowed when they need replacing and is to be done when the unit is in the no-cycle condition and by using PPE devices.
- The topping up or pouring of chemical from one tank to another is not allowed; by means of RFID technology, the system checks the batch, the expiry date and the original quantity of such chemical, therefore blocking any attempt to use combinations of different or expired batches and /or to carry out any top-ups.
- Before ever replacing chemicals, wear suitable personal protective gear; use extreme caution and carefully read the technical data and safety information sheets of the chemical in question.
- Chemicals must only be replaced by personnel who have been instructed about their inherent risks.
- Chemicals can be flammable. Refer to the instructions in the technical data and safety information sheets of each product.
- The unit's performance is subject to the use of chemicals supplied by the Manufacturer of the unit. The Manufacturer cannot guarantee that the unit will provide the same performance with different products.

4.4.6.1 Replacement of chemical product container

To replace the chemical product container perform the following procedure:

For the chemical product replacement, it is better to move the product container out of the machine.

- Take the new product container and open the chemical compartment.
- Replace the chemical product container removing the level sensor from the empty tank and put into the new one.
- Close the topper of the chemical product container and place it in the area for the storage of chemical substances.
- Close the chemical compartment.





ATTENTION

The chemical product that is used is dangerous if touched or inhaled.


Before the use, read carefully the safety information supplied by the manufacturer of the chemical product and the label on the package.

During the operations of replacement of chemical product container, use the appropriate tools for individual protection (chemical protective gloves, face masks for breathing, etc.).


The access to the technical compartment, where are located the chemical product containers, is permitted only with keys and to the authorized personal.

4.4.6.2 Water intake filter installation

1. Assemble and fasten the 0.45 μm filter onto the right coupling, making sure that the arrows on the filter are facing upwards;




2. Assemble and fasten the 0.1 μm filter onto the left coupling, making sure that the arrows on the filter are facing downwards;



3. Connect the 2 filters by using the pipe supplied;



4. Assemble and fasten the 3 x 45 μm filters, making sure that the arrows are facing upwards;



4.4.6.3 Calibration procedure

Required equipment:

No. 1 graduated cylinder with capacity about 300ml, with graduated scale and maximum graduation of 2 ml.

Then follow the instructions below:

1. From Main Menu push the button “**MAINTENANCE**” (pic 5.1 A);
2. Enter the access password (pic.5.1 B);
3. Push the button “**CALIBRATION**” on the Maintenance Menu” (pic.5.1.1);
4. Click the tab “**CALIBRATION FLOW METER**” right on the top of the page. Open the table of figure 5.1.2;
5. Choose the chemical or type of water to dispense, e.g. disinfectant (see point 1 of picture 5.2). [The chemical](#) chosen will appear in blue;
6. Set the number of control impulses (we recommend 300) in the input field “**IMP.PROG.**” (see point 2 of pic.5.2);
7. Open the manual valve in the corresponding chemicals’ compartment (in this example it is yellow – pic. 5.2.1 point 1) remembering to plug in the CPC male connector to open the safety valve which is under the Teflon valve supplied with the machine (one per valve) pic. 5.2.2. if you want to check the filling up water you will have to remove the pipe from the right of the machine, see pic.5.2.2.1;
8. Place a graduated cylinder under the open valve (to make this easier and avoid any chemical leaks, insert a long enough pipe in the CPC container, supplied with the machine, to enter the container) pic.5.2.2;
9. Press the “**DOSING**” key and wait until the machine has reached the limit set (point 3 pic.5.2); a difference in the values on the display of the two dispensers is normal (point 4 Pic. 5.2).
10. Check the quantity of chemical product dispensed in the graduated cylinder and input the value in the field “**QUANT.ML.**” (see point 5 of pic.5.2);
11. Press “**CONFIRM**” button (see point 6 of pic.5.2)
12. Repeat this twice to 3 times or until the dispensing in millilitres are the same or have a maximum difference of 2 ml (normally the first time is to fill the line correctly and the second and third to set the parameters correctly). In the event of an error, press the (see point 7 of pic.5.2).
13. [At the end of the calibration procedure, close the manual Teflon valve and remove the CPC male connector to close the safety valve which is under the manual Teflon valve.](#)

ATTENTION


For the normal operating conditions, the manual Teflon valve must be in closed position and without the male CPC connector.




Closed valve position




CALIBRATION OF CHEMICAL PRODUCTS WATER USED TO FILL THE WASHING CHAMBER




Pic. 5.1 A




Pic. 5.1 B



Pic. 5.1.1



Pic. 5.1.2




Pic. 5.2

Closed valve position



Pic. 5.2.2

Turn clockwise to open



Pic. 5.2.1




4.5 Connectivity

This device is equipped with a router NAT, which allows the connection to an external network. This device allows to provide just an IP address to the device, on which you must forward all the requests addressed to itself. That IP address is associated to the door marked by the *internet* sign. The door 1 (*internet*) must be connected to the customer network, mentre while the doors from 2 to 5 are reserved for the internal network to the device. The network data configuration inside the router is performed just with the software provided by Steelco. (Please see the Specification of Router NAT programming inside annex E).

5. CHECKING THE PARAMETERS AND SCREEN PAGE FUNCTIONS ON THE MONITOR

Control panel (Loading side)

The control panel is composed of a “Touch-screen” monitor and a membrane keypad with six setting keys. The monitor, depending on the programme/management status position, displays keys that can be selected by pressing the display area and which control specific control functions.





For the 6-key keypad, the functions are the following:

REFERENCE	DESCRIPTION
1	Selection for the access to functions for the adjustment of monitor display parameters: <ul style="list-style-type: none"> • Contrast • Brightness • Etc.
2	Always goes back to the previous page.
3	Always goes back to the main menu.
4	Reset the acoustic alarm signal.
5	Door closing.
6	Door opening.

Control panel (Unloading side - if present)

The control panel is composed of a “Touch-screen” monitor with the setting keys that allow to open/close the unloading door and reset the alarms.

Emergencies and checks

	<p>MAIN POWER SUPPLY SWITCH (optional)</p> <p>The machine is equipped with a main ON-OFF switch (optional), located on the dirty side, which disconnects the electrical power supply from all the auxiliary controls.</p>
	<p>EMERGENCY PUSH BUTTONS (optional)</p> <p>The machine is equipped with two emergency buttons, not automatically reset kind:</p> <ul style="list-style-type: none"> • Nr. 1 emergency button located on loading side; • Nr. 1 emergency button located on unloading side (in case of machine with pass-through door). <p>The red coloured emergency push-button, is marked with the word “EMERGENCY”. It is located in an easily accessible area but sufficiently protected against accidental activation.</p>

The EW 2 system already has the basic parameters stored, entered during the final inspection and test in the factory; they are usually valid for the majority of installations and, therefore, do not need to be changed. If they do need changing, follow the instructions below:


OPERATOR AREA

PARAMETERS DISPLAYABLE WITH THE USER PASSWORD (GREEN)

Password: call Steelco for technical assistance


5.1 Main maintenance area parameters

Press the “**MAINTENANCE**” key on the main screen page to gain access to the menu



Pic. 5.3

Enter the user code.



Pic.5.4

5.2 Clock and date update

To change the time and date press the “**CLOCK UPDATE**” key.




Pic. 5.5


Now press on the values to change and change them on the keypad that appears (sequence from Pic.5.6 – 5.7 – 5.8.).




Pic. 5.6



Pic. 5.7



Pic. 5.8



5.3 Restoring the parameters saved

To restore the backup parameters, press the “RESTORE” key on the menu.


Press “RESTORE” to restore the data saved or “MANUFACT. RESTORE” to restore the factory default parameters; confirm and wait. If ok press



to go to the previous page. Repeat if not ok.



Pic. 5.9



Pic. 5.9


5.4 Manual control of the mechanics

To operate the EW2 system manually press **“MAN.CONTROL”** on the menu.



Pic. 5.11


System status can be seen in the first screen page that appears (Pic.5.12 system working properly) and it is possible to go from the automatic to the manual phase by pressing on the word **“AUTOMATIC”**.



Pic. 5.12


The various **“FUNCTION”** screen pages can be scrolled by selecting the indexes at the top of the page.

Whether working in the manual or automatic mode, this screen page lets you see the status of the system's inputs; a green led on corresponds to an active input to the PLC.




Pic. 5.13

Whether working in the manual or automatic mode, this screen page lets you see the status of the outputs; a red led on corresponds to an active output from the plc while in the manual mode it is possible to activate or deactivate the corresponding function and test it (e.g. no. 177, the pump starts by pressing on "DRAIN PUMP").



Pic.5.14

Whether working in the manual or automatic mode, this screen page lets you see the temperature of the system's probes, the water conductivity and quantity in litres/min on the rotors. The purpose of the value by the side of the chamber temperatures, manually variable, is for setting the temperature of the chamber's heating elements to be activated on the "STATUS GRAPH." screen page by pressing on the drawing of the heating elements. this of course after having put at least 12 litres of water in the chamber. This function is normally used only to bring the chemicals up to the right temperature for manually washing the chamber (anti-limescale).



Pic. 5.15 - "PT1000+PUMPS FL"

Whether working in the manual or automatic mode, this screen page only lets you see the status of the pressures and flows in real time inside the channels connected and the leak test.


By pressing on the arrow at the bottom right you go from the pressure to the flow screen page.

19 ALARM !! MACHINES STATUS 26/09/14 10:45					
STATUS PLC	STATUS INPUT	STATUS OUTPUT	PT1000+ PUMPS FL	CHANNELS PR.+FL.	STATUS GRAPH
Press.transducer	mbar	Set 1	Set 2	Set 3	
Colector:	0	0	0		
Channel 3:	0	0	0		
Channel 4:	0	0	0		
Leak test:	3	0	0		
Leak test tank:				0	

19 ALARM !! MACHINES STATUS 05/09/14 09:41					
STATUS PLC	STATUS INPUT	STATUS OUTPUT	PT1000+ PUMPS FL	CHANNELS PR.+FL.	STATUS GRAPH
flow transducer			Instr.1	Instr.2	
Channel 1 flow ml/'	:	0		0	
Channel 2 flow ml/'	:	0		0	
Channel 3 flow ml/'	:	0		0	
Channel 4 flow ml/'	:	0		0	
Channel 5 flow ml/'	:	0		0	
Channel 6 flow ml/'	:	0		0	
Channel 7 flow ml/'	:	0		0	

Pic.5.16 – Pic.5.17 “CHANNELS PR.+FL.”

With this screen page you can interact with the system manually. Here it is possible to manually activate or deactivate a series of pumps, valves and heating elements, displaying their values. Another 4 control screen pages can be accessed from this one: on the left **“LEAK/ PURG.”** PIC. 5.19, on the right **“WATER/CHEM.”** PIC. 5.20, below **“DRAINS PUMP”/COMPRESSED, AIR** PIC.5.21, above **“CHANNELS”** PIC. 5.22 – (see table).




REFERENCE	DEVICE
1	Leak test pressure instr.1
2	Leak test pressure instr.2
3	Basket lock
4	L/Min. spiral/w.arm water
5	L/Min. w.arm water
6	Chamber heating elements
7	Quantity of water in the washing chamber
8	Machine cold point temperature
9	Chamber temperature control sensor
10	Chamber temperature adjustment sensor
11	Drying blower
12	Drying heating elements
13	Chamber opening valve for drying
14	Blower temperature

Pic. 5.18

With this screen page you can interact with the system manually. Here it is possible to manually activate or deactivate the valves and pump for the leak test and for flushing to conduct the tests, displaying the values. Another 3 control screen pages can be accessed from this one: on the right "MACHINE." – PIC. 5.18, below "DRAINS PUMP"/COMPRESSED AIR" PIC.5.21, above "CHANNELS" PIC. 5.22 – (see table).

1 2 3 4

Press to change from automatic to manual




REFERENCE	DEVICE
1	Leak test pump
2	Pressure value of leak test pump
3	Air loading valve of instrument 1
4	Instrument 1 pressure value
5	Air draining valve of leak test cylinder
6	Leak test cycl. expansion
7	Air loading valve of instrument 2
8	Instrument 2 pressure value

Pic. 5.19

5 6 7 8


With this screen page you can interact with the system manually. Here it is possible to manually activate or deactivate the valves and chemical pumps for conducting tests, displaying the values. Another 3 control screen pages can be accessed from this one: on the left **"MACHINE"** – PIC. 5.18, below **"PUMPS/DRAIN"** – PIC.5.21, above **"CHANNELS"** – PIC. 5.22 – (see table)



Pic. 5.20

REFERENCE	DEVICE
1	Displays of the quantity of water at inlet to the chamber (dual control system)
2	Flowmeters to control water loading in the chamber (dual control system)
3	0.1 micron water filter
4	0.45 micron water filter
5	Mains water filling valve
6	Chemical 3 loading control flowmeters (optional – dual control system)
7	Chemical 3 load pump (optional)
8	Displays of the quantity of chemical 1 (disinfectant) at inlet to the chamber (dual control system)
9	Flowmeters to control chemical 1 loading (disinfectant) in the chamber (dual control system)
10	Chemical 1 loading pump (disinfectant)
11	Displays of the quantity of chemical 2 (detergent) at inlet to the chamber (dual control system)
12	Chemical 2 loading pump (detergent)
13	Flowmeters to control chemical 2 loading (detergent) in the chamber (dual control system)
14	Displays of the quantity of chemical 3 (NEUTRALISER/ACTUATOR) at inlet to the chamber (dual control system)


With this screen page you can interact with the system manually. Here it is possible to manually activate or deactivate the valves and pumps that interact with chamber draining, spirals and w.arms, and the spurg with compressed air and channels, for conducting tests, displaying the values. Another 3 control screen pages can be accessed from this one: above “**MACHINE.**” – PIC. 5.18, left “**LEAK/PURG**” – PIC.5.19, right “**WATER/CHEM.**” – PIC. 5.20 – (see table).



Pic. 5.21

REFERENCE	DEVICE
1	Channels valve purging of instrument 1
2	Channels valve purging of instrument 2
3	Pneumatic valve of recicular pump
4	Channel separation valve
5	Flowmeter to control washing arms water
6	Tank heating element
7	Displays of water temperature of tank
8	Displays of water quantity of tank
9	Pneumatic valve, sterilisation filter/supply water circuit
10	Pressure sensor of medical air
11	Purging valve
12	Recirculation pump
13	Pneumatic draining valve

With this screen page you can interact with the system manually. Here it is possible to manually activate or deactivate the endoscopic channel pumps for conducting tests, displaying the values. To activate them the spiral and rotor water valves must be active, the door(s) must be closed and the basket locking pistons active besides having the 12 litres of water in the washing chamber (if required for the tests). Another 3 control screen pages can be accessed from this one: under “**MACHINE**” – PIC. 5.18, left “**LEAK/PURG.**” – PIC.5.19, right “**WATER/CHEM.**” – PIC. 5.20 – (see table).




REFERENCE	DEVICE
1	Display of channels by colour, instrument position 1
2	Pumps channels 3 to 4 instrument 1
3	Display in real time of the channels flow of instrument position 1
4	Display in real time of the channels flow of instrument position 2
5	Pumps channels 3 to 4 instrument 2
6	Display of channels by colour, instrument position 2

Pic. 5.22

5.5 Cycles historical visualization

To display the “CYCLES HISTOR.” press the button with the same name on the “MANUAL CONTROL MENU”.




Pic. 5.23


There are two possible ways to search for cycles:

- Enter the data of the cycle(s) you want to display (point 1)
- OR**
- Enter the cycle number, if known, to display (e.g. if you want to display cycle 90 enter number 90 at points 2 and 3).

THEN PRESS THE “READ” KEY



Pic. 5.24



The screen page that appears will be like the one shown in figure 5.25. by selecting the cycle wanted, scrolling with the arrows (point 4) and pressing the “DETAILS” key, you will be able to see and print all the parameters of the cycle selected (see sequence in figures 5.26 and 5.27) in figure 5.25.

“CANCEL” key: press it to return to the screen page of figure 5.23.

“LIST” key shows a list of all the cycles found in the time period requested


“PRINT” key prints a list of the cycles found in the time period requested.

By pressing “DETAILS” button it is displayed the picture 5.26.

REFERENCE	DESCRIPTION
1	It displays the date and time of the start and finish of the cycle selected
2	It displays how many cycles have been carried out on the dates entered for the search
3	It displays the number in sequence of the cycle carried out on the dates entered for the search
4	Arrows to select the cycle
5	Result of the cycle selected
6	Number of the cycle selected (not the progressive number of the cycle)
7	Type of programme executed for the cycle selected

Pic. 5.25

All the parameters stored by the machine can be seen on this screen page (one set every 4 seconds – standard setting):




REFERENCE	DESCRIPTION
1	Cycle number (progressive).
2	“DATA PROGR.” it shows the sequence of the cycle phases.
3	“EVENTS” It shows all the events that occur in the cycle performed.
4	“SAMPLES” it shows sampling time and the two chamber temperatures.
5	“CONSUMPT.” it shows the consumption of water and chemicals during the cycle.
6	Type of material used for the cycle.
7	Serial number and model of the reprocessed instrument.

TO CREATE A PRINTOUT OF THE CYCLE, PRESS THE “STAMPA DATI” BUTTON (POINT 8) AND PROCEED AS IN FIGURE 5.27.

By pressing the “CHEMICAL” button (point 9) you can see the activation status of the control rfid for the chemicals during the cycle, (point 10) data position second instrument (in the example given the cycle did not have the second instrument in the chamber).

Pic. 5.26

You can select what you want to print on this screen page; by touching the corresponding white boxes you can print the following information:




REFERENCE	DESCRIPTION
1	It prints a summary of the cycle phases and information about the instruments reprocessed, date, cycle type, etc.
2	It prints cycle detail plus consumption and times.
3	It prints the consumption of chemicals in detail.
4	It prints any alarms that may have occurred in the cycle and the reference phases.
5	It prints the consumption of water in the various phases of the
6	It resets printing and goes back to the screen page in figure 6.26.
7	It prints two lines for each point set on point 10, resuming pressure and flow of each channel (a reading every 4 seconds).
8	it lets you choose from which point stored to print the data gathered in the cycle (one every 4 seconds).
9	It displays the number of points stored in the cycle (one every 4 seconds); keeping within these values you can change them in order to select where you want printing to start. For instance, if you want to print just the central phases of the cycle, simply change number 1 with number 300 and number 834 with number 600 (when you touch the number a keypad appears on which you enter the value wanted).
10	See point 7.

Pic. 5.26


5.6 Parameter configuration

As far as concerns parameter configuration, proceed as explained below: press the “**SETUP**” key on the main screen page to gain access to the menu.



Pic. 5.28

Enter the user code (Green password).



Pic. 5.29

5.7 User data configuration


You gain access to the user data from the configuration menu; to see them press the “USER DATA” key (pic. 5.30) you gain access to the screen page of figure 5.31.

NB. THE PARAMETERS DISPLAYED ON THE MACHINE ARE THE MANUFACTURER DEFAULT SETTINGS



Pic. 5.30

As already mentioned, the EW2 system is programmed in the factory and some functions might not be active depending on the model. the following settings can be made on this screen page: **TO CHANGE THE PARAMETER, PRESS ON THE WORDING INSIDE THE WHITE WINDOW CORRESPONDING TO THE PARAMETER TO CHANGE.**




Pic. 5.31

REFERENCE	DESCRIPTION	
1	THIS PARAMETER MAKES IT POSSIBLE TO SET THE UNLOADING DOOR. THE POSSIBLE OPTIONS INCLUDE:	
	CYCLE OK	IT IS POSSIBLE TO OPEN THE DOOR ONLY WHEN THE CYCLE HAS BEEN CARRIED OUT CORRECTLY
	ALWAYS	IT IS POSSIBLE TO ALWAYS OPEN THE DOOR AT THE END OF THE CYCLE (IT IS NOT ADVISABLE TO CHOOSE THIS PARAMETER)
2	THIS PARAMETER MAKES IT POSSIBLE TO SET THE AUTOMATIC OPENING OF THE DOOR AT THE END OF THE CYCLE. THE POSSIBLE OPTIONS INCLUDE:	
	CYCLE OK	THE AUTOMATIC OPENING TAKES PLACE ONLY IF THE CYCLE HAS BEEN CARRIED OUT CORRECTLY
	ALWAYS	THE AUTOMATIC OPENING ALWAYS TAKES PLACE (IT IS NOT ADVISABLE TO CHOOSE THIS PARAMETER)
	NO	THE AUTOMATIC OPENING HAS NOT BEEN ACTIVATED
3	THIS OPTION MAKES IT POSSIBLE TO SET THE DATA ENTRY MODE (WE RECOMMEND KEEPING THE FACTORY DEFAULT SETTING). POSSIBLE OPTIONS INCLUDE:	
	INST.ID	THE DATA CAN ONLY BE ENTERED WITH THE BARCODE READER (WITH THE BARCODE READER INSTALLED)
	KEYBOARD	DATA CAN ONLY BE ENTERED MANUALLY VIA THE KEYBOARD
	MIXED	IT IS POSSIBLE TO ENTER DATA WITH BOTH THE KEYBOARD AND THE BARCODE READER
4	SET LIKE THIS THE SYSTEM PROMPTS THE OPERATOR ID BEFORE STARTING THE CYCLE; POSITIONING IT ON NO (SIMPLY TOUCH THE WORD YES TO CHANGE TO NO), THE CYCLE STARTS EVEN WITHOUT THE OPERATOR ID.	
5	IF PRESSED AFTER HAVING CHANGED A PARAMETER IT WILL SAVE THE NEW CONFIGURATION OTHERWISE THE PARAMETER WILL NOT BE CHANGED.	
6	SET LIKE THIS THE OPERATOR CAN SKIP THE CYCLE PHASES AND GO TO THE END OF THE CYCLE AFTER START-UP.	
7	THIS PARAMETER MAKES IT POSSIBLE TO ENABLE THE RFID OR THE BARCODE READER (WE RECOMMEND KEEPING THE FACTORY DEFAULT SETTING). POSSIBLE OPTIONS INCLUDE:	
	TASTE	THIS PARAMETER ENABLES DATA TO BE ENTERED ONLY BY USING THE KEYBOARD
	MIXED	THIS PARAMETER ENABLES DATA TO BE ENTERED BY USING BOTH THE KEYBOARD AND THE BARCODE/RFID READER
8	IN THIS WAY THE SELF-SANITISATION CYCLE CAN BE STARTED MANUALLY. IT WILL BE NECESSARY TO CHANGE “NO” TO “YES” EVERY TIME THAT YOU WANT TO START THE AUTOMATIC SELF-DISINFECTION OR PROGRAMMED DURING THE WEEK “DISINF. BY TIME” (POINT 11); TO ACTIVATE THE FUNCTION AT AN HOUR OR DAY DESIRED, ENTER THE “GREEN” OR “YELLOW” OPERATOR OR TECHNICIAN ID CODE ON THE “UTILITY” MASK ON THE MAIN MENU.	

9	POINT FOR CHANGING THE “ GREEN ” OPERATOR PASSWORD, KEYING THE NEW PASSWORD IN TWICE ON THE ALPHANUMERICAL KEYPAD THAT APPEARS. AFTER PRESSING ON THE ASTERISKS AND SAVING THE DATA (POINT 5) THE PASSWORD IS CHANGED.
10	BY PRESSING THIS KEY, YOU PRINT ALL THE USER DATA SET.

The following settings are possible from this screen page: to change a parameter press on the wording inside the white box that corresponds to the parameter to change and then press “save data” to store the changes made.
THE INFORMATION TO PRINT IS DEFINED ON THIS SCREEN PAGE.



Pic. 5.32

REFERENCE	DESCRIPTION
1	IT PRINTS THE EVENTS OF THE CYCLE, THE ALARMS, WATER FILLING, LEAK TEST RESETTING, ETC.
2	IT ENABLES PROGRESSIVE PRINTING DURING THE CYCLE.
3	SPEED OF THE PAPER WHEN PRINTING; AS IT IS THERMAL PAPER, THE QUICKER THE PRINTING THE WORSE THE QUALITY.
4	IT ENABLES THE PRINTER INTEGRATED IN THE MACHINE.
5	IF PRESSED AFTER HAVING CHANGED A PARAMETER IT WILL SAVE THE NEW CONFIGURATION OTHERWISE THE PARAMETER WILL NOT BE CHANGED.
6	IT PRINTS THE EVENTS OF THE CYCLE.
7	IT PRINTS THE QUANTITY OF WATER AND CHEMICALS DIVIDED BY CYCLE PHASES.
8	IT PRINTS IN DETAIL THE VALUES OF THE CHEMICAL USED PER CYCLE DIVIDED BY SENSOR.
9	IT CREATES A SUMMARISED COPY OF THE CYCLE WHICH IS NORMALLY ATTACHED TO THE PATIENT'S MEDICAL RECORDS.
10	BY PRESSING THIS KEY YOU PRINT ALL THE USER DATA SET.


The following settings are possible on this screen page: **to change a parameter press on the wording inside the white box that corresponds to the parameter to change and then press save data to store the changes made.**



Pic. 5.33

REFERENCE	DESCRIPTION	
1	THIS PARAMETER MAKES IT POSSIBLE TO SET THE MACHINE BEHAVIOUR IN CASE IT IS NECESSARY TO MANUALLY FORCE A STAGE OF THE PROCESS OR WHEN AN ALARM HAS BEEN TRIGGERED DURING THE CYCLE. POSSIBLE OPTIONS INCLUDE:	
	END CYCLE	THE MACHINE SKIPS A STAGE AT THE END OF THE CYCLE BY CARRYING OUT A RINSING STAGE WITH ONLY WATER IN ORDER TO REMOVE THE CHEMICAL SUBSTANCES FROM BOTH THE CHAMBER AND THE INSTRUMENT
	CONTINUE	THE MACHINE WILL SKIP THE STAGE IN PROGRESS IN ORDER TO PASS ONTO THE FOLLOWING ONE, OR IN THE EVENT OF AN ALARM RESET, IT WILL CONTINUE THIS STAGE (IT IS NOT ADVISABLE TO CHOOSE THIS PARAMETER)
2	THIS PARAMETER MAKES IT POSSIBLE TO SET THE MACHINE BEHAVIOUR IN CASE A POWER FAILURE OCCURS DURING THE CYCLE. POSSIBLE OPTIONS INCLUDE:	
	END CYCLE	THE MACHINE SKIPS A STAGE AT THE END OF THE CYCLE BY CARRYING OUT A RINSING STAGE WITH ONLY WATER IN ORDER TO REMOVE THE CHEMICAL SUBSTANCES FROM BOTH THE CHAMBER AND THE INSTRUMENT
	CONTINUE	THE MACHINE WILL CONTINUE TO CARRY OUT THE CYCLE FROM WHERE IT WAS INTERRUPTED (IT IS NOT ADVISABLE TO CHOOSE THIS PARAMETER)
3	IF POSITIONED ON “YES” THE SYSTEM PROMPTS THE OPERATOR CODE TO RESET THE ALARM, OPEN THE DOOR OR ABORT THE CYCLE.	
4	IF POSITIONED ON “YES” IT IS POSSIBLE TO ENABLE THE ACOUSTIC SIGNAL DURING THE MACHINE FUNCTIONING (ALARM).	
5	THIS OPTION APPEARS ONLY IF THE BARCODE READER OR RFID ARE ACTIVATED; IT IS THE OPERATOR’S ID PREFIX TO BE PUT EITHER ON THE BARCODE OR TAG.	
6	THIS OPTION APPEARS ONLY IF THE BARCODE READER OR RFID ARE ACTIVATED; IT IS THE INSTRUMENTS’ ID PREFIX TO BE PUT EITHER ON THE BARCODE OR TAG.	
7	NOT ACTIVATED	
8	ALL DATA ARE PRINTED BY PRESSING THIS KEY.	
9	IF PRESSED AFTER HAVING CHANGED A PARAMETER IT WILL SAVE THE NEW CONFIGURATION OTHERWISE THE PARAMETER WILL NOT BE CHANGED.	


The following settings are possible on this screen page: **to change a parameter press on the wording inside the white box that corresponds to the parameter to change and then press save data to store the changes made.**



REFERENCE	DESCRIPTION
1	With the parameter setting on “NO” it is not possible to modify the parameters of saved endoscopes and memorize the datas present on machine memory.
2	Door opening sensor.

Pic. 5.34

The following settings are possible on this screen page: **to change a parameter press on the wording inside the white box that corresponds to the parameter to change and then press save data to store the changes made.**




The names of the chemicals and waters used in the machine can be changed on this screen page:

	CHEMICAL 1	CAP	CHEMICAL 2	CAP	CHEMICAL 3	CAP
PERACETIC ACID	Neodisher Septo PAC	yellow	Neodisher SC	blue	Neodisher MEDIKLAR (opt.)	black
GLUTARALDEHYDE	Neodisher Septo GDA	yellow	Neodisher SC	blue	Neodisher MEDIKLAR (opt.)	black
HYDROGEN PEROXIDE	SteelcoXide-B	yellow	SteelcoXide-DT	blue	SteelcoXide-A	red

Pic. 5.35

5.8 Thermaldisinfection cycle

The following settings are possible on this screen page: to change a parameter press on the wording inside the white box that corresponds to the parameter to change and then press save data to store the changes made.



Granted that it is possible to carry out a self-disinfection cycle also manually, the parameters for automatic cycles can be set on this screen page.

REFERENCE	DESCRIPTION
1	Example of setting the time and day of the week to automatically start the self-disinfection cycle. in this example, if the start-up conditions have been met (doors closed, leak test closed and basket introduced). the automatic disinfection cycle will start at 2.30 in the morning.
2	Number 10 given in the example indicates the minutes the machine waits after the time set. in effect if, by chance, there is a blackout or an interruption in the water supply, etc.. The machine will wait, besides the time set, the minutes given in the box in point 3.
3	All data are printed by pressing this key.

Pic. 5.36

CAUTION!


BEFORE CARRYING OUT THE SELF-DISINFECTION CYCLE, EITHER MANUALLY OR SET TO START AUTOMATICALLY, CLOSE THE LEAK TESTS WITH THE CAPS SUPPLIED AND REMOVE ALL THE INSTRUMENTS AND ACCESSORIES FROM THE CHAMBER.

5.9 Self-disinfection procedure with a 12% sodium hypochlorite solution


According to ISO 15883-4, a self-disinfection cycle shall be provided to ensure that the endoscope washer does not become a focus for contamination of the load, and also to provide a means for disinfecting it after interventions for maintenance, repair or testing. Thermal disinfection using moist heat is the preferred method.

Steelco recommends to run the thermal self-disinfection cycle every day to make sure that the chamber of the unit, the overall hydraulic circuit and the bacteriological filters (0,45 + 0,1 micron) are disinfected on a daily basis. The thermal self-disinfection cycle has been designed in order to reach an A0 value of at least 600.

This procedure should be performed in the system only in the event of suspected contamination of the machine with the prion responsible for the transmission of spongiform encephalopathies, such as the new variant of Creutzfeldt Jakob disease. This procedure must not be used for periodic machine self-disinfection procedures.

	<p>ATTENTION</p> <p>The efficiency of this procedure as regards the treatment of the new variant of Creutzfeldt Jakob disease has not yet been proven. However, this is a procedure that has been recommended by several opinion-makers and research groups in the event of the outbreak of such infection.</p> <p>It has also been recommended to destroy all the materials that have come into contact with the fluid.</p> <p>In the event of a suspected new variant of Creutzfeldt Jakob disease, local recommendations and regulations must be complied with.</p>
---	---

- Open the machine door, remove and eliminate the transparent leak test pipe including the machine connector. The disposal of the (potentially contaminated) transparent leak test pipe and the 0.45 and 0.1 micron water filters must be carried out in compliance with the relevant hospital procedure. Check that all the other channel tubes are present in the chamber.
- Insert the leak test cap onto the appropriate chamber connector.
- Replace the disinfectant tank with the one containing the 12% sodium hypochlorite solution in order to obtain a 3% working concentration.
- Start the prion programme in order to achieve a 60-minute contact time as recommended in Ministerial Circular n° 138 issued by the French Directorate General for Health.
- At the end of the programme, remove the leak test cap and insert a new transparent pipe together with a new leak test connector in the chamber.


	<p>ATTENTION</p> <p>Make sure that all materials to be connected, such as the channel separators, have been reprocessed during this procedure.</p>
---	---

5.10 Table management

Access is gained to table management from the setup menu. to see them press the “**TABLES MANAG.**” key (pic.5.37); you will access the screen page of picture 5.38.



Pic. 5.37




REFERENCE	DESCRIPTION
1	By pressing the “ENDOSCOPE TYPE” key you gain access to the screen page for entering the models of the various endoscopic instruments (e.g. gastroscopes, bronchoscopes, etc.), pic. 5.39.
2	By pressing the “MANUFACTURERS” key you gain access to the screen page for entering the manufacturers’ names (Storz, Olympus, etc.), pic. 5.41.

Pic. 5.38

5.10.1 Endoscooper type table

The following settings are possible on this screen page: **to change a parameter press on the wording inside the white box that corresponds to the parameter to change and then press save data to store the changes made.**




REFERENCE	DESCRIPTION
1	It gives the position (on the progressive list – see pic. 5.40) where the text added in point 2 will be stored.
2	Space reserved for entering the type of endoscope you wish to store (colonscope, bronchoscope, etc.).
3	If pressed it displays the list of endoscope types entered (see pic.5.40).
4	If pressed it prints all the data of the endoscope table.

ALL WHAT IS STORED IN THIS TABLE WILL THEN BE USED AS ONE OF THE DETAILS FOR CREATING NEW ENDOSCOPIC INSTRUMENTS.

Pic. 5.39

All the types of enscope entered and their positions can be seen on this screen page.



The screenshot shows a window titled 'ENDOSCOPES TYPE TABLE' with a timestamp of '20/09/14 09:22'. Below the title bar is a header 'ENDOSCOPES TYPES LIST' with a red 'X' icon. The main content is a table with two columns: 'nr.' and 'description'. The table contains the following data:


nr.	description
0001	GASTROSCOPE
0002	COLONSCOPE
0003	DUODENOSCOPE
0004	ECOENDOSCOPE
0005	BRONCSCOPE
0006	
0007	
0008	

To change the parameters, press twice on the writing inside the “**DESCRIPTION**” window corresponding to the parameter to change. you return automatically to pic. 5.39.

Pic. 5.40

5.10.2 Manufacturers table

The following settings are possible on this screen page: **to change a parameter press on the wording inside the white box that corresponds to the parameter to change and then press save data to store the changes made.**



REFERENCE	DESCRIPTION
1	It gives the position (on the progressive list – see pic.5.42) where the text added in point 2 will be stored.
2	Space reserved for entering the name of the manufacturer you wish to store (olympus, pentax etc.).
3	If pressed it displays the list of manufacturers entered.
4	If pressed it prints all the data of the manufacturers table.

ALL WHAT IS STORED IN THIS TABLE WILL THEN BE USED AS ONE OF THE DETAILS FOR CREATING NEW ENDOSCOPIC INSTRUMENTS.

Pic. 5.41

Enter the names of the endoscope manufacturers and their positions can be seen on this screen page.



nr.	description
0001	FUJINON
0002	OLYMPUS
0003	STORZ
0004	PENTAX
0005	MACHITA
0006	
0007	
0008	

To change the parameters, press twice on the writing inside the “DESCRIPTION” window corresponding to the parameter to change. you return automatically to pic. 5.41.

Pic. 5.42


5.11 Operator setup

To enter the names of the operators who will be interacting with the unit, press the “**OPERATORS**” key of pic. 5.43 still on the “**SETUP**” menu and a screen page like the one in pic. 5.44 will appear.



Pic. 5.43

The following settings are possible on this screen page: **to change a parameter press on the wording inside the white box that corresponds to the parameter to change and then press save data to store the changes made.**




REFERENCE	DESCRIPTION
1	It gives the position (on the progressive list – see pic.5.45) where the text added in point 3 will be stored. Max. 300
2	It indicates where the operator access code has to be entered.
3	Space reserved for entering the name and last name of the operator you wish to store.
4	If pressed it displays the list of operators entered.
5	If pressed it prints all the names of the operators.

ALL WHAT HAS BEEN STORED IN THIS TABLE WILL THEN BE USED FOR RECOGNITION WHEN THE CYCLE IS STARTED.

Pic. 5.44

The list of operators with their access codes and positions can be displayed on this screen page.



pos. code	description
001 01	NAME OPERATOR 1
002	
003	
004	
005	
006	
007	
008	

To change the parameters, press twice on the writing inside the **“DESCRIPTION”** window corresponding to the parameter to change. you return automatically to pic. 5.44.

Pic. 5.45

5.12 Endoscope setup

To enter the new endoscopes with their pressure and flow parameters which will interact with the unit, proceed as follows: press the “ENDOSCOPES” key on the “SETUP” menu of pic. 5.46 and a screen page like the one in pic. 5.47 will appear.




NB. THE PARAMETERS DISPLAYED ON THE MACHINE ARE THE MANUFACTURER DEFAULT SETTINGS.

Pic. 5.46

To insert a new endoscope, push the button “LIST” in pic. 5.47 and double click on an empty line (pic.5.48). the picture pic. 5.49 is displayed.



Pic. 5.47



Pic. 5.48


5.12.1 Endoscope configuration parameter setting



Press “**DETAILS**” to start entering the information and parameters of the endoscope to add; an image like the one in pic. 5.49.1 will appear.


Pic. 5.49

From this screen page it is possible to set the working cycles of the instrument (point 1) and the type of instrument: single or multichannel, bronchoscope or similar, (point 2) or transesophageal probes (point 3). NB. It is necessary the special connector.
 The instruments can be maximum 500.



REFERENCE	DESCRIPTION
1	IN THE RED FIELD IT IS POSSIBLE TO SET THE MAIN PROGRAM (EX. COMPLETE CYCLE), WHEREAS IN THE BLACK FIELD FOR PROGRAM 2 YOU CAN SET THE SECONDARY CYCLE (EX. HIGH LEVEL DISINFECTION WITHOUT WASHING) OR NO CYCLE IS SET.
2	SET TO YES TO ACTIVATE TOE BOX.
3	ENDOSCOPE TYPE SETTING: IF THE ENDOSCOPE IS A SINGLE-CHANNEL INSTRUMENT, THE PARAMETER IS SET TO YES; IF THE ENDOSCOPE IS A MULTI-CHANNEL INSTRUMENT THEN THE PARAMETER IS SET TO NO.
4	SET THE DRYING TIME OF INSTRUMENT FOR DRYER CABINET IN CASE THE ARES SOFTWARE IS ACTIVATED (IN MINUTES).


Pic. 5.49.1



REFERENCE	DESCRIPTION
1	PRESS ON THE QUESTION MARKS AND A SCREEN PAGE APPEARS WHERE YOU CAN SELECT THE TYPE OF ENDOSCOPE TO ENTER (SEE PIC. 5.40). BY PRESSING TWICE ON THE TYPE OF ENDOSCOPE TO ENTER: THE NAME REPLACES THE QUESTION MARKS.
2	INSERT THE SERIAL NUMBER AND MODEL OF THE ENDOSCOPE ARE USUALLY ENTERED TO ENSURE CORRECT TRACEABILITY.
3	BY PRESSING ON THE QUESTION MARKS A SCREEN PAGE APPEARS WHERE YOU CAN SELECT THE NAME OF THE ENDOSCOPE MANUFACTURER (SEE PIC. 5.42). PRESS TWICE ON THE NAME OF THE MANUFACTURER: THE NAME REPLACES THE QUESTION MARKS.

Pic. 5.49.2

INSERTING MULTICHANNEL ENDOSCOPES




On this screen it is possible to set the multichannels instruments as gastroscopie, colonscopie, ect., fibro and video.



Insert the request data.

Press "SAVE" button to save and press "CHANNELS" button to continue the programming. the system displays the screen as shown on pic. 5.50.

Pic. 5.49.3

To change the parameter, press on the writing corresponding to the parameter to change.



REFERENCE	DESCRIPTION
1	PRESSING ON THE NUMBER WILL EITHER ACTIVATE OR DEACTIVATE THE CHANNELS TO CONNECT TO THE ENDOSCOPE WITH THE FOLLOWING CHOICE CRITERIA: CHANNEL 1 - COLOUR BROWN – BIOPSY 1 CHANNEL 2 - COLOUR GREEN – BIOPSY 2 CHANNEL 3 - COLOUR BLUE - WATER CHANNEL 4 - COLOUR YELLOW - AIR CHANNEL 5 - COLOUR GREY - SUCTION CHANNEL 6 - COLOUR RED – AUXILIARY WATER CHANNEL 7 - COLOUR BLACK – CHANNELS < A 1MM (OPTIONAL)
2	BY PRESSING THE SYMBOL IT IS POSSIBLE TO SELECT THE TYPE OF PUMP FUNCTIONING. IN PARTICULAR, THE  SYMBOL REPRESENTS THE NON-STOP PUMP FUNCTIONING WHILE THE  SYMBOL REPRESENTS THE “ON-OFF” PUMP FUNCTIONING.

Pic. 5.50

By pressing “PRESS.”button, the system displays the screen of pic.5.51 to insert the pressures parameters.

NB: The leak test is not displayed because it is automatic and always active.

STANDARD CHANNEL FLOW AND PRESSURE SETUP PARAMETERS:

THE STANDARD PARAMETERS ARE DISPLAYED BELOW WHICH ARE ABLE TO MEET THE SETUP REQUIREMENTS OF THE MAJORITY OF INSTRUMENTS CURRENTLY ON THE MARKET. THE INSTRUMENTS SUCH AS ENTEROSCOPES, PEDIATRIC GASTOSCOPES, PEDIATRIC COLONSCOPES OR ECOENDOSCOPES MUST BE ASSESSED INDIVIDUALLY.

	Min. Press.	Max. Press.	Min.F.	Max.F.
Channel 1	---	---	200 i/''	1200 i/''
Channel 2	---	---	200 i/''	1200 i/''
Channel 3	1490 mbar	1500 mbar	100 i/''	1000 i/''
Channel 4	1490 mbar	1500 mbar	100 i/''	1000 i/''
Channel 5	---	---	200 i/''	1200 i/''
Channel 6	---	---	50 i/''	1000 i/''
Channel 7	---	---	50 i/''	1000 i/''

Tab.5.10

PRESS "SAVE" TO CONFIRM WHAT HAS BEEN ENTERED.

STANDARD SETTING EXAMPLE FOR PRESSURES (SIMULATION OF A COLONSCOPE VIDEO) ACCORDING TO WHAT IS EXPRESSED IN TABLE.5.10.
 BY PRESSING "FLOWS" BUTTON, THE SYSTEM DISPLAYS THE SCREEN OF PIC.5.51 TO INSERT THE FLOWS PARAMETERS.

84 ENDOSCOPES MANAG.		27/04/15 12:05	
Position: 015		Code: 113	
Description: OLYMPUS GIF			
Type: GASTROSCOPIO		OLYMPUS	
Progr.1: 003		STANDARD PERACET	
Progr.2: 002		AUTODIS. CH 55 °C	
<div style="display: flex; justify-content: space-around;"> 1 2 3 4 5 6 7 </div> <p>Channel pressure mbar</p>			
Min:	1490	1490	
Max:	1500	1500	
SAVE	GROUP	CHANNELS	FLAWS



Pic. 5.51

STANDARD SETTING EXAMPLE FOR FLOWS (SIMULATION OF A COLONSCOPE VIDEO) ACCORDING TO WHAT IS EXPRESSED IN TABLE.5.10.
 PRESS "SAVE" TO CONFIRM WHAT HAS BEEN ENTERED.
 TO INSERT A NEW INSTRUMENT, REPEAT THE PROCEDURE FROM PIC.5.47 TO PIC.5.52.

Showned line/field only if the partial obstruction parameters are activeted.

84 ENDOSCOPES MANAG.		27/04/15 12:06					
Position: 015		Code: 113					
Description: OLYMPUS GIF							
Type: GASTROSCOPIO		OLYMPUS					
Progr.1: 003		STANDARD PERACET					
Progr.2: 002		AUTODIS. CH 55 °C					
<div style="display: flex; justify-content: space-around;"> 1 2 3 4 5 6 7 </div> Channels flow limit in l/'							
All:	0000	0000	0000	0000	0000	0000	0000
Min:	0050	0050	0050	0050	0050	0050	0050
Max:	0250	0250	0250	0250	0250	0250	0250
SAVE	GROUP	CHANNELS	PRESS.	FLAWS			

Pic. 5.52


5.13 Creating a backup of the settings

To create a Backup of the machine settings, to be recalled in the case of need, proceed as described below:
 press the **“BACKUP”** key of Pic. 5.54 on the **“SETUP”** menu, a screen page like the one in Pic. 5.55 will appear.

Press the highlighted **“BACKUP”** key, confirm your choice and wait a few seconds. all the data stored will be filed automatically and available for recalling.



Pic. 5.54




DO NOT USE THE “MANUFACT.BACKUP” KEY BECAUSE ALL THE “STANDARD” FACTORY DATA WILL BE OVERWRITTEN BY THE NEW ONES AND LOST AS A RESULT.

Pic. 5.55

5.14 Utility


The system is provided with a Utility Menu where it is possible to visualize the machine status, to enable manually the self-disinfection and, eventually, change the language. To access to the menu, proceed as follows:



Press the “UTILITY” button will appear the screen as on pic. 5.57.

Pic. 5.56

From this screen page it is possible to change the following parameters:



Pic. 5.57

REFERENCE	DESCRIPTION
1	BY PRESSING ON “SELECT LANGUAGE” YOU HAVE ACCESS TO A SUB MENU (PIC.5.58) WHERE IT IS POSSIBLE TO CHANGE THE MACHINE LANGUAGE WITH THE RELEVANT KEY.
2	BY PRESSING ON “MACHINE DATA” YOU HAVE ACCESS TO THE MAIN INFORMATION OF THE SYSTEM, SEE PIC. 5.59 – 5.60 – 5.61.
3	BY PRESSING ON “MACHINE STATUS” THE “FUNCTIONAL” SCREEN PAGES ARE DISPLAYED (ONLY VISUALIZATION) AS INDICATED IN THE PICTURES FROM PIC. 5.11 TO PIC.5.22.
4	DISABLED
5	DISABLED
6	BY PRESSING THE KEY “CF-CARD STATUS” THE AMOUNT OF MEMORY AVAILABLE ON THE REPLACEABLE CF-CARD IS DISPLAYED TOGETHER WITH INFORMATION ON THE CORRECT FUNCTIONING OF THE CARD.

From this screen it is possible to modify the language, pressing on the flag corresponding to the wanted one. pressing the “confirm” button, the modification will be saved, and on each screen all the phrases will be automatically changed on the selected language.



Pic. 5.58

Point 2 pic. 5.57




Pic. 5.59



Pic. 5.60

Point 2 pic. 5.57



Pic. 5.61

Point 6 pic.5.57

91 ALARM !!! CYCLE START 05/09/14 09 55

LOAD UNLOCKED OUT OF CYCLE UNLOAD UNLOCKED

WAITING DATAS INPUT!!!

Operator: ANDREA

Choice KB+READ. Pr.type AUTODIS.

Program 000

Work order:

CONFIRM CHEMICALS DELETE

DO NOT INSERT OPERATOR CODE, the system has already memorized the operator that have manually activated the self-disinfection function from the "UTILITY" menu.

ATTENTION: AS INDICATED, PAY ATTENTION THAT NO ONE INSTRUMENT HAVE TO BE INSIDE THE CHAMBER. IF THE CYCLE WILL START WITH ANY ENDOSCOPE INSTRUMENT OR THEIR ACCESSORIZE INSIDE, THESE ONES WILL BE FATALLY DAMAGED.

THE SYSTEM STARTS THE AUTOMATICALLY CYCLE AT DATE AND TIME SET.

Pic. 5.62


TECHNICAL AREA

PARAMETERS DISPLAYABLE WITH TECHNICAL PASSWORD (YELLOW)

Password: call Steelco for technical assistance


5.15 Maintenance area parameters

Press the “**MAINTENANCE**” key on the main screen page to gain access to the menu.



Pic. 5.63

Enter the technical code.



Pic.5.64

5.16 Storing a technical intervention

In order to store a technical intervention and reset the maintenance warning inside the system an electronic worksheet must be filled in, as in pic.5.66 following the procedures in pic.5.63.




Pic. 5.65

MAINTENANCE INTERVENT. 08/09/14 16:06
 Cycles 0 Hours 0
 Notes:
 Total time needed (hh:mm) 00:00
 Next maintenance date 00/00/0000
 Next maintenance hours 1000
 Technician
 SAVE CANCEL


Pic. 5.66

REFERENCE	DESCRIPTION
1	Enter the work that has been done on the unit.
2	Enter the time taken for the technical intervention.
3	Enter the date scheduled for the next maintenance.
4	Enter the name of the technician.
5	By pressing the "save" key the system automatically files the technical intervention, resets the maintenance warning (if active as an indication on the monitor) and automatically prints the worksheet for signing, taking the system back to the screen page of pic. 5.66.

Press  button on the side of monitor to exit.

5.17 I/O statistics

Input and output statistics can be accessed from this maintenance menu page for viewing and/or printing.




Pic. 5.67

5.18 Calibrations

SEE PROCEDURES DESCRIBED IN PARAGRAPH 4.6.3 OF THIS MANUAL FOR CALIBRATION OF THE CHEMICAL PRODUCTS.


System calibrations can be accessed from this maintenance menu page.

NB. THE PARAMETERS DISPLAYED ON THE MACHINE ARE THE MANUFACTURER DEFAULT SETTINGS.




Pic. 5.67

On this screen it is possible to correct the temperature displayed and used from the system for the disinfection cycle. these values are calibrated during the final tests and set as default. If it is necessary to replace one or more of this temperature sensors, call steelco for technical assistance.



Pic. 5.68

IN THIS WINDOW, IT IS POSSIBLE TO SEE, IN REAL TIME, THE FLOW OF LIQUID FROM THE WASHING SPRAY ARMS (POINT 2), AND IT IS POSSIBLE TO REGULATE THIS VALUE THROUGH POINT 1 BY EITHER REDUCING OR INCREASING THE VALUE READ BY THE FLOWMETERS.




NOTE. Normally no adjustments are necessary.

REFERENCE	DESCRIPTION
1	BY TOUCHING THE VALUE SHOWN INSIDE THE WHITE WINDOW, A NUMERIC KEYPAD APPEARS IN ORDER TO CORRECT THE FLOWMETER READING OF THE WASHING SPRAY ARMS WITHIN THE WASHING CHAMBER.
2	IT DISPLAYS THE WASHING ARM FLOW (L/') IN REAL TIME
3	THIS PARAMETER DISPLAYS THE WATER CONDUCTIVITY VALUE.
4	IF PRESSED IT PRINTS ALL THE DATA OF THE CALIBRATIONS SET.
5	BY TOUCHING THE VALUE SHOWN INSIDE THE WHITE WINDOW, A NUMERIC KEYPAD APPEARS IN ORDER TO CORRECT THE WATER CONDUCTIVITY READING IN THE WASHING TANK.
6	IF PRESSED IT SAVES ALL THE DATA SET.

Pic. 5.69

IN THIS WINDOW IT IS POSSIBLE TO SEE, IN REAL TIME, THE ENDOSCOPIC INSTRUMENT CHANNEL PRESSURE IN POSITIONS 1, 2, AND 3 IT IS POSSIBLE TO REGULATE THIS VALUE THROUGH POINT 1 BY EITHER REDUCING OR INCREASING THE VALUE READ BY THE FLOWMETERS (ONLY SMALL ADJUSTMENTS ARE POSSIBLE).




NOTE. Normally no adjustments are necessary.

REFERENCE	DESCRIPTION
1	THIS PARAMETER DISPLAYS THE CHANNEL PRESSURE OF THE INSTRUMENTS (MBAR) IN REAL TIME.
2	THIS PARAMETER MAKES IT POSSIBLE TO CORRECT THE PRESSURE VALUE READ BY THE PRESSURE SENSOR ON EACH INSTRUMENT.
3	THIS PARAMETER MAKES IT POSSIBLE TO CORRECT THE PRESSURE VALUE READ BY THE PRESSURE SENSOR ON CHANNELS 3 AND 4.
4	THIS PARAMETER MAKES IT POSSIBLE TO CORRECT THE PRESSURE VALUE READ BY THE PRESSURE SENSOR RELATING TO THE LEAK TEST.
5	THIS PARAMETER MAKES IT POSSIBLE TO CORRECT THE PRESSURE VALUE READ BY THE PRESSURE SENSOR RELATING TO THE LEAK TEST OF TANK.
6	IF PRESSED IT SAVES ALL THE DATA SET.
7	IF PRESSED IT PRINTS ALL THE DATA OF THE CALIBRATIONS SET.

Pic. 5.70


The parameters (impulses per litre) of the water filling control flowmeters can be displayed on this screen page. they are calculated automatically during calibration, see paragraph 4.6.3.



REFERENCE	DESCRIPTION
1	IT DISPLAYS THE PARAMETER SET AUTOMATICALLY BY THE COMPUTER FOR ADJUSTING THE WATER TO FILL THE WASH CHAMBER (ADJUSTMENT FLOWMETER). THIS PARAMETER CAN BE CHANGED MANUALLY (ADVISED AGAINST).
2	IT DISPLAYS THE PARAMETER SET AUTOMATICALLY BY THE COMPUTER FOR ADJUSTING THE WATER TO FILL THE WASH CHAMBER (CONTROL FLOWMETER). THIS PARAMETER CAN BE CHANGED MANUALLY (ADVISED AGAINST).
3	IF PRESSED IT PRINTS ALL THE DATA OF THE CALIBRATIONS SET.
4	IF PRESSED IT SAVES ALL THE DATA SET.

Pic. 5.71

ATTENTION!!! The values entered in point 1 and in point 2 are created automatically by the system after calibration of the chemicals, according to the procedure described in paragraph 4.6.3. We advise against making any changes.




REFERENCE	DESCRIPTION
1	THE VALUES INSIDE THE BOXES ARE THE IMPULSES NECESSARY FOR COUNTING 1 ML OF PRODUCT IN THE CONTROL FLOWMETERS. (DO NOT MODIFY THE VALUE ENTERED)
2	THE VALUES INSIDE THE BOXES ARE THE IMPULSES NECESSARY FOR COUNTING 1 ML OF PRODUCT IN THE ADJUSTMENT FLOWMETERS. (DO NOT MODIFY THE VALUE ENTERED)
3	DO NOT ENTER PARAMETERS – IT IS USED TO ASSOCIATE TO EACH ML OF CHEMICALS A TIME FOR THE LOADING IN SECONDS. (DO NOT MODIFY THE VALUE ENTERED)
4	DO NOT ENTER PARAMETERS – IT IS USED TO SET A MINIMUM TEMPERATURE LIMIT BEFORE LOADING THE CHEMICAL. (DO NOT MODIFY THE VALUE ENTERED)
5	IF PRESSED IT PRINTS ALL THE DATA OF THE CALIBRATIONS SET.
6	IF PRESSED IT SAVES ALL THE DATA SET.

Pic. 5.72

5.19 Recalling a technical intervention from maintenance from history

To view one or more technical interventions stored in the system's memory, access the maintenance menu and select the "MAINT.HISTOR." key see pic.5.74. The screen page of pic.5.75 will be displayed.




Pic. 5.73

21 MAINT. HISTOR. ARCHIVE		08/09/14 16:07	
Date from 01/01/2000 to 08/09/2014			
Record 000 of 000			
Date 00/00/0000	Hour 00:00:00		
Cycles 0	Hours 0		
Notes:			
Total time needed (hh:mm)		00:00	
Next maintenance 00/00/0000		Hours 0	
Technician			
<--	-->	READ	CANCEL LIST PRINT

Enter the time interval to search for the technical interventions to be displayed, entering the day, month and year, pressing on them to change them. then press the "READ" key and after a few seconds the info on the technical interventions will appear. see Pic.5.75.

Pic. 5.74

The following operations are possible on this screen page:




REFERENCE	DESCRIPTION
1	It indicates that in the time interval used for the search, there are 4 worksheet stored.
2	Use the arrows to scroll the worksheets. in this case from the first to the fourth.
3	It cancels the operations and returns to the screen page of pic.5.74.
4	It displays the list with the worksheets found. press twice on the file wanted to see
5	It prints the worksheet displayed.


Pic. 5.75

5.20 Consult the alarms history

To see the alarms in the system's memory, access the maintenance menu and select the **"ALARM HISTOR."** key. see pic.5.76. the screen page of pic. 5.77 is displayed.




Pic. 5.76



Enter the time interval to search for the alarms to be displayed, entering the day, month and year, pressing on them to change them. then press the "READ" key and after a few seconds the info on the technical interventions will appear. see pic.5.78.

Pic. 5.77

The following operations are possible on this screen page:




REFERENCE	DESCRIPTION
1	IT INDICATES THAT IN THE TIME INTERVAL USED FOR THE SEARCH. THERE ARE 38 ALARMS STORED.
2	IT INDICATES THAT THERE ARE 159 DIFFERENT ALARM TYPES IN THE SYSTEM.
3	IT INDICATES THAT YOU ARE VIEWING ALARM NO. 1 OF THE 38 FOUND.
4	TYPE OF ALARM TRIGGERED IN THE CYCLE; NO. 144 IS THE ALARM CODE.
5	TYPE OF PROGRAMME ONGOING DURING THE ALARM.
6	PHASE THE ALARM OCCURRED IN.
7	IT PRINTS THE LIST OF ALARMS FOUND.
8	IT DISPLAYS THE LIST OF ALARMS FOUND. PRESS TWICE ON THE FILE WANTED TO SEE THE DATA.
9	IT CANCELS THE OPERATIONS AND RETURNS TO THE SCREEN PAGE OF PIC.5.76.
10	USE THE ARROWS TO SCROLL THE ALARMS STORED. IN THIS CASE FROM THE ELEVENTH TO THE 38 TH .

Pic. 5.78


5.21 Changing maintenance parameters

As far as configuration of the maintenance parameters are concerned, proceed as explained below: press the “**SETUP**” key on the main screen page to gain access to the menu.




Pic. 5.79

Enter the user code (yellow password).



Pic.5.80


You may gain access to the maintenance data from this screen page to set the EW2 system.



NB. THE PARAMETERS DISPLAYED ARE THE MANUFACTURER DEFAULT SETTINGS.

Pic. 5.81


The following parameters can be changed on this screen page by touching the inside of the corresponding white boxes:



Pic. 5.82

REFERENCE	DESCRIPTION
1	Username - the name of the department where the system is working is normally entered.
2	Condition of the machine drainage valve with cycle stopped “CLOSED” or “OPEN” .
3	If positioned on “YES” , at the end of the cycle – if the machine is the double through door version – the doors will both open automatically (only in the cycle executed normally condition).
4	Positioned on “YES” will let the operator block the cycle and reset it on the “CYCLE” screen page (the one that appears after start-up).
5	If pressed it saves the modified data in the system’s memory.
6	If this parameter is set to “YES” , when the door(s) are closed and the operator and instruments are entered, the cycle starts automatically (advised against).
7	By setting the parameter at yes, the operator is forced to open the door at the end of each cycle, before a new cycle is started.
8	By pressing this key all the parameters in the “MAINT. DATA” folders will be printed.
9	Point for changing the “YELLOW” technical password. keying in twice the new password on the alphanumerical keypad that appears. after pressing on the asterisks and saving the data (point 9) the password is changed.


The following parameters can be changed on this screen page by touching the inside of the corresponding white boxes:



REFERENCE	DESCRIPTION
1	Maximum time to activate the door closing micro switch (safety).
2	Maximum time to activate the door opening micro switch (safety).
3	Motor rotation time to lock the door.
4	Pause time before the water filling cycle is activated by the working water level signal (central level meter).
5	Maximum time for draining the washing tank, controlled by a sensor under the tank filter in the heating element compartment.
6	Maximum time to drain the machine (washing tank, filters, tank).
7	By pressing this key all the parameters in the "MAINT. DATA" folders will be printed.
8	Maximum temperature increase time from when the tank heating elements were activated.
9	Maximum temperature increase time from when the drying heating elements were activated.
10	If pressed it saves the modified data in the system's memory.

Pic. 5.83


The following parameters can be changed on this screen page by touching the inside of the corresponding white boxes:



Pic. 5.84


REFERENCE	DESCRIPTION
1	MAXIMUM WAITING TIME FOR THE FLOWMETER READING OF WATER DOSAGE IN WASHING CHAMBER
2	MAXIMUM WAITING TIME FOR THE FLOWMETER READING OF CHEMICAL PRODUCT DOSAGE IN WASHING CHAMBER
3	MAXIMUM WAITING TIME FOR THE FLOWMETER READING OF CHANNELS
4	FLOWMETER READING SAMPLING TIME
5	MAXIMUM TIME FOR THE FILLING OF 1 LITRE OF WATER IN ORDER TO ACTIVATE THE FILTER OBSTRUCTION ERROR MESSAGE
6	BY PRESSING THIS KEY ALL THE PARAMETERS IN THE “MAINT. DATA” FOLDERS WILL BE PRINTED.
7	IF PRESSED IT SAVES THE MODIFIED DATA IN THE SYSTEM’S MEMORY.

The following parameters can be changed on this screen page by touching the inside of the corresponding white boxes:



REFERENCE	DESCRIPTION
1	E.G. IF THE TEMPERATURE OF THE CYCLE PHASES IS SET AT 35°C, AT THIS STAGE IT IS POSSIBLE TO ENTER THE TEMPERATURE AT WHICH THE HEATING ELEMENTS MUST REACTIVATE TO HEAT THE LIQUID IN THE TANK. THEREFORE, BY SETTING THE PARAMETER AT POINT 1 AT -1.0°C, AFTER THE HEATING ELEMENTS HAVE TAKEN THE TEMPERATURE OF THE WATER TO 35°C THEY WILL REACTIVATE ONLY AT 34°C.
2	E.G. IF THE TEMPERATURE OF THE CYCLE PHASES IS SET AT 35°C, AT THIS STAGE IT IS POSSIBLE TO ENTER THE TEMPERATURE AT WHICH THE HEATING ELEMENTS MUST DEACTIVATE TO HEAT THE LIQUID IN THE TANK. THEREFORE, BY SETTING THE PARAMETER AT POINT 2 AT +1.0°C THE HEATING ELEMENTS WILL DEACTIVATE AT 36°C AND REACTIVATE AT 35°C (ATTENTION!!! IF THE PARAMETER SET IN THE CYCLE PHASES IS EXCEEDED BY 5°C THE SYSTEM GOES INTO THE ALARM STATE, IN THIS CASE AT 40.1°C)
3	BEI BETÄTIGUNG WERDEN ALLE GEÄNDERTEN DATEN IM SPEICHER DES SYSTEMS GESPEICHERT
4	DURCH BETÄTIGEN DIESER TASTE WERDEN ALLE PARAMETER IN DEN ORDNERN "WARTUNGSDATEN" GEDRUCKT
5	MAXIMUM PERMITTED IMPULSE VALUE OF FLOWMETERS FOR WATER INTAKE DURING STAGES WITHOUT WATER INTAKE CONTROL
6	MAXIMUM PERMITTED IMPULSE VALUE OF FLOWMETERS FOR CHEMICAL PRODUCT INTAKE DURING STAGES WITHOUT CHEMICAL PRODUCT INTAKE CONTROL

The following parameters can be changed on this screen page by touching the inside of the corresponding white boxes:



REFERENCE	DESCRIPTION
1	MINIMUM WATER FLOW OF WASHING ARMS (LITRES/MINUTE) BELOW WHICH THE SYSTEM ALARM IS TRIGGERED
2	MAXIMUM WATER FLOW OF WASHING ARMS (LITRES/MINUTE) AT WHICH THE SYSTEM ALARM IS TRIGGERED
3	MINIMUM SETTABLE LIMIT FOR WATER INTAKE CONDUCTIVITY/QUALITY
4	MAXIMUM SETTABLE LIMIT FOR WATER INTAKE CONDUCTIVITY/QUALITY
5	DURCH BETÄTIGEN DIESER TASTE WERDEN ALLE PARAMETER IN DEN ORDNERN "WARTUNGSDATEN" GEDRUCKT
6	BEI BETÄTIGUNG WERDEN ALLE GEÄNDERTEN DATEN IM SPEICHER DES SYSTEMS GESPEICHERT

Pic. 5.86


5.22 Changing cycle parameters

To change the cycle parameters, access the “**MAINT. DATA**” screen pages, see pic.5.81 and select the “**DATA CYCLE**” button, see pic. 5.87: a new set of screen pages will appear that can be selected like the previous ones, see pic.5.82.



Pic. 5.87


The following parameters can be changed on this screen page by touching the inside of the corresponding white boxes:



REFERENCE	DESCRIPTION
1	IN CASE OF AUTOMATIC RINSING (FOLLOWING ALARM RESET) THE SYSTEM, FOR THE TIME LIMIT SET, WILL BLOW MEDICAL AIR INTO BOTH THE ENDOSCOPIC CHANNELS AND THE WASHING ARMS.
2	IN CASE OF AUTOMATIC RINSING WITHOUT INSTRUMENTS, (FOLLOWING ALARM RESET) THE SYSTEM, FOR THE TIME LIMIT SET, WILL BLOW MEDICAL AIR INTO THE WASHING ARMS.
3	WATER LEVEL CHECKING DELAY TIME AFTER TANK FILLING.
4	AUTOMATIC RINSING TIME AFTER AN ALARM OR CYCLE RESET BY THE OPERATOR.
5	ALARM TIME IN CASE OF FLOW VALUE OUT OF RANGE INTO THE INSTRUMENTS.
6	BY SETTING A TIME LIMIT, THE CHECKING OF THE PARTIAL OBSTRUCTION OF THE CHANNELS WILL BE ACTIVATED
7	PRESSING THIS KEY IT SAVES THE MODIFIED DATA IN THE SYSTEM'S MEMORY.
8	AIR BLEED TIME FOLLOWING THE CHECKING OF THE PARTIAL OBSTRUCTION OF THE CHANNELS.
9	PRESSING THIS KEY ALL THE PARAMETERS IN THE "MAINT. DATA" FOLDERS WILL BE PRINTED.

Pic. 5.88


The following parameters can be changed on this screen page by touching the inside of the corresponding white boxes:



REFERENCE	DESCRIPTION
1	THE MAXIMUM TIME (SETTABLE) TO LOAD AN ENDOSCOPE BEFORE THE MACHINE GIVES AN ALARM.
2	DRAINING TIME AT END CYCLE FOR LEAK TEST AIR INTO THE INSTRUMENT.
3	MAXIMUM NUMBER OF REFILLING (SETTABLE) IN CASE OF LEAK TEST FAILURE BEFORE THE MACHINE GIVES AN ALARM (NOT SET NEVER LESS THAN 20 REFILLING).
4	BALANCE TIME (SETTABLE) OF COMPRESSED AIR BETWEEN INSTRUMENT AND LT EXPANSION CYLINDER BEFORE THE LEAK TEST.
5	KEEP 00.0 VALUE – THE VALUE EXPRESSES THE CHECKING OF THE INITIAL TOE BOX LEAK TEST (TRANSESOPHAGEAL PROBE CONTAINER) IN SECONDS
6	SYSTEM CONTROL TIME FOR CHECKING OF TOE BOX LEAK – KEEP MANUFACTURER PARAMETERS.
7	THE VALUE EXPRESSES THE LEAK TEST CHECK UPON CYCLE START IN SECONDS
8	SYSTEM CONTROL TIME FOR LEAK TEST CHECKING DURING THE CYCLE.
9	PRESSING THIS KEY ALL THE PARAMETERS IN THE “MAINT. DATA” FOLDERS WILL BE PRINTED.
10	PRESSING THIS KEY IT SAVES THE MODIFIED DATA IN THE SYSTEM’S MEMORY.

Pic. 5.89


The following parameters can be changed on this screen page by touching the inside of the corresponding white boxes:



REFERENCE	DESCRIPTION
1	TYPE OF RINSING WATER (THE SYSTEM FORESEES A STANDARD TYPE OF WATER INTAKE, PLUS A SECOND, OPTIONAL TYPE) AUTOMATIC SYSTEM RESTORE FOLLOWING ALARM RESET
2	QUANTITY OF WATER IN WASHING CHAMBER FOR FINAL RINSE, AUTOMATIC SYSTEM RESTORE FOLLOWING ALARM RESET
3	MAXIMUM ERROR FOR WATER FILLING
4	MAXIMUM ERROR FOR CHEMICALS LOADING.
5	MAXIMUM ERROR FOR TEMPERATURE PROBES.
6	CONTROL OF THE INSTRUMENT IN THE CHAMBER
7	NUMBER OF REPETITION FINAL RINSE PHASE IN CASE OF ALARM.
8	PRESSING THIS KEY ALL THE PARAMETERS IN THE "MAINT. DATA" FOLDERS WILL BE PRINTED.
9	PRESSING THIS KEY IT SAVES THE MODIFIED DATA IN THE SYSTEM'S MEMORY.

Pic. 5.89.1


The following parameters can be changed on this screen page by touching the inside of the corresponding white boxes:



Pic. 5.90

REFERENCE	DESCRIPTION
1	MAXIMUM PERMITTED ERROR RELATING TO THE MAXIMUM PRESSURE SET
2	MAXIMUM PERMITTED LEAKAGE DURING THE THREE INSPECTION STAGES OF THE LEAK TEST EXPRESSED IN % COMPARED TO THE VALUE IN POINT 9
3	MAXIMUM PERMITTED LEAKAGE DURING THE CYCLE IN THE LEAK TEST EXPRESSED IN % COMPARED TO THE VALUE IN POINT 9
4	MAXIMUM LEAK ON TOE BOX DURING THE CYCLE.
5	MAXIMUM PRESSURE VALUE INSIDE THE ENDOSCOPE INSTRUMENT AT THE END OF LEAK TEST (END CYCLE).
6	PRESSURE INCREASING IN THE LEAK TEST OF TOE BOX.
7	PERMITTED LEAKAGE EXPRESSED IN % SPECIFIED IN POINT 9 FOR SYSTEM RESTORE WITHOUT TRIGGERING ANY SYSTEM ALARM, THE MBAR PARAMETER VALUE CANNOT EXCEED THE PARAMETER SPECIFIED IN POINT 3
8	PARAMETER EXPRESSED IN % EXCEEDING WHAT IS SPECIFIED IN POINT 9 (WITH THIS CONFIGURATION: 30MBAR) PRIOR TO THE DEACTIVATION OF THE LEAK TEST PUMP
9	SET VALUE OF LEAK TEST.
10	INCREASE OF THE LEAK TEST EXPRESSED IN % RELATING TO THE SECOND STAGE OF THE LEAK INSPECTION (300 MBAR) COMPARED TO WHAT IS SPECIFIED IN POINTS 9 + 8
11	PRESSING THIS KEY ALL THE PARAMETERS IN THE "MAINT. DATA" FOLDERS WILL BE PRINTED.
12	PRESSING THIS KEY IT SAVES THE MODIFIED DATA IN THE SYSTEM'S MEMORY.

The following parameters can be changed on this screen page by touching the inside of the corresponding white boxes:



Pic. 5.90.1






REFERENCE	DESCRIPTION
1	MINIMUM PRESSURE ALARM TIME FOR CONTINUOUS OPERATION CHANNELS
2	CHANNELS 6 AND 7 PUMP OFF ALARM TIME DURING ON-OFF OPERATION
3	CHANNELS 6 AND 7 PUMP ON ALARM TIME DURING ON-OFF OPERATION
4	ALARM PERCENTAGE BEYOND PUMP OPERATION SETPOINT OF DURING ON-OFF OPERATION
5	MINIMUM MANIFOLD PRESSURE
6	MAXIMUM MANIFOLD PRESSURE
7	PRESSING THIS KEY ALL THE PARAMETERS IN THE "MAINT. DATA" FOLDERS WILL BE PRINTED.
8	PRESSING THIS KEY IT SAVES THE MODIFIED DATA IN THE SYSTEM'S MEMORY.



6. WASHING CYCLE

6.1 Preparation of the endoscopes

In order to be correctly reprocessed inside the EW 2 system, the endoscopes must undergo pre-cleaning and manual washing procedures. Follow the current applicable National Guideline instructions as well as any internal protocols in force.

Some of the main currently available Guidelines are specified below:

COUNTRY		GUIDELINES
	EUROPE	<i>ESGE±ESGENA guideline: Cleaning and disinfection in gastrointestinal endoscopy Update 2008.</i>
	ITALY	<i>ANOTE-ANIGEA - Linee guida Pulizia e disinfezione in endoscopia - Update 2011</i>
	FRANCE	<i>Guide de Bonne Pratique de désinfection des dispositifs médicaux - obligatoire depuis le 14 juin 1998. Conseil Supérieur d'Hygiène Publique de France, section prophylaxie des maladies transmissibles. Comité Technique Nationale des Infections Nosocomiales.</i>
	IRELAND	<i>PART 4: RECOMMENDED PRACTICES FOR ENDOSCOPY UNITS Health Service Executive Code of Practice for Decontamination of Reusable Invasive Medical Devices - Review date August 2008.</i>
	GERMANY	<i>Recommendation of the Commission for Hospital Hygiene and Infection Prevention at the Robert Koch Institute (RKI). Hygiene Requirements for Reprocessing Flexible Endoscopes and Additional Endoscopic Instrumentation - Published in the Federal Health Gazette in April 2002.</i>

	<p align="center">GREAT BRITAIN</p>	<p><i>National Endoscopy Programme - Decontamination Standards for Flexible Endoscopes - Updated March 2009.</i></p>
	<p align="center">UNITED STATES</p>	<p><i>SGNA Society of Gastroenterology Nurses and Associates, Inc. Standards of Infection Control in Reprocessing of Flexible Gastrointestinal Endoscopes - Revised in 2012.</i></p>


6.2 Check of endoscope instrument connectors

Connectors for the endoscope instruments supplied with the EW 2 machine must be checked daily by machine users and periodically by technicians in charge of maintenance.

The use of not original Steelco connectors or worn connectors can damage the endoscope instruments, activate machine alarms (ex. “channels disconnected”, “channel partially clogged” etc.) and most of all does not guarantee a correct disinfection process, endangering both patients’ and sanitary personnel’s wealth.

6.3 Start cycle


Once all the instruments settings have been carried out and all the machine operating parameters have been checked, by pressing the key **“WORKING CYCLE”** on the main screen page (Pic.5.91) it is possible to start a cycle. The screen page shown in Pic. 5.92 is displayed, where the parameter needed must be introduced.



Pic. 5.91

START CYCLE FOR GASTROSCOPES, COLONSCOPES OR OTHERS

From this screen page it is possible to change the following parameters:




REFERENCE	DESCRIPTION
1	PRESS INSIDE THE FIELD AND INSERT THE ID-CODE. THE OPERATOR NAME WILL APPEAR.
2	BY PRESSING THIS FIELD, IT IS POSSIBLE TO SELECT THE TYPE OF PROGRAMME TO BE CARRIED OUT: MULTI-CHANNEL ENDOSCOPES (FIG. 5.92), SINGLE-CHANNEL ENDOSCOPES (FIG.5.93) OR THERMAL SELF-DESINFECTON (FIG. 5.94).
3	SELECTION OF ENDOSCOPE 3 TO BE REPROCESSED.
4	SELECTION OF ENDOSCOPE 2 TO BE REPROCESSED.
5	CONFIRM THE PARAMETER AND START THE WASHING
6	DISPLAY THE CHEMICAL STATUS.
7	CANCEL AND RESET THE PAREMETER.
8	THIS PARAMETER MAKES IT POSSIBLE TO SELECT THE SECOND PROGRAMME SET BY THE ENDOSCOPE PARAMETERS.

Pic 5.92

START CYCLE FOR BRONCHOSCOPES, CYSTOSCOPES OR OTHERS


From this screen page it is possible to change the following parameters:



REFERENCE	DESCRIPTION
1	THIS PARAMETER CARRIES OUT THE SAME FUNCTION EXPLAINED PREVIOUSLY.
2	
3	BY PRESSING THIS FIELD, IT IS POSSIBLE TO INSERT THE SINGLE-CHANNEL ENDOSCOPE: THE FOLLOWING WINDOW WILL APPEAR (FIG. 5.94). ONCE THE ENDOSCOPE IN THIS FIELD HAS BEEN SELECTED, THE CODE AND NAME OF THE INSTRUMENT IS STATED.
4	THIS PARAMETER CARRIES OUT THE SAME FUNCTION EXPLAINED PREVIOUSLY.
5	
6	


FOLLOW THE SAME METHOD TO INSERT THE OTHER INSTRUMENTS IN THE TABLE TO COMPLETE THE POSITIONS AVAILABLE. THE MACHINE CAN ALSO BE LOADED WITH MIXED BATCHES "VIDEO" AND FIBROENDOSCOPES.

Pic. 5.93




REFERENCE	DESCRIPTION
1	IN THIS FIELD, IT IS POSSIBLE TO ENTER THE ENDOSCOPE CODE.
2	IN THIS FIELD, IT IS POSSIBLE TO ENTER THE POSITION WHERE THE ENDOSCOPE IS TO BE INSERTED.
3	IN THIS SECTION, SOME DATA RELATING TO THE ENDOSCOPE THAT HAS JUST BEEN INSERTED IS STATED, SUCH AS: NAME, MANUFACTURER, TYPE OF PUMP OPERATION AND ASSOCIATED CYCLE.

Pic. 5.94



REFERENCE	DESCRIPTION
1	IN THIS FIELD, THE TYPE OF PUMP OPERATION IS DISPLAYED.
2	IN THIS FIELD, THE CODE OF THE PROGRAM ASSOCIATED WITH THE ENDSCOPE IS STATED.
3	THE FIRST NUMBER IS THE SET AND THE SECOND NUMBER THE CHANNEL
4	THE COLOR OF THE CHANNEL

Pic. 5.95



REFERENCE	DESCRIPTION
1	THIS PARAMETER CARRIES OUT THE SAME FUNCTION EXPLAINED PREVIOUSLY.
2	
3	IN THIS FIELD, IT IS POSSIBLE TO SELECT THE SELF-DISINFECTION CYCLE TO BE CARRIED OUT.
4	THIS PARAMETER CARRIES OUT THE SAME FUNCTION EXPLAINED PREVIOUSLY.
5	
6	

Pic. 5.96

6.4 Free programs setting

Endoscope washers with CE marking (class IIb) pursuant to Directive 93/42/CEE are medical devices that are approved for use as part of a **CLOSED SYSTEM** that includes:


- Endoscope washers;
- Chemical agents that have been tested and approved in order to guarantee the effectiveness of the re-conditioning process (Type Tests in conformity with standard EN ISO 15883);
- Filters;
- Connectors,
- Proven compatibility with Medical Devices treated in the re-conditioning cycle;
- Any other accessory mentioned in the User and Technical Manuals.

The CE marking applied by STEELCO SPA is guaranteed and considered valid only when the endoscope washer is used as part of a CLOSED SYSTEM: any other accessory different from those specified in the manufacturer's Medical Device Manual makes the validity of both the abovementioned marking and the Test Types carried out with such device void.

Therefore the “**Process Management**”, the carrying out of the operational and performance endorsements established by the manufacturer become the sole responsibility of the user. Such endorsements must prove the effectiveness, the safety and the repeatability over time of the re-conditioning process within the **CLOSED SYSTEM** with reference to the initial validation conditions: **or rather, the use of the same products and the same cycles that have been tested and approved by the Medical Device manufacturer.**


Otherwise, STEELCO SPA shall not be held responsible for **any different and improper uses** of the system compared to those foreseen in the document of reference and in conformity with IQ, OQ e PQ protocols issued by the manufacturer itself.

You gain access to programme management from the configuration menu. to display and/or change programmes press the “PROGRAMS” key to access the screen page of picture 5.98.



Pic. 5.97

The following controls are possible on this screen page:



To recall the cycle from the memory press on the writing inside the white window and enter the number of the cycle type to change and then press “DETAILS” to see it (point 2) or press on point 2 and select the cycle of view from the list.

REFERENCE	DESCRIPTION
1	It gives the position (on the progressive list – see pic.5.99) where to store the cycle.
2	The key is enabled by changing colour (blue) when a valid position is entered at point 1. It is used to see the programme and make any necessary changes.
3	If pressed it displays the list of cycle types entered.
4	If pressed it prints all the data of the cycles table.

To display a cycle, and change it if required, either enter the number of the cycle in point 1 and then press “DETAILS” or press “LIST” and select it from a list (see pic.5.99) pressing it twice. This will take you back to the mask of picture 5.98 and pressing “DETAILS” you gain access to the contents of the programme.

The same applies if you want to create a new programme – either enter the number of an empty position and then press “DETAILS” or press “LIST” and select an empty space twice on the list (picture 5.99). this takes you back to the mask of picture 5.99 and pressing “details” you gain access to the contents of the programme which is empty.

ATTENTION!

THE FIRST 3 PROGRAMMES ARE FACTORY SET AND ARE VALIDATED CYCLES – DO NOT CHANGE THEM – .

Pic. 5.98

The cycles stored and their positions can be seen on this screen page

PROGRAMS MANAG. 08/09/14 16:11

PROGRAMS LIST X

nr.	type	description
008	ENDOSCOPES	COLLAUDO 45 °
009	AUTODIS.	TEST BARILOTTO
010		
011		
012		
013		
014		
015		

To create a new programme
press twice on a free
position.


To change the parameter, press twice on the writing inside the “**DESCRIPTION**” window corresponding to the parameter to change.
You return automatically to pic. 5.98.

ATTENTION!
**THE FIRST 3 PROGRAMMES ARE FACTORY SET AND ARE VALIDATED
CYCLES – DO NOT CHANGE THEM – .**

Pic. 5.99

CREATING A PROGRAMME FOR REPROCESSING THE ENDOSCOPES

After having followed the instructions from picture 5.97 to picture 5.99, picture 5.100 appears. Follow the instructions given below to programme a cycle, taking into consideration that the responsibility of a new programme lies with the organisation or company that did the programming. Do not enter parameters in the washing and disinfection phases that differ from those given in this technical manual.




Pic. 5.100

REFERENCE	DESCRIPTION
1	The position is red where the programme created is saved;
2	Space for entering the name of the programme (press in the white box to enter the text);
3	Select the type of programme – endoscopes – self-disinfection – instruments (not active);
4	Empty cycle phases for entering the type (see sequence of Pic.5.101 and Pic.5.102);
5	Press on the arrows on the sides to scroll the list of phases;
6	For deleting the phase selected; after having selected the phase the line changes colour turning blue. If you then press the “delete” key twice the prompt to confirm deletion appears. If you confirm the blue line appears with no writing inside which means the phase has been deleted;
7	It cancels the ongoing operation;
8	The “COPY” key is to duplicate one or more phases. Pressing it once the key starts blinking and you can now select the phases to duplicate, touching them and making sure they change colour, going from grey to blue.
9	Press the “INSER.” key to add an empty phase line above the one selected: select the line above the one where you wish to add a new phase, press the “INSER.” key and a new phase above the one selected is created automatically.

You will now be able to create a new programme.

Example for creating a cycle with initial discharge phase, rinse, wash, rinse, disinfection, double final rinse. Below is the cycle setup sequence:



REFERENCE	DESCRIPTION
1	Press in the white box and enter the name you want to give to the new cycle on the alphanumerical keypad that appears.
2	Press twice; a selection of "TREATMENT", "DRAIN", and "RINSE" phases appears, press "DRAIN" and the screen page of pic.5.105 appears.


Pic. 5.101

If the operation was correctly done we will have the following screen page:



REFERENCE	DESCRIPTION
1	Press on the next box (01- empty) to enter a new phase; you return to the screen page of picture 5.102; select "DRAIN" following the same procedure and the screen page of pic. 5.105 appears.

Pic. 5.102



REFERENCE	DESCRIPTION
2	It indicates the time in seconds for the medical air to go through the rotors at the end of draining.
4	It indicates the time in seconds for the medical air to go through the endoscopic channels at the end of draining.

- The "CONFIRM" key takes you back to pic.5.101 with the parameters saved.
- The "CHANGE PHASE" key takes you back to pic. 5.102.
- The "CANCEL" key takes you back to pic.5.102 without saving the data set.

Pic. 5.103

TO INSERT OF RINSING PHASE AFTER DRAINING, FOLLOW THE PROCEDURE




Pic. 5.106



1

REFERENCE	DESCRIPTION
1	Press on "RINSE" and the screen page of pic.5.108 appears.


Pic. 5.107



REFERENCE	DESCRIPTION
2	SET THE PARAMETERS AS SHOWN IN PICTURE 5.108. IF NEEDED ALTER THE PHASE TIME; THE NAME OF THE WATER COULD BE DIFFERENT FROM WHAT IS DISPLAYED.
3	THESE PARAMETERS ARE USED ONLY FOR "PUMP IMPULSE". IT IS POSSIBLE TO COOL THE WASHING CHAMBER WITHOUT STRESSING THE STRUCTURE (AFTER THE AUTOSANIFICATION CYCLE AT 90°C OR 80°C) MAKING SURE THAT THE WASHING PUMP WORK BY IMPULSES (ALTERNATING MOMENTS OF PAUSE AND OPERATION). FOR EXAMPLE, SETTING ON "PHASE.T." PARAMETER 3 SECONDS OFF AND 2 SECONDS ON, WITHIN 30 SECONDS, THE WASHING PUMP TURN ON FOR 2 SECONDS AFTER 3 SECONDS OF PAUSE.
4	WITH THESE PARAMETERS IT IS POSSIBLE TO CHECK THE WATER QUALITY DURING THE RINSING PHASES. NORMALLY, THIS CONTROL IS USED ON THE LAST RINSE. FOR EXAMPLE: SETTING 100 MS AND 2 ATTEMPTS, IF THE MACHINE FINDS THE CONDUCTIVITY VALUE OVER 100 MS, THE RINSING PHASE IS REPEATED 2 TIMES. IF THE CONDUCTIVITY VALUE IS STILL GREATER THAN 100 MS, THE MACHINE GIVES AN ALARM.

Pic. 5.108

The following screen page appears after confirming:




Pic. 5.109

The steps for entering the other parameters are the same. the "STANDARD" parameters are given below in correct construction sequence with their screen pages, taking into account the cycle given in pic.5.101.



Pic. 5.110

The following screen page appears after confirming:



Pic. 5.111



Pic. 5.112

PROGRAMS MANAG. 019		26/09/14 12:20	
Description: <input type="text"/>			
Door: OK LOAD.		NOT OK LOAD.	
Program type: ENDOSCOPES			
WASHING		STOP END-NO	
Water	ACQUA	I 12.0	Pumps YES
Chem.	DET	% 05.00	<input type="text"/>
Chem.	<input type="text"/>	% 00.00	<input type="text"/>
Temp.°C	35.0	Maint.°C	YES
Mix.T.s	000	T.Phase s	0000
CONFIRM		CHANGE PHASE	
CANCEL			

It is possible to modify times, names, temperature and chemicals quantity.

NOTE: The chemical percentage is expressed in ml. and in thousandths (ex. ‰ 5 means 5 ml for one liter of water loaded in the washing chamber).

If two chemical products are used and they have to be mixed, use this parameter to set the mixing time (in seconds).

Pic. 5.113

26 PROGRAMS MANAG. 011 09/09/14 08 20

Description:

Door: OK LOAD. NOT OK LOAD.

Program type: ENDOSCOPES

01	DRAIN	Purging 05.0 NORMAL	▲
02	RINSE	for s 0030 uS 000.0	
03	DRAIN	Purging 05.0 NORMAL	
04	WASHING	for s 0060 to35.0 oC	
05			
06			▼


SAVE DELETE CANCEL COPY INSER.

Pic. 5.114

TO INSERT THE DRAIN PARAMETER, FOLLOW THE PROCEDURE AS SHOWN IN PICTURE 5.103



Pic. 5.115



Pic. 5.116

TO INSERT OF RINSING PHASE AFTER CLEANING, FOLLOW THE PROCEDURE



Pic. 5.117



Pic. 5.118

PROGRAMS MANAG. 012		20/09/14 10:06	
Description: <input type="text"/>			
Door: OK <input type="text" value="LOAD."/>		NOT OK <input type="text" value="LOAD."/>	
Program type: <input type="text" value="ENDOSCOPES"/>			
RINSE			
Water	<input type="text" value="ACQUA"/>	I	<input type="text" value="12.0"/> Pumps <input type="text" value="YES"/>
T.Phase s		<input type="text" value="0030"/>	
T.Imp.OFFs	<input type="text" value="0.0"/>	T.Imp.ON s	<input type="text" value="0.0"/>
Conductivity µS:		<input type="text" value="000.0"/>	
Nr attempts:		<input type="text" value="0"/>	
CONFIRM		CHANGE PHASE	CANCEL

Pic. 5.119

SCROLL THE PHASE PAGE

If all rows are programmed, scroll the page by using the arrows as shown in the picture 5.120.

26 PROGRAMS MANAG, 011 09/09/14 08 21

Description:

Door: OK LOAD. NOT OK LOAD.

Program type: ENDOSCOPES

01	DRAIN	Purging	05.0	NORMAL	▲
02	RINSE	for s	0030	uS 000.0	■
03	DRAIN	Purging	05.0	NORMAL	■
04	WASHING	for s	0060	to 35.0 oC	■
05	DRAIN	Purging	05.0	NORMAL	■
06	RINSE	for s	0030	uS 000.0	▼

SAVE DELETE CANCEL COPY INSE.



Pic. 5.120

26 PROGRAMS MANAG, 010 08/09/14 16 22

Description:

Door: OK LOAD. NOT OK LOAD.

Program type: ENDOSCOPES

07					▲
08					■
09					■
10					■
11					■
12					▼


SAVE DELETE CANCEL COPY INSE.

Pic. 5.121

TO INSERT THE DRAIN PHASE AFTER RINSE, FOLLOW THE PROCEDURE AS SHOWN IN PICTURE 5.103.




Pic. 5.122



Pic. 5.123

TO INSERT THE DESINFECTATION/STERILIZATION PHASE, FOLLOW THE PROCEDURE



Pic. 5.124



Pic. 5.125

IT IS POSSIBLE TO MODIFY TIMES, NAMES, TEMPERATURE AND CHEMICALS QUANTITY. NOTE: THE CHEMICAL PERCENTAGE IS EXPRESSED IN ML. AND IN THOUSANDTHS (EX. ‰ 5 MEANS 5 ML FOR ONE LITER OF WATER LOADED IN THE WASHING CHAMBER).

PROGRAMS MANAG., 012 20/09/14 10 08

Description:

Door: OK NOT OK

Program type: ENDOSCOPES

DISINFECTION

Water ACQUA I 12.0 Pumps YES

Chem. PAA % 10.00

Chem. % 00.00

Temp.oC 35.0 Maint.oC YES

Mix.T.s 000 T.Phase s 0300

Pic. 5.126

PROGRAMS MANAG., 011 09/09/14 08 23

Description:

Door: OK NOT OK

Program type: ENDOSCOPES


07	DRAIN	Pumping 05.0 NORMAL	▲
08	DISINFECTION	Temp s 0300 to 35.0 oC	
09			
10			
11			
12			▼

Pic. 5.127

TO INSERT THE DRAIN PHASE AFTER DISINFECTION, FOLLOW THE PROCEDURE AS SHOWN IN PICTURE 5.103




Pic. 5.128



Pic. 5.129

TO INSERT OF RINSING PHASE AFTER DRAINING, FOLLOW THE PROCEDURE



Pic. 5.130



Pic. 5.131

PROGRAMS MANAG. 012 20/09/14 10:06

Description:

Door: OK LOAD. NOT OK LOAD.

Program type: ENDOSCOPES

RINSE

Water ACQUA I 12.0 Pumps YES

T.Phase s 0030

T.Imp.OFFs 0.0 T.Imp.ON s 0.0

Conductivity μ S: 000.0

Nr attempts: 0

CONFIRM CHANGE PHASE CANCEL

Pic. 5.132

PROGRAMS MANAG. 011 09/09/14 08:23

Description:

Door: OK LOAD. NOT OK LOAD.

Program type: ENDOSCOPES

07	DRAIN	Purging 05.0 NORMAL	▲
08	DISINFECTION	for s 0300 to 35.0 oC	
09	DRAIN	Purging 05.0 NORMAL	
10	RINSE	for s 0030 uS 000.0	
11			
12			▼

SAVE DELETE CANCEL COPY INSER.

Pic. 5.133

TO INSERT THE DRAIN PHASE AFTER RINSE, FOLLOW THE PROCEDURE AS SHOWN IN PICTURE 5.103



Pic. 5.134



Pic. 5.135

ONCE THE PROGRAMME IS COMPLETE PRESS “SAVE” TO STORE IT IN THE COMPUTER, OR INSER THE DRYING PHASE IF IT IS PRESENT AS OPTIONAL ON THE MACHINE. IT MAY NOW BE RECALLED FROM THE LIST FOR ANY CHANGES TO BE MADE OR TO ASSOCIATE IT WITH AN INSTRUMENT.

REPEAT THE PROCESS TO CREATE OTHER PROGRAMMES.



Pic. 5.136

7. PROGRAMMED MAINTENANCE

7.1 Operator maintenance

Components	ENDOSCOPES WASHER EW2								TIME	
	Programmed maintenance scheme									
	Step	months								Activity
make every.....	3	6	9	12	15	18	24			
Chamber filters	make every day								Take filters and clean.	10'
Leak test filter	make every.....		x						Replace.	1'
Chemical compartment	make every day								Check every day the cleaning of chemical compartment. In case of leakage call the service.	5'
Pre filter dryer F5 (optional)	make every 100 hours								Replace.	2'
Dryer HEPA filter (optional)	make every 300 hours								Replace.	1'
Water filters 0,45 micron	make every.....		x						Replace.	10'
Water filters 0,1 micron	make every.....		x						Replace.	10'
Channel filters 40 micron	make every.....		x						Replace.	10'
Temperature probes	make every.....				x				During periodic validation, check the sensor status.	5'
Safety thermostat	make every.....				x				Verify the sensor.	1'

Chemical/water flowmeter	make every.....		x					Check the impeller, the calibration and the presence of leakages.	30'
Chemical tank level sensor	make every.....		x					Check the functioning and visual control (calcium problems)	5'
Dosing pump connection pipe	make every.....		x					Check of any crashing, leakages or hardening.	2'
Washing arms or nozzles	every week							Check for free rotation. Open the cleaning caps and wash inside; check and in case clean the nozzles.	2'
Door gasket	make every...		x					Verify the gasket and replace after 6000 cycles.	20'
Washing pumps	make every...				x			Check for water leakage from the arm seal and correct flow (50 to 60 impuls)	5'
Water heating element	make every...				x			Check for water leakage from the gasket and check ampere consumption (10 A per unit)	1'
Dryer heating elements (optional)	none							Operation is checked by the control system and check ampere consumption	5'
Macchine nozzles of basket connection	make every...		x					Check the status and replace the oring.	2'
Drain pump	make every...				x			Check for any leakage, if necessary, remove and clean the membrane seat.	1'
General test	make every...							Operation is checked by the control system. Check for any leakage and the electrical connections.	15'
Unloading water pipe	make every...				x			Check the situation of the pipe and the seal.	5'
Loading water pipes	make every...				x			Check the situation of the pipe and the seal.	5'
Teflon shim inside the chamber	every month							Check the correct position. Fix the screws.	10'

Compressed air	make every...		x						Check the calibration of compressed air for channels purging and pneumatic valves.	2'
Basket/instrument connection pipes	make every day								Check the status of the silicon pipe for the connection of basket/instrument. If necessary replace them.	1'
Machine leakage	make every...		x						Check the leakage during functioning.	5'
Check the security on the doors	make every		x						Control the correct work of the motors and switch	5'
Silicon pipe basket/instrument	make every...				x				Replace.	20'
Luer connectors male/female basket	make every...						x		Replace.	30'
Endoscope oring connectors	make every day								Check the status of the oring. If necessary, replace them.	1'
Basket	make every		X						Check the status of screws and if necessary, fix it.	5'

7.2 Technician maintenance

To perform a correct maintenance proceeds as explained on the following points:

POINT 1 – Verify and re-enable any possible liquid leak from the machine, pay particular attention to the door gasket integrity (see Pic. 7.2) and to the Teflon pipe connected to the flow meters and channel pumps (see Pic. 7.1). Check all this parts when the machine is working (washing or rinsing phase).


During the gasket check, verify the edge integrity (see Pic. 7.3) using a pocket torch, aiming the light from the chamber internal plastic edge to the external side. The light doesn't have to pass through the plastic edge; otherwise seal again with some silicone from the internal, to the external side.



Pic. 7.1




Pic.7.2



Pic. 7.3



Pic. 7.4




Pic. 7.5


POINT 2 – Check and eventually clean the filter on the channel pumps suction circuit (see Pic. 7.4); this filter is accessible from the rear side of the machine or from the unloading side (only for the double door version).

POINT 3 – Check and eventually re-enable the door/s lock/s safety and, for a correct functioning, check that the closure have to be perfectly lean to the end-switch block (see Pic.7.5).

POINT 4 – Check the usury condition of the coloured pipe on the endoscope basket (see Pic.7.6) and replace them apart from their condition, following the instruction on the table on section 6.8.1 and replace the luer-lock too



Pic. 7.6




Pic. 7.7

POINT 5 –Check if the basket can rotate correctly inside the chamber and eventually regulate the position by 4 registers (2 pair of centering roller – see Pic. 7.7). Follow the procedure:


- Loosen the 4 regulator nuts.
- Insert the basket into the washing chamber and activate the connection piston manually.
- Fix the 2 nuts (in case of double doors machine, fix all 4 nuts).
- Extract the basket and fix firmly the 4 regulator nuts.

POINT 6 – Clean the chamber filter.



Pic. 7.8

POINT 7 – Check the water and chemical flowmeters calibration (see chapter 6.4.1), verifying the suction nozzle and the electrical and hydraulic connections, if one or more flowmeters can't maintain the calibration, try to open and clean them paying attention, because they are used with chemical products. Replace the O-ring. – **PROVIDE THE PERSONNEL WITH INDIVIDUAL PROTECTION FOR THE EYES, RESPIRATORY WAY AND SKIN** – if necessary, replace them (see pic. 7.9).



Pic. 7.9


POINT 8 – To check the channels pumps correct functioning proceed as follows:

- Select “manual” for the machine functioning;
- Fill 12 litres of water and simulate the door on the position “closed”;
- Insert the basket into the washing chamber and activate the connection blocks;
- Close the leak test with the opposite closure provided with the machine
- Enable the water pneumatic valve for the washing arms, recirculated pump and instrument 1 and 2;
- Manually enable the channels pumps, without any connection on the outlet (colored silicone pipe);

Close one by one the endoscope channels after 60” of functioning (colored pipe on basket into washing chamber) checking if the pressure value on the monitor is higher than 2 bar; in this way it is possible to verify the pumps integrity and the channels max pressure alarm. If it is necessary, clean one of the pumps: on the following photos are indicated the 2 types of pumps with the correct sequence for disassembly the components (see Pic. 7.10).

THE MAXIMUM PRESSURE OF KNF 30 PUMP IS 600 kPa (6 bar). IF IT IS USED A PRESSURE HIGHER THE PUMP IS PERMANENTLY DAMAGED.

PUMP FOR CHANNEL NUMBER 3 and 4




Pic. 7.10

POINT 9 – Replace the water and chemical filters as indicated on table 7.3A.


POINT 10 – Check and calibrate the machine sensors (leak test, capacitive sensor chemical tank 1, flowmeters, temperature probe etc.. . For the configurations and calibrations see chapter 6).

POINT 11 – [Replace the basket connection pistons o-ring](#) (see pic.7.11 – the 2 pistons o-ring are not visible, so is necessary to disassembly the connection block for the replacement).



Pic. 7.11

POINT 12 – Check and eventually fix the basket screw, of the guide between the pistons and of the basket lock (in case of rotating basket).



The basket rotation has to be locked without force, but it hasn't the clearance on the upper part; otherwise the basket will not hook into the machine or the basket will hooked badly, damaging the connections.


Pic. 7.12

POINT 13 – – Check the heating element electrical input (10A for each one).



Pic. 7.13

POINT 14 – Check the basket lock on the door(s) Pic. 7.14, and the basket lock in INOX (used for loading/unloading basket). They must be intact and work correctly. If necessary, replace them.



Pic. 7.14




Pic. 7.15

POINT 15 – Check and eventually calibrate the capacitive sensor for draining and filling water control – Pic. 7.15.

POINT 16 – Check and eventually calibrate the pistons movement of basket hooking. Use the regulator placed above the pistons (Pic. 7.16) to calibrate the movement. In the models that don't use lubricant chemicals on the last rinse, before the calibration, clean and lubricate the pistons with sterile oil used on surgical instruments or spray silicone used for surgical instruments. The work correctly the two pistons have to entry at the same time into the washing chamber (remove the basket). The return phase could be not equal on the movement (unlock basket).

Check and eventually replace the 2 sensor of piston movement by using the appropriate instruments (pic. 7.16).



Pic. 7.16

POINT 17 - Check and eventually calibrate (every six months) the safety valve of leak test cylinder. The intervention value must be between 350 mbar and 400 mbar (pic. 7.17).




Pic. 7.17

POINT 18 – Check the correct calibration chemical 1 level and the correct reading of RFID (Pic. 7.18 – Chemical 1 sensor). Follow the procedure:

- **Calibration of chemical 1 level sensor:** Fill an empty tank (same used normally for the chemical 1) with 300 ml of water; place it on the balance and calibrate the specific screw until the system displays the “CHEM. 1 LOW LEVEL” message (Warning). Wait the system delay for the identification of warning and the visualization on the display. Repeat the procedure until the correct functioning.


This procedure allows the machine to notify an alarm for chemical lack, 2 cycle before, for the disinfection/sterilization phase.

- **Reading check of chemical 1 antenna RFID:** place correctly the chemical tank on the specific compartment. Enter utility menu pressing “UTILITY” button on the touch-screen and press “RFID STATE” button. On this page it is possible to display all informations of chemical tank (s) that must be correct for type, quantity, lot, expiration date, etc...



Pic. 7.18

POINT 19 – ATTENTION: check the correct status of channels connection; they have to be movable to avoid the leak or not correct work to avoid false alarm on the endoscope channels activated (pic. 7.19).



Pic. 7.19

POINT 20 – Connectors/channels separators. Check the O-Ring status every week



Pic. 7.20

7.3 Necessary materials

Below are described the necessary materials and the replacement time of the components usury-prone, that have to be checked during the standard maintenance:

Tab. 7.3A CONSUMABLE		
DESCRIPTION	CODE	REPLACEMENT TIME
0,45 μ WATER FILTER	660033	Replace max. every 6 months (independently from the cycles done)
0,1 μ WATER FILTER	660032	Replace max. every 6 months (independently from the cycles done)
AIR FILTER LEAK TEST	660010	Replace max. every 6 months (independently from the cycles done)
CHANNEL FILTER 40 μ	661050	Replace max. every 6 months (independently from the cycles done)
AIR COMPRESSOR FILTER OPTIONAL	660196	Replace max. every 6 months (independently from the cycles done)
5 KG STEELCO SC	9992037-9992038	Duration 2 years
10 KG STEELCO SEPTO PAC	9992039-9992040	Duration 1 year, with limit 100 days for the opening
5 KG MEDIKLAR	9992021	Duration 2 years
5 KG STEELCO XIDE A	-	Duration 1 year, with limit 100 days for the opening
5 KG STEELCO XIDE B	-	Duration 1 year, with limit 100 days for the opening
5 KG STEELCO DT	-	Duration 1 year, with limit 100 days for the opening
Set colorate tubing for basket	9991344	Duration 1 year
Kit o-ring gasket for the piston block (22+2)	660017 - 660013	Duration 6 months max
Door gasket	660516 - 660517	Duration 1 year max

8. MACHINE ALARMS

IN CASE OF ALARM, AFTER IT HAS BEEN CLEARED, THE MACHINE WILL CARRY OUT 2 FINAL RINSES AND DEPENDING ON THE RELATED PARAMETER IT CONTINUES THE CYCLE, RESTART IT OR ABORT IT.

8.1 Trouble shooting

In the following table are listed the alarms for the machine EW2, together with relevant possible causes and solutions. The explanations given below must not be considered exhaustive. Only the more frequent causes have been considered.

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
00 MOTHERBOARD PROBLEMS			- seek technical assistance.			X
01 PLC FAILURES (XOB 10-12 VTEST)			- seek technical assistance.			X
02 FAILUER ON TRANSD. OF LT TANK	During the leak test the pressure sensor, which control the air expansion cylinder, doesn't read correctly the pressure value.	- pressure sensor failure. - PLC board failure (see electrical scheme).	- seek technical assistance.			X
03 WATER LEAK	The system detects the water leak under the machine during the functioning normally.	- sensor failure. - leak on hydraulic circuit.	- sensor replacement. - check hydraulic circuit.	X in part	X in part	X
04 FAILURE ON DRYING PROBE	During the external drying phase or during the normally functioning, the temperature sensor doesn't read correctly.	- sensor failure.	- replace the sensor. - seek technical assistance.		X	X
05 PLC BLOCKED			- seek technical assistance			X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
06 ENERGY LACK ON CYCLE	During the cycle the power supply has failed.	- machine circuit breaker intervention. - no power supply.	- open the machine electrical panel, reset the circuit breaker and the running cycle. - restore the power supply and reset the running cycle.			X
07 EMERGENCY STOP	During the cycle a failure on the auxiliary has occurred.	- machine auxiliary circuit breaker intervention, auxiliary on failure.	- open the machine electrical panel, reset the circuit breaker and the running cycle. Check that the auxiliary is working.			X
08 MAGNETOTHERM. INTERV.	During the cycle a failure on one of the electrical components has occurred.	- fault in one of the pumps or in a solenoid valve (earth leakage).	- check the pumps windings and possible earth leakages of the machine solenoid valves.			X
09 HEAT.EL. THERM.SAFETY	During heating, protection against the heating elements overheating has occurred.	- sensor failure. - lack of water in the chamber. - temperature probes failure. - heating elements relay failure.	- check and/or replace the heating elements protection sensor. - check and/or replace the chamber temperature probes. - check and/or replace the heating elements. - check for possible water leakages from the washing chamber.			X
10 CONNECTION LACK: AIR/WATER	The pressure of the incoming compressed air is below the expected 6 bar or the water pressure is below the expected 2 bar.	- failure of the pressure switch that controls the minimum compressed air pressure. - compressed air not available. - failure of the pressure switch that controls the minimum water pressure. - water not available.	- check and/or replace the control pressure switch. - check and restore the inlet compressed air. - check and/or replace the control pressure switch. - check and/or replace the control pressure switch.			X
11 TANK TOO FULL	The washing chamber has been filled with more than 16 liters of water	- water flow meter failures. - PLC failure.	- check the calibration and the functioning of the two water flow meters. - PLC check.			X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
12 LOADING DOORS FAILURE TIME	The loading door doesn't close.	<ul style="list-style-type: none"> - door motor failure. - door microswitch failure. - door chain 	<ul style="list-style-type: none"> - check and/or replace the door motor. - check and/or replace the door microswitch. - check and/or replace the door chain. 			X
13 UNLOADING TIME FAILURE	The unloading door doesn't close.	<ul style="list-style-type: none"> - door motor failure. - door microswitch failure. - door chain 	<ul style="list-style-type: none"> - check and/or replace the door motor. - check and/or replace the door microswitch. - check and/or replace the door chain. 			X
14 L. DOOR NOT CLOSED ON CYCLE	The loading door has opened during the cycle.	<ul style="list-style-type: none"> - door position microswitch failure. - door chain 	<ul style="list-style-type: none"> - check and/or replace the door microswitch. - check and/or replace the door chain. 			X
15 UNL. DOOR NOT CLOSED ON CYCLE	The unloading door has opened during the cycle.	<ul style="list-style-type: none"> - door position microswitch failure. - door chain 	<ul style="list-style-type: none"> - check and/or replace the door microswitch. - check and/or replace the door chain. 			X
16 NORMAL DRAIN FAILURES	There is still water in the washing chamber or a fault in the minimum level sensor during the draining phase.	<ul style="list-style-type: none"> - drain pump failure (if drain pump is fitted). - chamber minimum level sensor failure. - failure on the drain valve pilot valve. - pneumatic drain valve failure. 	<ul style="list-style-type: none"> - check the drain pump. - check for water in the washing chamber and eventually check the correct functioning of the drain valve and its pilot valve. 			X
17 SPECIAL DRAIN FAILURES	N.C.					
18 BLOWER FAILURES (PR. SWITCH)	During the external drying phase of instrument, the pressure switch (control) has indicated pressure lack on outside of blower.	<ul style="list-style-type: none"> - pressure switch failure. - blower failure. - blower filter obstructed. - problem on air circuit 	<ul style="list-style-type: none"> - check that the aspiration filters of blower are clean. - check blower motor. - check blower pressure switch. - check air circuit 		X	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
19 TANK HEAT. FAILURES (TIME)	The set temperature has not been reached during the heating phase	<ul style="list-style-type: none"> - washing chamber heating elements faulty. - heating element remote control switch faulty. - PLC command fault (relay board). 	<ul style="list-style-type: none"> - check the washing chamber heating elements. - check the operation of the heating element remote control switch. - check the command board controlling the heating element remote control switch. 			X
20 DRYER HEAT. FAILURES (TIME)	The temperature did not increase 1°C in the set time during the external drying phase of instrument.	<ul style="list-style-type: none"> - dryer heating elements faulty. - missing permission to activate dryer heating element. - heating element remote control switch/relay faulty. 	<ul style="list-style-type: none"> - check the dryer heating elements. - check that the permissions to activate the dryer heating elements are OK. - check the command board controlling the heating element remote control switch. 		X	X
21 TANK HEAT. EMERGENCY	The water temperature has exceeded 60°C during the heating phase.	<ul style="list-style-type: none"> - PLC failure - heating element remote control switch/relay faulty. 	<ul style="list-style-type: none"> - check the command board controlling the heating element remote control switch. - check the command board controlling the heating element remote control switch. 			X
22 DRYER HEAT. EMERGENCY	The mechanic sensor has tripped to protect the maximum temperature of heating element during the external drying phase of instrument in the chamber.	<ul style="list-style-type: none"> - blower failure/not working. - mechanical sensor failure. - heating element remote control switch/relay faulty. 	<ul style="list-style-type: none"> - check the correct functioning of blower. - check the correct functioning of mechanical sensor placed on heating element. - check the command board controlling the heating element remote control switch. 		X	X
23 EXTREME WATER TEMP	The water temperature has exceeded the set temperature by 5°C during the heating phase.	<ul style="list-style-type: none"> - heating element remote control switch faulty. - PLC command fault (relay board) 	<ul style="list-style-type: none"> - check the command board controlling the heating element remote control switch. - check and/or replace the heating elements remote control switch. 			X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
24 WATER FILL. 1 CHAMBER (TIME)	The washing chamber has not been filled with the set quantity of water for the running phase.	<ul style="list-style-type: none"> - water inlet pneumatic valve fault. - water inlet valve pilot valve fault. - PLC command fault (relay board). - failure in flowmeters controlling the water when filling the washing chamber. - water filters clogged 	<ul style="list-style-type: none"> - check and/or replace the machine water input pneumatic valve. - check and/or replace the water input command valve pilot valve. - check and/or replace the water load command valve relay board or water load pump command. - check and/or replace the water load pump. - check and/or replace the water chamber quantity flow meters. - check and/or replace the machine water filters. 			X
25 WATER FILL. 2 CHAMBER (TIME)	The washing chamber has not been filled with the set quantity of water for the running phase.	<ul style="list-style-type: none"> - water inlet pneumatic valve fault. - water inlet valve pilot valve fault. - PLC command fault (relay board). - failure in flowmeters controlling the water when filling the washing chamber. - water filters clogged 	<ul style="list-style-type: none"> - check and/or replace the machine water input pneumatic valve. - check and/or replace the water input command valve pilot valve. - check and/or replace the water load command valve relay board or water load pump command. - check and/or replace the water load pump. - check and/or replace the water chamber quantity flow meters. - check and/or replace the machine water filters. 			X
26 TEMP.SENS. FAILURE (CONTROL)	Incorrect temperature reading in the washing chamber (control).	<ul style="list-style-type: none"> - PT1000 washing chamber faulty. - PLC reader card faulty. 	<ul style="list-style-type: none"> - check and/or replace the chamber temperature probe (control). - check and/or replace the temperature probe reader card. 			X
27 TEMP.SENS. FAILURE (RECORD)	Incorrect temperature reading in the washing chamber (recording).	<ul style="list-style-type: none"> - PT1000 washing chamber faulty. - PLC reader card faulty. 	<ul style="list-style-type: none"> - check and/or replace the chamber temperature probe (recording). - check and/or replace the temperature probe reader card. 			X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
28 LOADING TIME CHEM.1 EXCESSIVE	During the chemical loading phase, the time limit for reaching the set quantity parameter has been exceeded.	<ul style="list-style-type: none"> - fault or malfunction in chemical 1 load pump. - fault or malfunction in chemical 1 load control flow meters. 	<ul style="list-style-type: none"> - clean and/or replace chemical 1 load pump. - clean and/or replace chemical 1 load control flow meters, then carry out the chemical calibration procedure. 			X
29 LOADING TIME CHEM.2 EXCESSIVE	During the chemical loading phase, the time limit for reaching the set quantity parameter has been exceeded.	<ul style="list-style-type: none"> - fault or malfunction in chemical 2 load pump. - fault or malfunction in chemical 2 load control flow meters. 	<ul style="list-style-type: none"> - clean and/or replace chemical 2 load pump. - clean and/or replace chemical 2 load control flow meters, then carry out the chemical calibration procedure. 			X
30 LOADING TIME CHEM.3 EXCESSIVE	During the chemical loading phase, the time limit for reaching the set quantity parameter has been exceeded.	<ul style="list-style-type: none"> - fault or malfunction in chemical 3 load pump. - fault or malfunction in chemical 3 load control flow meters. 	<ul style="list-style-type: none"> - clean and/or replace chemical 3 load pump. - clean and/or replace chemical 3 load control flow meters, then carry out the chemical calibration procedure. 			X
31 LOADING TIME CHEM.4 EXCESSIVE	During the chemical loading phase, the time limit for reaching the set quantity parameter has been exceeded.	<ul style="list-style-type: none"> - fault or malfunction in chemical 4 load pump. - fault or malfunction in chemical 4 load control flow meters. 	<ul style="list-style-type: none"> - clean and/or replace chemical 4 load pump. - clean and/or replace chemical 4 load control flow meters, then carry out the chemical calibration procedure. 			X
32 WASH.ARMS FLOW SENSOR BREACKDOWN	There is no flow in the washing chamber wash.arm circuit for more than 5 seconds.	<ul style="list-style-type: none"> - fault or malfunction in washing chamber impeller control flow meters. - recirculation pump faulty. - breakage in recirculation pump electric power supply remote control switch. - PLC control card faulty. 	<ul style="list-style-type: none"> - check and/or replace the washing chamber impeller control flow meters. - check recirculation pump operation or replace it in the event of faulty. - check and/or replace the recirculation pump electric power supply remote control switch command board. - check and/or replace the recirculation pump electric power supply remote control switch. 			X
33 CONTROL TANK CONNECTION LEAK TEST	During the first phase of cycle, the leak test pressure was below the set value.	<ul style="list-style-type: none"> - leak test pump broken. - set 1 disconnected. - set 2 disconnected. - set 3 disconnected. 	<ul style="list-style-type: none"> - check and/or replace the leak test pump. - check the set 1 connection. - check the set 2 connection. - check the set 3 connection. 			X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
34 CONTROL CONNECTION SET 1	During the cycle the pressure on the collector was below the set value.	<ul style="list-style-type: none"> - failure of electronic sensor for pressure collector 1. - failure of mechanical sensor for pressure collector 1. - PLC card faulty. 	<ul style="list-style-type: none"> - check and/or replace the electronic sensor. - check and/or replace the mechanical sensor. - check and/or replace sensor control card/s and channel pump command. 		X	X
35 CONTROL CONNECTION CHA.3 SET 1	During the cycle there was no minimum working pressure for channel 3 instrument 1.	<ul style="list-style-type: none"> - breakage in drum connection pipe (red). - disconnection in relative connector. - channel 3 instrument 1 pump faulty. - channel 3 instrument 1 pressure sensor faulty. - PLC card faulty. - pump broken 	<ul style="list-style-type: none"> - check pipe integrity and eventually replace it. - check if the connector is properly fastened to the instrument. - check and/or replace channel 3 instrument 1 pump. - check and/or replace channel 3 instrument 1 pressure sensor. - check and/or replace sensor control card/s and channel pump command. - check and/or replace the pump 	X in part	X in part	X
36 CONTROL CONNECTION CHA.4 SET 1	During the cycle there was no minimum working pressure for channel 4 instrument 1.	<ul style="list-style-type: none"> - breakage in drum connection pipe (black). - disconnection in relative connector. - channel 4 instrument 1 pump faulty. - channel 4 instrument 1 pressure sensor faulty. - PLC card faulty. - pump broken 	<ul style="list-style-type: none"> - check pipe integrity and eventually replace it. - check if the connector is properly fastened to the instrument. - check and/or replace channel 4 instrument 1 pump. - check and/or replace channel 4 instrument 1 pressure sensor. - check and/or replace sensor control card/s and channel pump command. - check and/or replace the pump 	X in part	X in part	X
37 CONTROL CONNECTION SET 2	During the cycle the pressure on the collector was below the set value.	<ul style="list-style-type: none"> - failure of electronic sensor for pressure collector 2. - failure of mechanical sensor for pressure collector 2. - PLC card faulty. 	<ul style="list-style-type: none"> - check and/or replace the electronic sensor. - check and/or replace the mechanical sensor. - check and/or replace sensor control card/s and channel pump command. 		X	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">38 CONTROL CONNECTION CHA.3 SET 2</p>	During the cycle there was no minimum working pressure for channel 3 instrument 2.	<ul style="list-style-type: none"> - breakage in drum connection pipe (red). - disconnection in relative connector. - channel 3 instrument 2 pump faulty. - channel 3 instrument 2 pressure sensor faulty. - PLC card faulty. - pump broken 	<ul style="list-style-type: none"> - check pipe integrity and eventually replace it. - check if the connector is properly fastened to the instrument. - check and/or replace channel 3 instrument 2 pump. - check and/or replace channel 3 instrument 2 pressure sensor. - check and/or replace sensor control card/s and channel pump command. - check and/or replace the pump 	X in part	X in part	X
<p style="text-align: center;">39 CONTROL CONNECTION CHA.4 SET 2</p>	During the cycle there was no minimum working pressure for channel 4 instrument 2.	<ul style="list-style-type: none"> - breakage in drum connection pipe (black). - disconnection in relative connector. - channel 4 instrument 2 pump faulty. - channel 4 instrument 2 pressure sensor faulty. - PLC card faulty. - pump broken 	<ul style="list-style-type: none"> - check pipe integrity and eventually replace it. - check if the connector is properly fastened to the instrument. - check and/or replace channel 4 instrument 2 pump. - check and/or replace channel 4 instrument 2 pressure sensor. - check and/or replace sensor control card/s and channel pump command. - check and/or replace the pump 	X in part	X in part	X
<p style="text-align: center;">40 CONTROL CONNECTION SET 3</p>	During the cycle the pressure on the collector was below the set value.	<ul style="list-style-type: none"> - failure of electronic sensor for pressure collector 3. - failure of mechanical sensor for pressure collector 3. - PLC card faulty. 	<ul style="list-style-type: none"> - check and/or replace the electronic sensor. - check and/or replace the mechanical sensor. - check and/or replace sensor control card/s and channel pump command. 		X	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">41 CONTROL CONNECTION CHA.3 SET 3</p>	<p>During the cycle there was no minimum working pressure for channel 3 instrument 3.</p>	<ul style="list-style-type: none"> - breakage in drum connection pipe (red). - disconnection in relative connector. - channel 3 instrument 3 pump faulty. - channel 3 instrument 3 pressure sensor faulty. - PLC card faulty. - pump broken 	<ul style="list-style-type: none"> - check pipe integrity and eventually replace it. - check if the connector is properly fastened to the instrument. - check and/or replace channel 3 instrument 3 pump. - check and/or replace channel 3 instrument 3 pressure sensor. - check and/or replace sensor control card/s and channel pump command. - check and/or replace the pump 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">42 CONTROL CONNECTION CHA.4 SET 3</p>	<p>During the cycle there was no minimum working pressure for channel 4 instrument 3.</p>	<ul style="list-style-type: none"> - breakage in drum connection pipe (black). - disconnection in relative connector. - channel 4 instrument 3 pump faulty. - channel 4 instrument 3 pressure sensor faulty. - PLC card faulty. - pump broken 	<ul style="list-style-type: none"> - check pipe integrity and eventually replace it. - check if the connector is properly fastened to the instrument. - check and/or replace channel 4 instrument 3 pump. - check and/or replace channel 4 instrument 3 pressure sensor. - check and/or replace sensor control card/s and channel pump command. - check and/or replace the pump 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">43 CONTROL CONNECTION SET 1 LEAK TEST</p>	<p>At the beginning of cycle, the leak test has not reached the minimum set pressure.</p>	<ul style="list-style-type: none"> - pump broken. - electronic sensor for pressure broken. - PLC card faulty. 	<ul style="list-style-type: none"> - check and/or replace the pump. - check and/or replace the electronic sensor. - check pipe integrity and eventually replace it. - check connections and eventually replace it. - check oring integrity and eventually replace it. - check and/or replace sensor control card/s and channel pump command. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
44 CONTROL CONNECTION SET 2 LEAK TEST	At the beginning of cycle, the leak test has not reached the minimum set pressure.	<ul style="list-style-type: none"> - pump broken. - electronic sensor for pressure broken. - PLC card faulty. 	<ul style="list-style-type: none"> - check and/or replace the pump. - check and/or replace the electronic sensor. - check pipe integrity and eventually replace it. - check connections and eventually replace it. - check oring integrity and eventually replace it. - check and/or replace sensor control card/s and channel pump command. 	X in part	X in part	X
45 CONTROL CONNECTION SET 3 LEAK TEST	At the beginning of cycle, the leak test has not reached the minimum set pressure.	<ul style="list-style-type: none"> - pump broken. - electronic sensor for pressure broken. - PLC card faulty. 	<ul style="list-style-type: none"> - check and/or replace the pump. - check and/or replace the electronic sensor. - check pipe integrity and eventually replace it. - check connections and eventually replace it. - check oring integrity and eventually replace it. - check and/or replace sensor control card/s and channel pump command. 	X in part	X in part	X
46 EVENT REGISTR. FAILURES (TIME)	It is not possible to record cycle events.	<ul style="list-style-type: none"> - PLC fault/malfunction. 	<ul style="list-style-type: none"> - check and eventually reset set the system. 			X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
47 CHA.3 SET 3 OBSTRUCTED	During the cycle an obstruction was found in the endoscopic channel connected to channel 3 instrument 3 with ON-OFF function	<ul style="list-style-type: none"> - organic material or of a different nature obstructing the endoscopic channel of the instrument or the connection pipe to the instrument. - channel 3 instrument 3 pressure sensor faulty. - incorrect connector connection or bent connector pipe. - endoscopic instrument maintenance carried out badly (reduction in channel diameter different from the original). - set parameters incorrect. - endoscope code insertion different from that attributed. 	<ul style="list-style-type: none"> - verify openness of the channel and the connection pipe. - check and/or replace the pressure sensor. - check that the pipe is correctly positioned inside the drum and the connector. - check that the set maximum pressure parameters are correct. - check that the code recalled for the instruments matches the inserted instrument. - as a final analysis have the instrument checked by an authorised technical assistance. 	X in part	X in part	X
48 CHA.4 SET 3 OBSTRUCTED	During the cycle an obstruction was found in the endoscopic channel connected to channel 4 instrument 3 with ON-OFF function	<ul style="list-style-type: none"> - organic material or of a different nature obstructing the endoscopic channel of the instrument or the connection pipe to the instrument. - channel 4 instrument 3 pressure sensor faulty. - incorrect connector connection or bent connector pipe. - endoscopic instrument maintenance carried out badly (reduction in channel diameter different from the original). - set parameters incorrect. - endoscope code insertion different from that attributed. 	<ul style="list-style-type: none"> - verify openness of the channel and the connection pipe. - check and/or replace the pressure sensor. - check that the pipe is correctly positioned inside the drum and the connector. - check that the set maximum pressure parameters are correct. - check that the code recalled for the instruments matches the inserted instrument. - as a final analysis have the instrument checked by an authorised technical assistance. 	X in part	X in part	X
49 INCONGRUITY LEAK TEST CAP 3	When the control is activated, the instrument presence sensor is active with the endoscope in the position 3.	<ul style="list-style-type: none"> - sensor failure - wrong connection of operator - wrong insertion of operator on touch-screen 	<ul style="list-style-type: none"> - replace or calibrate the sensor - check the correct connection of endoscope 	X in part	X	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
50 W.ARMS FLOW TR. OUT OF LIM.-	During the cycle the flow sensor for controlling the passage of liquid to the washing impellers has read a value lower than the set minimum.	<ul style="list-style-type: none"> - recirculation pump malfunction. - no filter in chamber (recirculation pump cavitation - sucks air). - flow sensor broken. - control board electric contact or flow sensor disconnected. 	<ul style="list-style-type: none"> - check the operational state of the wash pump and/or replace it. - check that the filter is in the tank and is properly clean. - check and/or replace the flow sensor. - check and/or replace the flow sensor electric connections 	X in part	X in part	X
51 W.ARMS FLOW TR. OUT OF LIM.+	During the cycle the flow sensor for controlling the passage of liquid to the washing impellers has read a value higher than the set maximum.	<ul style="list-style-type: none"> - flow sensor broken. - set parameter less than 90 impulses. - spiral water pneumatic valve deactivated/faulty. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check the set parameters (default 140). - check that the spiral water pneumatic valve is open during washing. 	X in part	X in part	X
52 CHA.1 SET 3 OBSTRUCTED	During the cycle in the flow control stage, the system has detected an obstruction in channel 1 instrument 3.	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 1 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 1 connection pipe (brown). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">53 CHA.2 SET 3 OBSTRUCTED</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 2 instrument 3.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 2 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 2 connection pipe (green). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">54 CHA.3 SET 3 OBSTRUCTED</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 3 instrument 3.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 3 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 3 connection pipe (blue). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">55 CHA.4 SET 3 OBSTRUCTED</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 4 instrument 3.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 4 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 4 connection pipe (yellow). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">56 CHA.5 SET 3 OBSTRUCTED</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 5 instrument 3.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 5 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 5 connection pipe (grey). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">57 CHA.6 SET 3 OBSTRUCTED</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 6 instrument 3 when pump is continuous working.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 6 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 6 connection pipe (red). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">58 CHA.7 SET 3 OBSTRUCTED</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 7 instrument 3 when pump is continuous working.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 7 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 7 connection pipe (black). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">59 CHA.3 SET 1 DISCONNECTED</p>	<p>During the cycle the connector in channel 3 instrument 1 has detached from the instrument</p>	<ul style="list-style-type: none"> - connector not connected to the instrument perfectly. - connector worn. - channel 3 instrument 1 pressure sensor faulty. - channel 3 instrument 1 pipe broken (red). 	<ul style="list-style-type: none"> - wear personal protective equipment and check that the joint is connected correctly, in case reconnect and restart the cycle. - replace the connector if worn or has a broken o ring seal. - check and/or replace the pressure sensor. - check and/or replace the channel pipe. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
60 CHA.4 SET 1 DISCONNECTED	During the cycle the connector in channel 4 instrument 1 has detached from the instrument.	<ul style="list-style-type: none"> - connector not connected to the instrument perfectly. - connector worn. - channel 4 instrument 1 pressure sensor faulty. - channel 4 instrument 1 pipe broken (black). 	<ul style="list-style-type: none"> - wear personal protective equipment and check that the joint is connected correctly, in case reconnect and restart the cycle. - replace the connector if worn or has a broken o ring seal. - check and/or replace the pressure sensor. - check and/or replace the channel pipe. 	X in part	X in part	X
61 DRYING HEATING EMERGENCY	The air temperature has exceeded 110°C during the drying phase into the heating elements.	<ul style="list-style-type: none"> - PLC failure - heating element remote control switch/relay faulty. 	<ul style="list-style-type: none"> - check the command board controlling the heating element remote control switch. - check the command board controlling the heating element remote control switch. 			X
62 SECUR. INTERVENTION LOADING DOOR						
63 SECUR. INTERVENTION DRAIN DOOR						
66 CHA.3 SET 2 DISCONNECTED	During the cycle the connector in channel 3 instrument 2 has detached from the instrument	<ul style="list-style-type: none"> - connector not connected to the instrument perfectly. - connector worn. - channel 3 instrument 2 pressure sensor faulty. - channel 3 instrument 2 pipe broken (red). 	<ul style="list-style-type: none"> - wear personal protective equipment and check that the joint is connected correctly, in case reconnect and restart the cycle. - replace the connector if worn or has a broken o ring seal. - check and/or replace the pressure sensor. - check and/or replace the channel pipe. 	X in part	X in part	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">67 CHA.4 SET 2 DISCONNECTED</p>	<p>During the cycle the connector in channel 4 instrument 2 has detached from the instrument.</p>	<ul style="list-style-type: none"> - connector not connected to the instrument perfectly. - connector worn. - channel 4 instrument 2 pressure sensor faulty. - channel 4 instrument 2 pipe broken (black). 	<ul style="list-style-type: none"> - wear personal protective equipment and check that the joint is connected correctly, in case reconnect and restart the cycle. - replace the connector if worn or has a broken o ring seal. - check and/or replace the pressure sensor. - check and/or replace the channel pipe. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">68 CHA.3 SET 3 DISCONNECTED</p>	<p>During the cycle the connector in channel 3 instrument 3 has detached from the instrument</p>	<ul style="list-style-type: none"> - connector not connected to the instrument perfectly. - connector worn. - channel 3 instrument 3 pressure sensor faulty. - channel 3 instrument 3 pipe broken (red). 	<ul style="list-style-type: none"> - wear personal protective equipment and check that the joint is connected correctly, in case reconnect and restart the cycle. - replace the connector if worn or has a broken o ring seal. - check and/or replace the pressure sensor. - check and/or replace the channel pipe. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">69 CHA.4 SET 3 DISCONNECTED</p>	<p>During the cycle the connector in channel 4 instrument 3 has detached from the instrument.</p>	<ul style="list-style-type: none"> - connector not connected to the instrument perfectly. - connector worn. - channel 4 instrument 3 pressure sensor faulty. - channel 4 instrument 3 pipe broken (black). 	<ul style="list-style-type: none"> - wear personal protective equipment and check that the joint is connected correctly, in case reconnect and restart the cycle. - replace the connector if worn or has a broken o ring seal. - check and/or replace the pressure sensor. - check and/or replace the channel pipe. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
70 CHA.1 SET 3 DISCONNECTED	During the cycle the connector of channel 1 instrument 3 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X
71 CHA.2 SET 3 DISCONNECTED	During the cycle the connector of channel 2 instrument 3 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X
72 CHA.3 SET 3 DISCONNECTED	During the cycle the connector of channel 3 instrument 3 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X
73 CHA.4 SET 3 DISCONNECTED	During the cycle the connector of channel 4 instrument 3 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
74 CHA.5 SET 3 DISCONNECTED	During the cycle the connector of channel 5 instrument 3 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X
75 CHA.6 SET 3 DISCONNECTED	During the cycle the connector of channel 6 instrument 3 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X
76 CHA.7 SET 3 DISCONNECTED	During the cycle the connector of channel 7 instrument 3 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X
77 EXCESSIVE PRESSURE LT SET 3	During the cycle the seal test has exceeded the maximum pressure parameter admitted within the endoscope on instrument 3.	<ul style="list-style-type: none"> - seal test air opener electrovalve seeping (dirty/worn) with contemporary fault in pressure sensor or seal test air pump command relay faulty. - maximum inserted parameter less than 400 mbar. - seal test in alarm during 90°C thermal disinfection. 	<ul style="list-style-type: none"> - check and/or replace the seal test air opener electrovalve shutter. - check and/or replace the seal test pump command board. - change the seal test maximum pressure parameter. 			X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p align="center">78 EXCESSIVE PRESSURE LT SET 1</p>	<p>During the cycle the seal test has exceeded the maximum pressure parameter admitted within the endoscope on instrument 1.</p>	<ul style="list-style-type: none"> - seal test air opener electrovalve seeping (dirty/worn) with contemporary fault in pressure sensor or seal test air pump command relay faulty. - maximum inserted parameter less than 400 mbar. - seal test in alarm during 90°C thermal disinfection. 	<ul style="list-style-type: none"> - check and/or replace the seal test air opener electrovalve shutter. - check and/or replace the seal test pump command board. - change the seal test maximum pressure parameter. 			<p>X</p>
<p align="center">79 EXCESSIVE PRESSURE LT SET 2</p>	<p>During the cycle the seal test has exceeded the maximum pressure parameter admitted within the endoscope on instrument 2.</p>	<ul style="list-style-type: none"> - seal test air opener electrovalve seeping (dirty/worn) with contemporary fault in pressure sensor or seal test air pump command relay faulty. - maximum inserted parameter less than 400 mbar. - seal test in alarm during 90°C thermal disinfection. 	<ul style="list-style-type: none"> - check and/or replace the seal test air opener electrovalve shutter. - check and/or replace the seal test pump command board. - change the seal test maximum pressure parameter. 			<p>X</p>
<p align="center">80 CHA.3 SET 1 OBSTRUCTED PR+</p>	<p>During the cycle in the pressure control stage, the system has detected an obstruction in channel 3 instrument 1 when pump is continuous working.</p>	<ul style="list-style-type: none"> - pressure sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 3 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the pressure sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 3 connection pipe (red). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">81 CHA.4 SET 1 OBSTRUCTED PR+</p>	<p>During the cycle in the pressure control stage, the system has detected an obstruction in channel 4 instrument 1 when pump is continuous working.</p>	<ul style="list-style-type: none"> - pressure sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 4 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the pressure sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 4 connection pipe (black). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">82 CHA.3 SET 2 OBSTRUCTED PR+</p>	<p>During the cycle in the pressure control stage, the system has detected an obstruction in channel 3 instrument 2 when pump is continuous working.</p>	<ul style="list-style-type: none"> - pressure sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 3 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the pressure sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 3 connection pipe (red). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">83 CHA.4 SET 2 OBSTRUCTED PR+</p>	<p>During the cycle in the pressure control stage, the system has detected an obstruction in channel 4 instrument 2 when pump is continuous working.</p>	<ul style="list-style-type: none"> - pressure sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 4 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the pressure sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 4 connection pipe (black). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">84 SET 1 LEAK FAILURES</p>	<p>During the cycle the endoscopic instrument in position 1 has lost more than the value set in mbars in the control time set in seconds.</p>	<ul style="list-style-type: none"> - leaking endoscope. - leak test connector worn. - leak test oring worn or broken. - connection pipe (neutral colour) damaged/torn. - oring (pair - position 8) for connection to endoscope drum worn or broken. - machine-drum connection joint not leaking. - charge or discharge electrovalve seeping. - pressure sensor malfunction. 	<ul style="list-style-type: none"> - manually test the endoscope according to the procedures indicated by the manufacturer. - check if the leak test connector is worn or has a worn/broken leaking oring. - check and/or replace the leak test connection pipe. - check and/or replace the o rings on the machine-drum connection connector. - check the shutters of the two leak tests 1 unit electrovalves. - check and/or replace the leak test 1 pressure sensor 			<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">85 SET 2 LEAK FAILURES</p>	<p>During the cycle the endoscopic instrument in position 2 has lost more than the value set in mbars in the control time set in seconds.</p>	<ul style="list-style-type: none"> - leaking endoscope. - leak test connector worn. - leak test oring worn or broken. - connection pipe (neutral colour) damaged/torn. - oring (pair - position 8) for connection to endoscope drum worn or broken. - machine-drum connection joint not leaking. - charge or discharge electrovalve seeping. - pressure sensor malfunction. 	<ul style="list-style-type: none"> - manually test the endoscope according to the procedures indicated by the manufacturer. - check if the leak test connector is worn or has a worn/broken leaking oring. - check and/or replace the leak test connection pipe. - check and/or replace the o rings on the machine-drum connection connector. - check the shutters of the two leak tests 2 unit electrovalves. - check and/or replace the leak test 2 pressure sensor 			X
<p style="text-align: center;">86 FAILURE DRAINING LEAK TEST</p>	<p>At the end of the cycle the pressure remaining inside the endoscope is higher than that set (10 mbar)</p>	<ul style="list-style-type: none"> - leak test 1 pressure sensor (expansion cylinder) faulty. - leak test 1 electrovalve discharge fault. - leak test 1 electrovalve discharge command card faulty. 	<ul style="list-style-type: none"> - check and/or replace the pressure sensor. - check and/or replace the leak test air discharge shutter coil. - check that during the air discharge phase the card feeds the electrovalve, replace it in the case of malfunction. 			X
<p style="text-align: center;">87 REFILL ALARM LT SET 1</p>	<p>The system has exceeded the number of refill leak test set by parameter in set 1.</p>	<ul style="list-style-type: none"> - problem on leak test instrument. - leak from idraulic circuit of leak test. - wrong value of parameter. 	<ul style="list-style-type: none"> - check the instrument leak test. - check the o-ring and connection of leak test. - check the pipe connection. - check the value parameter. 		X	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">88 CHA.3 SET 3 OBSTRUCTED</p>	<p>During the cycle in the pressure control stage, the system has detected an obstruction in channel 3 instrument 3 when pump is continuous working.</p>	<ul style="list-style-type: none"> - pressure sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 3 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the pressure sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 3 connection pipe (red). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">89 CHA.4 SET 3 OBSTRUCTED</p>	<p>During the cycle in the pressure control stage, the system has detected an obstruction in channel 4 instrument 3 when pump is continuous working.</p>	<ul style="list-style-type: none"> - pressure sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 4 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the pressure sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 4 connection pipe (black). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">90 SET 3 LEAK FAILURES</p>	<p>During the cycle the endoscopic instrument in position 3 has lost more than the value set in mbars in the control time set in seconds.</p>	<ul style="list-style-type: none"> - leaking endoscope. - leak test connector worn. - leak test oring worn or broken. - connection pipe (neutral colour) damaged/torn. - oring (pair - position 8) for connection to endoscope drum worn or broken. - machine-drum connection joint not leaking. - charge or discharge electrovalve seeping. - pressure sensor malfunction. 	<ul style="list-style-type: none"> - manually test the endoscope according to the procedures indicated by the manufacturer. - check if the leak test connector is worn or has a worn/broken leaking oring. - check and/or replace the leak test connection pipe. - check and/or replace the o rings on the machine-drum connection connector. - check the shutters of the two leak tests 1 unit electrovalves. - check and/or replace the leak test 3 pressure sensor 			X
<p style="text-align: center;">91 BOTH DOORS OPEN</p>	<p>During the cycle one or both the machine doors have opened</p>	<ul style="list-style-type: none"> - breakage or fault in the door lock 	<ul style="list-style-type: none"> - check that the door lock is working properly. - check that the chamber filter is working properly. 			X
<p style="text-align: center;">92 CHEM. DOS. TEMP TOO HIGH</p>	<p>During the chemical loading phase the detected temperature is higher than that set by 5°C</p>	<ul style="list-style-type: none"> - temperature of water entering the machine is too high. - resistance power supply control switch faulty (connect relay broken or faulty). - incorrect setting of the cycle parameters. 	<ul style="list-style-type: none"> - check the machine water feed mixer. - check and/or replace the resistance control switch and the command relay board. - check cycle settings in the chemical extraction phases, these should not be lower than the temperature of the water entering the machine 			X
<p style="text-align: center;">93 NO HEAT. FOR LOW LEVEL</p>	<p>During the heating phase the water level has dropped below that set.</p>	<ul style="list-style-type: none"> - tank level sensor faulty. - loss of water from hydraulic circuit 	<ul style="list-style-type: none"> - check/clean and/or replace the wash tank minimum water level. - check and/or replace the wash tank water discharge valve. - check and remedy any leaks in the hydraulic circuit. 			X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
94 WATER LEAKS 1	The system detects a water flow during a phase of not water filling.	- filling water valve and filling water pump faulty. - filter sanification valve faulty.	- check/clean and/or replace the filling water valve. - check/clean and/or replace the filling water pump. - check/clean and/or replace the filter sanification valve.			X
95 WATER LEAKS 2	The system detects a water flow during a phase of not water filling.	- filling water valve and filling water pump faulty. - filter sanification valve faulty.	- check/clean and/or replace the filling water valve. - check/clean and/or replace the filling water pump. - check/clean and/or replace the filter sanification valve.			X
96 CHEMICAL LEAKS 1	The system detects a chemical flow during a phase of not chemical 1 loading.	- dosing pump in function - flowmeter failure	- check chemical pump command - check and replace the flowmeter		X	X
97 CHEMICAL LEAKS 2	The system detects a chemical flow during a phase of not chemical 2 loading.	- dosing pump in function - flowmeter failure	- check chemical pump command - check and replace the flowmeter		X	X
98 CHEMICAL LEAKS 3	The system detects a chemical flow during a phase of not chemical 3 loading.	- dosing pump in function - flowmeter failure	- check chemical pump command - check and replace the flowmeter		X	X
99 CHEMICAL LEAKS 4	The system detects a chemical flow during a phase of not chemical 4 loading.	- dosing pump in function - flowmeter failure	- check chemical pump command - check and replace the flowmeter		X	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">100 CHA.1 SET 1 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 1 instrument 1.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 1 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 1 connection pipe (brown). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">101 CHA.2 SET 1 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 2 instrument 1.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 2 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 2 connection pipe (green). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">102 CHA.3 SET 1 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 3 instrument 1.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 3 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 3 connection pipe (blue). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">103 CHA.4 SET 1 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 4 instrument 1.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 4 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 4 connection pipe (yellow). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">104 CHA.5 SET 1 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 5 instrument 1.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 5 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 5 connection pipe (grey). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">105 CHA.6 SET 1 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 6 instrument 1 when pump is continuous working.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 6 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 6 connection pipe (red). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">106 CHA.7 SET 1 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 7 instrument 1 when pump is continuous working.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 7 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 7 connection pipe (black). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">107 CHA.1 SET 1 DISCONNECTED FL+</p>	<p>During the cycle the connector of channel 1 instrument 1 has detached.</p>	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">108 CHA.2 SET 1 DISCONNECTED FL+</p>	<p>During the cycle the connector of channel 2 instrument 1 has detached.</p>	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
109 CHA.3 SET 1 DISCONNECTED FL+	During the cycle the connector of channel 3 instrument 1 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X
110 CHA.4 SET 1 DISCONNECTED FL+	During the cycle the connector of channel 4 instrument 1 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X
111 CHA.5 SET 1 DISCONNECTED FL+	During the cycle the connector of channel 5 instrument 1 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X
112 CHA.6 SET 1 DISCONNECTED FL+	During the cycle the connector of channel 6 instrument 1 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p>113 CHA.7 SET 1 DISCONNECTED FL+</p>	<p>During the cycle the connector of channel 7 instrument 1 has detached.</p>	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p>114 CHA.1 SET 2 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 1 instrument 2.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 1 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 1 connection pipe (brown). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p>115 CHA.2 SET 2 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 2 instrument 2.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 2 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 2 connection pipe (green). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">116 CHA.3 SET 2 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 3 instrument 2.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 3 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 3 connection pipe (blue). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">117 CHA.4 SET 2 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 4 instrument 2.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 4 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 4 connection pipe (yellow). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">118 CHA.5 SET 2 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 5 instrument 2.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 5 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 5 connection pipe (grey). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">119 CHA.6 SET 2 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 6 instrument 2 when pump is continuous working.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 6 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 6 connection pipe (red). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">120 CHA.7 SET 2 OBSTRUCTED FL-</p>	<p>During the cycle in the flow control stage, the system has detected an obstruction in channel 7 instrument 2 when pump is continuous working.</p>	<ul style="list-style-type: none"> - flow sensor malfunction. - the channel obstruction (organic material or of another nature). - channel pipe 7 bent/badly positioned. - error in the instrument connector connection. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the flow sensor. - check both the endoscopic channel (with pipe-cleaner) and channel 7 connection pipe (black). - check that the pipe is correctly positioned in the drum (avoid necking or 90° bends). - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">121 CHA.1 SET 2 DISCONNECTED FL+</p>	<p>During the cycle the connector of channel 1 instrument 2 has detached.</p>	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">122 CHA.2 SET 2 DISCONNECTED FL+</p>	<p>During the cycle the connector of channel 2 instrument 2 has detached.</p>	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
123 CHA.3 SET 2 DISCONNECTED FL+	During the cycle the connector of channel 3 instrument 2 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X
124 CHA.4 SET 2 DISCONNECTED FL+	During the cycle the connector of channel 4 instrument 2 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X
125 CHA.5 SET 2 DISCONNECTED FL+	During the cycle the connector of channel 5 instrument 2 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X
126 CHA.6 SET 2 DISCONNECTED FL+	During the cycle the connector of channel 6 instrument 2 has detached.	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	X in part	X in part	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
<p style="text-align: center;">127 CHA.7 SET 2 DISCONNECTED FL+</p>	<p>During the cycle the connector of channel 7 instrument 2 has detached.</p>	<ul style="list-style-type: none"> - disconnected channel connector. - breakage in connector pipe. - error in the connection of the connector to the instrument. - selected endoscope different from that connected. 	<ul style="list-style-type: none"> - check and/or replace the connector. - check and/or replace the connector pipe. - check that the connector is connected to the correct endoscopic channel. - check that the correct endoscope has been selected on the cycle start monitor. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>
<p style="text-align: center;">128 CHA.3 SET 1 OBSTRUCTED PR.OFF</p>	<p>During the cycle an obstruction was found in the endoscopic channel connected to channel 3 instrument 1 with ON-OFF function</p>	<ul style="list-style-type: none"> - organic material or of a different nature obstructing the endoscopic channel of the instrument or the connection pipe to the instrument. - channel 3 instrument 1 pressure sensor faulty. - incorrect connector connection or bent connector pipe. - endoscopic instrument maintenance carried out badly (reduction in channel diameter different from the original). - set parameters incorrect. - endoscope code insertion different from that attributed. 	<ul style="list-style-type: none"> - verify openness of the channel and the connection pipe. - check and/or replace the pressure sensor. - check that the pipe is correctly positioned inside the drum and the connector. - check that the set maximum pressure parameters are correct. - check that the code recalled for the instruments matches the inserted instrument. - as a final analysis have the instrument checked by an authorised technical assistance. 	<p>X in part</p>	<p>X in part</p>	<p>X</p>

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
129 CHA.4 SET 1 OBSTRUCTED PR.OFF	During the cycle an obstruction was found in the endoscopic channel connected to channel 4 instrument 1 with ON-OFF function	<ul style="list-style-type: none"> - organic material or of a different nature obstructing the endoscopic channel of the instrument or the connection pipe to the instrument. - channel 4 instrument 1 pressure sensor faulty. - incorrect connector connection or bent connector pipe. - endoscopic instrument maintenance carried out badly (reduction in channel diameter different from the original). - set parameters incorrect. - endoscope code insertion different from that attributed. 	<ul style="list-style-type: none"> - verify openness of the channel and the connection pipe. - check and/or replace the pressure sensor. - check that the pipe is correctly positioned inside the drum and the connector. - check that the set maximum pressure parameters are correct. - check that the code recalled for the instruments matches the inserted instrument. - as a final analysis have the instrument checked by an authorised technical assistance. 	X in part	X in part	X
130 CHA.3 SET 2 OBSTRUCTED PR.OFF	During the cycle an obstruction was found in the endoscopic channel connected to channel 3 instrument 2 with ON-OFF function	<ul style="list-style-type: none"> - organic material or of a different nature obstructing the endoscopic channel of the instrument or the connection pipe to the instrument. - channel 3 instrument 2 pressure sensor faulty. - incorrect connector connection or bent connector pipe. - endoscopic instrument maintenance carried out badly (reduction in channel diameter different from the original). - set parameters incorrect. - endoscope code insertion different from that attributed. 	<ul style="list-style-type: none"> - verify openness of the channel and the connection pipe. - check and/or replace the pressure sensor. - check that the pipe is correctly positioned inside the drum and the connector. - check that the set maximum pressure parameters are correct. - check that the code recalled for the instruments matches the inserted instrument. - as a final analysis have the instrument checked by an authorised technical assistance. 	X in part	X in part	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
131 CHA.4 SET 2 OBSTRUCTED PR.OFF	During the cycle an obstruction was found in the endoscopic channel connected to channel 4 instrument 2 with ON-OFF function	<ul style="list-style-type: none"> - organic material or of a different nature obstructing the endoscopic channel of the instrument or the connection pipe to the instrument. - channel 4 instrument 2 pressure sensor faulty. - incorrect connector connection or bent connector pipe. - endoscopic instrument maintenance carried out badly (reduction in channel diameter different from the original). - set parameters incorrect. - endoscope code insertion different from that attributed. 	<ul style="list-style-type: none"> - verify openness of the channel and the connection pipe. - check and/or replace the pressure sensor. - check that the pipe is correctly positioned inside the drum and the connector. - check that the set maximum pressure parameters are correct. - check that the code recalled for the instruments matches the inserted instrument. - as a final analysis have the instrument checked by an authorised technical assistance. 	X in part	X in part	X
132 CHEM.1 DISPARITY	During chemical 1 extraction phase the difference between the two control and recording flow meters has exceeded the maximum set value.	<ul style="list-style-type: none"> - flow meters faulty. - no chemical in canister. 	<ul style="list-style-type: none"> - clean and/or replace both flow meters then repeat the calibration procedure. - check that the quantity of chemical in the canister is sufficient (min 150 ml). 			X
133 CHEM.2 DISPARITY	During chemical 2 extraction phase the difference between the two control and recording flow meters has exceeded the maximum set value.	<ul style="list-style-type: none"> - flow meters faulty. - no chemical in canister. 	<ul style="list-style-type: none"> - clean and/or replace both flow meters then repeat the calibration procedure. - check that the quantity of chemical in the canister is sufficient (min 150 ml). 			X
134 CHEM.3 DISPARITY	During chemical 3 extraction phase the difference between the two control and recording flow meters has exceeded the maximum set value.	<ul style="list-style-type: none"> - flow meters faulty. - no chemical in canister. 	<ul style="list-style-type: none"> - clean and/or replace both flow meters then repeat the calibration procedure. - check that the quantity of chemical in the canister is sufficient (min 150 ml). 			X
135 CHEM.4 DISPARITY	During chemical 4 extraction phase the difference between the two control and recording flow meters has exceeded the maximum set value.	<ul style="list-style-type: none"> - flow meters faulty. - no chemical in canister. 	<ul style="list-style-type: none"> - clean and/or replace both flow meters then repeat the calibration procedure. - check that the quantity of chemical in the canister is sufficient (min 150 ml). 			X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
136 WATER 1 DISPARITY	During water loading phase the difference between the two control and recording flow meters has exceeded the maximum set value.	<ul style="list-style-type: none"> - flow meters faulty. - high flow meter temperature 	<ul style="list-style-type: none"> - clean and/or replace both flow meters then repeat the calibration procedure. - wait for the flow meters to cool - in the case of cycle after thermo disinfection at 90°C. 			X
137 CONDUCTIVITY CHECK FAILED	During the last rinse phase, in machines with water quality control, the maximum number of automatic rinses has been exceeded.	<ul style="list-style-type: none"> - water quality lower than that set for the final rinse. - water quality control conduct meter faulty. 	<ul style="list-style-type: none"> - check that the water entering the machine is of the quality desired. - check the settings of the parameters linked to machine quality. - check and/or replace the conduct meter. 			X
138 WATER 2 DISPARITY	During water loading phase the difference between the two control and recording flow meters has exceeded the maximum set value.	<ul style="list-style-type: none"> - flow meters faulty. - high flow meter temperature 	<ul style="list-style-type: none"> - clean and/or replace both flow meters then repeat the calibration procedure. - wait for the flow meters to cool - in the case of cycle after thermo disinfection at 90°C. 			X
139 TANK LEVEL FAILURES	The capacitive sensor doesn't read the water level during the water loading phase when it is reached the first 8 liters (heating element protection).	<ul style="list-style-type: none"> - capacitive sensor of water level failure. 	<ul style="list-style-type: none"> - calibrate the capacitive sensor of washing chamber level. - check and/or replace the capacitive sensor of washing chamber level. 		X	X
140 TEMP.CHECK SENS.CHAMBER FAILURE	Incorrect temperature reading in the washing chamber	<ul style="list-style-type: none"> - washing chamber PT1000 faulty. - PLC reader card faulty. 	<ul style="list-style-type: none"> - check and/or replace the tank temperature probe. - check and/or replace the temperature probe reader card. 			X
141 CHAMBER TEMP. INCONGRUITY	Failure on the PT1000 probe reading - the difference between the two PT1000 probe is upper than 2°C	<ul style="list-style-type: none"> - PT1000 chamber probe failure. - PLC reading board failure. 	<ul style="list-style-type: none"> - check and/or replace the tank temperature probe. - check and/or replace the temperature probe reader card. 			X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
142 HEATING WITH PUMP OFF ?	Error on the phase programming.		- check the phases programming and correct the washing pump parameter (set ON).			X
144 PLC REC. FAILURES	It is impossible to record the parameters of the cycle	- Recording PLC failure.	- replace the recording PLC.			X
145 CHANGE TANK CHEMICAL 1	Only with RFID active - disinfectant tank empty		- replace the disinfectant tank	X	X	X
146 CHANGE TANK CHEMICAL 2	Only with RFID active - detergent tank empty		- replace the detergent tank	X	X	X
147 CHANGE TANK CHEMICAL 3	Only with RFID active - chemical 3 tank empty.		- replace the chemical 3 tank	X	X	X
148 CHANGE TANK CHEMICAL 4	Only with RFID active - chemical 4 tank empty.		- replace the chemical 4 tank	X	X	X
149 HAKKO TERMINAL NOT ACTIVE	It is visualized during the PLC programming or with the HAKKO terminal failure	- PLC programming failure. - Terminal failure. - Entry in visualization mode after insertion new HAKKO parameters.	- switch on and switch off the machine. - program again the terminal. - replace the terminal - reset the alarm	X	X	X
150 LOW PRESS. COLLECTOR SET 1	During the cycle the pressure of collector set 1 is below the set value.	- electronic pressure sensor failure. - mechanical pressure sensor failure. - washing pump failure - PLC card faulty.	- check and/or replace the pressure sensor. - check and/or replace the pressure sensor. - check and/or replace the washing pump. - check and/or replace sensor control card/s.	X in part	X in part	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
151 LOW PRESS. COLLECTOR SET 2	During the cycle the pressure of collector set 2 is below the set value.	<ul style="list-style-type: none"> - electronic pressure sensor failure. - mechanical pressure sensor failure. - washing pump failure - PLC card faulty. 	<ul style="list-style-type: none"> - check and/or replace the pressure sensor. - check and/or replace the pressure sensor. - check and/or replace the washing pump. - check and/or replace sensor control card/s. 	X in part	X in part	X
152 LOW PRESS. COLLECTOR SET 3	During the cycle the pressure of collector set 3 is below the set value.	<ul style="list-style-type: none"> - electronic pressure sensor failure. - mechanical pressure sensor failure. - washing pump failure - PLC card faulty. 	<ul style="list-style-type: none"> - check and/or replace the pressure sensor. - check and/or replace the pressure sensor. - check and/or replace the washing pump. - check and/or replace sensor control card/s. 	X in part	X in part	X
153 HIGH PRESS. COLLECTOR SET 1	During the cycle the pressure of collector set 1 is over the set value.	<ul style="list-style-type: none"> - electronic pressure sensor failure. - PLC card faulty. 	<ul style="list-style-type: none"> - check and/or replace the pressure sensor. - check and/or replace sensor control card/s. 	X in part	X in part	X
154 HIGH PRESS. COLLECTOR SET 2	During the cycle the pressure of collector set 1 is below the set value.	<ul style="list-style-type: none"> - electronic pressure sensor failure. - PLC card faulty. 	<ul style="list-style-type: none"> - check and/or replace the pressure sensor. - check and/or replace sensor control card/s. 	X in part	X in part	X
155 HIGH PRESS. COLLECTOR SET 3	During the cycle the pressure of collector set 1 is below the set value.	<ul style="list-style-type: none"> - electronic pressure sensor failure. - PLC card faulty. 	<ul style="list-style-type: none"> - check and/or replace the pressure sensor. - check and/or replace sensor control card/s. 	X in part	X in part	X
156 FAILURE COLLECTOR PRESS.SWITCH SET 1	During the startup machine, the system has detected a failure on the reading of electronic pressure switch.	<ul style="list-style-type: none"> - electronic pressure sensor failure. - PLC card faulty. 	<ul style="list-style-type: none"> - check and/or replace the pressure sensor. - check and/or replace sensor control card/s. 	X in part	X in part	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
157 FAILURE COLLECTOR PRESS.SWITCH SET 2	During the startup machine, the system has detected a failure on the reading of electronic pressure switch.	- electronic pressure sensor failure. - PLC card faulty.	- check and/or replace the pressure sensor. - check and/or replace sensor control card/s.	X in part	X in part	X
158 FAILURE COLLECTOR PRESS.SWITCH SET 3	During the startup machine, the system has detected a failure on the reading of electronic pressure switch.	- electronic pressure sensor failure. - PLC card faulty.	- check and/or replace the pressure sensor. - check and/or replace sensor control card/s.	X in part	X in part	X
160 INCONGRUITY LEAK TEST CAP 1	When the control is active, the instrument presence sensor is active with the endoscope in the position 1 or is not active without instrument on position 1.	- sensor failure - wrong connection of operator - wrong insertion of operator on touch-screen	- replace or calibrate the sensor - check the correct connection of endoscope	X in part	X	X
161 INCONGRUITY LEAK TEST CAP 2	When the control is activated, the instrument presence sensor is active with the endoscope in the position 2.	- sensor failure - wrong connection of operator - wrong insertion of operator on touch-screen	- replace or calibrate the sensor - check the correct connection of endoscope	X in part	X	X
162 FAILURE ON LT PUMP	During the leak test the pressure set hasn't been reached.	- air pump for the leak test failure. - leak on the leak test circuit. - endoscope basket reversed. - endoscope connection lack. - oring of the basket locking system broken (on the chamber). - leak test pipe broken.	- leak test pump replacement. - check and restore the leak test circuit. - check the endoscope basket position (connection blocks on the left side of the loading side). - check the oring integrity on the basket connection block inside the chamber. - check the instrument connection pipe integrity (near the metallic clamp).	X in part	X in part	X
163 REFILL ALARM LT SET 2	The system has exceeded the number of refill leak test set by parameter.	- problem on leak test instrument. - leak from idraulic circuit of leak test. - wrong value of parameter.	- check the instrument leak test. - check the o-ring and connection of leak test. - check the pipe connection. - check the value parameter.		X	X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
164 REFILL ALARM LT SET 3	The system has exceeded the number of refill leak test set by parameter.	- problem on leak test instrument. - leak from idraulic circuit of leak test. - wrong value of parameter.	- check the instrument leak test. - check the o-ring and connection of leak test. - check the pipe connection. - check the value parameter.		X	X
165 FAILURE LS LOCKING/ UNLOCKING	Failure limitswitch locking/unlocking 53SQ25/53SQ25A/53SQ26/ 53SQ26A	-Limitswitch not fed or damaged	- Replace the limitswitch - Check the electrical circuit		X	X
166 FAILURE LOCKING/ UNLOCKING	Failure locking/unlocking;	Pressure regulator on pilot valves group is not calibrated.	-Check the pressure regulator of piston - Adjust pressure to 4 bar.		X	X
167 LOADING DOOR LS DISAPRITY	Loading door limitswitch disparity;	- Electrical contact; - Both limitswitches are intercept.	- Check the micro;		X	X
168 UNLOADING DOOR LS DISPARITY	Unloading door limitswitch disparity;	- Electrical contact; - Both limitswitches are intercept.	- Check the micro;		X	X
WARNINGS						
176 PLC LOW BATTERY			- replace buffer battery.			X
177 DISPLAY LOW BATTERY			- replace buffer battery.			X
178 ASKING MAINTENANCE			- call technical assistance for ordinary maintenance.	X	X	X
179 'HEPA' FILTER OBSTRUCTED			- replace the chamber filter.			X

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
180 UNLOD. SIDE TERMINAL FAIL.			- check and/or replace the unload.side terminal in the case of passing machine.			X
181 PROBLEM COMUNICATION WITH RFID	During the cycle start or during the washing cycle the RFID control system has stopped to communicate with the PLC.	- RFID control unit failure. - PLC connection failure. - RFID antenna failure.	- check and/or replace the reading RFID antenna. - check and/or replace the RFID antenna. - check the connection between the RFID control unit and the PLC.			X
182 PROBLEM READ RFID CH.1	During the cycle start the disinfectant RFID antenna has stopped to communicate with the PLC.	- RFID control unit failure. - PLC connection failure. - RFID antenna failure.	- check and/or replace the reading RFID antenna of the disinfectant or try to replace the chemical tank (with the Steelco original RFID). - check and/or replace the RFID antenna. - check the connection between the RFID control unit and the PLC.			X
183 PROBLEM READ RFID CH. 2	During the cycle start the detergent RFID antenna has stopped to communicate with the PLC.	- RFID control unit failure. - PLC connection failure. - RFID antenna failure.	- check and/or replace the reading RFID antenna of the detergent or try to replace the chemical tank (with the Steelco original RFID). - check and/or replace the RFID antenna. - check the connection between the RFID control unit and the PLC.			X
184 PROBLEM READ RFID CH. 3	During the cycle start the chemical 3 RFID antenna has stopped to communicate with the PLC.	- RFID control unit failure. - PLC connection failure. - RFID antenna failure.	- check and/or replace the reading RFID antenna of the chemical 3 or try to replace the chemical tank (with the Steelco original RFID). - check and/or replace the RFID antenna. - check the connection between the RFID control unit and the PLC.			X
N.C.						

HAKKO ALARM MESSAGE	ALARM DESCRIPTION	POSSIBLE CAUSES	SOLUTION	OPER.	SUPERV.	MANUF.
185 PROBLEM READ RFID CH. 4 N.C.	During the cycle start the chemical 4 RFID antenna has stopped to communicate with the PLC.	<ul style="list-style-type: none"> - RFID control unit failure. - PLC connection failure. - RFID antenna failure. 	<ul style="list-style-type: none"> - check and/or replace the reading RFID antenna of the chemical 4 or try to replace the chemical tank (with the Steelco original RFID). - check and/or replace the RFID antenna. - check the connection between the RFID control unit and the PLC. 			X
187 FAILURE ON WATER FILLING (TIME)	During one of the water loading phases the time limit to reach the water quantity (12 lt) has been exceeded	<ul style="list-style-type: none"> - water inlet tap closed. - water filter obstructed. - no pressure on the hydraulic circuit (1 bar minimum). - pneumatic valve or pilot valve failure. - PLC command board failure (see electrical scheme). 	<ul style="list-style-type: none"> - check if the water inlet tap is open. - check if there is pressure on the external hydraulic circuit. - check the functioning of the water loading pump inside the machine under the break tank. - check the water loading pneumatic valve and the pilot valve. - check the water filter. - check the PLC control board. 			X
188 LOW LEVEL/ NOT OK CHEM. 1	Indicates that 3 cycles remain to the end of the disinfectant chemical		- replace the disinfectant chemical tank (chemical 1)	X		
189 LOW LEVEL/ NOT OK CHEM. 2	Indicates that 3 cycles remain to the end of the detergent chemical		- replace the detergent chemical tank (chemical 2)	X		
190 LOW LEVEL/ NOT OK CHEM. 3	Indicates that 3 cycles remain to the end of the chemical 3		- replace the chemical tank (chemical 3)	X		
191 LOW LEVEL/ NOT OK CHEM. 4	Indicates that 3 cycles remain to the end of the chemical 4		- replace the chemical tank (chemical 4)	X		

8.2 Machine alarms concerning mechanical parts

ALARM! – 02
FAILURE ON TRANS. OF LT TANK



ALARM! – 162
FAILURE ON LT PUMP



ALARM ! –09
HEAT.EL.THERM.SAFETY



ALARM ! –10
AIR COMPRESSED LACK



ALARM ! – 26 TEMP.SENS.FAILURE (CONTROL).
ALARM! – 27 TEMP.SENS.FAILURE (RECORD)

ALARM! - 32 WASH.ARMS FLOW SENSOR BREAKDOWN
ALARM! - 50 W.ARMS FLOW TR. OUT OF LIM.-
ALARM! - 51 W.ARMS FLOW TR. OUT OF LIM.+



ALARM! - 86 FAILURE DRAINING LEAK TEST



ALARM! - 88 UNCOUPLING/COUPLING ES FAILURES
ALARM! - 89 UNCOUPLING/COUPLING FAILURES

ALARM! - 132 CHEM.1 DISPARITY
ALARM! - 133 CHEM.2 DISPARITY
ALARM! - 134 CHEM.3 DISPARITY



ALARM! - 136 WATER DISPARITY



ALARM! - 137 CONDUCTIVITY CHECK FAILED



ALARM! - 139 TANK LEVEL FAILURES



ALARM! - 140 TANK PT1000 OUT OF BOUND

ALARM! - 141 CHAMBER TEMP. INCONGRUITY



ALARM! - 143 FAILURE WATER DRAIN BREAK-TANK



ALARM! - 187 FAILURE ON WATER FILLING (TIME)



ALARM! - 182 PROBLEM READ RFID CH. 1
ALARM! - 183 PROBLEM READ RFID CH. 2



9. DEMOLITION AND DISPOSAL

For a correct disposal of the machine follow the next indications:

- Do not dispose these machines as solid waste but make a separated collection.
- The reuse or the correct recycling of the Electrical and Electro-mechanics equipment (AEE) is useful for the environment preservation and the human health.
- According to the European Directive WEEE 2012/19/EC specific collection centre are available; it is possible to deliver this electrical equipment to this centre, and it is possible to return the equipment to the Distributor when a new one machine is bought.
- The public administration and the AEE manufacturer are committed to favour the reuse and the reclamation process of the RAEE (Electrical and electro-mechanics equipment waste) through the organization of the refuse collection activity, and through the use of opportune planning solution.
- **The law penalizes with opportune sanctions who illegally dispose the RAEE.**

ATTENTION: PLEASE CONSULT ANNEX C IN THIS MANUAL.

10. QUALIFICATION

10.1 Installation qualification (IQ)

Provided on request

10.2 Operative qualification (OQ)

Provided on request

10.3 Performance qualification (PQ)

As specified in Attachment A.3 of Standard ISO 15883:4 2009, from the moment in which the EW2 is installed upon the final site of use, the responsibility for guaranteeing that the equipment has been correctly installed and is fully operational falls on the purchaser/user.

This includes:

- Installation check and tests;
- Operational tests;
- Operational tests;
- Periodic tests;
- The use of process chemicals recommended by the manufacturer;
- The use of the EW2 in accordance with the manufacturer's instructions (limiting the re-conditionable devices in the EW2 to those specified by the manufacturer).



Certificate N° 632


Certificate of Participation

We certify that Mr. **XXXXXXXX** of the company **YYYYYYYY** has attended the training course for

EW2

Riese Pio X, 03/09/2014

Person in charge for training



Steelco S.p.A.

Via Balegante, 27 - 31039 - Riese Pio X (TV) - Italy
Tel. +39 0423.7561 Fax +39 0423.75528
Website: www.steelcospa.com



ANNEX A – TRAINING CERTIFICATE

ANNEX B – WATER SAMPLING DURING THE FINAL RINSING PHASE

Annex B provides both the instructions and the methods to take the microbiological sample by using the Steelco - Q water BSK (Professional Sampling Kit for Bacterial Check) as well as the relative key for the decoding and the reading of the results obtained.

In any case, please refer to current Legislation, National Guidelines and/or internal protocols.



Q water BSK
Professional Sampling Kit for Bacterial Check
(code 99911268)

Endoscopes can be re-processed in the cycle dedicated to biological sampling. The only difference compared to a normal disinfection cycle refers to the interruptions during the final rinsing phase.

Water samples are taken during the final rinsing stage following the disinfection stage. The final rinsing cycle consists of two stages. The sample must be taken during the final rinsing stage.

It is possible to take a water sample from the washing chamber by carrying out the following procedure:

- Set the parameter **CYCLE STOP ENABL.** to **YES**.
- Note: at the end of the water sample-taking cycle, restore the value to **NO**.

19 ALARM 11 CONFIG, MANUT, DATA 05/09/14 09 59

DATA CONFIG. TIMES CONFIG. 1 TIMES CONFIG. 2 DATA VARIOUS1 DATA VARIOUS2 DATA CYCLE

User: _____

Drain v. out of cycle: CLOSED

Contem. door open.: NO

Cycle stop enabl.: YES

Autom. doors lock.: NO

Automatic start cycle: NO

Must unload basket: NO

Password: *****

SAVE DATA PRINT DATA

- Check if it is activated the **STOP END = YES** on the selected program for the water sampling and start the program.

26 PROGRAMS MANAG. 023 05/09/14 14 36

Description: CAMPIONATURA

Door: OK UNLOAD. NOT OK LOAD.

Program type: ENDOSCOPES

RISCIACQUO STOP END=YES

Water WATER I 12.0 Pumps YES

Chem. % 00.00

Chem. % 00.00

Temp. oC 00.0 Maint. oC NO

Mix. T. s 000 T. Phase s 0060

CONFIRM CHANGE PHASE CANCEL

- When the machine displays the message shown on the following picture, open the door of the washing chamber and pick up the water sampling from the sump by using the Steelco - Q water BSK (Professional Sampling Kit for Bacterial Check). Be sure to perform the sampling in an aseptic way.



- Close the door and press **CONFIRM** button to continue the washing program.




ATTENTION



- Adequate measures must be implemented to avoid contamination of the sample during the water sample-taking stage. It is recommended to use the Steelco - Q water BSK (Professional Sampling Kit for Bacterial Check).
- The water samples for the microbiological test must be taken only during the final rinsing stage. Samples must not be taken during other stages of the process.
- Correct, periodic maintenance of the equipment prevents the risk of contamination. Follow the manufacturer's instructions as regards the replacement of bacteriological filters as well as any pre-filters.
- It is important to schedule appropriate thermal disinfection and chemical self-disinfection cycles in accordance with the manufacturer's instructions and internal protocols.

SAMPLE TAKING

In order to take water samples during the final rinsing stage use the Steelco - Q water BSK (Professional Sampling Kit for Bacterial Check - code 99911268) and follow the instructions below:



1. Insert the 80 cm PVC pipe **(5)** into the sump of the washing chamber;
2. Suck the water from the washing chamber using the 60 ml syringe **(7)**;
3. Push the plunger of the syringe **(7)** to drain the water into the collection bag for biological fluids **(1)**;
4. Disconnect the collection bag for biological fluids **(3)** from the Kit and hand it to the microbiology laboratory quickly.
5. Compile the fields on the bag with the appropriate information:
 - Model and serial number of the machine;
 - Name and ID of the operator (that takes the water sampling);
 - Date.

REPROCESSING CYCLE EFFECTIVENESS TEST CARRIED OUT IN THE SUCTION CHANNEL

It is also possible to test the effectiveness of the entire re-conditioning cycle of the endoscope suction channel in the following way:

- Put 20 ml of sterile water in the suction channel (by using a sterile syringe) collecting it from a sterile container situated at the distal part of the tool;
- Clean the same channel using a special cleaning brush by cutting it directly inside the sterile water container.
WARNING: the cleaning brush **MUST** be sterile.
- Remember to close all air and water channel valves before carrying out this procedure.
- Every single container must be accompanied not only by the request form but also by the endoscope type and serial number, the type of channel on which the brushing has been carried out, the name of the operator performing the sampling procedure as well as the report relating to the re-conditioning cycle carried out.

INTERPRETATION OF THE RESULTS

As for the interpretation of the results, please refer to the following table and criteria.

NUMBER OF COLONIES/100 ml	INTERPRETATION	ACTION
0	SATISFACTORY	-----
1-9 (carried out on a regular basis)	ACCEPTABLE	Count under reasonable control
10-100	UNSATISFACTORY	Investigate by implementing opportune thermal disinfection and chemical self-disinfection cycles.
>100	UNACCEPTABLE	Stop the endoscope washer. Start two thermal disinfection and two chemical self-disinfection cycles. Repeat the test of control and take necessary actions as regards decontamination.

After interpreting the results, take opportune measures in accordance with the 2 attached tables.

NUMBER OF COLONIES (NOT PSEUDOMONAS)		
TVC LEVEL		ACTION
SATISFACTORY	< 1cfu/100 ml	<ul style="list-style-type: none"> No action required.
ACCEPTABLE	1 – 9 cfu/100 ml	<ul style="list-style-type: none"> The personnel must have carried out a self-disinfection cycle in the morning in accordance with the instructions provided by the endoscope washer manufacturer as well as internal protocols. The carrying out of a daily self-disinfection cycle.
UNSATISFACTORY	10 – 50 cfu/ ml	<ul style="list-style-type: none"> The personnel must carry out two thermal disinfection cycles (one after another - 80°C for 10 min) and a special chemical disinfection cycles (35°C for 10 min double chemical)
	51 – 100 cfu/ ml	<ul style="list-style-type: none"> The personnel must carry out a special chemical disinfection cycle as well as a thermal disinfection cycle (one after another 80°C for 10 min one after another 55°C for 10 min double chemical). Take another water sample.



UNACCEPTABLE		<p>STOP USING THE ENDOSCOPE WASHER.</p> <p>The person in charge of the endoscope washer must:</p>
	> 100 cfu/100 ml	<ul style="list-style-type: none"> • Carry out a special chemical disinfection cycle (55° double chemical for 10 min). • Carry out 2 thermal disinfections cycles (one after another to 90°C for 10 min). • Take another water sample. <p>The personnel must:</p> <ul style="list-style-type: none"> • Do not use the endoscope washer until it has been confirmed that the water sample has a contamination value of < 100 cfu/100 ml. • Continue to carry out a self-disinfection cycle on a daily basis. • Carry out a special thermal disinfection cycle on a daily basis. <p>Note: Advice will be obtained from the Lead Doctor for Infection Prevention and Control if there are recurring unacceptable TVC levels.</p>

NUMBER OF COLONIES (PSEUDOMONAS)		
TVC LEVEL		ACTION
SATISFACTORY	< 1cfu/100 ml	<ul style="list-style-type: none"> No action required.
ACCEPTABLE	1 – 9 cfu/100 ml	<ul style="list-style-type: none"> The personnel must have carried out a self-disinfection cycle in the morning in accordance with the instructions provided by the endoscope washer manufacturer as well as internal protocols. The carrying out of a daily self-disinfection cycle. Do not re-process endoscopes until it has been confirmed that the water sample has a contamination value of < 6 cfu/100 ml.
UNSATISFACTORY	10 – 50 cfu/ ml	<ul style="list-style-type: none"> The personnel must carry out two thermal disinfection cycles (one after another - 80°C for 10 min) and a special chemical disinfection cycles (35°C for 10 min double chemical) Do not re-process endoscopes until it has been confirmed that the water sample has a contamination value of < 6 cfu/100 ml. Do not re-process cystoscopes or bronchoscopes until it has been confirmed that water sample has a contamination value of < 10 cfu/100 ml.
	51 – 100 cfu/ ml	<ul style="list-style-type: none"> The personnel must carry out two special chemical disinfection cycles (55°C for 10 min double chemical) and two thermal disinfection cycle (one after another 80°C for 10 min). Take another water sample. Do not re-process endoscopes until it has been confirmed that the water sample has a contamination value of < 6 cfu/100 ml. Do not re-process cystoscopes or bronchoscopes until it has been confirmed that water sample has a contamination value of < 10 cfu/100 ml. <p>Note: Advice will be obtained from the Lead Doctor for Infection Prevention and Control if there are recurring unacceptable TVC levels.</p>

UNACCEPTABLE		<p>STOP USING THE ENDOSCOPE WASHER.</p> <p>The person in charge of the endoscope washer must:</p>
	<p>> 100 cfu/100 ml</p>	<ul style="list-style-type: none"> • Change all internal water filters. Carry out 2 special chemical disinfection cycles (55° double chemical for 10 min). • Carry out 2 thermal disinfection cycles (one after another to 90°C for 10 min). • Take another water sample. <p>The personnel must:</p> <ul style="list-style-type: none"> • Do not use the endoscope washer until it has been confirmed that the water sample has a contamination value of < 100 cfu/100 ml. • Continue to carry out the thermal disinfection cycle on a daily basis. • Carry out a chemical self-disinfection cycle on a daily basis. • Do not re-process endoscopes until it has been confirmed that the water sample has a contamination value of < 6 cfu/100 ml. Do not re-process cystoscopes or bronchoscopes until it has been confirmed that water sample has a contamination value of < 10 cfu/100 ml. <p>Note: Advice will be obtained from the Lead Doctor for Infection Prevention and Control if there are recurring unacceptable TVC levels.</p>

	ATTENTION
	THE PREVIOUSLY DESCRIBED CYCLES ARE STORED INTO THE MEMORY OF THE MACHINE.

ANNEX C – TEST PROCEDURE WHEN INSTALLING

	ATTENTION
	<p>During the first installation of the equipment upon the site of use, with the aim of eliminating the risk of contamination of the endoscope washer, following maintenance operations that influence critical parts of the endoscope washer (wash pumps, etc.) or after transferring the endoscope washer to another operational centre, the following procedure MUST be carried out:</p> <ul style="list-style-type: none"> • 3 empty cycles carried out using only detergent; • 2 complete cycles with detergent + disinfectant; • 1 thermal disinfection cycle carried out at 80°C.
	ATTENTION
	<p>USE ONLY CHEMICAL PRODUCTS THAT HAVE BEEN TESTED AND APPROVED BY THE MANUFACTURER AND IN USE WITH THIS SYSTEM.</p>

ANNEX D – BARCODE READER SETTING

In case of barcode replacement or memory loss, set the barcode reader by reading the following codes.



**RS-232 Serial Interface
Quick Set Command**



PROGRAM

Host Interface Selection




for **Gold/Jade/Diamond** series



1




END (Exit)



PROGRAM


Record Suffix



3



END (Exit)




PROGRAM

Baud Rate (BPS)



1



END (Exit)

ANNEX E: SPECIFICATION OF ROUTER NAT PROGRAMMING

 MQ205-02 Rev.00	<p style="text-align: center;">ROUTER NAT INSTRUCTION</p> <p style="text-align: center;">ROUTER WIZARD TOOL USER</p> <p style="text-align: center;">GUIDE</p>	Ref.: n.d.	
		Code: n.d.	
		Rev.: 04	Page: 1 of 9


ROUTER WIZARD TOOL USER GUIDE FOR NAT CONFIGURATION

Scope: Configuration guide of Router NAT.
Actual software version 1.1.0




Steelco Group Member



 MQ205-02 Rev.00	ROUTER NAT INSTRUCTION ROUTER WIZARD TOOL USER GUIDE		Ref.: n.d.
			Code: n.d.
	Rev.: 04	Page: 2 of 9	


HISTORICAL OF DOCUMENT REVISIONS

Rev. No.	Date	Comments
00	27/11/2019	Document creation.
01	17/12/2019	Updating 'Software revisions paragraph
02	20/12/2019	Updating 2.2, 2.3 paragraphs
03	03/02/2020	Enabling single machine configuration
04	30/03/2020	Screens updating and comments simplification.
05		
06		
07		
08		
09		
10		
11		
12		
13		
14		
15		
16		
17		
18		
19		
20		

 MQ205-02 Rev.00	ROUTER NAT INSTRUCTION ROUTER WIZARD TOOL USER GUIDE	Ref.: n.d.	
		Code: n.d.	
	Rev.: 04	Page: 3 of 9	

INDEX

1.	SOFTWARE REVISIONS LIST	4
1.1	VERSION 1.0.....	4
1.2	VERSION 1.0.1.....	4
1.3	VERSION 1.0.2.....	4
1.4	VERSION 1.0.3.....	4
1.4	VERSION 1.1.0 (ACTUAL VERSION).....	4
2.	SOFTWARE GUIDE USE	5
2.1	NEW MANUFACTURING ROUTER CONFIGURATION	5
2.2	ROUTER RECONFIGURATION STILL PROGRAMMED	8
2.3	ROUTER HARD RESET PROCEDURE.....	9

 MQ205-02 Rev.00	ROUTER NAT INSTRUCTION ROUTER WIZARD TOOL USER GUIDE	Ref.: n.d.	
		Code: n.d.	
	Rev.: 04	Page: 4 of 9	

1. SOFTWARE REVISIONS LIST

1.1 Version 1.0

First version released with the following functionalities:

- IP addresses configuration, network and gateway mask for EW1 S rack machines;
- Enabling of SteelcoData ARES traceability system;
- Configuration of credentials reserved for router configuration access;
- Fabrication router reset configuration.

1.2 Version 1.0.1

The following updating introduces some modifications on the router configuration. The programming procedure remains the same.

- DHCP-server function disabled by all router doors sides;
- Display on the current Router OS version present on the device;
- OS router updating during the configuration in the case in which a previous version than 6.46 has been detected;
- Tools menu added with functionalities relative to device firmware.

1.3 Version 1.0.2

In this updating:

- Paragraph added on router reprogramming already configurated;
- Hard reset procedure added.

1.4 Version 1.0.3


In this updating:

- Added dialogue windows useful to configurate correctly the PC during the programming in case of wrong configurations.
- Possibility of programming single machines enabled.

1.4 Version 1.1.0 (actual version)

In this updating:

- Added ATS among the configurable machines;
- Automatic identification preloaded configuration on the router for all the machines;
- ROSI router management;
- Guided procedure for router programming in test area;
- Steelco signature management;
- Protected with password sensitive operations;
- Added *Empty* key to empty windows still filled;
- Guide coherently updated to 1.1.0 version.

 MQ205-02 Rev.00	ROUTER NAT INSTRUCTION ROUTER WIZARD TOOL USER GUIDE	Ref.: n.d.	
		Code: n.d.	
			Rev.: 04


2. SOFTWARE GUIDE USE

2.1 New manufacturing router configuration

To set the router, it is necessary to connect it by using an Ethernet cable with one numbered door of the router (please do not connect it to the door signed by Internet).



When the led, which is situated on the upper side of the router corresponding to the door to which the Ethernet cable has been connected, is on, it is possible to start up the software for the configuration.

Set the IP address of your PC to **192.168.88.222** (network mask 255.255.255.0).




From the dropdown menu, which is situated in the middle, it is possible to select the machine type for which you should set the router. From the file → settings submenu, set the IP address and the access credentials as indicated in the picture below on the right and here below:

IP: **192.168.88.1**
 Default username: **admin**
 Default password: **(empty)**



Once the fields have been set, press *Test connection* key, in order to verify the connection of the router. If the connection is correctly set, a message will appear with the following text: *Test successful*.


When the configurations have been correctly set, press the *Start configuration* key, in order to start the router programming. The PC will try to connect to the router in this way and a similar window will appear until the connection has not been established yet.

 MQ205-02 Rev.00	ROUTER NAT INSTRUCTION ROUTER WIZARD TOOL USER GUIDE	Ref.: n.d.	
		Code: n.d.	
	Rev.: 04	Page: 6 of 9	



When the connection has been established, in the following window it is possible to set all data related to the machine configuration.

In case a single type machine is selected (ED200, EW2, EW1 S) a window will appear, which will allow to add all the data related to the network connection of a single machine.




In this window it is necessary to insert all the data to be associated to the machine: *Machine IP address*, *Gateway IP address* e *Subnet mask*. Just in case the machine must be connected to the ARES traceability system, the relative field must be enabled and the IP address field of the PC, in which the traceability system is installed, must be filled.


The previous data should be provided by IT responsible of the installation, where the machines must be installed.

Hakko IP address and *Subnet mask* fields must not be modified if it is not required. The default values are always filled and they have to be the ones indicated here below:

Hakko IP address: **192.168.1.20**
Subnet mask: **255.255.255.0**

In case a double type machine is selected (EW1 S in rack version) a window will appear, in which it will be possible to insert the data related to the network connection of two machines.

 MQ205-02 Rev.00	ROUTER NAT INSTRUCTION ROUTER WIZARD TOOL USER GUIDE	Ref.: n.d.	
		Code: n.d.	
			Rev.: 04



On the left side it is necessary to insert the data related to the machine on the upper side of the rack (IP address, gateway address and network mask). On the right side, similar data are requested for the machine on the lower side of the rack.

In case the machine must be connected to ARES system, and only in this case, it is necessary to flag the field on the left lower side. In this case it is necessary to add the IP address corresponding to the PC on which the traceability system will be installed.

The previous data must be provided by the IT installation responsible, in which the machines are installed.

Hakko IP address and *Subnet mask* fields are still filled and they must not be modified if not in case on a specific request. The default values are indicated here below:

Upper machine:


Hakko IP address: **192.168.1.20**
Subnet mask: **255.255.255.0**


Lower machine:

Hakko IP address: **192.168.1.21**
Subnet mask: **255.255.255.0**


When all the fields are correctly filled, press the *Set configuration* key. In this way the information will be saved on the router. It will start after this operation and it is necessary to wait for about 60 seconds, until the process ends. During this operation, the waiting window will appear.

Do not switch off or disconnect the router during this operation.

 MQ205-02 Rev.00	ROUTER NAT INSTRUCTION ROUTER WIZARD TOOL USER GUIDE	Ref.: n.d.	
		Code: n.d.	
	Rev.: 04	Page: 8 of 9	



If the operation ends with a positive result, on the display the screen shown below appears with the positive result of the configuration.



Press ok and close the program.

In this way the configuration will be completed, and it is possible to disconnect the router from the pc and connect it to the machine with the usual conventions:

- **The “blue” door marked by the Internet sign should be connected to the wall plug external from the machine and therefore to the customer network;**
- **The “white” doors numerated from 2 to 5 are used to connect the router to the internal network of the machine.**


2.2 Router reconfiguration still programmed


In case you must reset a router, which has been programmed, the procedure to be followed is the one indicated below. First of all, set the IP address of the pc to **192.168.1.222** (network mask 255.255.255.0). therefore, open the *File* → *settings* menu and set the parameters as follows:

IP: **192.168.1.1**
Default username: **steelco or admin**
Default password: **(to be not modified)** *the password field is automatically filled*

Once these fields are filled in, press the *Test connection* key to verify if the connection to the router is possible. In positive case, press the *Save* key and go on with the following step. In case the test result is not positive, please have a look at the following chapter, where is indicated how to perform a hard router reset.

If the connection test has been positively performed, go on with the restore of the fabrication router settings. In order to do that, select the *Tools* → *Factory reset* menu, as it is shown in the following picture.

 MQ205-02 Rev.00	ROUTER NAT INSTRUCTION ROUTER WIZARD TOOL USER GUIDE	Ref.: n.d.	
		Code: n.d.	
		Rev.: 04	Page: 9 of 9



Once this procedure has been completed, change the settings into the *File* → *settings* menu with the

following parameters: IP:**192.168.88.1**

Default username: **admin**

Default password: **(empty)**

Select the machine model and complete the setting as illustrated in the previous paragraph.

2.3 Hard router reset procedure

In case the connection to the router is not possible or for some reasons the setting procedure has not a positive result, it is possible to perform a router hardware reset to restore it to the manufacturing settings.

In order to do that, follow the points indicated below:

- When the router is switched off, press the RES key and keep it pressed on the frontal part of the router (1);
- Supply the router (2) and keep the RES key pressed until the USR led (3) starts to blink and not over (it is important not to keep the key pressed beyond this point for more than 5 seconds).
- Leave the router in standby for 60 seconds.
- At this point the router has come back to its fabrication origins and it is possible to program it by following the procedure indicated on paragraph 2.1.



