\delta Schröder

EC DECLARATION OF CONFORMITY

Date of issue: 03.04.2023

Conformity declaration issued by:

The Legal and Actual Manufacturer	Schroder Medikal Ticaret Anonim Şirketi
Registered address	Başpınar (Organize) O.S.B Mah. O.S.B 5. Bölge 83548 Nolu Cad. No:6/1, Sehitkamil Gaziantep Türkive
Telephone	+90 342 230 1583 / +90 850 202 7750
Fax	+90 342 230 1583
E-mail	info@schroderhealth.de

Brand

Products

	Schröder:
	PDB-1130 PEDIATRIC BED 2 MOTORS, PDB-1120 PEDIATRIC BED 2 MOTORS, PDB-1140 PEDIATRIC BED 3 MOTORS, BC-11 BABY COTE, BC-12 BABY COTE WITH CABINET, MT-04 VISCOELASTIC MATTRESS, MT-01
	FOAM MATTRESS, MT-02 FOAM MATTRESS SOFT SENSE, BSC-311 BEDSIDE CABINET, BSC-341 BEDSIDE CABINET, BSC-351 BEDSIDE CABINET, BSC-351 BEDSIDE CABINET, BSC-352
	3EDSIDE CABINET, BSC-332 BEDSIDE CABINET, BSC-322 BEDSIDE CABINET, OBT-161 OVERBED TABLE, OBT- 162 OVERBED TABLE, OBT-160 OVERBED TABLE, OBT-163 OVERBED TABLE, AC-03 ATTENDANT COUCH,
F F T Y	AC-02 ATTENDANT COUCH, AC-01 ATTENDANT COUCH WITH CASTORS, DT-21 ELITE DELIVERY BED, ES-110 FUNCTIONAL EMERGENCY STRETCHER, ES-110 CYCLONE TRAUMA STRETCHER, ST-120 GENERAL PURPOSE PATIENT STRETCHER, OT-140 OPERATION ROOM TRANSFER STRETCHER, DT-21 ELECTRONIC DELIVERY FABLE 3 MOTORS, DIA-2 ELECTRONIC DIALYSIS CHAIR 2 MOTORS, DIA-4 ELECTRONIC DIALYSIS CHAIR 4 MOTORS, MPC-14 ELECTRONIC BLOOD TRANSFUSION CHAIR 4 MOTORS, ENT-7 ELECTRONIC ENT CHAIR 4
1 (MOTORS, BDC-12 ELECTRONIC BLOOD DRAWING CHAIR 2 MOTORS, ARTE FUNCTIONAL PATIENT IRANSFER CHAIR, SW-14 SPACE SAVER STACKABLE WHEELCHAIR SS, EXC-01 EXAMINATON COUCH WITH DRAWERS AND CABINET, EXC-02 EXAMINATION COUCH WITH 3 DRAWERS, EXC-04 EXAMINATION COUCH, EXC-05 EXAMINATION COUCH, EXC-06 EXAMINATION COUCH FOLDABLE, EXC-07 PEDIATRIC
	EXAMINATION COUCH, WS-122 WAITING ROOM SEAT DOUBLE, WS 123 WAITING ROOM SEAT TRIPLE, WS- 124 WAITING ROOM SEAT TRIPLE, WS-125 WAITING ROOM SEAT DOUBLE, WS-132 WAITING ROOM SEAT
F	DOUBLE, WS-133 WAITING ROOM SEAT TRIPLE, WS-143 WAITING ROOM SEAT TRIPLE, WS-151 WAITING ROOM SEAT SINGLE, WS-152 WAITING ROOM SEAT DOUBLE, WS-153 WAITING ROOM SEAT TRIPLE, WS- 161 WAITING ROOM SEAT SINGLE, WS-162 WAITING ROOM SEAT DOUBLE, WS-163 WAITING ROOM SEAT FRIPLE, WS 171 WAITING ROOM SEAT SINGLE, WS-172 WAITING ROOM SEAT DOUBLE, WS-173 WAITING
F	ROOM SEAT TRIPLE, SWS-101 WAITING ROOM SEAT SINGLE, SWS-102 WAITING ROOM SEAT DOUBLE,
S	SWS-103 WAITING ROOM SEAT TRIPLE, SWS-102-S WAITING ROOM SEAT DOUBLE WITH MIDDLE TABLE, SWS-111 WAITING ROOM SEAT SINGLE, SWS-112 WAITING ROOM SEAT DOUBLE, SWS-113 WAITING ROOM SEAT TRIPLE, ST-06 STOOL, ST-07 STOOL WITH BACKREST, IVP-06 IV STAND MOBILE, IVP-07 IV
1	STAND MOBILE SS, EQ-05 ESCABO DOUBLE, EQ-07 INSTRUMENT AND MEDICINE CABINET, NT-110 NEGATOSCOPE SINGLE SLIM, NT-120 NEGATOSCOPE DOUBLE SLIM, NT-130 NEGATOSCOPE TRIPLE SLIM,
۵	FS-14 FOLDING SCREEN METAL, DC-101 LOCKER SINGLE, DC-102 LOCKER DOUBLE, DC-103 SHOWCASE, DC-104 LOCKER QUADRUPLE, DC-105 STEEL LOCKER DOUBLE, DC-106 STEEL LOCKER SINGLE, DC-107 STEEL LOCKER DOUBLE, DC-108 STEEL LOCKER TRIPLE, DC-109 STEEL LOCKER, QUADRUPLE, DC-110 STEEL
١	LOCKER SEXTETTE, SSV-50-T SINGLE BAY SCRUB SINK WITH FOTOCELL, SSV-50-D, SINGLE BAY SCRUB SINK WITH KNEE OPERATED, SSV-130-D DOUBLE BAY SCRUB SINK WITH KNEE OPERATED, SSV-130-M DOUBLE BAY SCRUB SINK WITH FOTOCELL, SSV-205-F TRIPLE BAY SCRUB SINK WITH FOTOCELL, SSV-205-D TRIPLE
F C V	BAY SCRUB SINK WITH KNEE OPERATED, SSD-01 OPERATION ROOM CUPBOARD, SSD-020 PERATION ROOM CUPBOARD, SSM-01 OPERATION ROOM WORKDESK, SSM-02 OPERATION ROOM WORKDESK, SSW- 01 TOOLS WASHING SINK WITH TABLE, SSW-02 TOOLS WASHING SINK, SSW-03 TOOLS WASHING SINK WITH TABLE, ESB-01 ESCABO, SINGLE, ESB-02 ESCABO DOUBLE, SST-01 INSTRUMENT TROLLEY 2 SHELVES, SST-02 INSTRUMENT TROLLEY 3 SHELVES, MM-01 MAYO TABLE HYDRAULIC, MM-02 MAYO TABLE
י ז	MANUAL, SOT-01 AUTOPSY TABLE WITH SINK, SOT-02 AUTOPSY TABLE HEIGHT ADJUSTABLE, SMU-02 MORTUARY UNIT DOUBLE, PK-01 SINGLE BOWL STAND, PK-02 DOUBLE BOWL STAND, SDT-01 DRESSING
ŀ	IROLLEY, AVANGARD 5000 HOSPITAL ELECTRIC BED 4 MOTORS, ALUMINUM BASE, AVANGARD 4040 HOSPITAL ELECTRIC BED 4 MOTORS, ALUMINUM BASE, AVANGARD 3050 HOSPITAL ELECTRIC BED 3 MOTORS, ALUMINUM BASE, AVANGARD 2060 HOSPITAL ELECTRIC BED 2 MOTORS, ALUMINUM BASE,
C	2018 5000 HOSPITAL ELECTRIC BED, 5 MOTORS, CURA 4060 HOSPITAL ELECTRIC BED, 4 MOTORS, CURA 4050 HOSPITAL ELECTRIC BED, 4 MOTORS, CURA 4040 HOSPITAL ELECTRIC BED, 4 MOTORS, CURA 3060
ŀ	HOSPITAL ELECTRIC BED, 3 MOTORS, CURA 3050 HOSPITAL ELECTRIC BED, 3 MOTORS, CURA 3060 HOSPITAL ELECTRIC BED, 3 MOTORS, CURA 3050 HOSPITAL ELECTRIC BED, 3 MOTORS, CURA 2060 HOSPITAL ELECTRIC BED, 2 MOTORS, CURA 2050 HOSPITAL ELECTRIC BED, 2 MOTORS, CURA 2040
	HOSPITAL ELECTRIC BED, 2 MOTORS, CURA 240 DX MECHANICAL OPERATED HOSPITAL BED, 4 CRANKS, CURA 240 PL MECHANICAL HOSPITAL BED, 4 CRANKS, CURA 230 PL MECHANICAL OPERATED HOSPITAL
E F	3ED, 3 CRANKS, CURA 220 PL PARALLEL MECHANICAL OPERATED HOSPITAL BED, 2 CRANKS, CURA 220 DX PLASTIC MECHANICAL OPERATED HOSPITAL BED, 2 CRANKS, CURA 120 PL MECHANICAL OPERATED HOSPITAL BED, 1 CRANK, PDB 1130 PEDIATRIC HOSPITAL BED, 3 MOTORISED, PDB 1120 PEDIATRIC HOSPITAL BED, 2 ADJUSTMENTS, PDB 1100 PEDIATRIC HOSPITAL BED, 2 MOTORISED, PDB 1020 PEDIATRIC
	HOSPITAL BED, 2 ADJOSTNELATS, PDB 1100 PEDIATRIC HOSPITAL BED, 2 MOTORISED, PDB 1020 PEDIATRIC HOSPITAL BED, 2 MOTORISED, PDA 3020 PEDIATRIC HOSPITAL BED, 3 MOTORISED ALUMINUM BASE,



A) Declaration

We declare that the said products are safe under the conditions of common use in compliance with the instructions and that measures have been taken to ensure the conformity of all the products brought to market with basic requirements of directives related thereto, stated in paragraph B.

B) Fulfilled technical requirements

These products' characteristics comply with the technical parameters related to it and stated in MDD 93/42/EC as amended, which stipulates the technical parameters for healthcare products.

C) Means of assessing conformity

Conformity was assessed by the procedure stated MDD 93/42/EEC

D) Used standards

The said products fulfill the requirements of these harmonized technical standards which were used for assessing of conformity: -EN60601-2-41:2009, -EN 60601-1-2:2015, -EN 60601-1:2006 +A1:2013, -TS EN 60601-1, -EN 60601-2-52:2010

