

VERIFICATION OF COMPLIANCE


No.: LVD SHES2106012445ME

Applicant: Zhongke Meiling Cryogenics Company Limited
No.1862 Zishi Road, Economic & Technological Development
Zone, 230601 Hefei, China

Manufacturer: Zhongke Meiling Cryogenics Company Limited
No.1862 Zishi Road, Economic & Technological Development
Zone, 230601 Hefei, China

Product Name: Pharmacy Refrigerator

Model No.: See Page 2

Trade Mark: 

Rating: See Page 2

Protection against Electric Shock: Class I

Additional Information: N/A

Sufficient samples of the product have been tested and found to be in conformity with

Test Standard: EN 61010-1:2010+A1:2019

as shown in the

Test Report Number(s): SHES210601244501

This Verification of Compliance has been granted to the applicant based on the results of tests, performed by Laboratory of SGS-CSTC Standards Technical Services Co., Ltd. on sample of the above-mentioned product in accordance with the provisions of the relevant harmonized standards under the Low Voltage Directive 2014/35/EU. The CE marking as shown below can be affixed, under the responsibility of the manufacturer, after completion of an EC Declaration of Conformity and compliance with all relevant EC Directives. The affixing of the CE marking presumes in addition that the conditions in annexes III and IV of the Directive are fulfilled.



Andrew Zhai
Technical Manager
SGS-CSTC



2022-01-27

This verification is issued by the company under its General Conditions of Services accessible at <https://www.sgs.com/en/terms-and-conditions>. Attention is drawn to the limitations of liability defined therein and in the Test Report here above mentioned which findings are reflected in this verification. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Member of SGS Group (Société Générale de Surveillance)

No.:

LVD SHES2106012445ME

Other information added:

Model and ratings list:

| Model | Ratings |
|-------------------|-----------------------------|
| YC-55L, YC-55EL | 220-240 VAC, 50 Hz, 0.9 VA |
| YC-75L, YC-75EL | 220-240 VAC, 50 Hz, 0.92 VA |
| YC-130L, YC-130EL | 220-240 VAC, 50 Hz, 0.98 VA |
| YC-315L, YC-315EL | 220-240 VAC, 50 Hz, 1.35 VA |
| YC-330L, YC-330EL | 220-240 VAC, 50 Hz, 1.31 VA |
| YC-395L, YC-395EL | 220-240 VAC, 50 Hz, 1.8 VA |
| YC-525L, YC-525EL | 220-240 VAC, 50 Hz, 2.49 VA |
| YC-725L, YC-725EL | 220-240 VAC, 50 Hz, 3.90 VA |

Andrew Zhai

Technical Manager

SGS-CSTC

This verification is issued by the company under its General Conditions of Services accessible at <https://www.sgs.com/en/terms-and-conditions>. Attention is drawn to the limitations of liability defined therein and in the Test Report here above mentioned which findings are reflected in this verification. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Member of SGS Group (Société Générale de Surveillance)



2022-01-27