
Declaration letter

To whom it may concern

We, SHENZHEN COMEN MEDICAL INSTRUMENTS CO., LTD

**Do confirm that offered model L5(L5B+L5A) for the tender 21080839
LOT 10 Surgical lamp with 2 satellites (medium characteristics) has the
following technical specification:**

Working distance 0.7-1.5 m Yes

Vertical adjustment, cm ≥ 70 Yes

Temperature increase in the operative field $<1^{\circ}$ C Yes

Average lifetime of LEDs min. 50000 h Yes

Removable sterilizable handles Yes

Bridge height 3-4 m Yes

Signature:

Name: Shaun Yeung

Date: 2023/06/29

