

# Test Report

## – Eppendorf PCR clean & sterile –

### ep Dualfilter T.I.P.S.<sup>®</sup> 50 - 1000 µL



10x 96 Eppendorf Tips  
PCR clean / sterile

LOT L207928Q

2027-11-28

STERILE R

Order no.: 0030 078.578  
Eppendorf SE  
22331 Hamburg · Germany  
Email: eppendorf@eppendorf.com

Cat. no.: 0030078578  
Distributed in North America by  
Eppendorf North America, Inc. · USA  
Email: Info@eppendorf.com

- Human DNA**  
Method: *real-time* PCR acc. to SOP 99880/V2  
Limit: < 2 pg; less than one human cell
- PCR Inhibition**  
Method: *real-time* PCR acc. to SOP 100559/V1  
Limit: less than 10 targets amplifiable
- DNase**  
Method: DNA Digestion acc. to SOP 100570/V1  
Limit: Not detectable  
(LOD: 1.0 x 10<sup>-6</sup> Kunitz units)
- Sterility**  
Method: Sterilization with SAL 10<sup>-6</sup> by  
 Irradiation according DIN EN ISO 11137:2015\* or  
 EO: Ethylene oxide\*  
\*Please see attached label for the standard method of sterilization for this product
- RNase**  
Method: RNA Digestion acc. to SOP 100571/V1  
Limit: Not detectable  
(LOD: 1.0 x 10<sup>-9</sup> Kunitz units)
- Test method: USP, Ph. Eur. 8.0 / 2.6.12
- Endotoxin**  
Method: Kinetic-chromogenic LAL test (Ph. Eur. 8.0 / 2.6.14)  
Limit: < 0.001 EU/mL

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 15.11.22

Testing of the above described lot showed conformity within the limits of detection.

18.11.22

Date of release

Eppendorf - 2254

Registration number

F. Weigel

Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.

page 1 of 1