



# EU Declaration of Conformity

In accordance with EU Regulation 2017/746 of the European Parliament and of the Council of 5 April 2017 on *in vitro* diagnostic medical devices.

Manufacturer: **Ventana Medical Systems Inc.**  
**1910 E Innovation Park Drive**  
**Tucson, AZ 85755, USA**

Single Registration Number (SRN) **US-MF-000016993**  
 Manufacturer:

Authorized Representative: **Roche Diagnostics GmbH**  
**Sandhofer Strasse 116**  
**68305 Mannheim**  
**Germany**

Single Registration Number (SRN) **DE-AR-000006262**  
 Authorized Representative:

This declaration is issued under the sole responsibility of Ventana Medical Systems Inc.

## Product Information

<b>Part Number:</b>	<b>Product Name:</b>	<b>Basic UDI-DI:</b>
05929903001 (alternative P/N: 790-4451)	CONFIRM anti-CD5 (SP19) Rabbit Monoclonal Primary Antibody	761333601857BH

**Intended Purpose:** CONFIRM anti-CD5 (SP19) Rabbit Monoclonal Primary Antibody is intended for laboratory use in the qualitative immunohistochemical detection of CD5 by light microscopy in sections of formalin fixed, paraffin embedded tissue stained on a BenchMark IHC/ISH instrument.

**Risk Class:** Class C

**Common Specifications:** Not applicable as no Common Specifications exist for the concerned device.

**Name, Address and Identification number of the Notified Body:** TÜV SÜD Product Service GmbH (No. 0123)  
 Ridlerstraße 65  
 80339 MÜNCHEN  
 Germany

Conformity Assessment was established through the procedure described in Annex IX of EU Regulation 2017/746, including an assessment of the Technical Documentation as described in Annex IX, Chapter II. This declaration is supported by the following certificate(s):

EU Quality Management System Certificate: V12 096981 0003



Conformity of the product with EU Regulation 2017/746 and other applicable EU legislation has been established.

On behalf of Ventana Medical Systems Inc.

Place: Tucson, AZ 85755, USA

Date: 22-Jun-2022

Place: Tucson, AZ 85755, USA

Date: 21-Jun-2022

*Jeff Boone*

**Jeff Boone**

Site Head of Quality Function

*Ben Curson*

**Benjamin Curson**

Site Head of Regulatory Affairs Function