

CERTIFICATE OF ANALYSIS

MICRO SSP COMPONENT

Product Name: 5XTB Buffer with Ethidium Bromide
Catalog Number: 5XTBE100
Lot Number: 143
Expiration Date: See printed expiration date on package.

Quality Control Requirement: Each production lot is validated by HLA typing concordance test using reference samples, and must meet the following acceptance criteria:

1. Each lot when tested against a control lot must be comparable for overall signal intensity, number of non-amplified samples, and quality.
2. Number of False Positives must be zero.
3. Number of False Negatives must be zero.
4. Good optimal visual qualities must be exhibited for unambiguous recognition of gel patterns

Method of Analysis: PCR typing method in accordance with the standard protocol described in product insert, and HLA Fusion software.

Results: Passed

This product is prepared in accordance to GMP standards.

This certificate is applicable to all batches under the specified lot that meet the QC requirements as stated herein.


Liliana Lioe/Product Release Specialist

22 Nov 2021
Date issued